


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0043 Accredited to ISO/IEC 17025:2017	Roxspur Measurement and Control Ltd Issue No: 072 Issue date: 07 May 2026	
	2 Downgate Drive Sheffield South Yorkshire S4 8BT	Contact: Mr M Donnelly Tel: +44 (0)114 244 2521 Fax: +44 (0)114 243 4838 E-Mail: Mark.Donnelly@ttelectronics.com Website: www.roxspur.com

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 2 Downgate Drive Sheffield South Yorkshire S4 8BT Local contact Mr Mark Donnelly Tel: +44 (0)114 244 2521 Fax: +44 (0)114 243 4838 Email: Mark.Donnelly@ttelectronics.com Website: www.roxspur.com	Electrical Flow Pressure Temperature	P1

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. Local contact Mr Mark Donnelly Tel: +44 (0)114-244 2521 Fax: +44 (0)114-243 4838 Email: Mark.Donnelly@ttelectronics.com Website: www.roxspur.com	Electrical Pressure Temperature	S



0043

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 072 Issue date: 07 May 2026

Calibration performed by the Organisation at the locations specified

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL				
DC Voltage	0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	Q[15 μ V/V, 0.70 μ V] Q[10 μ V/V, 0.60 μ V] 10 μ V/V 15 μ V/V 15 μ V/V	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	0 V to 30 V	5.0 mV		S
DC Current	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	Q[50 μ A/A, 2.0 nA] Q[75 μ A/A, 15 nA] 75 μ A/A 100 μ A/A 200 μ A/A	Measured using digital multimeter. The capability includes generation and measurement of this quantity	P1
	0 mA to 100 mA	0.010 mA		S
Generation	320 mA to 3.2 A 3.2 A to 10 A 10 A to 20 A	Q[550 μ A/A, 150 μ A] Q[500 μ A/A, 1.1 mA] Q[0.11 %, 5.2 mA]	For the calibration of clampmeters and similar devices using multi-turn coil method.	P1
	20 A to 1000 A	0.37 %		
DC Resistance Measurement	0 Ω to 20 Ω 20 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω	Q[100 μ Ω / Ω , 2.0 m Ω] Q[20 μ Ω / Ω , 2.0 m Ω] 20 μ Ω / Ω 30 μ Ω / Ω 50 μ Ω / Ω 200 μ Ω / Ω 700 μ Ω / Ω 0.10 %	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
AC Voltage	10 mV to 200 mV 20 Hz to 10 kHz 10 kHz to 100 kHz	Q[280 μ V/V, 8.0 μ V] Q[550 μ V/V, 10 μ V]	Measured using digital multimeter. The capability includes generation and measurement of this quantity.	P1
	200 mV to 200 V 20 Hz to 30 kHz 30 kHz to 100 kHz	400 μ V/V 480 μ V/V		
	200 V to 1100 V 45 Hz to 10 kHz 10 kHz to 30 kHz	450 μ V/V 510 μ V/V		



0043

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 072 Issue date: 07 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)				
AC Current	20 Hz to 1 kHz		Measured using digital multimeter. The capability includes generation and measurement of this quantity	
	10 μ A to 200 μ A 200 μ A to 200 mA 200 mA to 2 A	Q[250 μ A/A, 41 nA] 350 μ A/A 600 μ A/A		P1
	1 kHz to 5 kHz			
	10 μ A to 200 μ A 200 μ A to 200 mA 200 mA to 2 A	Q[800 μ A/A, 30 nA] 700 μ A/A 0.13 %		P1
Generation	10 Hz to 3 kHz 2 A to 20 A	Q[0.25 %, 7.0 mA]		
	40 Hz to 100 Hz 20 A to 1000 A	0.50 %	For the calibration of clampmeters and similar devices using multi-turn coil method.	
Frequency	0.1 Hz to 300 MHz	0.70 μ Hz/Hz	Using counter timer.	P1
Time interval	30 s to 8 hrs 30 s to 8 hrs	0.15 s 0.25 s	Using counter timer.	P1 S
Electrical calibration of temperature indicators, controllers and recorders for the following sensors:				
Noble metal thermocouples			Using millivolt injection or measurement.	
Types R and S	0 °C to 200 °C 200 °C to 1700 °C	0.50 °C 0.40 °C	with cold junction Compensation.	P1
Type B	600 °C to 1700 °C	0.40 °C	with cold junction compensation.	P1
Type R and S	0 °C to 200 °C 200 °C to 1700 °C	0.60 °C 0.60 °C	with cold junction compensation.	S
Type B	600 °C to 1700 °C	1.0 °C	with cold junction compensation.	S
Base metal thermocouples	-200 °C to 0.0 °C 0 °C to 1372 °C	0.26 °C 0.20 °C	with cold junction compensation.	P1
	-200 °C to 0.0 °C 0 °C to 1372 °C	0.80 °C 0.60 °C	with cold junction compensation.	S



0043

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 072 Issue date: 07 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Electrical calibration of temperature indicators, controllers and recorders for the following sensors: (continued)				
Pt 100	-200 °C to +850 °C	0.050 °C	Using DC Resistance measurement or injection.	P1
	-200 °C to +850 °C	0.50 °C		S
FLOW				
Flow rate calibration using air, oxygen, argon, helium and nitrogen	5 ml/min to 40 ml/min 40 ml/min to 200 ml/min 200 ml/min to 6 l/min 6 l/min to 10 l/min 10 l/min to 20 l/min 20 l/min to 30 l/min 30 l/min to 100 l/min 100 l/min to 200 l/min 200 l/min to 600 l/min	2.0 % 1.1 % 1.2 % 1.1 % 1.2 % 1.7 % 1.1 % 1.2 % 1.4 %	Calibration of flowmeters measuring volumetric or mass flowrates under ambient conditions.	P1
PRESSURE				
Gas pressure (gauge)			Methods consistent with EURAMET CG17. Absolute pressures (P1) across these gauge ranges can be generated – this will attract an additional uncertainty of 72 Pa (P1) and 0.52 kPa (S).	
Calibration of pressure indicating instruments and gauges	-90 kPa to -1.5 kPa 1.5 kPa to 100 kPa 100 kPa to 2.5 MPa	0.014 % 0.011 % 0.011 %	Using deadweight tester Using deadweight tester Using deadweight tester	P1
	-90 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa	90 Pa 9.4 Pa	Using digital pressure standards	P1
Hydraulic pressure (gauge)	-90 kPa to +2.0 MPa	5.0 kPa	Using digital pressure standards	S
Calibration of pressure indicating instruments and	600 kPa to 6 MPa 6 MPa to 120 MPa	0.019 % 0.019 %	Using deadweight tester Using deadweight tester	P1
	0 to 70 MPa	0.30 MPa	Using digital pressure standards	S



0043

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 072 Issue date: 07 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Resistance thermometers	-196 °C 0 °C to -80 °C Ice Point (0 °C) Triple Point of water (0.01 °C) Gallium Melt Point (29.7646 °C) 0 °C to 300 °C	0.040 °C 0.050 °C 0.031 °C 0.012 °C 0.010 °C 0.050 °C	Calibration performed within Liquid Baths.	P1
Platinum thermocouples	0 °C to 1100 °C 1100 °C to 1600 °C 1064.18 °C 1553.5 °C 0 °C to 1100 °C 1100 °C to 1600 °C Fixed point calibrations Gallium Melt Point (29.7646 °C) Tin (231.9 °C) Zinc (419.5 °C)	0.50 °C 1.8 °C 0.50 °C 1.8 °C 2.0 °C 4.0 °C 0.35 °C 0.50 °C 0.50 °C	Calibration performed within Tube Furnace or Metrology Well. Au and Pd wire bridge measurements.	P1 P1
Other thermocouples	-196 °C 0 °C to -80 °C Gallium Melt Point (29.7646 °C) 0 °C to 300 °C 300 °C to 660 °C 660 °C to 1100 °C 1100 °C to 1300 °C -80 °C to +200 °C 200 °C to 1000 °C 1000 °C to 1300 °C	0.26 °C 0.15 °C 0.10 °C 0.15 °C 0.25 °C 1.0 °C 2.5 °C 1.0 °C 3.0 °C 5.0 °C	Calibration performed within Liquid Bath, Metrology Well or Furnace.	P1 S
Compensating and extension cables for Noble metal thermocouples Base metal thermocouples	0 °C to 40 °C 0 °C to 40 °C	0.26 °C 0.15 °C	Calibration performed within Liquid Baths	P1
Liquid-in-glass thermometers	-80 °C to 100 °C 100 °C to 300 °C	0.04 °C 0.050 °C	Uncertainties will increase for partial immersion thermometers	P1
Electronic thermometers with sensors	Range as per sensors	As for sensors		P1
Electronic thermometers with sensors Analogue Digital	0 °C to 200 °C 200 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1300 °C	0.30 °C 1.3 °C 2.0 °C 3.4 °C		S



0043

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 072 Issue date: 07 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE (continued)				
Metal block calibrators	-100 °C to +300 °C 300 °C to 1100 °C	0.050 °C 1.0 °C		P1
Temperature surveys				
Autoclaves, incubators, fridges/refrigerators and freezers	-80 °C to +200 °C	1.0 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping.	S
Ovens, furnaces, fridges/refrigerators and environmental chambers	0 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1600 °C	1.1 °C 1.9 °C 3.5 °C		
END				



0043

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 072 Issue date: 07 May 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$