

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0066

Accredited to
ISO/IEC 17025:2017

Denison Mayes Group Limited

Issue No: 068

Issue date: 17 November 2023

98 Church Street
Hunslet
Leeds
West Yorkshire
LS10 2AZ

Contact: Mr R Adkin
Tel: +44 (0)113 270 8011
Fax: +44 (0)113 271 2850
E-Mail: r.adkin@denisonmayesgroup.com
Website: www.denisonmayesgroup.com

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers' sites or premises The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Force Hardness Mass	S



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
FORCE				S
UNIVERSAL MATERIAL TESTING MACHINES				
Verification and calibration of the force measuring system by force proving instruments in tension	0.10 kN to 500 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.32 %		
	0.05 kN to 4 MN For Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.38 %		
	0.05 kN to 4 MN to ASTM E4-21	0.38 %		
Verification and calibration of the force measuring system by force proving instruments in compression	0.10 kN to 500 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.32 %		
	0.05 kN to 5 MN For Class 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.38 %		
	0.05 kN to 10 MN for ASTM E4-21	0.38 %		
Verification and calibration of the force measuring system by calibrated masses in tension	0.01 N to 500 N for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.10 %		
Verification and calibration of the force measuring system by calibrated masses in compression	0.01 N to 500 N for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2018	0.10 %		
CONCRETE CUBE TESTING MACHINES				S
Verification and calibration of the force measuring system by force proving instruments in Compression	0.05 kN to 5 MN for Class 1, 2 and 3 machines to BS EN ISO 7500-1:2018 See note 1	0.38 %	Notes 1. Calibration also includes the alignment and restraint of the upper machine platen required by BS EN 12390-4:2019	



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
FORCE (continued)				S
Rate of application of force (Load pacer rate)	As BS EN 12390-4:2019 3 kN/min to 1300 kN/min	1.0 %		
Flatness of platens (Machine platens, auxiliary platens and spacing blocks)	As BS EN 12390-4:2019 40 mm to 300 mm	0.010 mm		
CREEP TESTING MACHINES				
Verification and calibration of the force measuring system by force proving instruments in tension	0.1 kN to 100 kN for Class 0.5, 1 and 2 machines to BS EN ISO 7500-2:2006	0.32 %		
Verification and calibration of the force measuring system by calibrated masses in tension	5 N to 100 N for Class 0.5, 1 and 2 machines to BS EN ISO 7500-2:2006	0.10 %		
FORCE MEASURING DEVICES				
Calibration of force measuring devices used in soil testing machines	0.1 kN to 500 kN As BS 1377:Part 1:2016	0.38 %		
IMPACT TESTING MACHINES				
Charpy	Absorbed Energy (joules) 1 J to 600 J BS EN ISO 148-2:2016 ASTM E23-18 but excluding proof test using certified specimens	0.70 J		
Izod	BS 131:Part 4:1972			



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
MASS Masses for generating force on tension creep testing machines	100 g to 5 kg 5 kg to 33 kg	0.340 g 2.1 g		S
HIP SIMULATORS ISO 14242-1:2014+A1:2018 (section 6 – Apparatus)	Angular displacement (protractor) 0 to 50 degrees of arc Angular displacement (clinometer) 0 to 60 degrees of arc Test cycle frequency 0 to 1 Hz Linear displacement (see note) 0 to 10 mm	0.25 degrees of arc 0.50 degrees of arc 0.0040 Hz 0.025 mm	In house procedures A11.16 A11.16 A11.17 A11.15 Note: Measurement of linear displacement is not required by the standard but has been included for completeness	S
LENGTH Extensometers	As BS EN ISO 9513:2012 for the following classes and gauge lengths: Class 0.2 from 25 mm Class 0.5 from 10 mm Up to Class 1 from 5 mm Class 2 from 5 mm As ASTM E83-16 for the following classes and gauge lengths: B-1 from 20 mm B-2 from 10 mm C from 5 mm Displacements 0.02 mm to 2.5 mm 2.5 mm to 100 mm	 0.50 μm + (0.50 μm per mm) 5.0 μm + (0.25 μm per mm)	 .	S
Testing machine cross head and actuator displacement by use of linear encoders	0.01 mm to 50 mm 50 mm to 300 mm 300 mm to 450 mm 450 mm to 500 mm	5.5 μm + (700 μm per m) 20 μm + (700 μm per m) 25 μm + (700 μm per m) 45 μm + (700 μm per m)		



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HARDNESS TESTING MACHINES IN SERVICE Direct verification of Rockwell Hardness Testing Machines Indirect verification of Rockwell Hardness Testing Machines	Rockwell scales: A, B, C, D, E, F, G, H, K, L, M, N , P, R, S, T, V, W, X & Y Force Length Time Rockwell scales: A, B, C, D, E, F, G, H, K, L, M, N , P, R, S, T, V, W, X & Y HRA Scale 80 to 85 70 to 79 60 to 69 HRB Scale 80 51 to 79 10 to 50 HRC Scale 60 to 70 40 to 59 20 to 39 HRD Scale 70 to 80 50 to 69 40 to 49 HRE Scale 89 75 to 88 65 to 87 HRF Scale 87 70 to 86 40 to 69 HRG Scale 80 40 to 79 10 to 39 HRH Scale 90 80 to 89 60 to 79	See Note 2 0.24 % 0.20 μm 0.10 s See note 2 0.15 HRA 0.16 HRA 0.28 HRA 0.42 HRB 0.87 HRB 1.36 HRB 0.31 HRC 0.32 HRC 0.37 HRC 0.17 HRD 0.25 HRD 0.27 HRD 0.54 HRE 0.54 HRE 0.54 HRE 0.40 HRF 0.40 HRF 0.54 HRF 0.30 HRG 0.30 HRG 0.76 HRG 0.40 HRH 0.40 HRH 0.68 HRH	2. The calibration/ verification shall be in accordance with the requirements of BS EN ISO 6508- 2:2015 & ASTM E18-20	S



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HARDNESS TESTING MACHINES IN SERVICE (continued) Indirect verification of Rockwell Hardness Testing Machines (cont'd)				S
	HRK Scale 70 30to69 10to29	0.40 HRK 0.40 HRK 0.64 HRK		
	HRL Scale 115 90 to 114	0.35 HRL 0.35 HRL		
	HRM Scale 100 70 to 99	0.56 HRM 0.56 HRM		
	HRP Scale 85 40 to 84	0.65 HRP 0.91 HRP		
	HRR Scale 120 100 to 119	0.23 HRR 0.40 HRR		
	HRS Scale 112 110 to 111	0.19 HRS 0.91 HRS		
	HRV Scale 104 80 to 103	0.20 HRV 0.61 HRV		
	HR15N Scale 90to95 80to89 40to79	0.18 HR15N 0.18 HR15N 0.39 HR15N		
	HR15T Scale 88 to 100 80 to 87 20 to 79	0.21 HR15T 0.21 HT15T 0.37 HR15T		
	HR15W Scale 89 to 100 80 to 88	0.53 HR15W 0.44 HR15W		
	HR15X Scale 88 to 100 80 to 87	0.33 HR15X 0.62 HR15X		
	HR15Y Scale 94 to 100 85 to 93	0.63 HR15Y 1.30 HR15Y		



0066
Accredited to
ISO/IEC 17025:2017


Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HARDNESS TESTING MACHINES IN SERVICE (continued) Indirect verification of Rockwell Hardness Testing Machines (cont'd)	HR30N Scale 77 to 85 60 to 76 40 to 59 HR30T Scale 57 to 85 50 to 56 20 to 49 HR30W Scale 65 to 100 40 to 64 HR30X Scale 79 to 100 60 to 78 HR30Y Scale 88 to 100 60 to 87 HR45N Scale 67 to 75 50 to 66 10 to 49 HR45T Scale 50 to 75 40 to 49 10 to 39 HR45W Scale 49 to 100 10 to 47 HR45X Scale 69 to 100 40 to 68 HR45Y Scale 82 to 100 60 to 81	0.27 HR30N 0.27 HR30N 0.55 HR30N 0.39 HR30T 0.66 HR30T 0.90 HR30T 0.76 HR30W 0.90 HR30W 0.15 HR30X 0.99 HR30X 0.37 HR30Y 0.82 HR30Y 0.18 HR45N 0.21 HR45N 0.43 HR45N 0.40 HR45T 0.40 HR45T 0.73 HR45T 0.12 HR45W 0.29 HR45W 0.34 HR45X 0.81 HR45X 0.29 HR45Y 0.94 HR45Y		S
Direct verification of Brinell Hardness Testing Machines	Brinell scales: HB 10/3000 to HB1/1 Force Time Length	See note 3 0.24 % 0.10 s 6 μm	3. The calibration/ verification shall be in accordance with the requirements of BS EN ISO 6506- 2:2018 and ASTM E10-18.	



issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
HARDNESS TESTING MACHINES IN SERVICE (continued) Indirect verification of Vickers Hardness Testing Machines	Vickers scales: HV 100 200 HV 100 400 HV 100 700 HV 50 200 HV 50 400 HV 50 700 HV 30 200 HV 30 400 HV 30 700 HV 20 200 HV 20 400 HV 20 700 HV 10 200 HV 10 400 HV 10 700 HV5 200 HV5 400 HV5 700 HV3 200 HV3 400 HV3 700 HV1 200 HV1 400 HV1 700 HV 0.5 200 HV 0.5 400 HV 0.5 700 HV 0.3 200 HV 0.3 400 HV 0.3 700 HV 0.2 200 HV 0.2 400 HV 0.2 700 HV 0.1 200 HV 0.1 400	See note 4 1.2 HV 3.4 HV 4.1 HV 1.9 HV 3.5 HV 6.3 HV 2.0 HV 4.4 HV 9.3 HV 2.5 HV 6.2 HV 11.0 HV 3.1 HV 7.7 HV 14.9 HV 3.9 HV 11.0 HV 19.7 HV 6.9 HV 16.3 HV 31.0 HV 8.7 HV 21.4 HV 44.0 HV 5.0 HV 15.0 HV 17.0 HV 6.0 HV 16.0 HV 19.0 HV 7.0 HV 17.0 HV 20.0 HV 10.0 HV 30.0 HV		S
END				



0066
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Denison Mayes Group Limited
Issue No: 068 Issue date: 17 November 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$