


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|---|---|
|  0080 Accredited to ISO/IEC 17025:2017 | Southern Calibration Laboratories Limited | |
| | Issue No: 052 Issue date: 14 March 2025 | |
| | Unit 7 Solent Industrial Estate Hedge End Southampton SO30 2FX | Contact: Mr S C Sparks Tel: +44 (0)1489 790296 Fax: +44 (0)1489 790294 E-Mail: info@southcal.co.uk Website: www.southcal.co.uk |
| Calibration performed by the Organisation at the locations specified | | |

Locations covered by the organisation and their relevant activities

Laboratory location:

| Location details | | Activity | Location code |
|--|---|--|---------------|
| Address Unit 7 Solent Industrial Estate Hedge End Southampton SO30 2FX | Local contact Mr S C Sparks Tel: +44 (0)1489 790296 Fax: +44 (0)1489 790294 E-Mail: info@southcal.co.uk Website: www.southcal.co.uk | Electrical Temperature Dimensional | Lab |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|--|-------------------------|---------------|
| The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Temperature Humidity | Site |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--------------------|---|---|------------------|
| ELECTRICAL CALIBRATION | | | | |
| DC VOLTAGE | | | By comparison with a voltage standard, using voltage ratio techniques. | Lab |
| Specific Values | 1.018 V | 1.5 μ V | This uncertainty can only be achieved with temperature controlled cells of suitable stability. | |
| | 0.1 V | 1.9 μ V/V | | |
| | 1 V | 0.80 μ V/V | | |
| | 10 V | 0.80 μ V/V | | |
| | 100 V | 0.80 μ V/V | | |
| | 1 kV | 0.95 μ V/V | | |
| Other Values | 0 mV to 100 mV | 0.37 μ V | | |
| | 100 mV to 1.1 V | 1.0 μ V | | |
| | 1.1 V to 11 V | 0.77 μ V/V + 2.4 μ V | | |
| | 11 V to 1 kV | 0.95 μ V/V | | |
| | 1 kV to 2 kV | 0.20 % + 0.12 V | Using DC kilovolt meter. | |
| | 2 kV to 40 kV | 0.20 % + 12V | | |
| DC VOLTAGE RATIO | Unity to 10^{-7} | 2.6×10^{-7} of input | Input voltage range 1 V to 100 V. Using voltage ratio techniques. | Lab |
| DC RESISTANCE | | | | |
| Measurement and Generation | 0.01 Ω | 28 $\mu\Omega/\Omega$ | Fixed values for the calibration of measuring instruments. Potentiometric comparison with standard resistors. | Lab |
| | 0.1 Ω | 7.2 $\mu\Omega/\Omega$ | | |
| | 1 Ω | 3.8 $\mu\Omega/\Omega$ | | |
| | 10 Ω | 2.9 $\mu\Omega/\Omega$ | | |
| | 100 Ω | 2.1 $\mu\Omega/\Omega$ | | |
| | 1 k Ω | 2.7 $\mu\Omega/\Omega$ | | |
| | 10 k Ω | 2.6 $\mu\Omega/\Omega$ | | |
| | 100 k Ω | 4.7 $\mu\Omega/\Omega$ | | |
| | 1 M Ω | 10 $\mu\Omega/\Omega$ | | |
| | 10 M Ω | 25 $\mu\Omega/\Omega$ | | |
| Generation only | 100 M Ω | 200 $\mu\Omega/\Omega$ | Fixed resistance values for calibration of measuring instruments. The applied voltage will normally be in the range 100 V to 500 V. | Lab |
| | 1 G Ω | 0.45 % | | |
| | 10 G Ω | 0.45 % | | |
| | 100 G Ω | 0.75 % | | |
| | 1 T Ω | 2.75 % | | |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|--|------------------|
| DC RESISTANCE (continued) | | | | |
| Measurement only | 0 Ω to 20 Ω 20 Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 100 M Ω | 110 $\mu\Omega$ 7.0 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 300 $\mu\Omega/\Omega$ | Using digital multimeter. | Lab |
| DC CURRENT | 0 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 0.1 A to 1 A 1 A to 10 A 10 A to 100 A 100 A to 1000 A | 3.5 μ A/A + 300 pA 3.5 μ A/A + 2.5 nA 3.0 μ A/A + 26 nA 13 μ A/A + 0.26 μ A 18 μ A/A + 3.0 μ A 25 μ A/A + 30 μ A 0.12 % + 2.0 mA 0.15 % + 250 mA | Voltage and resistance method. For the calibration of clamp meters and similar devices, using multi-turn coil technique. | Lab |
| Specific Value | 100 A | 0.032 % | Using current shunt. | |
| AC VOLTAGE | <i>10 Hz to 200 kHz</i> 0.9 mV to 3.5 mV 3.5 mV to 10 mV 10 mV to 35 mV 35 mV to 90 mV <i>200 kHz to 500 kHz</i> 0.9 mV to 3.5 mV 3.5 mV to 10 mV 10 mV to 35 mV 35 mV to 90 mV <i>0.5 MHz to 1 MHz</i> 0.9 mV to 3.5 mV 3.5 mV to 10 mV 10 mV to 35 mV 35 mV to 900 mV <i>10 Hz to 30 kHz</i> 90 mV to 100 V 100 V to 1.1 kV 90 mV to 350 mV <i>30 kHz to 200 kHz</i> <i>200 kHz to 500 kHz</i> <i>500 kHz to 1 MHz</i> | 0.20 % 0.10 % 0.055 % 0.035 % 0.40 % 0.21 % 0.16 % 0.090 % 0.75 % 0.52 % 0.40 % 0.25 % 60 μ V/V 80 μ V/V 200 μ V/V 450 μ V/V 0.12 % | Using alternating voltage measurement standard. Both measurement and generation may be undertaken, however generation will be limited to a minimum of 45 Hz above 100 V. | Lab |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|---|--|------------------|
| AC VOLTAGE (continued) | 0.35 V to 35 V 30 kHz to 200 kHz | 110 μ V/V | | Lab |
| | 0.35 V to 35 V 200 kHz to 500 kHz 500 kHz to 1 MHz | 0.040 % 0.11 % | | |
| | 35 V to 100 V 30 kHz to 100 kHz | 85 μ V/V | | |
| | 50 Hz 1 kV to 2 kV 2 kV to 20 kV | 0.40 % + 0.12 V 0.40 % + 12 V | Using AC kilovolt meter. | |
| AC RESISTANCE | 50 Hz 0.06 Ω , 0.1 Ω , 0.2 Ω , 0.5 Ω , 1 Ω , 5 Ω 10 Ω , 100 Ω and 1 k Ω | 0.60 % + 30 m Ω | For calibration of loop testers, using dedicated calibrator. | Lab |
| | 50 Hz 0.05 Ω , 0.1 Ω , 0.2 Ω , 0.5 Ω | 1.5 % | Earth bond resistance using dedicated calibrator. | |
| AC CURRENT | 40 Hz to 3 kHz 1 μ A to 2 mA | 0.020 % + 50 nA | Voltage and resistance method. | Lab |
| | 40 Hz to 1 kHz 2 mA to 6 mA | 0.070 % | | |
| | 40 Hz to 10 kHz 6 mA to 10 mA 18 mA to 30 mA 60 mA to 100 mA 180 mA to 300 mA 600 mA to 1 A 1.8 A to 3 A 6 A to 10 A | 0.025 % 0.025 % 0.025 % 0.025 % 0.025 % 0.025 % 0.025 % | Using AC/DC transfer technique. Other values within this range may also be calibrated but with an uncertainty of: 0.060 %. | |
| | 50 Hz 10 A to 100 A | 0.15 % + 20 mA | Using current shunt. | |
| | 50 Hz 100 A to 1000 A | 0.30 % + 250 mA | For the calibration of clamp meters and similar devices, using multi-turn coil technique. | |
| Specific Value | 50 Hz 100 A | 0.12 % | Using current shunt. | Lab |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|------------------------|---|--|------------------|
| FREQUENCY | 1 mHz to 10 kHz | 1.0 in 10^7 | Using off air receiver and counter timer | Lab |
| | 10 kHz to 100 MHz | 1.2 in 10^8 | | |
| Tachometer calibration | 10 RPM to 90000 RPM | 0.010 % + 0.010 RPM | Calibration of optical tachometers. | Lab |
| TIME INTERVAL | 1 ns to 100 ms | 4.0 ns | Additional uncertainties may be necessary depending on the specific characteristics of the input waveform. | Lab |
| | 100 ms to 1 s | 4.0 in 10^8 | | |
| | 1 s to 100 s | 1.4 in 10^8 | | |
| Timer and stopwatch calibrations | 10 s to 9 999.99 s | 0.15 s | Automatic triggering | Lab |
| | 10 s to 999.99 s | 0.050 s | Manual triggering | |
| | 999.99 s to 9 999.99 s | 0.075 s | Manual triggering | |
| CALIBRATIONS IN SUPPORT OF 17 TH AND 18 TH EDITION TEST EQUIPMENT | | | Using dedicated calibrator. | Lab |
| RESISTANCE | | | | |
| Continuity | 100 mΩ to 4.99 Ω | 0.30 % + 25 mΩ | 2 wire | |
| | 5 Ω to 29.9 Ω | 0.20 % + 25 mΩ | | |
| | 30 Ω to 199.9 Ω | 0.20 % + 25 mΩ | | |
| | 200 Ω to 499 Ω | 0.20 % | | |
| | 500 Ω to 1.999 kΩ | 0.20 % | | |
| | 2 kΩ to 4.99 kΩ | 0.20 % | | |
| | 5 kΩ to 10 kΩ | 0.20 % | | |
| | 100 mΩ to 4.99 Ω | 0.30 % + 10 mΩ | 4 wire | |
| | 5 Ω to 29.9 Ω | 0.20 % + 10 mΩ | | |
| | 30 Ω to 199.9 Ω | 0.20 % + 10 mΩ | | |
| | 200 Ω to 499 Ω | 0.20 % | | |
| | 500 Ω to 1.999 kΩ | 0.20 % | | |
| | 2 kΩ to 4.99 kΩ | 0.20 % | | |
| | 5 kΩ to 10 kΩ | 0.20 % | | |
| Insulation | 10 kΩ to 999.9 kΩ | 0.20 % | | |
| | 1 MΩ to 9.999 MΩ | 0.30 % | | |
| | 10 MΩ to 200 MΩ | 0.50 % | | |
| | 200 MΩ to 999.9 MΩ | 0.80 % + 6.0 kΩ | | |
| | 1 GΩ to 10 GΩ | 1.3 % | | |
| | 100 GΩ | 4.5 % | | |
| Terminal Voltage loaded | 0 V to 2 kV, >1 MΩ | 1.0 % + 5.0 V | | |
| | 0 V to 2 kV, <1 MΩ | 1.0 % + 2.0 V | | |
| Test Current | 1 μA to 9.9 mA | 1.5 % | | |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|---|---|------------------|
| CALIBRATIONS IN SUPPORT OF 17 TH AND 18 TH EDITION TEST EQUIPMENT (continued) | | | | |
| Earth Bond Resistance | | | | Lab |
| Nominal values | 25 mΩ, 50 mΩ and 100 mΩ | 5.0 mΩ | | |
| | 330 mΩ | 7.0 mΩ | | |
| | 500 mΩ | 8.0 mΩ | | |
| | 1 Ω | 10 mΩ | | |
| | 1.8 Ω | 18 mΩ | | |
| | 5 Ω | 30 mΩ | | |
| | 10 Ω | 60 mΩ | | |
| | 18 Ω | 100 mΩ | | |
| | 50 Ω | 300 mΩ | | |
| | 100 Ω | 500 mΩ | | |
| | 180 Ω | 1.0 Ω | | |
| | 500 Ω | 2.5 Ω | | |
| | 1 kΩ | 5.0 Ω | | |
| | 500 Ω | 10 Ω | | |
| Earth bond current | 25 mA to 30 A DC 20 Hz to 400 Hz AC | 1.6% + 2.0 mA | Offset may be increased to 0.70 A depending on the load | |
| Line/Loop Impedance | 25 mΩ, 50 mΩ and 100 mΩ | 5.0 mΩ | | |
| | 330 mΩ | 7.0 mΩ | | |
| | 500 mΩ | 8.0 mΩ | | |
| | 1 Ω | 10 mΩ | | |
| | 1.8 Ω | 18 mΩ | | |
| | 5 Ω | 30 mΩ | | |
| | 10 Ω | 60 mΩ | | |
| | 18 Ω | 100 mΩ | | |
| | 50 Ω | 300 mΩ | | |
| | 100 Ω | 500 mΩ | | |
| | 180 Ω | 1.0 Ω | | |
| | 500 Ω | 2.5 Ω | | |
| | 1 kΩ | 5.0 Ω | | |
| | 500 Ω | 10 Ω | | |
| Leakage Current | 30 mA to 100 mA | 0.40 % + 2.5 μA | | |
| RCD Current | 3 mA to 3000 mA 3 mA to 1500 mA 3 mA to 600 mA | 1.2 % 2.3 % 5.8 % | | |
| RCD Trip time | 10 ms to 5 s | 0.020 % + 0.25 ms | | |
| Line/Touch Voltage | Nominal 250 V | 5.0 % + 3.0 V | | |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|--|--|------------------|
| VOLTAGE | | | Using dedicated calibrator | Lab |
| AC Voltage | 40 Hz to 400 Hz 3 V to 30 V 30 V to 100 V 100 V to 300 V 300 V to 600 V | 0.10 % + 9.0 mV 0.10 % + 30 mV 0.10 % + 90 mV 0.10 % + 180 mV | Output values for calibration of measuring instruments | |
| DC Voltage | 3 V to 30 V 30 V to 150 V 150 V to 600 V | 0.10 % + 9.0 mV 0.10 % + 45 mV 0.10 % + 180 mV | | |
| FREQUENCY | 40 Hz to 400 Hz | 0.020 % | | |
| DC Voltage | 0 V to 10 V 10 V to 100 V 100 V to 1100 V | 0.16 % + 5.0 mV 0.20 % + 50 mV 0.20 % + 640 mV | For the measurement of electrical outputs. | |
| AC Voltage | 20 Hz to 400 Hz 0 V to 10 V (rms) 10 V to 100 V (rms) 100 V to 1 100 V (rms) | 0.18 % + 6.0 mV 0.20 % + 55 mV 0.20 % + 640 mV | | |
| CURRENT | | | | |
| DC Current | 0.1 mA to 300 mA 300 mA to 3 A 3 A to 30 A | 0.16 % + 0.15 mA 0.20 % + 1.5 mA 0.40 % + 15 mA | | |
| AC Current | 20 Hz to 400 Hz 0.1 mA to 300 mA 300 mA to 3 A 3 A to 30 A | 0.20 % + 0.17 mA 0.20 % + 1.7mA 0.40 % + 18 mA | | |
| APPARENT POWER | | | | |
| DC and 20 Hz to 400 Hz | 0 kVA to 33 kVA | The sum of the corresponding Voltage and Current uncertainties. | | |
| HiPoT LEAKAGE CURRENT | | | | |
| DC | 0 µA to 300 µA 300 µA to 3 mA 3 mA to 30 mA 30 mA to 300 mA | 0.40 % + 2.5 µA 0.40 % + 2.5 µA 0.40 % + 2.5 µA 0.20 % + 170 uA | | |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|--|---|------------------|
| HiPoT LEAKAGE CURRENT AC | 20 Hz to 400 Hz 2.5 μ A to 300 μ A 300 μ A to 3 mA 3 mA to 30 mA 30 mA to 300 mA | 0.40 % + 2.5 μ A 0.40 % + 2.5 μ A 0.40 % + 2.5 μ A 0.20 % + 170 μ A | | Lab |
| TEMPERATURE SIMULATION Temperature indicators and simulators: Calibration by electrical simulation | | | The temperature range of each thermocouple type is limited to values listed in the prevailing ITS 90 tables | |
| Base metal thermocouples K, J, T and E N | -200 °C to -100 °C -200 °C to -100 °C | 0.13 °C 0.15 °C | Excluding UUT internal reference junction. | Lab |
| K, J, T, E and N | -100 °C to +1300 °C | 0.060 °C | | |
| Noble Metal Thermocouples | 0 °C to 1760 °C | 0.25 °C | | |
| K, J, T and E N | -200 °C to -100 °C -200 °C to -100 °C | 0.15 °C 0.20 °C | Including UUT internal reference junction. | Lab |
| K, J, T, E and N | -100 °C to +1300 °C | 0.15 °C | | Lab |
| Noble Metal Thermocouples | 0 °C to 1760 °C | 0.70 °C | | Lab |
| Resistance thermometer (Pt 100) | -200 °C to 420 °C 420 °C to 850 °C | 0.015 °C 0.018 °C | Resistance method. | Lab |
| Calibration of internal reference junctions | 15 °C to 25 °C | 0.11 °C | The UUT will be stabilised to the prevailing ambient temperature. | Lab |
| Base metal thermocouple Depending on type | -200 °C to +1372 °C | 0.35 °C | Including cold junction compensation | Site |
| Noble metal thermocouple R and S type | -50 °C to +1786 °C | 0.50 °C | Including cold junction compensation | Site |
| Resistance thermometer (Pt 100) | -200 °C to +850 °C | 0.15 °C | Resistance method. | Site |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|--|------------------|
| DIMENSIONAL CALIBRATION by comparison with a reference instrument | | | | |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETERS, UNLESS OTHERWISE STATED | | | | |
| In addition to the items listed other similar items, including parts of measuring instruments and machines, may be calibrated in accordance with the stated uncertainties. Where the item or part calibrated is of lower quality due to wear, errors in geometry or form, poor surface texture, or where any other factor that adversely affects the measurement capability, greater uncertainties may be assigned. All linear calibrations can be also be provided in inch units. | | | | |
| LENGTH | | | | |
| Plain plug gauges (parallel) cylindrical setting standards, and rollers | Diameter: 0.1 to 10 10 to 150 150 to 200 200 to 250 250 to 300 | 0.50 1.0 1.3 1.6 1.9 | | Lab |
| Plain ring gauge (parallel) and setting standards | Diameter: 2 to 50 50 to 150 150 to 250 250 to 300 | 1.5 2.0 3.0 4.0 | | Lab |
| Length gauges, flat and spherical setting standards | 0 to 575 | 1.0 + (8.0 x length in m) | | Lab |
| Plain gap gauges (parallel) | As BS 969:2008 0.5 to 100 100 to 200 200 to 300 | 3.0 5.0 8.0 | | Lab |
| Screw plug gauges (parallel) including check and setting plugs | Diameter: 1 to 100 | Pitch diameter 3.0 | Single start, symmetrical thread forms only. | Lab |
| Screw ring gauges (parallel) | 2.5 to 100 100 to 150 | 5.0 7.0 | Single start, symmetrical thread forms only. | Lab |
| Screw thread pitch | 0.2 to 8 | 1.5 | | Lab |
| Screw thread flank angle | 0° to 5° | 5.0 minutes of arc | | Lab |
| Graduated rules | As BS 4372:1968 and above 0 to 2000 | 5.0 + (10 x length in m) | | Lab |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|---|---|---|------------------|
| LENGTH (continued) | | | | |
| Parallels | As BS 906:1972 5 to 50 x 100 x 400 | Dependent on size and grade 1.5 to 5.0 | | Lab |
| Receiver and position gauges, jigs, fixtures | Maximum dimensions 0 to 400 x 200 x 75 | Dependent on size and features Minimum per coordinate $3.0 + (20 \times \text{length in m})$ | | Lab |
| ANGLE | | | | |
| Sine bars and tables | As BS 3064:1978 100 to 300 | Linear dimensions $1.0 + (10 \times \text{length in m})$ Overall performance 3.0 seconds of arc. | | Lab |
| Angle plates and box angle plates | As BS 5535:1978 50 to 600 | Squareness: $3.0 + (1.0 \text{ per } 100 \text{ mm})$ Parallelism: $1.0 + (1.0 \text{ per } 100 \text{ mm})$ | The quoted CMCs are for the departure from flatness, straightness, parallelism, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration. | Lab |
| MEASURING INSTRUMENTS AND MACHINES | | | | |
| Micrometers External | As BS 870:2008 0 to 600 | Heads: 2.0 between any two points. Setting and extension rods: $1.0 + (8.0 \times \text{lengthin m})$ | | Lab |
| Internal | As BS 959:2008 0 to 1000 | | | Lab |
| Depth | As BS 6468:2008 0 to 300 | | | Lab |
| Height setting micrometer | 0 to 300 | Heads 1.5 Stepped column 2.5 Overall performance 3.0 | | Lab |
| Riser blocks for above | 150 300 | 2.5 5.0 | | Lab |
| Vernier gauges Caliper | As BS 887:2008 0 to 1000 | Overall performance: $10 + (30 \times \text{length in m})$ | | Lab |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|---|------------------|
| MEASURING INSTRUMENTS AND MACHINES (continued) | | | | |
| Calliper gauges including vernier, dial and digital types | ISO 13385-1 2019 Partial surface contact error (E) 0 to 1000 mm | 2 + (10 x length in m) | Calibration by comparison to length standards | Lab |
| | Shift error (S) internal jaws 3 to 50 mm | 4 | The stated uncertainty has been calculated in accordance With ISO 14253-5 and relates to the test value uncertainty. | |
| | Shift error (S) depth and step 3 to 50 mm | 4 | The uncertainty quoted Excludes contributions relating the instrument under test. | |
| Height | As ISO13225:2012 BS 1643:2008 (withdrawn) 0 to 1000 | Overall performance: 15 + (8.0 x length in m) | | Lab |
| Depth | As BS 6365:2008 0 to 600 | Overall performance: 10 + (30 x length in m) | | Lab |
| Dial gauges and dial test indicators | As BS 907:2008 and BS 2795:1981 0 to 50 | 1.0 | | Lab |
| Bench micrometers | As NPL MOY/SCMI 22 0 to 100 | Overall performance 2.0 | | Lab |
| Comparators (external) | As BS 1054:1975 250 to 10 000 magnifications | 1.0 % of range; minimum 0.20 | | Lab |
| Feeler gauges | As BS 957:2008 0.02 to 1 | 3.0 | | Lab |
| Bore Indicators | 0 to 150 diameter | Overall performance 5.0 | | Lab |
| ANCILLARY MEASUREMENTS | | | | |
| Flatness | | 0.7 | Ancillary measurements made for completeness of calibration. Best CMC's are dependent on methodology and range. | Lab |
| Parallelism | | 1.8 | | |
| Squareness | | 1.8 | | |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|---|------------------|
| TEMPERATURE calibration by comparison with a reference instrument | | | | |
| Resistance Thermometers | -95 °C to +140 °C 140 °C to 425 °C 425 °C to 660 °C | 0.067 °C 0.14 °C 0.19 °C | Calibrating by comparison in a Dry Block | Lab |
| | 0 °C | 0.020 °C | Ice point | Lab |
| Platinum Thermocouples | 0 °C to 140 °C 140 °C to 660 °C | 1.4 °C 1.1 °C | Calibrating by comparison in a Dry Block | Lab |
| | 300 °C to 1100 °C | 3.3 °C + 0.020 % of value | Calibrating by | |
| | 1100 °C to 1200 °C | 4.0 °C + 0.020 % of value | comparison in a furnace. | |
| Base Metal Thermocouples | -95 °C to -50 °C -50 °C to 0 °C 0 °C to 425 °C 425 °C to 660 °C | 0.75 °C 0.55 °C 0.40 °C 0.45 °C | Calibrating by comparison in a Dry Block. | Lab |
| | 300 °C to 1100 °C | 3.3 °C + 0.050 % of value | Calibrated by comparison in a furnace | |
| | 1100 °C to 1200 °C | 4.0 °C + 0.050 % of value | | |
| Electronic thermometers with sensors | Ranges as for above sensors | as for sensor | | Lab |
| Temperature controlled, incubators, ovens, environmental chambers, fridges/refrigerators, freezers and liquid baths. | -90 °C to -50 °C -50 °C to 0 °C 0 °C to 250 °C | 0.90 °C 0.75 °C 0.70 °C | Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping. | Site |
| Temperature indicators and recorders, load probes and monitoring thermometers. Autoclaves oven, liquid baths, freezers fridges etc | 0°C -50 °C to +140 °C 140 °C to 400 °C 400 °C to 650 °C 650 °C to 1050 °C | 0.020 °C 0.10 °C 0.15 °C 2.3 °C 3.6 °C | Calibrated within various dry block calibrators, ice point equipment or in a chamber. For a customer supplied environment the uncertainty will depend on the stability of the environment. | Site |
| Time Interval calibration by comparison | 0 s to 5400 s | 2.0 s | Relating to the timer functions of chambers and autoclaves. | Site |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|---|---|---------|------------------|
| RELATIVE HUMIDITY Relative humidity, calibration by comparison | 20 °C to 25 °C 26 %rh to 98 %rh 25 °C to 35 °C 20 %rh to 98 %rh 35 °C to 40 °C 11 %rh to 98 %rh 40 °C to 85 °C 8 %rh to 88 %rh | 3.4% of reading +0.50 %rh 3.3% of reading 3.3% of reading 3.1% of reading | | Site |
| Dew Point | 0 °C to 82 °C | 0.40 °C | | Site |
| Temperature measurement Associated with dew point meter | 0 °C to 82 °C | 0.40 °C | | Site |
| END | | | | |



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Southern Calibration Laboratories Limited

Issue No: 052 Issue date: 14 March 2025

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \cdot 0.01 \cdot q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$