# **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



# Locations covered by the organisation and their relevant activities

## Laboratory locations:

Location details		Activity	Location code
Address Standards Laboratory 14 Rose Way Blaby Leicester LE8 4BY	Local contact Hazel Horn Tel: +44 (0)116-275 0123 Fax: +44 (0)116-275 0262 Email: sales@uiscal.co.uk Website: www.uiscal.com	Dimensional Electrical Pressure Temperature Humidity Torque	Permanent Laboratory

## Site activities performed away from the locations listed above:

Location details	Activity	Location code
Calibrations may be performed in an air-conditioned vehicle taken to the customers' sites or in suitable areas within the customers' premises.	<u>Dimensional</u> <u>Electrical</u> <u>Pressure</u> <u>Temperature</u>	Site Ca
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.		alibration

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Ur Issı	niversal Instrument Serv ne No: 076 Issue date: 16	<b>vices Ltd</b> 6 May 2025	
	Calibration performed by the O	rganisation at the locations spe	cified	
	Calibration and Meas	urement Capability (CMC)		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty $(k = 2)$	Remarks	Location Code
DIMENSIONAL CALIBRATION				
RANGES IN MILLIM	ETRES AND UNCERTAINTIES IN	I MICROMETRES UNLESS STAT	ED OTHERWISE	
LENGTH Plain plug gauges (parallel) Cylindrical setting standards and rollers	0.2 to 10 10 to 50 50 to 100 100 to 150 150 to 200	1.0 1.0 1.5 2.0 2.0	By comparison with reference standards	
Plain ring gauges (parallel)	1.5 to 10 10 to 50 50 to 100 100 to 150 150 to 200	1.0 1.0 1.5 1.7 2.5		
Length gauges, flat and spherical ended	0 to 1000 Parallelism	1.0 + (8.0 x length in m) 0.5		
Measuring Instruments				υ
Micrometers External	BS 870:2008 - 0 to 600 Traverse of micrometer screw Flatness of anvils Parallelism of anvils Setting rods	2.0 between any points 0.20 0.25 1.0 + (8.0 x length in m)		ermanent Laboratory
Internal	BS 959:2008 0 to 900	points Setting and extension rods: 1.0 + (8.0 x length in m)		
Depth	BS 6468:2008 0 to 300	Heads: 2.0 between any points Setting and extension rods: 1.0 + (8.0 x length in m)		
Micrometer heads	BS 1734:1951 0 to 50 Squareness	1.0 0.5		
Bore micrometers (three points)	3 to 50 diameter 50 to 100 100 to 150	3.0 3.5 5.0	Using setting rings	

	So United K 2 Pine Trees,	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
0125 Accredited to ISO/IEC 17025:2017	Ur Issu	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	organisation at the locations spe	cified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code	
RANGES IN MILLIMETRES AND UNCERTAINTIES IN MICROMETRES UNLESS STATED OTHERWISE					
LENGTH (continued)					
Vernier/digital gauges Caliper,	BS 887:2008 - 0 to 1000 Linear error Flatness Parallelism Squareness Coplanar of jaws	10 + (30 x length in m) 2.5 3.5 5.0 15.0			
Height	As ISO 13225:2012 BS 1643:2008 0 to 1000	15 + (10 x length in m)			
Depth	BS 6365:2008 0 to 600	10 + (30 x length in m)			
Dial gauges and dial test indicators	BS 907:2008 0 to 50	1.0		Permar	
	BS 2795:1981 0 to 1	0.40		ient Lab	
Height setting micrometer	0 to 300	Heads 1.0 Stepped column 1.0 + (5.0 x length in m) Overall performance 3.0		oratory	
Riser blocks for above	150 300	2.5 3.0			
Protractors	BS 1685:2008 0° to 360°	2.0 minutes of arc 6.0 minutes of arc	1 minute of arc resolution instruments 5 minute of arc resolution		
Feeler gauges	BS 957:2008 0.02 to 1.00	3.0	instruments		
Receiver, position and profile gauges, jigs and fixtures	0 to 500 x 400 x 400 0° to 360°	3.0 + (10 x length in m) µm 2 minutes of arc See Note 1	Notes 1. Features and associated parts of these gauges / fixtures can be measured to the uncertainties given for equivalent items listed in this schedule.		
Profile projectors	10 to 100 magnifications Angular Linear 0 to 50 Linear 50 to 300	125 at the screen, 5.0 minutes of arc 2.5 6.0		Site Calibration	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Un Issu	iversal Instrument Serv le No: 076 Issue date: 16	<b>vices Ltd</b> 6 May 2025	
	Calibration performed by the O	rganisation at the locations spec	cified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
TORQUE CALIBRATION				
Wrenches and Screwdrivers	BS EN 6789-2:2017 0.1 N·m to 1000 N·m	1.0 %	The uncertainty quoted is for both the application of the calibration torque and the	
	BS EN 6789:2003 (withdrawn) 0.1 N⋅m to 1000 N⋅m	1.6 %	characteristics of the device being calibrated. Calibrations may also be given in lbf-in and lbf-ft.	
ELECTRICAL CALIBRATION				
DC RESISTANCE				
Measurement	0 $\Omega$ to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 200 $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$	25 μΩ/Ω + 20 μΩ 16 μΩ/Ω + 60 μΩ 15 μΩ/Ω + 0.60 mΩ 15 μΩ/Ω + 6.0 mΩ 16 μΩ/Ω + 60 mΩ 27 μΩ/Ω + 1.4 Ω 50 μΩ/Ω + 80 Ω 0.040 % + 9.0 kΩ	Using digital multimeter.	
Generation			Known values of DC	Perm
Specific Values	100 μΩ 1 mΩ, 10 mΩ and 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ	0.030 % 0.010 % 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 15 μΩ/Ω 15 μΩ/Ω 140 μΩ/Ω	resistance measuring instruments, in a two terminal or four terminal configuration, as appropriate.	anent Laboratory
	100 ΜΩ 1 GΩ 10 GΩ 100 GΩ 1 ΤΩ	0.050 % 0.50 % 0.50 % 2.0 % 3.0 %	Applied voltages in the range 100 V to 500 V	
	10 ΜΩ 100 ΜΩ 1 GΩ 10 GΩ	0.60 % 0.60 % 0.60 % 0.65 %	Applied voltages up to 5 kV	
DC VOLTAGE Voltage Reference Values	1 V 10 V	3.1 μV/V 2.5 μV/V	For calibration of voltage references of suitable stability.	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Un Issu	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	rganisation at the locations spe	cified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
DC VOLTAGE (continued)					
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V 1 kV to 15 kV 15 kV to 40 kV	12 μV/V + 1.0 μV 9.0 μV/V + 1.0 μV 8.0 μV/V + 8.0 μV 13 μV/V + 8.0 μV 13 μV/V + 0.80 mV 0.30 % + 1.0 V 0.30 % + 10 V	Using digital multimeter. Using high voltage divider. Generation of DC voltages for this range can also be undertaken.		
Generation	0 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V 220 V to 1100 V	12 μV/V + 1.0 μV 10 μV/V + 1.5 μV 10 μV/V + 10 μV 10 μV/V + 120 μV 12 μV/V + 0.80 mV	Known values of DC voltage for application to voltage measuring instruments.		
DC CURRENT					
Measurement	0 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 3 A 3 A to 10 A 10 A to 100 A	19 μA/A + 0.16 nA 20 μA/A + 13 nA 20 μA/A + 0.16 μA 60 μA/A + 13 μA 100 μA/A + 16 μA 80 μA/A + 160 μA	Using digital multimeter and current shunts	Permanent La	
Generation	10 μA to 220 μA 220 μA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 100 A	70 μA/A + 11 nA 70 μA/A + 12 nA 70 μA/A + 120 nA 100 μA/A + 1.2 μA 180 μA/A + 32 μA 300 μA/A + 21 mA	Known values of DC current for application to current measuring instruments.	boratory	
	100 A to 2 kA	0.030 % + 100 mA	Using multi-turn coil technique; for the calibration of clamp-on ammeters.		
AC RESISTANCE					
Specific Values	50 Hz 0.2 Ω, 0.5 Ω, 1 Ω, 5 Ω, 47 Ω, 100 Ω and 1 kΩ	0.10 % + 0.010 Ω	For the calibration of loop testers		
	1 Ω 100 Hz to 1 kHz 10 kHz	0.40 % 0.60 %	Resistance values available for the calibration of resistance measuring devices using four terminal		
	10 Ω 100 Hz to 1 kHz 10 kHz	0.25 % 0.40 %	pair connections.		

	Sc United K 2 Pine Trees,	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Un Issu				
	Calibration performed by the O	rganisation at the locations spe	cified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
AC RESISTANCE (continued)					
Specific values (continued)	100 Ω 100 Hz to 1 kHz 10 kHz	0.15 % 0.40 %			
	1 kΩ 100 Hz to 1 kHz 10 kHz	0.15 % 0.20 %			
	10 kΩ 100 Hz to 1 kHz 10 kHz	0.15 % 0.20 %			
	100 kΩ 100 Hz to 10 kHz	0.20 %			
	1 MΩ 100 Hz to 10 kHz	0.30 %			
	10 MΩ 100 Hz to 1 kHz	1.3 %		Perr	
Other Values	10 mΩ to 1.1 Ω 50 Hz	0.10 % + 0.50 mΩ		nanent L	
AC VOLTAGE				abora	
Measurement	20 mV to 200 mV 50 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	290 μV/V + 4.0 μV 530 μV/V + 8.0 μV 930 μV/V + 20 μV	Using digital multimeter.	tory	
	200 mV to 2 V 50 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	200 μV/V + 20 μV 300 μV/V + 40 μV 620 μV/V + 0.20 mV			
	2 V to 20 V 50 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	190 µV/V + 0.20 mV 300 µV/V + 0.40 mV 620 µV/V + 2.0 mV			
	20 V to 200 V 50 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	200 μV/V + 2.0 mV 310 μV/V + 4.0 mV 650 μV/V + 20 mV			
	200 V to 300 V 50 Hz to 10 kHz	200 μV/V + 20 mV			

	Scl United Ki 2 Pine Trees, Cl	nedule of Accredit issued by ngdom Accreditat nertsey Lane, Staines-upon-Tha	tation tion Service mes, TW18 3HR, UK	
O125 Accredited to ISO/IEC 17025:2017	Univ Issue	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025		
	Calibration performed by the Org	anisation at the locations speci	fied	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
AC VOLTAGE (continued)				
Measurement (continued)	50 Hz to 10 kHz 300 V to 1000 V 500 V 1000 V	0.15% + 20 mV 250 μV/V 500 μV/V		
	<i>At 50 Hz:</i> 1 kV to 5 kV 5 kV to 28 kV	1.3 % + 10 V 1.3 % + 40 V	Using high voltage divider. Generation of AC voltages over these ranges can also be undertaken.	
Generation	100 mV to 2.2 V 40 Hz to 20 kHz 20 kHz to 500 kHz	110 μV/V + 9.0 μV 0.14 % + 410 μV	Known values of AC voltage for application to voltage measuring instruments.	
	2.2 V to 22 V 40 Hz to 20 kHz 20 kHz to 500 kHz	110 μV/V + 82 μV 0.20 % + 6.0 mV		П
	22 V to 220 V 40 Hz to 20 kHz 20 kHz to 50 kHz	120 μV/V + 2.0 mV 0.030 % + 5.0 mV		<sup>9</sup> ermanent
	220 V to 1000 V <i>50 Hz to 1 kHz</i>	120 μV/V + 5.0 mV		Labora
AC CURRENT				atory
Measurement	20 μA to 1 mA 50 Hz to 1 kHz	0.040 % + 33 nA	Using digital multimeter and current shunts.	
	1 mA to 10 mA <i>50 Hz to 1 kHz</i>	0.040 % + 0.33 μA		
	10 mA to 100 mA 50 Hz to 1 kHz	0.030 % <b>+</b> 3.3 μA		
	100 mA to 3 A <i>50 Hz to 1 kHz</i>	0.030 % + 330 μA		
	3 A to 10 A 50 Hz to 1 kHz	0.030 % + 330 μA		
	10 A to 100 A 50 Hz to 400 Hz	0.030 % + 3.3 mA		

	Sc United K 2 Pine Trees, (	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Un Issu Calibration performed by the Or	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
AC CURRENT (continued)					
Generation	100 μA to 220 μA 50 Hz to 1 kHz	250 μΑ/Α + 20 nA	Known values of AC current for application to current measuring instruments.		
	50 Hz to 1 kHz	270 μA/A + 40 nA			
	2.2 mA to 22 mA 50 Hz to 1 kHz	200 µA/A + 410 nA			
	22 mA to 220 mA 50 Hz to 1 kHz	400 μA/A + 4.0 μA			
	220 mA to 2.2 A 50 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.080 % + 40 μA 0.090 % + 100 μA 1.1 % + 200 μA			
	2.2 A to 100 A 50 Hz to 400 Hz	0.050 %		Per	
	100 A to 1 kA <i>50 Hz</i>	0.20 % + 120 mA	Using multi turn coil technique; for the calibration of clamp-on ammeters and Rogowski coils.	manent Labo	
AC POWER	1 W to 100 kW 50 Hz to 400 Hz	0.15 %	Using phantom load technique. The canabilities are for	oratory	
	100 kW to 1 MW <i>50 Hz</i>	0.25 %	unity power factor only. Voltages will be in the range 10 V to 1000 V and currents will be in the range 100 mA to 1000 A. The applied frequency is limited to 50 Hz at currents above 100 A.		
CAPACITANCE					
Generation	<i>At 1 kHz</i> 100 pF 1 nF 10 nF 100 nF 1 μF	0.15 % 0.030 % 0.030 % 0.030 % 0.030 %	Known values of capacitance for application to capacitance measuring instruments.		
Measurement	<i>At 1 kHz, 1 V</i> 10 pF to 1 μF	0.68 % + 3 digits	Using capacitance meter with a 3½ digit display		

UKAS CALIBRATION 0125 Accredited to	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
ISO/IEC 17025:2017	Calibration nonformed by the Ora		ti1	
	Calibration performed by the Org	anisation at the locations speci		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
INDUCTANCE	<i>At 1 kHz</i> 10 μH 100 μH 1 mH 10 mH 100 mH 1 H	0.60 % 0.10 % 0.10 % 0.10 % 0.10 % 0.10 %	For the calibration of inductance measuring devices	
FREQUENCY				
Measurement	0.1 Hz to 2 GHz	2.0 in 10 <sup>6</sup>	Using counter timer. The periodic time of repetitive waveforms may be expressed in terms of 1/f or as events per unit time such as RPM.	F
Generation	0.01 Hz to 2.4 GHz	2.0 in 10 <sup>9</sup>	Using frequency synthesiser.	<sup>9</sup> ermaner
TIME INTERVAL				nt Lab
Single events	5 ms to 50 ms 50 ms to 2.2 s	0.20 ms 2.5 ms	For the calibration of RCD testers	oratory
Stopwatches (Mechanical)	5 s to 3 hours 3 hours to 30 hours	0.25 s 0.50 s	By comparison with calibrated counter timer.	
Temperature indicators, calibration by electrical simulation				
Base metal thermocouples	-200 °C to +1600 °C	0.20 °C	including cold junction compensation	
Noble metal thermocouples	-50 °C to +1760 °C	0.25 ℃	including cold junction compensation	
Resistance thermometers (Pt 100)	-200 °C to +800 °C	0.030 °C		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Un Issu	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	rganisation at the locations spe	cified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code	
DC RESISTANCE					
Measurement	0 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 300 MΩ	$\begin{array}{l} 120 \ \mu\Omega/\Omega + 4.6 \ m\Omega \\ 120 \ \mu\Omega/\Omega + 15 \ m\Omega \\ 120 \ \mu\Omega/\Omega + 150 \ m\Omega \\ 120 \ \mu\Omega/\Omega + 1.5 \ \Omega \\ 120 \ \mu\Omega/\Omega + 1.5 \ \Omega \\ 120 \ \mu\Omega/\Omega + 500 \ \Omega \\ 950 \ \mu\Omega/\Omega + 60 \ k\Omega \\ 2.1 \ \% + 0.10 \ M\Omega \end{array}$	Using digital multimeter.		
Generation					
Specific values	10 ΜΩ 100 ΜΩ 1 GΩ 10 GΩ	0.60 % 0.60 % 0.60 % 1.0 %	Applied voltages up to 5 kV	Site Ca	
Other values	0 $\Omega$ to 11 $\Omega$ 11 $\Omega$ to 33 $\Omega$ 33 $\Omega$ to 330 $\Omega$ 330 $\Omega$ to 3.3 k $\Omega$ 3.3 k $\Omega$ to 33 k $\Omega$ 3.3 k $\Omega$ to 330 k $\Omega$ 330 k $\Omega$ to 3.3 M $\Omega$ 3.3 M $\Omega$ to 11 M $\Omega$ 11 M $\Omega$ to 33 M $\Omega$ 33 M $\Omega$ to 110 M $\Omega$		Known values of DC resistance for application to resistance measuring instruments.	libration	
DC VOLTAGE					
Measurement	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	60 μV/V + 5.0 μV 50 μV/V + 10 μV 50 μV/V + 70 μV 60 μV/V + 0.80 mV 60 μV/V + 13 mV	Using digital multimeter.		

	Sc United K 2 Pine Trees, 0	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Un Issu				
	Calibration performed by the O	rganisation at the locations spe	cified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty $(k = 2)$	Remarks	Location Code	
DC VOLTAGE (continued)					
Generation	0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V	70 μV/V + 4.0 μV 60 μV/V + 6.0 μV 60 μV/V + 60 μV 65 μV/V + 0.60 mV 65 μV/V + 2.0 mV	Known values of DC voltage for application to voltage measuring instruments.		
DC CURRENT					
Measurement	0 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 3 A 3 A to 10 A 10 A to 100 A	0.015 % + 12 nA 0.010 % + 20 nA 0.010 % + 1.6 μA 0.010 % + 110 μA 0.012 % + 160 μA 0.012 % + 1.6 mA	Using digital multimeter.		
Generation	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 2.2 A 2.2 A to 11 A 11 A to 100 A	160 μA/A + 0.060 μA 130 μA/A + 0.31 μA 130 μA/A + 4.0 μA 400 μA/A + 45 μA 710 μA/A + 340 μA 300 μA/A + 21 mA	Known values of DC current for application to current measuring instruments.	Sit	
	100 A to 1 kA	0.030 % + 21 mA	Using multi-turn coil technique; for the calibration of clamp-on ammeters.	e Calibrati	
AC RESISTANCE	0.2 Ω to 1 kΩ 50 Hz	0.10 % + 0.50 mΩ	For the calibration of loop testers	on	
AC VOLTAGE					
Measurement	10 mV to 100 mV 45 Hz to 20 kHz 20 kHz to 50 kHz	0.080 % + 50 μV 0.30 % + 450 μV	Using digital multimeter.		
	100 mV to 1 V 45 Hz to 20 kHz 20 kHz to 50 kHz	0.080 % + 60 μV 0.25 % + 100 μV			
	1 V to 10 V 45 Hz to 20 kHz 20 kHz to 50 kHz	0.080 % + 4.0 mV 0.14 % + 6.0 mV			
	10 V to 100 V 45 Hz to 20 kHz 20 kHz to 50 kHz	0.080 % + 35 mV 0.25 % + 210 mV			
	200 V to 750 V 45 Hz to 1 kHz	0.080 % + 250 mV			

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
O125 Accredited to ISO/IEC 17025:2017	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	rganisation at the locations spe	cified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
AC VOLTAGE (continued)				
Generation	3 mV to 33 mV 50 Hz to 10 kHz	0.18 % + 24 μV	Known values of AC voltage for application to voltage measuring instruments.	
	33 mV to 330 mV 50 Hz to 10 kHz	0.065 % + 24 μV		
	330 mV to 3.3 V 50 Hz to 10 kHz 10 kHz to 50 kHz	0.045 % + 80 μV 0.18 % + 0.40 mV		
	3.3 V to 33 V 50 Hz to 10 kHz 10 kHz to 50 kHz	0.055 % + 0.80 mV 0.23 % + 7.0 mV		
	33 V to 330 V 50 Hz to 1 kHz 1 kHz to 10 kHz	0.065 % + 9.0 mV 0.10 % + 18 mV		
	330 V to 1020 V 50 Hz to 1 kHz	0.065 % + 90 mV		
AC CURRENT				Site (
Measurement	50 Hz to 100 Hz 3 mA to 30 mA 30 mA to 100 mA 100 mA to 1 A 1 A to 3 A 100 mA to 2 A	0.60 % + 12 μA 0.60 % + 120 μA 0.19 % + 0.46 mA 0.24 % + 0.66 mA	Using digital multimeter.	Calibration
	50 Hz to 100 Hz 100 Hz to 1 kHz	0.50 % + 1.0 mA 0.50 % + 1.0 mA		
	2 A to 10 A 50 Hz to 60 Hz	1.1 % + 12 mA		
	10 A to 100 A 50 Hz to 60 Hz	0.13 % + 40 mA		
Generation	0.3 mA to 3.3 mA 50 Hz to 1 kHz	0.14 % + 0.32 μA	Known values of AC current for application to current measuring instruments.	
	3.3 mA to 33 mA 50 Hz to 1 kHz	0.13 % + 3.2 μA		
	33 mA to 330 mA 50 Hz to 1 kHz	0.13 % + 32 μA		
	330 mA to 2.2 A 50 Hz to 1 kHz	0.14 % + 320 μA		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	rganisation at the locations spec	cified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
AC CURRENT (continued)				
Generation (continued)	2.2 A to 11 A 50 Hz to 400 Hz	0.14 % + 2.1 mA		
	11 A to 100 A 50 Hz to 400 Hz	0.30 %		
	100 A to 1 kA <i>50 Hz</i>	0.20 % + 120 mA	Using multi turn coil technique; for the calibration of clamp-on ammeters and Rogowski coils.	
FREQUENCY				
Measurement and generation	0.1 Hz to 160 MHz	2 in 10 <sup>6</sup>	Using counter timer. The periodic time of repetitive waveforms may be expressed in terms of 1/f or as events per unit time such as RPM.	Site C
Generation	0.02 Hz to 2.1 GHz	2 in 10 <sup>9</sup>	Using Signal Generator	alibrat
TIME INTERVAL				tion
Single events	5 ms to 50 ms 50 ms to 2.2 s	0.20 ms 2.5 ms	For the calibration of RCD testers	
Stopwatches (Mechanical)	0 s to 3 hours 3 hours to 30 hours	0.25 s 0.50 s	By comparison with calibrated counter timer.	
Temperature indicators, calibration by electrical simulation				
Base metal thermocouples	-200 °C to +1600 °C	0.57 °C	Including cold junction compensation	
Noble metal thermocouples	-50 °C to +1760 °C	1.4 °C	Including cold junction compensation	
Resistance thermometers (Pt 100)	-200 °C to 800 °C	0.15 °C		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025 Calibration performed by the Organisation at the locations specified			
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
PRESSURE CALIBRATIONGas pressure (gauge)Calibration of pressure indicating instruments and gauges.Gas pressure (absolute)Calibration of pressure indication instruments and gauges.Hydraulic pressure (gauge)Calibration of pressure indicating instruments and gauges.Hydraulic pressure (absolute)Calibration of pressure indicating instruments and gauges.Hydraulic pressure (absolute)Calibration of pressure indicating instruments and gauges.	-100 kPa to -7 kPa -7 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa 2.5 kPa to 7 kPa 7 kPa to 70 kPa 70 kPa to 70 kPa 700 kPa to 4.3 MPa 4.3 MPa to 6 MPa 6 MPa to 10 MPa 30 Pa to 75 kPa 75 kPa to 120 kPa 120 kPa to 170 kPa 170 kPa to 800 kPa 800 kPa to 4.3 MPa 0 kPa to 54 kPa 54 kPa to 60 kPa 60 kPa to 600 kPa 6 MPa to 120 MPa	$\begin{array}{l} 58 \ Pa/MPa + 13 \ Pa \\ 68 \ Pa/MPa + 5.1 \ Pa \\ 0.010 \ \% + 2.0 \ Pa \\ 68 \ Pa/MPa + 5.1 \ Pa \\ 58 \ Pa/MPa + 13 \ Pa \\ 66 \ Pa/MPa \\ 0.010 \ \% + 33 \ Pa \\ 0.12 \ \% + 4.2 \ kPa \\ 0.12 \ \% + 4.0 \ kPa \\ \hline \\ 58 \ Pa/MPa + 17 \ Pa \\ 10 \ Pa \\ 58 \ Pa/MPa + 17 \ Pa \\ 10 \ Pa \\ 58 \ Pa/MPa + 10 \ Pa \\ 0.010 \ \% + 35 \ Pa \\ \hline \\ 0.030 \ \% + 110 \ Pa \\ 0.032 \ \% + 20 \ Pa \\ 0.0066 \ \% + 29 \ Pa \\ 0.010 \ \% \\ \hline \\ 0.010 \ \% \\ \hline \\ 0.00066 \ \% + 32 \ Pa \\ 0.010 \ \% + 12 \ Pa \\ \hline \\ 0.00066 \ \% + 12 \ Pa \\ 0.010 \ \% + 12 \ Pa \\ \hline \end{array}$	Methods consistent with EURAMET CG17. Calibrations of devices with an electrical output may be undertaken	Permanent Laboratory
Gas pressure (gauge) Calibration of pressure indicating instruments and gauges. Hydraulic pressure (gauge) Calibration of pressure indicating instruments and gauges.	-100 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa 2.5 kPa to 200 kPa 200 kPa to 2 MPa 2 MPa to 7 MPa 0 MPa to 20 MPa 20 MPa to 70 MPa	0.005 8 % + 0.22 kPa 0.010 % + 2.0 Pa 96 Pa 0.010% + 0.80 kPa 0.010% + 7.0 kPa 0.010% + 5.8 kPa 0.050 % + 22 kPa	Absolute pressures within these ranges can also be generated, subject to an additional uncertainty of 90 Pa.	Site Calibration

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	rganisation at the locations spec	cified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
TEMPERATURE CALIBRATION				
Liquid-in-glass thermometers	Ice-point (0 °C) -80 °C to +5 °C 5 °C to 50 °C 50 °C to 250 °C	0.013 °C 0.020 °C 0.016 °C 0.025 °C	By comparison with reference standards Calibration performed within liquid baths	
Resistance thermometers	Nitrogen boiling point (-196 °C) lce point (0 °C) Triple point of water (0.01 °C) -80 °C to 5 °C 5 °C to 50 °C 50 °C to 250 °C 250 °C to 660 °C	0.013 °C 0.012 °C 0.005 °C 0.010 °C 0.015 °C 0.014 °C 0.095 °C	-196 °C & 0 °C FP by comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	
Platinum thermocouples	-50 °C to 0 °C 0 °C to 80 °C 80 °C to 250 °C 250 °C to 650 °C 650 °C to 1100 °C	1.3 °C 0.67 °C 0.52 °C 0.54 °C 1.6 °C	By comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	
Base metal thermocouples	Nitrogen boiling point (-196 °C) -80 °C to 0 °C 0 °C to 250 °C 250 °C to 650 °C 650 °C to 1100 °C	0.23 °C 0.18 °C 0.17 °C 0.19 °C 1.6 °C	-196 °C & 0 °C FP by comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	Permanent L
Electronic Thermometers with probes	Nitrogen boiling point (-196 °C) -80 °C to +250°C Ice point (0 °C) Triple point of water (0.01 °C) 250 °C to 650 °C 650 °C to 1100 °C	0.014 °C 0.013 °C 0.014 °C 0.007 0 °C 0.12 °C 1.7 °C	-196 °C & 0 °C FP by comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	aboratory
Calibration of Temperature Data loggers, Electronic Temperature Indicators and Recorders in Air	-40 °C to +80 °C	0.47 °C	Calibration within air chamber	
Block calibrators	-50 °C to +660 °C	0.12 °C	By comparison with reference standards	
Cold Junction Compensation	Ice-point (0 °C) 15 °C to 25 °C	0.13 °C 0.17 °C	In stirred liquid bath	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	Universal Instrument Services Ltd Issue No: 076 Issue date: 16 May 2025			
	Calibration performed by the O	rganisation at the locations spe	cified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty $(k = 2)$	Remarks	Location Code
HUMIDITY CALIBRATION	In the temperature range $0 \circ C$ to 20 $\circ C$ $5 \circ rh$ to 50 $\circ rh$ $50 \circ rh$ to 90 $\circ rh$ In the temperature range $20 \circ C$ to 40 $\circ C$ $5 \circ rh$ to 95 $\circ rh$ In the temperature range $40 \circ C$ to 60 $\circ C$ $5 \circ rh$ to 90 $\circ rh$ $0 \circ C$ to 60 $\circ C$	2.0 %rh 2.5 %rh 2.5 %rh 2.5 %rh 0.20 °C	By comparison with reference in controlled chamber Calibrations of devices with an electrical output may be undertaken	Permanent Laboratory
TEMPERATURE CALIBRATION				
Temperature sensors with indicators	-20 °C to 120 °C -25 °C to +140 °C 140 °C to 350 °C Ice-point (0 °C)	0.089 °C 0.061 °C 0.15 °C 0.041 °C	By comparison with reference standards In stirred liquid baths By comparison with reference standards within metal block baths Calibration performed within	Site Cali
Liquid baths			liquid baths	bratio
Temperature measurement Gradient measurement	-40 °C to 250 °C -40 °C to 250 °C	0.044 °C 0.020 °C	By comparison with reference standards	ō
Temperature controlled chambers, ovens, fridges/refrigerators and freezers	-80 °C to +300 °C 300 °C to 1100 °C	0.84 °C 3.7 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	
END				



## Appendix - Calibration and Measurement Capabilities

## Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$