Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

UKAS CALIBRATION 0125 Accredited to ISO/IEC 17025:2017	**Universal Instrument Services Ltd** **Issue No:** 078 **Issue date:** 10 May 2026	
	Standards Laboratory **14 Rose Way** **Blaby** **Leicester** **LE8 4BY**	**Contact: Hazel Horn** **Tel: +44 (0)116 275 0123** **Fax: +44 (0)116 275 0262** **E-Mail: sales@uiscal.co.uk** **Website: www.uiscal.com**

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Standards Laboratory 14 Rose Way Blaby Leicester LE8 4BY	**Local contact** Hazel Horn Tel: +44 (0)116-275 0123 Fax: +44 (0)116-275 0262 Email: sales@uiscal.co.uk Website: www.uiscal.com	Dimensional Electrical Pressure Temperature Humidity Torque	Permanent Laboratory

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Calibrations may be performed in an air-conditioned vehicle taken to the customers' sites or in suitable areas within the customers' premises. The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Contact as above	Dimensional Electrical Pressure Temperature	Site Calibration

0125 Accredited to ISO/IEC 17025:2017	**Schedule of Accreditation** **issued by** **United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	Universal Instrument Services Ltd **Issue No:** 078 **Issue date:** 10 May 2026

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
DIMENSIONAL CALIBRATION				Permanent Laboratory
RANGES IN MILLIMETRES AND UNCERTAINTIES IN MICROMETRES UNLESS STATED OTHERWISE				
LENGTH				
Plain plug gauges (parallel) Cylindrical setting standards and rollers	0.2 to 10 10 to 50 50 to 100 100 to 150 150 to 200	1.0 1.0 1.5 2.0 2.0	By comparison with reference standards	
Plain ring gauges (parallel)	1.5 to 10 10 to 50 50 to 100 100 to 150 150 to 200	1.0 1.0 1.5 1.7 2.5		
Length gauges, flat and spherical ended	0 to 1000 Parallelism	Q[1.0, (8.0 x length in m)] 0.5		
Measuring Instruments				
Micrometers External	BS 870:2008 - 0 to 600 Traverse of micrometer screw Flatness of anvils Parallelism of anvils Setting rods	 2.0 between any points 0.20 0.25 Q[1.0, (8.0 x length in m)]		
Internal	BS 959:2008 0 to 900	Heads: 2.0 between any points Setting and extension rods: Q[1.0, (8.0 x length in m)]		
Depth	BS 6468:2008 0 to 300	Heads: 2.0 between any points Setting and extension rods: Q[1.0 + (8.0 x length in m)]		
Micrometer heads	BS 1734:1951 0 to 50 Squareness	 1.0 0.5		
Bore micrometers (three points)	3 to 50 diameter 50 to 100 100 to 150	3.0 3.5 5.0	Using setting rings	

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
RANGES IN MILLIMETRES AND UNCERTAINTIES IN MICROMETRES UNLESS STATED OTHERWISE				Permanent Laboratory
LENGTH (continued)				
Vernier/digital gauges Caliper,	BS 887:2008 - 0 to 1000 Linear error Flatness Parallelism Squareness Coplanar of jaws	 Q[10, (30 x length in m)] 2.5 3.5 5.0 15.0		
Height	As ISO 13225:2012 BS 1643:2008 0 to 1000	 Q[15, (10 x length in m)]		
Depth	BS 6365:2008 0 to 600	 Q[10, (30 x length in m)]		
Dial gauges and dial test indicators	BS 907:2008 0 to 50	 1.0		
	BS 2795:1981 0 to 1	 0.40		
Height setting micrometer	0 to 300	Heads 1.0 Stepped column Q[1.0, (5.0 x length in m)] Overall performance 3.0		
Riser blocks for above	150 300	2.5 3.0		
Protractors	BS 1685:2008 0º to 360º	2.0 minutes of arc 6.0 minutes of arc	1 minute of arc resolution instruments 5 minute of arc resolution instruments	
Feeler gauges	BS 957:2008 0.02 to 1.00	3.0		
Receiver, position and profile gauges, jigs and fixtures	0 to 500 x 400 x 400 0º to 360º	Q[3.0, (10 x length in m)] µm 2 minutes of arc See Note 1	Notes 1. Features and associated parts of these gauges / fixtures can be measured to the uncertainties given for equivalent items listed in this schedule.	
Profile projectors	10 to 100 magnifications Angular Linear 0 to 50 Linear 50 to 300	125 at the screen, 5.0 minutes of arc 2.5 6.0		Site Calibration

0125 Accredited to ISO/IEC 17025:2017	**Schedule of Accreditation** **issued by** **United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	Universal Instrument Services Ltd **Issue No:** 078 **Issue date:** 10 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
TORQUE CALIBRATION				Permanent Laboratory
Wrenches and Screwdrivers	BS EN 6789-2:2017 0.1 N·m to 1000 N·m	1.0 %	The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated. Calibrations may also be given in lbf·in and lbf·ft.	
	BS EN 6789:2003 (withdrawn) 0.1 N·m to 1000 N·m	1.6 %		
ELECTRICAL CALIBRATION				
DC RESISTANCE				
Measurement	0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ	Q[25 μΩ/Ω, 20 μΩ] Q[16 μΩ/Ω, 60 μΩ] Q[15 μΩ/Ω, 0.60 mΩ] Q[15 μΩ/Ω, 6.0 mΩ] Q[16 μΩ/Ω, 60 mΩ] Q[27 μΩ/Ω, 1.4 Ω] Q[50 μΩ/Ω, 80 Ω] Q[0.040 %, 9.0 kΩ]	Using digital multimeter.	
Generation			Known values of DC resistance for application to resistance measuring instruments, in a two terminal or four terminal configuration, as appropriate.	
Specific Values	100 μΩ 1 mΩ, 10 mΩ and 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ	0.030 % 0.010 % 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 6.0 μΩ/Ω 15 μΩ/Ω 15 μΩ/Ω 140 μΩ/Ω		
	100 MΩ 1 GΩ 10 GΩ 100 GΩ 1 TΩ	0.050 % 0.50 % 0.50 % 2.0 % 3.0 %	Applied voltages in the range 100 V to 500 V	
	10 MΩ 100 MΩ 1 GΩ 10 GΩ	0.60 % 0.60 % 0.60 % 0.65 %	Applied voltages up to 5 kV	
DC VOLTAGE Voltage Reference Values	1 V 10 V	3.1 μV/V 2.5 μV/V	For calibration of voltage references of suitable stability.	

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
DC VOLTAGE (continued)				Permanent Laboratory
Measurement	0 mV to 200 mV	Q[12 µV/V, 1.0 µV]	Using digital multimeter.	
	200 mV to 2 V	Q[9.0 µV/V, 1.0 µV]		
	2 V to 20 V	Q[8.0 µV/V, 8.0 µV]		
	20 V to 200 V	Q[13 µV/V, 80 µV]		
	200 V to 1000 V	Q[13 µV/V, 0.80 mV]		
	1 kV to 15 kV	Q[0.30 %, 1.0 V]	Using high voltage divider. Generation of DC voltages for this range can also be undertaken.	
	15 kV to 40 kV	Q[0.30 %, 10 V]		
Generation	0 mV to 220 mV	Q[12 µV/V, 1.0 µV]	Known values of DC voltage for application to voltage measuring instruments.	
	220 mV to 2.2 V	Q[10 µV/V, 1.5 µV]		
	2.2 V to 22 V	Q[10 µV/V, 10 µV]		
	22 V to 220 V	Q[10 µV/V, 120 µV]		
	220 V to 1100 V	Q[12 µV/V, 0.80 mV]		
DC CURRENT				
Measurement	0 mA to 1 mA	Q[19 µA/A, 0.16 nA]	Using digital multimeter and current shunts	
	1 mA to 10 mA	Q[20 µA/A, 13 nA]		
	10 mA to 100 mA	Q[20 µA/A, 0.16 µA]		
	100 mA to 3 A	Q[60 µA/A, 13 µA]		
	3 A to 10 A	Q[100 µA/A, 16 µA]		
	10 A to 100 A	Q[80 µA/A, 160 µA]		
Generation	10 µA to 220 µA	Q[70 µA/A, 11 nA]	Known values of DC current for application to current measuring instruments.	
	220 µA to 2.2 mA	Q[70 µA/A, 12 nA]		
	2.2 mA to 22 mA	Q[70 µA/A, 120 nA]		
	22 mA to 220 mA	Q[100 µA/A, 1.2 µA]		
	220 mA to 2.2 A	Q[180 µA/A, 32 µA]		
	2.2 A to 100 A	Q[300 µA/A, 21 mA]		
	100 A to 2 kA	Q[0.030 %, 100 mA]	Using multi-turn coil technique; for the calibration of clamp-on ammeters.	
AC RESISTANCE				
Specific Values	*50 Hz* 0.2 Ω, 0.5 Ω, 1 Ω, 5 Ω, 47 Ω, 100 Ω and 1 kΩ	Q[0.10 %, 0.010 Ω]	For the calibration of loop testers	
	1 Ω *100 Hz to 1 kHz*	0.40 %	Resistance values available for the calibration of resistance measuring devices using four terminal pair connections.	
	10 kHz	0.60 %		
	10 Ω *100 Hz to 1 kHz*	0.25 %		
	10 kHz	0.40 %		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
AC RESISTANCE (continued)				Permanent Laboratory
Specific values (continued)	100 Ω			
	100 Hz to 1 kHz	0.15 %		
	10 kHz	0.40 %		
	1 kΩ			
	100 Hz to 1 kHz	0.15 %		
	10 kHz	0.20 %		
	10 kΩ			
	100 Hz to 1 kHz	0.15 %		
	10 kHz	0.20 %		
	100 kΩ			
	100 Hz to 10 kHz	0.20 %		
	1 MΩ			
	100 Hz to 10 kHz	0.30 %		
	10 MΩ			
	100 Hz to 1 kHz	1.3 %		
Other Values	10 mΩ to 1.1 Ω			
	50 Hz	Q[0.10 % + 0.50 mΩ]		
AC VOLTAGE				
Measurement	20 mV to 200 mV		Using digital multimeter.	
	50 Hz to 10 kHz	Q[290 μV/V, 4.0 μV]		
	10 kHz to 30 kHz	Q[530 μV/V, 8.0 μV]		
	30 kHz to 100 kHz	Q[930 μV/V, 20 μV]		
	200 mV to 2 V			
	50 Hz to 10 kHz	Q[200 μV/V, 20 μV]		
	10 kHz to 30 kHz	Q[300 μV/V, 40 μV]		
	30 kHz to 100 kHz	Q[620 μV/V, 0.20 mV]		
	2 V to 20 V			
	50 Hz to 10 kHz	Q[190 μV/V, 0.20 mV]		
	10 kHz to 30 kHz	Q[300 μV/V, 0.40 mV]		
	30 kHz to 100 kHz	Q[620 μV/V, 2.0 mV]		
	20 V to 200 V			
	50 Hz to 10 kHz	Q[200 μV/V, 2.0 mV]		
	10 kHz to 30 kHz	Q[310 μV/V, 4.0 mV]		
	30 kHz to 100 kHz	Q[650 μV/V, 20 mV]		
	200 V to 300 V			
	50 Hz to 10 kHz	Q[200 μV/V, 20 mV]		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
AC VOLTAGE (continued)				Permanent Laboratory
Measurement (continued)	*50 Hz to 10 kHz*			
	300 V to 1000 V	Q[0.15%, 20 mV]		
	500 V	250 µV/V		
	1000 V	500 µV/V		
	At 50 Hz:		Using high voltage divider. Generation of AC voltages over these ranges can also be undertaken.	
	1 kV to 5 kV	Q[1.3 %, 10 V		
	5 kV to 28 kV	Q[1.3 %, 40 V		
Generation	100 mV to 2.2 V		Known values of AC voltage for application to voltage measuring instruments.	
	40 Hz to 20 kHz	Q[110 µV/V, 9.0 µV]		
	20 kHz to 500 kHz	Q[0.14 %, 410 µV]		
	2.2 V to 22 V			
	40 Hz to 20 kHz	Q[110 µV/V, 82 µV]		
	20 kHz to 500 kHz	Q[0.20 %, 6.0 mV]		
	22 V to 220 V			
	40 Hz to 20 kHz	Q[120 µV/V, 2.0 mV]		
	20 kHz to 50 kHz	Q[0.030 %, 5.0 mV]		
	220 V to 1000 V			
	50 Hz to 1 kHz	Q[120 µV/V, 5.0 mV]		
AC CURRENT				
Measurement	20 µA to 1 mA		Using digital multimeter and current shunts.	
	50 Hz to 1 kHz	Q[0.040 %, 33 nA]		
	1 mA to 10 mA			
	50 Hz to 1 kHz	Q[0.040 %, 0.33 µA]		
	10 mA to 100 mA			
	50 Hz to 1 kHz	Q[0.030 %, 3.3 µA]		
	100 mA to 3 A			
	50 Hz to 1 kHz	Q[0.030 %, 330 µA]		
	3 A to 10 A			
	50 Hz to 1 kHz	Q[0.030 %, 330 µA]		
	10 A to 100 A			
	50 Hz to 400 Hz	Q[0.030 %, 3.3 mA]		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
AC CURRENT (continued)				Permanent Laboratory
Generation	100 μA to 220 μA *50 Hz to 1 kHz*	Q[250 μA/A, 20 nA]	Known values of AC current for application to current measuring instruments.	
	220 μA to 2.2 mA *50 Hz to 1 kHz*	Q[270 μA/A, 40 nA]		
	2.2 mA to 22 mA *50 Hz to 1 kHz*	Q[200 μA/A, 410 nA]		
	22 mA to 220 mA *50 Hz to 1 kHz*	Q[400 μA/A, 4.0 μA]		
	220 mA to 2.2 A *50 Hz to 1 kHz* *1 kHz to 5 kHz* *5 kHz to 10 kHz*	 Q[0.080 %, 40 μA] Q[0.090 %, 100 μA] Q[1.1 %, 200 μA]		
	2.2 A to 100 A *50 Hz to 400 Hz*	0.050 %		
	100 A to 1 kA *50 Hz*	Q[0.20 %, 120 mA]	Using multi turn coil technique; for the calibration of clamp-on ammeters and Rogowski coils.	
AC POWER	1 W to 100 kW *50 Hz to 400 Hz*	0.15 %	Using phantom load technique. The capabilities are for unity power factor only. Voltages will be in the range 10 V to 1000 V and currents will be in the range 100 mA to 1000 A. The applied frequency is limited to 50 Hz at currents above 100 A.	
	100 kW to 1 MW *50 Hz*	0.25 %		
CAPACITANCE				
Generation	*At 1 kHz* 100 pF 1 nF 10 nF 100 nF 1 μF	 0.15 % 0.030 % 0.030 % 0.030 % 0.030 %	Known values of capacitance for application to capacitance measuring instruments.	
Measurement	*At 1 kHz, 1 V* 10 pF to 1 μF	Q[0.68 %, 3 digits]	Using capacitance meter with a 3½ digit display	

0125 Accredited to ISO/IEC 17025:2017	**Schedule of Accreditation** **issued by** **United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	Universal Instrument Services Ltd **Issue No:** 078 **Issue date:** 10 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
INDUCTANCE	*At 1 kHz* 10 μH 100 μH 1 mH 10 mH 100 mH 1 H	 0.60 % 0.10 % 0.10 % 0.10 % 0.10 % 0.10 %	For the calibration of inductance measuring devices	Permanent Laboratory
FREQUENCY				
Measurement	0.1 Hz to 2 GHz	2.0 in 10^6	Using counter timer. The periodic time of repetitive waveforms may be expressed in terms of 1/f or as events per unit time such as RPM.	
Generation	0.01 Hz to 2.4 GHz	2.0 in 10^9	Using frequency synthesiser.	
TIME INTERVAL				
Single events	5 ms to 50 ms 50 ms to 2.2 s	0.20 ms 2.5 ms	For the calibration of RCD testers	
Stopwatches (Mechanical)	5 s to 3 hours 3 hours to 30 hours	0.25 s 0.50 s	By comparison with calibrated counter timer.	
Temperature indicators, calibration by electrical simulation				
Base metal thermocouples	-200 °C to +1600 °C	0.20 °C	including cold junction compensation	
Noble metal thermocouples	-50 °C to +1760 °C	0.25 °C	including cold junction compensation	
Resistance thermometers (Pt 100)	-200 °C to +800 °C	0.030 °C		

0125 Accredited to ISO/IEC 17025:2017	**Schedule of Accreditation** **issued by** **United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	Universal Instrument Services Ltd **Issue No:** 078 **Issue date:** 10 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
DC RESISTANCE				Site Calibration
Measurement	0 Ω to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ	Q[85 μΩ/Ω, 0.20 mΩ] Q[85 μΩ/Ω, 0.30 mΩ] Q[85 μΩ/Ω, 2.5 mΩ] Q[75 μΩ/Ω, 12 mΩ] Q[75 μΩ/Ω, 0.12 Ω] Q[75 μΩ/Ω, 1.6 Ω] Q[100 μΩ/Ω, 21 Ω] Q[400 μΩ/Ω, 120 Ω] Q[0.20 %, 4.0 kΩ]	Using digital multimeter.	
Generation				
Specific values	10 MΩ 100 MΩ 1 GΩ 10 GΩ	0.60 % 0.60 % 0.60 % 1.0 %	Applied voltages up to 5 kV	
Other values	0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 330 Ω 330 Ω to 3.3 kΩ 3.3 kΩ to 33 kΩ 33 kΩ to 330 kΩ 330 kΩ to 3.3 MΩ 3.3 MΩ to 11 MΩ 11 MΩ to 33 MΩ 33 MΩ to 110 MΩ	Q[150 μΩ/Ω, 6.2 mΩ] Q[160 μΩ/Ω, 12 mΩ] Q[120 μΩ/Ω, 12 mΩ] Q[120 μΩ/Ω, 72 mΩ] Q[120 μΩ/Ω, 0.72 Ω] Q[150 μΩ/Ω, 7.2 Ω] Q[180 μΩ/Ω, 67 Ω] Q[650 μΩ/Ω, 670 Ω] Q[0.13%, 10 kΩ] Q[0.55%, 11 kΩ]	Known values of DC resistance for application to resistance measuring instruments.	
DC VOLTAGE				
Measurement	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	Q[30 μV/V, 5.0 μV] Q[25 μV/V, 15 μV] Q[25 μV/V, 0.13 mV] Q[40 μV/V, 1.5 mV] Q[40 μV/V, 16 mV]	Using digital multimeter.	

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
DC VOLTAGE (continued)				Site Calibration
Generation	0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V	Q[70 µV/V, 4.0 µV] Q[60 µV/V, 6.0 µV] Q[60 µV/V, 60 µV] Q[65 µV/V, 0.60 mV] Q[65 µV/V, 2.0 mV]	Known values of DC voltage for application to voltage measuring instruments.	
DC CURRENT				
Measurement	0 mA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 10 A	Q[0.045 %, 20 nA] Q[0.045 %, 100 nA] Q[0.020 %, 1.0 µA] Q[0.020 %, 13 µA] Q[0.040 %, 220 µA] Q[0.050 %, 500 µA] Q[0.22 %, 5.0 mA]	Using digital multimeter.	
	10 A to 20 A 20 A to 100 A	Q[0.030 %, 1.0 mA] Q[0.060 %, 10 mA]	Using active shunt	
Generation	0 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 2.2 A 2.2 A to 11 A 11 A to 100 A	Q[160 µA/A, 0.060 µA] Q[130 µA/A, 0.31 µA] Q[130 µA/A, 4.0 µA] Q[400 µA/A, 45 µA] Q[710 µA/A, 340 µA] Q[300 µA/A, 21 mA]	Known values of DC current for application to current measuring instruments.	
	100 A to 1 kA	Q[0.030 %, 21 mA]	Using multi-turn coil technique; for the calibration of clamp-on ammeters.	
AC RESISTANCE	0.2 Ω to 1 kΩ *50 Hz*	Q[0.10 %, 0.50 mΩ]	For the calibration of loop testers	
AC VOLTAGE				
Measurement	10 mV to 100 mV *40 Hz to 20 kHz*	Q[0.060 %, 40 µV]	Using digital multimeter.	
	100 mV to 1 V *40 Hz to 20 kHz* *20 kHz to 50 kHz*	Q[0.060 %, 0.4 mV] Q[0.12 %, 2.0 mV]		
	1 V to 10 V *40 Hz to 20 kHz* *20 kHz to 50 kHz*	Q[0.060 %, 4.0 mV] Q[0.12 %, 30 mV]		
	10 V to 100 V *40 Hz to 20 kHz*	Q[0.060 %, 35 mV]		
	100 V to 750 V *40 Hz to 1 kHz*	Q[0.060 %, 250 mV]		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC VOLTAGE (continued)				Site Calibration
Generation	3 mV to 33 mV *50 Hz to 10 kHz*	Q[0.18 %, 24 μV]	Known values of AC voltage for application to voltage measuring instruments.	
	33 mV to 330 mV *50 Hz to 10 kHz*	Q[0.065 %, 24 μV]		
	330 mV to 3.3 V *50 Hz to 10 kHz* *10 kHz to 50 kHz*	Q[0.045 %, 80 μV] Q[0.18 %, 0.40 mV]		
	3.3 V to 33 V *50 Hz to 10 kHz* *10 kHz to 50 kHz*	Q[0.055 %, 0.80 mV] Q[0.23 %, 7.0 mV]		
	33 V to 330 V *50 Hz to 1 kHz* *1 kHz to 10 kHz*	Q[0.065 %, 9.0 mV] Q[0.10 %, 18 mV]		
	330 V to 1020 V *50 Hz to 1 kHz*	Q[0.065 %, 90 mV]		
AC CURRENT				
Measurement	*50 Hz to 1 kHz* 0 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A 3 A to 10 A	 Q[0.10 %, 0.10 μA] Q[0.10 %, 0.80 μA] Q[0.10 %, 5.0 μA] Q[0.10 %, 65 μA] Q[0.10 %, 1.1 mA] Q[0.15 %, 2.6 mA] Q[0.40 %, 11 mA]	Using digital multimeter.	
	50 Hz to 1 kHz 10 A to 20 A 20 A to 100 A	 Q[0.13 %, 4.0 mA] Q[0.13 %, 40 mA]		
Generation	0.3 mA to 3.3 mA *50 Hz to 1 kHz*	Q[0.14 %, 0.32 μA]	Known values of AC current for application to current measuring instruments.	
	3.3 mA to 33 mA *50 Hz to 1 kHz*	Q[0.13 %, 3.2 μA]		
	33 mA to 330 mA *50 Hz to 1 kHz*	Q[0.13 %, 32 μA]		
	330 mA to 2.2 A *50 Hz to 1 kHz*	Q[0.14 %, 320 μA]		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
AC CURRENT (continued)				Site Calibration
Generation (continued)	2.2 A to 11 A *50 Hz to 400 Hz*	Q[0.14 %, 2.1 mA]		
	11 A to 100 A *50 Hz to 400 Hz*	0.30 %		
	100 A to 1 kA *50 Hz*	Q[0.20 %, 120 mA]	Using multi turn coil technique; for the calibration of clamp-on ammeters and Rogowski coils.	
FREQUENCY				
Measurement and generation	0.1 Hz to 160 MHz	2 in 10^6	Using counter timer. The periodic time of repetitive waveforms may be expressed in terms of 1/f or as events per unit time such as RPM.	
Generation	0.02 Hz to 2.1 GHz	2 in 10^9	Using Signal Generator	
TIME INTERVAL				
Single events	5 ms to 50 ms 50 ms to 2.2 s	0.20 ms 2.5 ms	For the calibration of RCD testers	
Stopwatches (Mechanical)	0 s to 3 hours 3 hours to 30 hours	0.25 s 0.50 s	By comparison with calibrated counter timer.	
Temperature indicators, calibration by electrical simulation				
Base metal thermocouples	-200 °C to +1600 °C	0.57 °C	Including cold junction compensation	
Noble metal thermocouples	-50 °C to +1760 °C	1.4 °C	Including cold junction compensation	
Resistance thermometers (Pt 100)	-200 °C to 800 °C	0.15 °C		

0125 Accredited to ISO/IEC 17025:2017	**Schedule of Accreditation** **issued by** **United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	Universal Instrument Services Ltd **Issue No:** 078 **Issue date:** 10 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
PRESSURE CALIBRATION			Methods consistent with EURAMET CG17. Calibrations of devices with an electrical output may be undertaken	Permanent Laboratory
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges.	-100 kPa to -7 kPa -7 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa 2.5 kPa to 7 kPa 7 kPa to 70 kPa 70 kPa to 700 kPa 700 kPa to 4.3 MPa 4.3 MPa to 6 MPa 6 MPa to 10 MPa	Q[0.0058 %, 17 Pa] Q[0.0068 %, 5.1 Pa] Q[0.010 % , 2.0 Pa] Q[0.0068 %, 5.1 Pa] Q[0.0058 %, 13 Pa] 0.0066 % Q[0.010 %, 33 Pa] Q[0.18 %, 4.2 kPa] Q[0.12 %, 4.0 kPa]		
Gas pressure (absolute)				
Calibration of pressure indication instruments and gauges.	30 Pa to 75 kPa 75 kPa to 120 kPa 120 kPa to 170 kPa 170 kPa to 800 kPa 800 kPa to 4.3 MPa	Q[0.0058 %, 17 Pa] 10 Pa Q[0.0058 %, 17 Pa] Q[0.0066 %, 10 Pa] Q[0.010 %, 35 Pa]		
Hydraulic pressure (gauge)				
Calibration of pressure indicating instruments and gauges.	0 kPa to 54 kPa 54 kPa to 60 kPa 60 kPa to 600 kPa 600 kPa to 6 MPa 6 MPa to 120 MPa	Q[0.030 %, 110 Pa] Q[0.032 %, 220 Pa] Q[0.0066 %, 29 Pa] 0.010 % 0.010 %		
Hydraulic Pressure (absolute)				
Calibration of pressure indicating instruments and gauges.	170 kPa to 800 kPa 800 kPa to 6.1 MPa 6.1 MPa to 120 MPa	Q[0.0066 %, 32 Pa] Q[0.010 %, 12 Pa] Q[0.010 %, 12 Pa]		
Gas pressure (gauge)				Site Calibration
Calibration of pressure indicating instruments and gauges.	-100 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa 2.5 kPa to 200 kPa 200 kPa to 2 MPa 2 MPa to 7 MPa	Q[0.005 8 %, 0.22 kPa] Q[0.010 %, 2.0 Pa] 96 Pa Q[0.010%, 0.80 kPa] Q[0.010%, 7.0 kPa]	Absolute pressures within these ranges can also be generated, subject to an additional uncertainty of 90 Pa.	
Hydraulic pressure (gauge)				
Calibration of pressure indicating instruments and gauges.	0 MPa to 20 MPa 20 MPa to 70 MPa	Q[0.010%, 5.8 kPa] Q[0.050 %, 22 kPa]		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (*k* = 2)	Remarks	Location Code
TEMPERATURE CALIBRATION				Permanent Laboratory
Liquid-in-glass thermometers	Ice-point (0 °C) -80 °C to +5 °C 5 °C to 50 °C 50 °C to 250 °C	0.013 °C 0.020 °C 0.016 °C 0.025 °C	By comparison with reference standards Calibration performed within liquid baths	
Resistance thermometers	Nitrogen boiling point (-196 °C) Ice point (0 °C) Triple point of water (0.01 °C) -80 °C to 5 °C 5 °C to 50 °C 50 °C to 250 °C 250 °C to 660 °C	0.013 °C 0.012 °C 0.005 0 °C 0.010 °C 0.015 °C 0.014 °C 0.095 °C	-196 °C & 0 °C FP by comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	
Platinum thermocouples	-50 °C to 0 °C 0 °C to 80 °C 80 °C to 250 °C 250 °C to 650 °C 650 °C to 1100 °C	1.3 °C 0.67 °C 0.52 °C 0.54 °C 1.6 °C	By comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	
Base metal thermocouples	Nitrogen boiling point (-196 °C) -80 °C to 0 °C 0 °C to 250 °C 250 °C to 650 °C 650 °C to 1100 °C	0.23 °C 0.18 °C 0.17 °C 0.19 °C 1.6 °C	-196 °C & 0 °C FP by comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	
Electronic Thermometers with probes	Nitrogen boiling point (-196 °C) -80 °C to +250°C Ice point (0 °C) Triple point of water (0.01 °C) 250 °C to 650 °C 650 °C to 1100 °C	0.014 °C 0.013 °C 0.014 °C 0.007 0 °C 0.12 °C 1.7 °C	-196 °C & 0 °C FP by comparison with reference standards Calibration within liquid baths up to 650 °C and above 650 °C in a furnace	
Calibration of Temperature Data loggers, Electronic Temperature Indicators and Recorders in Air	-40 °C to +80 °C	0.58 °C	Calibration within air chamber	
Block calibrators	-50 °C to +660 °C	0.12 °C	By comparison with reference standards	
Cold Junction Compensation	Ice-point (0 °C) 15 °C to 25 °C	0.13 °C 0.17 °C	In stirred liquid bath	

0125 Accredited to ISO/IEC 17025:2017	**Schedule of Accreditation** **issued by** **United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	Universal Instrument Services Ltd **Issue No:** 078 **Issue date:** 10 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
HUMIDITY CALIBRATION			By comparison with reference in controlled chamber Calibrations of devices with an electrical output may be undertaken	Permanent Laboratory
	In the temperature range 0 °C to 20 °C			
	5 %rh to 50 %rh	2.0 %rh		
	50 %rh to 90 %rh	2.5 %rh		
	In the temperature range 20 °C to 40 °C			
	5 %rh to 95 %rh	2.5 %rh		
	In the temperature range 40 °C to 60 °C			
	5 %rh to 90 %rh	2.5 %rh		
Temperature in air	0 °C to 60 °C	0.20 °C		
TEMPERATURE CALIBRATION				Site Calibration
Temperature sensors with indicators	-20 °C to 120 °C	0.089 °C	By comparison with reference standards In stirred liquid baths	
	-25 °C to +140 °C	0.061 °C	By comparison with reference standards within metal block baths	
	140 °C to 350 °C	0.19 °C		
	Ice-point (0 °C)	0.041 °C	Calibration performed within liquid baths	
Liquid baths				
Temperature measurement	-40 °C to 250 °C	0.044 °C	By comparison with reference standards	
Gradient measurement	-40 °C to 250 °C	0.020 °C		
Temperature controlled chambers, ovens, fridges/refrigerators and freezers	-80 °C to +300 °C	0.84 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	
	300 °C to 1100 °C	3.7 °C		

END

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.
Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$