


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0138</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>DH Budenberg A division of WIKA Instruments Ltd</p> <p>Issue No: 055 Issue date: 21 April 2026</p>	
	<p>10 Huntsman Drive Northbank Industrial Estate Irlam Manchester M44 5EG</p>	<p>Contact: Mr M Parkinson Tel: +44 (0)161 777 2802 Fax: +44 (0)161 777 2801 E-Mail: malcolm.parkinson@wika.com Website: www.wika.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	
PRESSURE				
<u>Gas pressure (absolute)</u>				
Calibration of pressure indicating instruments and gauges	1 kPa to 10 kPa 10 kPa to 160 kPa 160 kPa to 400 kPa 400 kPa to 7 MPa 7 MPa to 40 MPa	Q [0.023 %, 2.1 Pa] Q [0.0045 %, 2.9 Pa] Q [0.0040 %, 15 Pa] Q [0.0035 %, 15 Pa] Q [0.0070 %, 15 Pa]	Methods consistent with EURAMET CG3 and CG17 Calibration of devices with an electrical output may be undertaken	
<u>Gas pressure (gauge)</u>				
Determination of effective area of Dead Weight Testers and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set)	1.8 kPa to 10 kPa 10 kPa to 7 MPa 7 MPa to 40 MPa	Q [0.0040 %, 0.090 Pa] 0.0040 % 0.0070 %		
Calibration of pressure indicating instruments and gauges	- 90 kPa to - 10 kPa - 10 kPa to - 1.8 kPa 1.8 kPa to 10 kPa 10 kPa to 4 MPa 4 MPa to 7 MPa 7 MPa to 40 MPa	0.0070 % 0.0080 % Q [0.0040 %, 0.090 Pa] 0.0040 % 0.0035 % 0.0070 %		
<u>Hydraulic pressure (absolute)</u>				
Calibration of pressure indicating instruments and gauges	100 kPa to 600 kPa 600 kPa to 6 MPa 6 MPa to 250 MPa 250 MPa to 400 MPa 400 MPa to 500 MPa	Q [0.0040 %, 25 Pa] Q [0.0045 %, 15 Pa] 0.0040 % 0.0045 % 0.0050 %		
<u>Hydraulic pressure (gauge)</u>				
Determination of effective area of Dead Weight Testers and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set)	600 kPa to 6 MPa 6 MPa to 50 MPa 50 MPa to 500 MPa	0.0045 % Q [0.0040 %, 0.11x10 ⁻⁶ /MPa] Q [0.0040 %, 0.065x10 ⁻⁶ /MPa]		



0138
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DH Budenberg
A division of WIKA Instruments Ltd
Issue No: 055 Issue date: 21 April 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
PRESSURE (cont'd) Calibration of pressure indicating instruments and gauges	10 kPa to 600 kPa 600 kPa to 6 MPa 6 MPa to 250 MPa 250 MPa to 400 MPa 400 MPa to 500 MPa	Q [0.0040 %, 20 Pa] 0.0045 % 0.0040 % 0.0045 % 0.0050 %	
MASS Mass determination of dead weight tester weights	0.1 g 0.2 g 0.5 g 1 g 2 g 5 g 10 g 20 g 50 g 50 g to 20 kg	0.035 mg 0.040 mg 0.050 mg 0.060 mg 0.080 mg 0.10 mg 0.12 mg 0.16 mg 0.20 mg 3.0 µg/g	Comparison methods using ABABA or ABA Intermediate values will be calibrated to an uncertainty interpolated from the next higher and lower values
END			



0138
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DH Budenberg
A division of WIKA Instruments Ltd
Issue No: 055 Issue date: 21 April 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$