

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0149

Accredited to
ISO/IEC 17025:2017

Testo Industrial Services Ltd

Issue No: 043 Issue date: 02 June 2026

Unit 1
Armstrong Point
Swan Lane
Hindley Green
Wigan
WN2 4AU

Contact: Mr Paul Miller
Tel: +44 (0)1942-882275
E-Mail: info@testotis.co.uk
Website: www.testotis.co.uk

Calibration performed at the above address only

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
Values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column			
ELECTRICAL MEASUREMENTS			
DC RESISTANCE			
Specific values (sourcing)	1 mΩ	35 μΩ/Ω	
	10 mΩ	12 μΩ/Ω	
	100 mΩ	8.0 μΩ/Ω	
	1 Ω	2.0 μΩ/Ω	
	10 Ω	2.5 μΩ/Ω	
	100 Ω	3.0 μΩ/Ω	
	1 kΩ	2.0 μΩ/Ω	
	10 kΩ	1.5 μΩ/Ω	
	100 kΩ	3.0 μΩ/Ω	
	1 MΩ	10 μΩ/Ω	
	10 MΩ	20 μΩ/Ω	
	100 MΩ	20 μΩ/Ω	
	1 GΩ	250 μΩ/Ω	
	10 GΩ	0.15 %	
Specific values (measurement)	1 mΩ	40 μΩ/Ω	
	10 mΩ	20 μΩ/Ω	
	100 mΩ	20 μΩ/Ω	
	1 Ω	4.0 μΩ/Ω	
	10 Ω	5.0 μΩ/Ω	
	100 Ω	3.0 μΩ/Ω	
	1 kΩ	2.0 μΩ/Ω	
	10 kΩ	2.0 μΩ/Ω	
	100 kΩ	5.0 μΩ/Ω	
	1 MΩ	7.0 μΩ/Ω	
	10 MΩ	15 μΩ/Ω	
	100 MΩ	28 μΩ/Ω	
	1 GΩ	220 μΩ/Ω	
	10 GΩ	0.14 %	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
DC RESISTANCE (continued)			
Other values (measurement)	0 $\mu\Omega$ to 200 $\mu\Omega$ 200 $\mu\Omega$ to 2 m Ω 2 m Ω to 20 m Ω 20 m Ω to 200 m Ω 200 m Ω to 2 Ω	40 n Ω 200 $\mu\Omega/\Omega$ 180 $\mu\Omega/\Omega$ 180 $\mu\Omega/\Omega$ 25 $\mu\Omega/\Omega$	
	2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 K Ω 20 K Ω to 200 k Ω	20 $\mu\Omega/\Omega$ 6.0 $\mu\Omega/\Omega$ 3.5 $\mu\Omega/\Omega$ 4.0 $\mu\Omega/\Omega$ 6.0 $\mu\Omega/\Omega$	
	200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω 2 G Ω to 20 G Ω	10 $\mu\Omega/\Omega$ 60 $\mu\Omega/\Omega$ 65 $\mu\Omega/\Omega$ 700 $\mu\Omega/\Omega$ 0.60 %	
DC VOLTAGE			
Specific values	100 mV 200 mV 1 V 2 V 10 V	6.0 $\mu V/V$ 6.0 $\mu V/V$ 3.0 $\mu V/V$ 4.0 $\mu V/V$ 4.0 $\mu V/V$	
	20 V 100 V 200 V 1 kV	4.0 $\mu V/V$ 4.0 $\mu V/V$ 5.0 $\mu V/V$ 6.0 $\mu V/V$	
Other values	0 mV to 20 mV 20 mV to 200 mV 200 mV to 2 V	0.60 μV 8.5 $\mu V/V$ 5.0 $\mu V/V$ 5.0 $\mu V/V$	
	2 V to 20 V 20 V to 200 V 200 V to 1 kV 1 kV to 30 kV 30 kV to 90 kV	7.0 $\mu V/V$ 7.0 $\mu V/V$ 7.0 $\mu V/V$ 0.12 % 0.15 %	
DC VOLTAGE RATIO 100 mV to 10 V reference	0.1 to unity	0.5 $\mu V/V$	
DC Voltage linearity	0 V to 10 mV 0 V to 100 mV	0.40 μV 0.60 μV	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (<i>k</i> = 2)	Remarks
DC CURRENT	0 μ A to 1 μ A 1 μ A to 10 μ A 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA and 10 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A 10 A to 150 A	100 μ A/A + 80 pA 120 μ A/A 35 μ A/A 30 μ A/A 20 μ A/A 30 μ A/A 20 μ A/A 30 μ A/A 60 μ A/A 500 μ A/A	
DC Current linearity	0 A to 1 μ A 0 A to 10 μ A	7.5 pA 12 pA	
DC POWER	1 W to 20 kW	The arithmetic sum of the individual uncertainties of the corresponding voltages and current measurements	
AC VOLTAGE Specific values at specific frequencies	10 mV at 1 kHz 100 mV 20 Hz, 55 Hz 305 Hz, 1 kHz, 10 kHz 30 kHz 60 kHz 100 kHz 1 V 100 Hz 20 Hz, 55 Hz, 305 Hz 1 kHz 3 kHz, 10 kHz 30 kHz 60 kHz 100 kHz 500 kHz 1 MHz 10 V 20 Hz, 55 Hz, 100 Hz, 305 Hz, 1kHz 3 kHz, 10 kHz 30 kHz 60 kHz 100 kHz 500 kHz 1 MHz	100 μ V/V 100 μ V/V 90 μ V/V 100 μ V/V 180 μ V/V 190 μ V/V 55 μ V/V 50 μ V/V 40 μ V/V 50 μ V/V 60 μ V/V 65 μ V/V 160 μ V/V 0.135 % 0.30 % 50 μ V/V 60 μ V/V 80 μ V/V 180 μ V/V 190 μ V/V 0.135 % 0.30 %	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC VOLTAGE (continued) Specific values at specific Frequencies (continued)	100 V 20 Hz, 55 Hz, 305 Hz 100 Hz, 1 kHz 3 kHz, 10 kHz 30 kHz 60 kHz 100 kHz	60 μ V/V 55 μ V/V 60 μ V/V 80 μ V/V 180 μ V/V 200 μ V/V	
	500 V 55 Hz 100 Hz 305 Hz 1 kHz 3 kHz, 10 kHz 30 kHz	80 μ V/V 90 μ V/V 80 μ V/V 70 μ V/V 80 μ V/V 150 μ V/V	
	1 kV 55 Hz 305 Hz, 1 kHz, 3 kHz, 10 kHz 30 kHz	80 μ V/V 80 μ V/V 200 μ V/V	
Specific values at other frequencies	1 V 20 Hz to 30 kHz 30 kHz to 100 kHz 100 kHz to 1MHz	70 μ V/V 160 μ V/V 0.30 %	
	10 V 20 Hz to 30 kHz 30 kHz to 100 kHz 100 kHz to 1MHz	90 μ V/V 180 μ V/V 0.30 %	
	100 V 20 Hz to 30 kHz 30 kHz to 100 kHz	85 μ V/V 150 μ V/V	
	1 kV 55 Hz to 10 kHz 10 kHz to 30 kHz	100 μ V/V 200 μ V/V	
Other values	50 Hz to 2 kHz 100 μ V to 1 mV 1 mV to 10 mV 10 mV to 100 mV	0.75 % 750 μ V/V 100 μ V/V	
	100 mV to 200 mV 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	150 μ V/V 360 μ V/V 850 μ V/V	
	200 mV to 1 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	160 μ V/V 250 μ V/V 0.13 %	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC VOLTAGE (continued) Other values (continued)	1 V to 2 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	120 µV/V 250 µV/V 650 µV/V	
	2 V to 10 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	160 µV/V 350 µV/V 0.13 %	
	10 V to 20 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	160 µV/V 300 µV/V 300 µV/V	
	20 V to 200 V 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	150 µV/V 150 µV/V 150 µV/V	
	200 V to 1000 V 40 Hz to 10 kHz 10 kHz to 30 kHz	200 µV/V 700 µV/V	
	1 kV to 40 kV 50 Hz	0.65 %	
Waveform analysis	3 µV to 300 V 20 Hz to 76 kHz	5.0 % of FSD*	* 15 ranges of 30 µV to 300 V FSD in 1-3-10 sequence
AC CURRENT Specific values and frequencies	100 µA 55 Hz, 305 Hz 1 kHz 5 kHz	150 µA/A 150 µA/A 200 µA/A	
	1 mA 55 Hz, 305 Hz 1 kHz 5 kHz 10 kHz	150 µA/A 150 µA/A 150 µA/A 160 µA/A	
	10 mA 55 Hz, 305 Hz 1 kHz, 5 kHz, 10 kHz	150 µA/A 150 µA/A	
	100 mA 55 Hz, 305 Hz 1 kHz, 5 kHz, 10 kHz	150 µA/A 150 µA/A	
	1 A 55 Hz, 305 Hz, 1 kHz, 5 kHz 10 kHz	150 µA/A 260 µA/A	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (<i>k</i> = 2)	Remarks
AC CURRENT (continued) Other Values (continued)	10 A 55 Hz, 305 Hz, 1 kHz	170 µA/A	
Other Values	5 kHz, 10 kHz	200 µA/A	
	20 µA to 100 µA 50 Hz to 5 kHz	0.12 %	
	100 µA to 200 µA 50 Hz to 5 kHz	400 µA/A	
	200 µA to 1 mA 55 Hz to 5 kHz	0.12 %	
	1 mA to 2 mA 50 Hz to 10 kHz	400 µA/A	
	2 mA to 10 mA 50 Hz to 10 kHz	0.12 %	
	10 mA to 20 mA 50 Hz to 10 kHz	400 µA/A	
	20 mA to 100 mA 50 Hz to 10 kHz	0.12 %	
	100 mA to 200 mA 40 Hz to 10 kHz	400 µA/A	
	200 mA to 1 A 1 kHz to 10 kHz	0.15 %	
	1 A to 2 A 55 Hz, 305 Hz, 1 kHz	750 µA/A	
	2 A to 10 A 50 Hz to 1 kHz 1 kHz to 10 kHz	0.15 % 0.32 %	
	10 A to 20 A 50 Hz to 1 kHz 1 kHz to 10 kHz	0.10 % 0.30 %	
	10 A to 150 A 50 Hz to 60 Hz	0.10 %	
AC RESISTANCE	At 40 Hz to 60 Hz		
	10 mΩ to 100 mΩ	300 µΩ/Ω	
	100 mΩ to 1Ω	300 µΩ/Ω	
	1 Ω to 100 k Ω	75 µΩ/Ω	
	100 kΩ to 10 MΩ	0.10 %	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
FREQUENCY			
Generation			
Specific values	100 kHz, 1 MHz, 5 MHz and 10 MHz	2.0 in 10^{10}	Sine wave generation
	0.02 Hz to 10 MHz in 2-5-10 sequence	2.0 in 10^{10}	Square wave generation
Range values	1 Hz to 100 kHz 100 kHz to 10 MHz	2.0 in 10^{10} 5.0 in 10^{11}	Sine wave generation
	1 Hz to 10 kHz 10 kHz to 100 kHz 100 kHz to 100 MHz 100 MHz to 1 GHz	1.2 in 10^8 1.2 in 10^9 1.2 in 10^9 1.2 in 10^{10}	Measurement of sources These values may also be reported as the reciprocal; seconds, for repetitive signals.
TIME INTERVAL	0 s to 500s 0 s to 500s	1.0 us 50 ms	Electronically triggered devices Mechanically triggered devices
Pulse period	1 μ s to 1 s	5.0 ns	
Rise time	1 ns to 1 ms	3.0 ns	Into 50 Ω
RCD testers			
Trip time	10 ms to 5 s	0.25 ms	
Trip Current	3 mA to 3 A	1.0 %	
Earth Loop	8 m Ω to 330 m Ω 330 m Ω to 500 m Ω 500 m Ω 10 1.8 Ω 1.8 Ω to 5 Ω 5 Ω to 10 Ω 10 Ω to 18 Ω 18 Ω 50 Ω 50 Ω to 100 Ω 100 Ω to 180 Ω 180 Ω to 500 Ω 500 Ω to 1 k Ω 1 k Ω to 1.8 k Ω	8.0 m Ω 10 m Ω 12 m Ω 36 m Ω 70 m Ω 120 m Ω 350 m Ω 600 m Ω 1.2 Ω 3.0 Ω 6.0 Ω 12 Ω	



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
Temperature indicators, calibration by electrical simulation			The claims below cover test items with a resolution of 10 m°C
ELECTRICAL SIMULATION OF TEMPERATURE			
Thermocouple capabilities listed below are given for type K Base and Type S Noble, using EMF sensitivity values as listed in BS EN 60584-1:2013. Other Thermocouple types can be calibrated, the uncertainties will correspond to the appropriate sensitivities listed. Calibrations which include the internal reference junction (CJC) are available for types: J, K, N, T, E, R, S, B & C			
Base Metal Thermocouples	-200 °C to -100 °C -100 °C to -50 °C -50 °C to 0 °C 0 °C to 100 °C 100 °C to 700 °C 700 °C to 900 °C 900 °C to 1370 °C	0.20 °C 0.15 °C 0.14 °C 0.14 °C 0.19 °C 0.18 °C 0.21 °C	Excluding automatic CJC
Noble Metal Thermocouples	0 °C to 1500 °C	0.35 °C	
Base Metal Thermocouples	-200 °C to -100 °C -100 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.40 °C 0.24 °C 0.31 °C 0.43 °C	Including automatic CJC
Noble Metal Thermocouples	0 °C to 1500 °C	0.50 °C	
Cold Junction Compensation	0 °C to 30 °C	0.10 °C	
Resistance thermometer (Pt 100)	- 200 °C to + 800 °C	0.020 °C	
Supporting temperature measurements for electrical simulation and cold junction verification	At Nominal 0 °C Nominal ambient between 17 °C to 23 °C	0.050 °C 0.30 °C	
Capacitance	350 pF 480 pF 600 pF 1 nF 3.0 nF 4.0 nF 10 nF 30 nF 40 nF 100 nF 120 nF 300 nF 400 nF 1.0 µF 3.0 µF 4.0 µF 10 µF 30 µF 100 µF 300 µF 800 µF	12 pF 12 pF 12 pF 13 pF 22 pF 26 pF 59 pF 150 pF 170 pF 310 pF 520 pF 1.1 nF 1.7 nF 3.1 nF 13 nF 16 nF 32 nF 280 nF 590 nF 2.5 µF 9.3 µF	These values are sourced by simulation.
END			



0149
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Testo Industrial Services Ltd
Issue No: 043 Issue date: 02 June 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$