

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0179

Accredited to
ISO/IEC 17025:2017

Michell Instruments Ltd

Issue No: 025 Issue date: 18 February 2022

48 Lancaster Business Park
Ely
Cambridgeshire
CB6 3NW

Contact: Mr C Parker/Mr P Newton
Tel: +44 (0)1353 658000
Fax: +44 (0)1353 658199
E-Mail: info@michell.co.uk
Website: www.michell-instruments.com

Calibration performed at the above address only

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
HUMIDITY			By comparison with dew-point hygrometer and Platinum Resistance Thermometers
Dewpoint	90 °C to 82 °C 82 °C to 20 °C 20 °C to - 60 °C - 60 °C to - 75 °C - 75 °C to - 90 °C	0.25 °C to 0.22 °C 0.22 °C 0.15 °C to 0.23 °C 0.23 °C to 0.38 °C 0.38 °C to 0.79 °C	Calibrations can be performed on instruments with an electrical output
Relative humidity			
Relative humidity instruments	10 %rh to 98 %rh, 19 °C to 90 °C 10 %rh to 98 %rh, 10 °C to 19 °C	2.1 % of reading 2.2 % of reading	
Temperature sensors incorporated in humidity instruments	10 °C to 90 °C	0.10 °C	
TEMPERATURE			
Electronic thermometers with sensors	0 °C to 90 °C	0.10 °C	
END			



0179
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Michell Instruments Ltd
Issue No: 025 Issue date: 18 February 2022

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$