


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0221 Accredited to ISO/IEC 17025:2017	Druck Limited Issue No: 080 Issue date: 01 June 2026	
	Fir Tree Lane Groby Leicester LE6 0FH	Contact: Mr S Berdej Tel: +44 (0)116-231 7100 Fax: +44 (0)116-231 7101 E-Mail: sensing.grobyukas@bakerhughes.com Website: www.druck.com

Calibration performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Fir Tree Lane Groby Leicester LE6 0FH Contact: Mr S Berdej Tel: +44 (0)116-231 7100 Fax: +44 (0)116-231 7101 Email: sensing.grobyukas@bakerhughes.com	Electrical Pressure Temperature Mass Humidity	UK

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Pressure	SITE



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
PRESSURE			Methods consistent with EURAMET CG3 and CG17	
<u>Gas pressure (absolute)</u>				
Calibration of pressure measuring instruments and gauges and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set) and Effective area calibration of Dead Weight Testers	0 kPa to 450 kPa 450 kPa kPa to 3.1 MPa 3.1 MPa to 21.1 MPa 21.1 MPa to 40.1 MPa	Q[0.0020 %, 0.40 Pa] Q[0.0020 %, 0.40 Pa] Q[0.0030 %, 0.40 Pa] Q[0.0039 %, 11 Pa]	Calibration of pressure measuring devices with an electrical output may be undertaken.	UK
Calibration of air data test sets manufactured by Druck Ltd	3.5 kPa to 12 kPa 12 kPa to 20 kPa 20 kPa to 30 kPa 30 kPa to 70 kPa 70 kPa to 100 kPa 100 kPa to 110 kPa 110 kPa to 160 kPa 160 kPa to 199.9 kPa 199.9 kPa to 200 kPa	Q[0.062 %, 2.2 Pa] Q[0.016 %, 1.9 Pa] Q[0.010 %, 2.1 Pa] Q[0.0084 %, 2.5 Pa] Q[0.0049 %, 3.4 Pa] Q[0.0045 %, 4.5 Pa] Q[0.0044 %, 4.9 Pa] Q[0.0046 %, 7.3 Pa] Q[0.0043 %, 8.5 Pa]	Pitot ranges (Ps) by comparison with a characterised reference air data test set	SITE
	3.5 kPa to 30 kPa 30 kPa to 100 kPa 100 kPa to 160 kPa 160 kPa to 200 kPa 200 kPa to 270 kPa 270 kPa to 349 kPa 349 kPa to 350 kPa	Q[0.097 %, 3.4 Pa] Q[0.012 %, 3.5 Pa] Q[0.0048 %, 4.8 Pa] Q[0.0051 %, 8.2 Pa] Q[0.0046 %, 9.1 Pa] Q[0.0046 %, 13 Pa] Q[0.0043 %, 15 Pa]	Pitot ranges (Pt) by comparison with a characterised reference air data test set	SITE
<u>Gas pressure (gauge)</u>				
Calibration of pressure measuring instruments and gauges and "Pressure equivalent" calibration of Dead Weight Testers (pressure balances supplied with an associated mass set) and Effective area calibration of Dead Weight Testers	-100 kPa to -3.5 kPa -3.5 kPa to 0 kPa 0 Pa to 2.5 Pa 2.5 Pa to 350 Pa 350 Pa to 1 kPa 1 kPa to 2 kPa 2 kPa to 2.5 kPa 2.5 kPa to 350 kPa 350 kPa to 3 MPa 3 MPa to 21 MPa 21 MPa to 40 MPa	Q[0.0043 %, 0.60 Pa] 0.75 Pa 0.15 Pa 0.10 Pa 0.11 Pa 0.12 Pa 0.14 Pa 0.0020 % 0.0020 % 0.0030 % 0.0039 %		UK
<u>Hydraulic pressure (gauge)</u>				
Calibration of pressure measuring instruments and gauges. "Pressure equivalent" calibration of Dead Weight Testers (Pressure balance with associated mass set). Effective area calibration of Dead Weight Testers.	0.5 MPa to 140 MPa 140 MPa to 500 MPa	Q[0.0036 %, (0.12p x 10 ⁻⁶)/MPa] Q[0.010 %, (0.30p x 10 ⁻⁶)/MPa] p = measured pressure	Absolute pressure calibrations can be undertaken using gauge pressure generation and the associated barometric pressure with the additional uncertainty of 11 Pa	UK



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL			Calibration by comparison with a reference standard	UK
DC Resistance measurement (At low current)	0 Ω to 2 Ω	Q[16 $\mu\Omega/\Omega$, 35 $\mu\Omega$]		
	2 Ω to 20 Ω	Q[7.0 $\mu\Omega/\Omega$, 30 $\mu\Omega$]		
	20 Ω to 200 Ω	Q[6.5 $\mu\Omega/\Omega$, 250 $\mu\Omega$]		
	200 Ω to 2 k Ω	Q[10.0 $\mu\Omega/\Omega$, 2.0 m Ω]		
	2 k Ω to 20 k Ω	Q[10.0 $\mu\Omega/\Omega$, 25 m Ω]		
	20 k Ω to 200 k Ω	Q[5.0 $\mu\Omega/\Omega$, 135 m Ω]		
	200 k Ω to 2 M Ω	Q[13 $\mu\Omega/\Omega$, 2.0 Ω]		
	2 M Ω to 20 M Ω	Q[72 $\mu\Omega/\Omega$, 120 Ω]		
	20 M Ω to 200 M Ω	Q[720 $\mu\Omega/\Omega$, 120 k Ω]		
	200 M Ω to 2 G Ω	Q[720 $\mu\Omega/\Omega$, 1.2 M Ω]		
	DC Resistance measurement (At higher current)	0 Ω to 2 Ω		
		2 Ω to 20 Ω	Q[7.0 $\mu\Omega/\Omega$, 30 $\mu\Omega$]	
20 Ω to 200 Ω		Q[7.0 $\mu\Omega/\Omega$, 70 $\mu\Omega$]		
	200 Ω to 2 k Ω	Q[3.4 $\mu\Omega/\Omega$, 0.70 m Ω]		
	2 k Ω to 20 k Ω	Q[3.5 $\mu\Omega/\Omega$, 5.0 m Ω]		
	20 k Ω to 200 k Ω	Q[4.3 $\mu\Omega/\Omega$, 75 m Ω]		
	200 k Ω to 2 M Ω	Q[9.5 $\mu\Omega/\Omega$, 1.5 Ω]		
	2 M Ω to 20 M Ω	Q[71 $\mu\Omega/\Omega$, 120 Ω]		
	20 M Ω to 200 M Ω	Q[700 $\mu\Omega/\Omega$, 12 k Ω]		
	200 M Ω to 2 G Ω	Q[720 $\mu\Omega/\Omega$, 1.2 M Ω]		
	DC Resistance Generation Specific values	1 m Ω	0.15%	
		10 m Ω	0.10 %	
100 m Ω		250 $\mu\Omega/\Omega$		
1 Ω		42 $\mu\Omega/\Omega$		
1.9 Ω		36 $\mu\Omega/\Omega$		
10 Ω		7.0 $\mu\Omega/\Omega$		
19 Ω		20 $\mu\Omega/\Omega$		
25 Ω		6.0 $\mu\Omega/\Omega$		
50 Ω		3.0 $\mu\Omega/\Omega$		
100 Ω		3.0 $\mu\Omega/\Omega$		
190 Ω	15 $\mu\Omega/\Omega$			
250 Ω	6.0 $\mu\Omega/\Omega$			
300 Ω	5.0 $\mu\Omega/\Omega$			
1 k Ω	3.0 $\mu\Omega/\Omega$			
1.9 k Ω	5.5 $\mu\Omega/\Omega$			
10 k Ω	3.0 $\mu\Omega/\Omega$			
19 k Ω	5.0 $\mu\Omega/\Omega$			



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
DC Resistance Generation Specific values (continued)				
	100 k Ω	6.0 $\mu\Omega/\Omega$		
	190 k Ω	8.5 $\mu\Omega/\Omega$		
	1 M Ω	8.5 $\mu\Omega/\Omega$		
	1.9 M Ω	18 $\mu\Omega/\Omega$		
	10 M Ω	12 $\mu\Omega/\Omega$		
	19 M Ω	35 $\mu\Omega/\Omega$		
	100 M Ω	50 $\mu\Omega/\Omega$		
	1 G Ω	250 $\mu\Omega/\Omega$		
	1 k Ω	0.25 %	With applied voltages of 50 V and 100 V	
	10 k Ω	0.25 %		
	100 k Ω	0.25 %		
	1 M Ω	0.25 %		
	10 M Ω	0.25 %		
	100 M Ω	0.25 %		
	1 G Ω	0.50 %		
	10 G Ω	1.5 %		
	1.0 M Ω	0.30 %	With applied voltages of 250 V and 500 V	
	10 M Ω	0.30 %		
	100 M Ω	0.30 %		
	1 G Ω	0.40 %		
	10 G Ω	1.20 %		
	100 G Ω	2.0 %	With an applied voltage of 500 V	
	1 T Ω	2.0 %		
	100 M Ω	0.50 %	With an applied voltage of 1000 V	
	1.0 G Ω	0.50 %		
	10 G Ω	1.0 %		
	100 G Ω	2.0 %		
	1 T Ω	2.0 %		
	1.0 G Ω	1.1 %	With an applied voltage of 5000 V	
	10 G Ω	1.0 %		
	100 G Ω	2.0 %		
	1 T Ω	2.0 %		
Other values	0.1 Ω to 10 Ω	1.0 m Ω		
	10 Ω to 100 Ω	18 $\mu\Omega/\Omega$		
	100 Ω to 1 k Ω	25 $\mu\Omega/\Omega$		
	1 k Ω to 10 k Ω	10 $\mu\Omega/\Omega$		
	10 k Ω to 100 k Ω	10 $\mu\Omega/\Omega$		
	100 k Ω to 1 M Ω	35 $\mu\Omega/\Omega$		
	1 M Ω to 10 M Ω	0.05 %		
	10 M Ω to 100 M Ω	0.08 %		
	100 M Ω to 1 G Ω	0.60 %		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
DC Voltage				
Voltage Reference Values	1.018 V 10 V	2.0 $\mu\text{V/V}$ 1.0 $\mu\text{V/V}$		
Measurement	0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	Q[6.5 $\mu\text{V/V}$, 1.3 μV] Q[6.5 $\mu\text{V/V}$, 1.4 μV] Q[2.5 $\mu\text{V/V}$, 5.5 μV] Q[4.0 $\mu\text{V/V}$, 59 μV] Q[7.0 $\mu\text{V/V}$, 650 μV]		
Generation	0 mV to 200 mV 0.2 V to 2 V 2 V to 11 V 11 V to 20 V 20 V to 200 V 200 V to 1100 V	Q[10.0 $\mu\text{V/V}$, 0.50 μV] Q[3.0 $\mu\text{V/V}$, 1.0 μV] Q[1.5 $\mu\text{V/V}$, 3.0 μV] Q[1.5 $\mu\text{V/V}$, 5.0 μV] Q[2.5 $\mu\text{V/V}$, 60 μV] Q[2.5 $\mu\text{V/V}$, 300 μV]		
DC Current				
Measurement	0 A to 20 μA 20 μA to 200 μA 200 μA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A 20 A to 30 A	Q[70.0 $\mu\text{A/A}$, 1.0 nA] Q[10.0 $\mu\text{A/A}$, 1.0 nA] Q[10.5 $\mu\text{A/A}$, 5.0 nA] Q[9.5 $\mu\text{A/A}$, 50.0 nA] Q[23 $\mu\text{A/A}$, 930 nA] Q[67 $\mu\text{A/A}$, 19 μA] Q[80 $\mu\text{A/A}$, 463 μA] Q[230 $\mu\text{A/A}$, 6.0 μA] Q[230 $\mu\text{A/A}$, 6.0 μA]		
Generation	0 μA to 200 μA 200 μA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A 20 A to 60 A	Q[5.0 $\mu\text{A/A}$, 0.7 nA] Q[7.0 $\mu\text{A/A}$, 1.5 nA] Q[8.0 $\mu\text{A/A}$, 20 nA] Q[8.0 $\mu\text{A/A}$, 175 nA] Q[12 $\mu\text{A/A}$, 2.0 μA] 1.2 mA 10 mA 25 mA		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
AC Voltage Measurement	50 µV to 20 mV 10 Hz to 40 Hz 40 Hz to 105 Hz 105 Hz to 1 kHz 1 kHz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz 20 mV to 200 mV 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 200 mV to 2 V 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz 2 V to 20 V 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz	5.5 µV 5.5 µV 5.5 µV 5.0 µV 5.0 µV 5.0 µV 7.5 µV 7.5 µV 12.0 µV 50.0 µV 80.0 µV 6.0 µV 6.0 µV 6.0 µV 6.5 µV 6.5 µV 10.0 µV 15.0 µV 15.0 µV 50.0 µV 20.0 µV 85.0 µV 60.0 µV 45.0 µV 40.0 µV 40.0 µV 75.0 µV 130.0 µV 130.0 µV 0.6 mV 2.0 mV 420 µV 360 µV 320 µV 320 µV 330 µV 450 µV 1.3 mV 1.5 mV 5.0 mV 40.0 mV		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued) AC Voltage Measurement (continued)	20 V to 200 V 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 60 kHz 60 kHz to 100 kHz 100 kHz to 200 kHz At 200 kHz 200 V to 1 kV 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 30 kHz At 30 kHz	6.0 mV 5.0 mV 5.0 mV 5.0 mV 8.0 mV 17 mV 25 mV 30 mV 90 mV 120 mV 60 mV 60 mV 65 mV 75 mV 130 mV 250 mV	Calibration by comparison with a reference standard	UK



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
AC Voltage Generation				
	<i>10 Hz to 40 Hz</i>			
	0.1 mV to 2 mV	7.0 μ V		
	2 mV to 20 mV	11.0 μ V		
	20 mV to 100 mV	26 μ V		
	100 mV to 1 V	100 μ V		
	1 V to 10 V	2.4 mV		
	10 V to 100 V	24 mV		
	<i>40 Hz to 500 Hz</i>			
	0.1 mV to 2 mV	7.0 μ V		
	2 mV to 20 mV	10.0 μ V		
	20 mV to 100 mV	14 μ V		
	100 mV to 1 V	50 μ V		
	1 V to 10 V	700 μ V		
	10 V to 100 V	7.0 mV		
	<i>500 Hz to 1 kHz</i>			
	0.1 mV to 2 mV	7.0 μ V		
	2 mV to 20 mV	9.0 μ V		
	20 mV to 100 mV	13 μ V		
	100 mV to 1 V	25 μ V		
	1 V to 10 V	300 μ V		
	10 V to 100 V	4.0 mV		
	<i>1 kHz to 10 kHz</i>			
	0.1 mV to 2 mV	7.0 μ V		
	2 mV to 20 mV	10.0 μ V		
	20 mV to 100 mV	11 μ V		
	100 mV to 1 V	21 μ V		
	1 V to 10 V	230 μ V		
	10 V to 100 V	3.5 mV		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
AC Voltage Generation (continued)	10 kHz 2 mV 20 mV 100 mV 1 V 10 V 100 V	7.0 μ V 9.0 μ V 13 μ V 21 μ V 230 μ V 3.5 mV		
	1000 V 55 Hz to 500 Hz 500 Hz to 1 kHz 1 kHz	62 mV 75 mV 75 mV		
AC Current Measurement	30 nA to 10 μ A 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz At 10 kHz	8.0 nA 8.0 nA 8.0 nA 8.0 nA 8.0 nA 8.0 nA		
	10 μ A to 200 μ A 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 30 kHz At 30 kHz	20.0 nA 20.0 nA 20.0 nA 20.0 nA 20.0 nA 40.0 nA 60.0 nA 90.0 nA		
	200 μ A to 2 mA 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 30 kHz At 30 kHz	150 nA 120 nA 120 nA 120 nA 120 nA 250 nA 330 nA 390 nA		
	2 mA to 20 mA 10 Hz to 55 Hz 55 Hz to 305 Hz 305 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 30 kHz At 30 kHz	1.5 μ A 1.5 μ A 1.5 μ A 1.5 μ A 1.5 μ A 1.5 μ A 2.0 μ A 3.0 μ A		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
AC Current Measurement (continued)				
	20 mA to 200 mA	15 μ A		
	10 Hz to 55 Hz	12 μ A		
	55 Hz to 305 Hz	12 μ A		
	305 Hz to 1 kHz	12 μ A		
	1 kHz to 3 kHz	12 μ A		
	3 kHz to 10 kHz	12 μ A		
	10 kHz to 20 kHz	22 μ A		
	20 kHz to 30 kHz	30 μ A		
	At 30 kHz	30 μ A		
	200 mA to 2 A	220 μ A		
	10 Hz to 55 Hz	175 μ A		
	55 Hz to 305 Hz	175 μ A		
	305 Hz to 1 kHz	300 μ A		
	1 kHz to 3 kHz	450 μ A		
	3 kHz to 10 kHz	850 μ A		
	10 kHz to 20 kHz	630 μ A		
	20 kHz to 30 kHz	230 μ A		
	At 30 kHz			
	2 A to 20 A	20 mA		
	10 Hz to 55 Hz	20 mA		
	55 Hz to 305 Hz	20 mA		
	305 Hz to 1 kHz	20 mA		
	1 kHz to 3 kHz	20 mA		
	3 kHz to 10 kHz	20 mA		
	At 10 kHz	27 mA		
	20 A to 30 A	20 mA		
	10 Hz to 55 Hz	20 mA		
	55 Hz to 305 Hz	20 mA		
	305 Hz to 1 kHz	20 mA		
	1 kHz to 3 kHz	20 mA		
	3 kHz to 10 kHz	20 mA		
	At 10 kHz	27 mA		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (continued)			Calibration by comparison with a reference standard	UK
AC Current Generation	<i>55 Hz to 400 Hz</i> 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A <i>60 Hz</i> 1 A to 2 A 2 A to 10 A 10 A to 20 A 20 A to 50 A <i>400 Hz to 1 kHz</i> 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 200 mA to 1 A <i>1 kHz to 5 kHz</i> 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A <i>5 kHz to 10 kHz</i> 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A <i>At 10 kHz</i> 10 μ A to 100 μ A 100 μ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	20 nA 140 nA 580 nA 7.5 μ A 85 μ A 0.15 % 0.30 % 0.5 % 0.5 % 21 nA 135 nA 600 nA 7.5 μ A 95 μ A 30 nA 200 nA 1.0 μ A 12 μ A 200 μ A 100 nA 1.0 μ A 7.5 μ A 22 μ A 300 μ A 120 nA 1.2 μ A 7.5 μ A 25 μ A 320 μ A		
Frequency	10 MHz 1 Hz to 2 GHz	7.0 in 10^{11} 0.20 μ Hz/Hz	Measurement and generation of repetitive waveforms.	UK



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Temperature indicators, calibration by electrical simulation				UK
Base metal thermocouples	Type E -210 °C to +0 °C 0 °C to +980 °C	0.030 °C 0.020 °C	Excluding cold junction compensation.	
	Type J -190 °C to +0 °C 0 °C to +1170 °C	0.030 °C 0.020 °C		
	Type K -200 °C to -100 °C -100 °C to 900 °C	0.050 °C 0.030 °C		
	Type L -190 °C to +0 °C 0 °C to +880 °C	0.030 °C 0.020 °C		
	Type N -200 °C to +0 °C 0 °C to +1240 °C	0.070 °C 0.040 °C		
	Type T -200 °C to +0 °C 0 °C to +390 °C	0.050 °C 0.020 °C		
	Type U -170 °C to +0 °C 0 °C to +590 °C	0.075 °C 0.020 °C		
Nobel metal thermocouples	Type B 250 °C to +900 °C 900 °C to 1790 °C	0.250 °C 0.015 °C	Excluding cold junction compensation.	
	Type C 250 °C to +600 °C 600 °C to 2250 °C	0.025 °C 0.015 °C		
	Type D 200 °C to +600 °C 600 °C to +1790 °C	0.020 °C 0.025 °C		
	Type R -30 °C to +0 °C 0 °C to +1740 °C	0.18 °C 0.08 °C		
	Type S -30 °C to +0 °C 0 °C to 1740 °C	0.18 °C 0.08 °C		
Cold junction compensation	At 0 °C	0.21 °C		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Temperature indicators, calibration by electrical simulation (continued)				UK
Resistance thermometers	PT1000 -190 °C to 0 °C 0 °C to 390 °C	0.010 °C 0.025 °C		
	PT500 -190 °C to 0 °C 0 °C to 830 °C	0.010 °C 0.030 °C		
	PT200 -190 °C to 0 °C 0 °C to 830 °C	0.010 °C 0.030 °C		
	PT100 -190 °C to 0 °C 0 °C to 830 °C	0.010 °C 0.030 °C		
	PT50 -190 °C to 0 °C 0 °C to 830 °C	0.020 °C 0.035 °C		
	Ni100 -50 °C to 230 °C	0.010 °C		
	Ni120 -70 °C to 250 °C	0.010 °C		
	Cu10 -190 °C to 0 °C 0 °C to 250 °C	0.025 °C 0.10 °C		
	D-100 -190 °C to 0 °C 0 °C to 490 °C	0.012 °C 0.020 °C		
Temperature simulators, calibration by electrical simulation				
Base metal thermocouples	Type E -200 °C to +0 °C 0 °C to +980 °C	0.055 °C 0.030 °C	Excluding cold junction compensation.	
	Type J -190 °C to +0 °C 0 °C to +1170 °C	0.055 °C 0.035 °C		
	Type K -200 °C to -100 °C -100 °C to 900 °C 900 °C to 1300 °C	0.100 °C 0.050 °C 0.060 °C		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Temperature simulators, calibration by electrical simulation (continued)				UK
Base metal thermocouples (continued)	Type L -190 °C to +0 °C 0 °C to +880 °C	0.050 °C 0.030 °C	Excluding cold junction compensation.	
	Type N -200 °C to +0 °C 0 °C to +1240 °C	0.070 °C 0.060 °C		
	Type T -200 °C to +0 °C 0 °C to +390 °C	0.090 °C 0.040 °C		
	Type U -170 °C to +0 °C 0 °C to +590 °C	0.090 °C 0.040 °C		
Nobel metal thermocouples	Type B 250 °C to +900 °C 900 °C to 1790 °C	0.460 °C 0.180 °C	Excluding cold junction compensation.	
	Type C 250 °C to +600 °C 600 °C to 2250 °C	0.100 °C 0.180 °C		
	Type D 200 °C to +600 °C 600 °C to +1790 °C	0.100 °C 0.120 °C		
	Type R -30 °C to +0 °C 0 °C to +1740 °C	0.350 °C 0.150 °C		
	Type S -30 °C to +0 °C 0 °C to 1740 °C	0.350 °C 0.150 °C		
Resistance thermometers	PT1000 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 850 °C	0.0012 °C 0.0035 °C 0.0068 °C		
	PT500 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 850 °C	0.0020 °C 0.0022 °C 0.0050 °C		
	PT200 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 850 °C	0.0020 °C 0.0022 °C 0.0050 °C		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Temperature simulators, calibration by electrical simulation (continued)				UK
Resistance thermometers (continued)	PT100 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 850 °C	0.0020 °C 0.0022 °C 0.0050 °C		
	PT50 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 850 °C	0.0022 °C 0.0042 °C 0.0093 °C		
	Ni100 -60 °C to 0 °C 0 °C to 180 °C	0.0015 °C 0.0018 °C		
	Ni120 -80 °C to 0 °C 0 °C to 260 °C	0.0015 °C 0.0015 °C		
	Cu10 -200 °C to 0 °C 0 °C to 260 °C	0.0025 °C 0.0042 °C		
	D-100 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 500 °C	0.0019 °C 0.0034 °C 0.0030 °C		



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
MASS	Nominal value (grams) 0.001 to 0.02 0.05 to 0.2 0.5 1 2 5 10 20 50 100 200 500 1 000 2 000 5 000 10 000 20 000 30 000	(mg) 0.012 0.015 0.018 0.020 0.024 0.030 0.040 0.050 0.060 0.10 0.20 0.50 1.0 2.0 5.0 10 20 30	Intermediate values under 100g can be calibrated with an uncertainty equal to the uncertainty of the next higher nominal value. Intermediate values over 100 g can be calibrated with an uncertainty of 1.0 mg/kg. Substitution Method.	UK
TEMPERATURE Resistance thermometers and electronic thermometers with PRT, thermocouple or thermistor sensors	-75 °C to -40 °C -40 °C to -30 °C -30 °C to 0 °C 0 °C (ice point) 0.01 °C (Triple Point of Water) 0 °C to 60 °C 60 °C to 150 °C 150 °C to 200 °C 200 °C to 250 °C	0.034 °C 0.028 °C 0.017 °C 0.018 °C 0.002 0 °C 0.012 °C 0.020 °C 0.026 °C 0.033 °C	Calibration performed in liquid baths.	UK
HUMIDITY Relative humidity meters	11 %rh 33 %rh 54 %rh 75 %rh 90 %rh For the temperature range 21 °C ± 3 °C	2.2 %rh 2.2 %rh 2.2 %rh 2.2 %rh 2.2 %rh	Calibrations by comparison with dew- point hygrometer and Platinum Resistance Thermometers or against salt solutions.	UK
END				



0221
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Druck Limited
Issue No: 080 Issue date: 01 June 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$