


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|---|--|--|
|  <p>0246</p> <p>Accredited to ISO/IEC 17025:2017</p> | <p>DE&S Deca</p> <p>Issue No: 051 Issue date: 15 May 2025</p> | |
| | <p>Deca Sealand Welsh Road Deeside Flintshire CH5 2LS</p> | <p>Contact: Mr Ian Ford Tel: +44 (0)1244 847242 Fax: +44 (0)1244 847058 E-Mail: Ian.Ford@deca.mod.uk Website: www.gov.uk/deca</p> |

Calibration performed at the above address only

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|---|--|--|--|
| ELECTRICAL | | | All electrical calibrations are performed as a comparison against a reference standard unless otherwise stated |
| FREQUENCY | | | |
| Specific Values | 100 kHz 1 MHz 5 MHz 10 MHz | 2.3 in 10^{12} 2.3 in 10^{12} 2.3 in 10^{12} 2.3 in 10^{12} | For the calibration of Frequency equipment, timers / counters |
| RESISTANCE | 0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω | 18 $\mu\Omega/\Omega + 1.0 \mu\Omega$ 11 $\mu\Omega/\Omega + 7.0 \mu\Omega$ 10 $\mu\Omega/\Omega + 43 \mu\Omega$ 10 $\mu\Omega/\Omega + 200 \mu\Omega$ 10 $\mu\Omega/\Omega + 2.0 m\Omega$ 10 $\mu\Omega/\Omega + 20 m\Omega$ 12 $\mu\Omega/\Omega + 200 m\Omega$ 27 $\mu\Omega/\Omega + 4.0 \Omega$ 210 $\mu\Omega/\Omega + 200 \Omega$ | For measurement of instrument outputs |
| DC/LF MULTIFUNCTION TRANSFER STANDARD SYSTEM | | | |
| DC RESISTANCE | 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω | 12 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 7.5 $\mu\Omega/\Omega$ 5.8 $\mu\Omega/\Omega$ 9.9 $\mu\Omega/\Omega$ 17 $\mu\Omega/\Omega$ 67 $\mu\Omega/\Omega$ 250 $\mu\Omega/\Omega$ | Sourcing resistance for measuring instruments |
| DC VOLTAGE | 100 mV 1 V 10 V 19 V 100 V 1000 V | 11 $\mu V/V$ 4.5 $\mu V/V$ 4.3 $\mu V/V$ 4.3 $\mu V/V$ 7.0 $\mu V/V$ 6.1 $\mu V/V$ | Sourcing and measurement capability for the calibration of voltage instruments |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca**Issue No: 051 Issue date: 15 May 2025****Calibration performed at main address only**

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---|---|--|
| DC/LF MULTIFUNCTION TRANSFER STANDARD SYSTEM (cont'd) | | | |
| DC CURRENT | 100 μ A 1 mA 10 mA 100 mA 1 A 10 A | 39 μ A/A 35 μ A/A 30 μ A/A 36 μ A/A 51 μ A/A 73 μ A/A | Sourcing and measurement capability for the calibration of current instruments |
| AC VOLTAGE | 1 mV 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz 50 kHz 100 kHz 10 mV 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz 50 kHz 100 kHz 100 mV 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz 50 kHz 100 kHz | 0.21 % 0.21 % 0.21 % 0.22 % 0.21 % 0.21 % 0.21 % 0.21 % 0.21 % 0.21 % 0.21 % 0.32 % 320 μ V/V 310 μ V/V 310 μ V/V 310 μ V/V 300 μ V/V 310 μ V/V 300 μ V/V 310 μ V/V 310 μ V/V 360 μ V/V 340 μ V/V 530 μ V/V 130 μ V/V 120 μ V/V 130 μ V/V 160 μ V/V 120 μ V/V 110 μ V/V 110 μ V/V 130 μ V/V 180 μ V/V 190 μ V/V 370 μ V/V | Sourcing and measurement capability for the calibration of voltage instruments |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 051 **Issue date:** 15 May 2025

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|---------|---|---------|
| AC VOLTAGE (cont'd) | 1 V | | |
| | 10 Hz | 49 $\mu\text{V/V}$ | |
| | 20 Hz | 51 $\mu\text{V/V}$ | |
| | 30 Hz | 46 $\mu\text{V/V}$ | |
| | 40 Hz | 41 $\mu\text{V/V}$ | |
| | 55 Hz | 41 $\mu\text{V/V}$ | |
| | 300 Hz | 39 $\mu\text{V/V}$ | |
| | 1 kHz | 45 $\mu\text{V/V}$ | |
| | 10 kHz | 41 $\mu\text{V/V}$ | |
| | 20 kHz | 40 $\mu\text{V/V}$ | |
| | 30 kHz | 40 $\mu\text{V/V}$ | |
| | 50 kHz | 47 $\mu\text{V/V}$ | |
| | 100 kHz | 78 $\mu\text{V/V}$ | |
| | 300 kHz | 180 $\mu\text{V/V}$ | |
| | 500 kHz | 470 $\mu\text{V/V}$ | |
| | 1 MHz | 0.16 % | |
| | 10 V | | |
| | 10 Hz | 55 $\mu\text{V/V}$ | |
| | 20 Hz | 48 $\mu\text{V/V}$ | |
| | 30 Hz | 53 $\mu\text{V/V}$ | |
| | 40 Hz | 41 $\mu\text{V/V}$ | |
| | 55 Hz | 40 $\mu\text{V/V}$ | |
| | 300 Hz | 40 $\mu\text{V/V}$ | |
| | 1 kHz | 45 $\mu\text{V/V}$ | |
| | 19 V | | |
| | 1 kHz | 39 $\mu\text{V/V}$ | |
| | 10 kHz | 36 $\mu\text{V/V}$ | |
| | 20 kHz | 38 $\mu\text{V/V}$ | |
| | 30 kHz | 46 $\mu\text{V/V}$ | |
| | 50 kHz | 57 $\mu\text{V/V}$ | |
| | 100 kHz | 91 $\mu\text{V/V}$ | |
| | 300 kHz | 160 $\mu\text{V/V}$ | |
| | 500 kHz | 710 $\mu\text{V/V}$ | |
| | 1 MHz | 0.14 % | |
| | 100 V | | |
| | 10 Hz | 55 $\mu\text{V/V}$ | |
| | 20 Hz | 51 $\mu\text{V/V}$ | |
| | 30 Hz | 50 $\mu\text{V/V}$ | |
| | 40 Hz | 52 $\mu\text{V/V}$ | |
| | 55 Hz | 48 $\mu\text{V/V}$ | |
| | 300 Hz | 44 $\mu\text{V/V}$ | |
| | 1 kHz | 42 $\mu\text{V/V}$ | |
| | 10 kHz | 43 $\mu\text{V/V}$ | |
| | 20 kHz | 44 $\mu\text{V/V}$ | |
| | 30 kHz | 45 $\mu\text{V/V}$ | |
| | 50 kHz | 51 $\mu\text{V/V}$ | |
| | 100 kHz | 100 $\mu\text{V/V}$ | |
| | 700 V | | |
| | 50 kHz | 190 $\mu\text{V/V}$ | |
| | 100 kHz | 480 $\mu\text{V/V}$ | |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca**Issue No:** 051 **Issue date:** 15 May 2025**Calibration performed at main address only**

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|--|---|--|
| AC VOLTAGE (cont'd) | 1000 V 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz | 53 $\mu\text{V/V}$ 51 $\mu\text{V/V}$ 52 $\mu\text{V/V}$ 49 $\mu\text{V/V}$ 60 $\mu\text{V/V}$ 63 $\mu\text{V/V}$ 89 $\mu\text{V/V}$ | Sourcing and measurement capability for the calibration of current instruments |
| AC CURRENT | 100 μA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz | 190 $\mu\text{A/A}$ 190 $\mu\text{A/A}$ 180 $\mu\text{A/A}$ 180 $\mu\text{A/A}$ 190 $\mu\text{A/A}$ 180 $\mu\text{A/A}$ 170 $\mu\text{A/A}$ 390 $\mu\text{A/A}$ | |
| | 1 mA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz | 140 $\mu\text{A/A}$ 150 $\mu\text{A/A}$ 130 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 170 $\mu\text{A/A}$ | |
| | 10 mA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz | 140 $\mu\text{A/A}$ 130 $\mu\text{A/A}$ 120 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 170 $\mu\text{A/A}$ | |
| | 100 mA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz | 140 $\mu\text{A/A}$ 130 $\mu\text{A/A}$ 130 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 110 $\mu\text{A/A}$ 170 $\mu\text{A/A}$ | |
| | 1 A 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz | 180 $\mu\text{A/A}$ 180 $\mu\text{A/A}$ 170 $\mu\text{A/A}$ 140 $\mu\text{A/A}$ 140 $\mu\text{A/A}$ 130 $\mu\text{A/A}$ 180 $\mu\text{A/A}$ 300 $\mu\text{A/A}$ | |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 051 Issue date: 15 May 2025

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|---|--|--|---|
| AC CURRENT (cont'd) | 10 A 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 10 kHz | 370 μ A/A 380 μ A/A 380 μ A/A 360 μ A/A 360 μ A/A 370 μ A/A 390 μ A/A 640 μ A/A 0.12 % | |
| DC/LF AUTOMATED CALIBRATION SYSTEM | | | |
| DC VOLTAGE | 0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V | 10 μ V/V + 0.60 μ V 7.9 μ V/V + 2.2 μ V 5.6 μ V/V + 6.7 μ V 7.3 μ V/V + 110 μ V 9.0 μ V/V + 0.30 mV | Sourcing and measurement capability for the calibration of voltage instruments |
| DC CURRENT | 0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A | 150 μ A/A + 2.3 nA 65 μ A/A + 12 nA 70 μ A/A + 120 nA 65 μ A/A + 1.2 μ A 150 μ A/A + 24 μ A | Sourcing and measurement capability for the calibration of current instruments |
| DC RESISTANCE | 0 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω | 10 m Ω (2-wire configuration) 35 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 16 $\mu\Omega/\Omega$ 36 $\mu\Omega/\Omega$ 70 $\mu\Omega/\Omega$ 250 $\mu\Omega/\Omega$ | The zero uncertainty for 4-wire configurations will be lower than that for 2-wire configurations and will largely depend on the resolution of the instrument being calibrated. |
| AC VOLTAGE | 2 mV to 200 mV 20 Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 40 kHz 40 kHz to 100 kHz 200 mV to 2 V 20 Hz to 300 Hz 300 Hz to 500 Hz 500 Hz to 30 kHz 30 kHz to 40 kHz 40 kHz to 100 kHz 2 V to 20 V 20 Hz to 300 Hz 300 Hz to 500 Hz 500 Hz to 30 kHz 30 kHz to 100 kHz | 190 μ V/V + 11 μ V 140 μ V/V + 11 μ V 150 μ V/V + 11 μ V 400 μ V/V + 11 μ V 530 μ V/V + 11 μ V 110 μ V/V + 47 μ V 65 μ V/V + 47 μ V 55 μ V/V + 47 μ V 100 μ V/V + 47 μ V 110 μ V/V + 47 μ V 110 μ V/V + 0.47 mV 65 μ V/V + 0.47 mV 55 μ V/V + 0.47 mV 100 μ V/V + 0.47 mV | Sourcing and measurement capability for the calibration of voltage instruments |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 051 Issue date: 15 May 2025

Calibration performed at main address only

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|--|--|--|--|
| AC VOLTAGE (cont'd) | 20 V to 200 V 20 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 30 kHz 30 kHz to 40 kHz 40 kHz to 100 kHz 200 V to 1000 V 500 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 30 kHz | 130 $\mu\text{V/V} + 4.7 \text{ mV}$ 76 $\mu\text{V/V} + 4.7 \text{ mV}$ 76 $\mu\text{V/V} + 4.7 \text{ mV}$ 78 $\mu\text{V/V} + 4.7 \text{ mV}$ 150 $\mu\text{V/V} + 4.7 \text{ mV}$ 160 $\mu\text{V/V} + 4.7 \text{ mV}$ 130 $\mu\text{V/V} + 47 \text{ mV}$ 170 $\mu\text{V/V} + 47 \text{ mV}$ 180 $\mu\text{V/V} + 47 \text{ mV}$ | |
| AC CURRENT | 2 μA to 200 μA 40 Hz to 5 kHz 200 μA to 2 mA 40 Hz to 5 kHz 2 mA to 20 mA 40 Hz to 5 kHz 20 mA to 200 mA 40 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 5 kHz 200 mA to 2 A 40 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 5 kHz | 540 $\mu\text{A/A} + 200 \text{ nA}$ 360 $\mu\text{A/A} + 280 \text{ nA}$ 370 $\mu\text{A/A} + 280 \text{ nA}$ 470 $\mu\text{A/A} + 170 \mu\text{A}$ 700 $\mu\text{A/A} + 300 \mu\text{A}$ 470 $\mu\text{A/A} + 300 \mu\text{A}$ 470 $\mu\text{A/A} + 490 \mu\text{A}$ 690 $\mu\text{A/A} + 690 \mu\text{A}$ 720 $\mu\text{A/A} + 690 \mu\text{A}$ | Sourcing and measurement capability for the calibration of voltage instruments |
| PRESSURE | | | Methods consistent with EURAMET CG17. |
| Gas pressure (absolute) | | | |
| Calibration of pressure measuring instruments and gauges | 3.5 kPa to 175 kPa 175 kPa to 700 kPa | 0.005 0 % + 2.2 Pa 0.004 6 % + 2.2 Pa | Calibration against a deadweight tester standard |
| Gas pressure (gauge) | | | |
| Calibration of pressure measuring instruments and gauges | -100 kPa to -20 kPa -20 kPa to 0 kPa 3.5 kPa to 175 kPa 175 kPa to 700 kPa 700 kPa to 3.5 MPa | 0.004 3 % 70 Pa 0.004 3 % 0.004 3 % 0.005 7 % | Calibration against a deadweight tester standard |
| Hydraulic pressure (gauge) | | | |
| Calibration of pressure measuring instruments and gauges | 0.35 MPa to 0.69 MPa 0.69 MPa to 6.9 MPa 6.9 MPa to 69 MPa | 0.011 % 0.008 5 % 0.005 8 % | Calibration against a deadweight tester standard |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca**Issue No: 051 Issue date: 15 May 2025****Calibration performed at main address only**

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks |
|---|---|---|---|
| FORCE Calibration of push strength testers in compression Calibration of force push pull Devices in compression only | 150 N to 500 N 50 N to 500 N | 6.0% of applied force 2.0 % | Calibrated using alignment guides |
| TEMPERATURE Temperature Indicators and/or recorders with temperature sensors | 5 °C to 60 °C 60 °C to 120 °C | 0.037 °C 0.057 °C | Calibration against a reference standard in stirred liquid baths |
| ROTATIONAL SPEED Optical Tachometers | 45 RPM to 1000 RPM 1000 RPM to 100,000 RPM | 0.011% + 0.050 RPM 0.011% + 0.50 RPM | Calibration against a reference standard |
| TORQUE Hand Torque Tools Hand Torque Tools | BS EN ISO 6789:2017 0.04 N·m to 1500 N·m For screwdrivers (Type II, Classes D, E and F 1.0 N·m to 10 N·m BS EN ISO 6789:2003 (withdrawn & superseded) 0.04 N·m to 1500 N·m | 1.0% of reading 1.0% of reading 1.0% of reading | The uncertainty quoted is both for the application of the calibration torque and the device being calibrated Calibrations may also be given in lb·in & lb·ft |
| END | | | |



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 051 Issue date: 15 May 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$