


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>0246</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>DE&S Deca</h3> <p>Issue No: 052 Issue date: 26 March 2026</p>	
	<p>Deca Sealand Welsh Road Deeside Flintshire CH5 2LS</p>	<p>Contact: Mr Ian Ford Tel: +44 (0)1244 847242 Fax: +44 (0)1244 847058 E-Mail: Ian.Ford@deca.mod.uk Website: www.gov.uk/deca</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL			All electrical calibrations are performed as a comparison against a reference standard unless otherwise stated
FREQUENCY			
Specific Values	100 kHz 1 MHz 5 MHz 10 MHz	$2.3 \text{ in } 10^{12}$ $2.3 \text{ in } 10^{12}$ $2.3 \text{ in } 10^{12}$ $2.3 \text{ in } 10^{12}$	For the calibration of Frequency equipment, timers / counters
RESISTANCE	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω	18 $\mu\Omega/\Omega + 1.0 \mu\Omega$ 11 $\mu\Omega/\Omega + 7.0 \mu\Omega$ 10 $\mu\Omega/\Omega + 43 \mu\Omega$ 10 $\mu\Omega/\Omega + 200 \mu\Omega$ 10 $\mu\Omega/\Omega + 2.0 \text{ m}\Omega$ 10 $\mu\Omega/\Omega + 20 \text{ m}\Omega$ 12 $\mu\Omega/\Omega + 200 \text{ m}\Omega$ 27 $\mu\Omega/\Omega + 4.0 \Omega$ 210 $\mu\Omega/\Omega + 200 \Omega$	For measurement of instrument outputs
DC/LF MULTIFUNCTION TRANSFER STANDARD SYSTEM			
DC RESISTANCE	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	12 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 7.5 $\mu\Omega/\Omega$ 5.8 $\mu\Omega/\Omega$ 9.9 $\mu\Omega/\Omega$ 17 $\mu\Omega/\Omega$ 67 $\mu\Omega/\Omega$ 250 $\mu\Omega/\Omega$	Sourcing resistance for measuring instruments
DC VOLTAGE	100 mV 1 V 10 V 19 V 100 V 1000 V	11 $\mu\text{V}/\text{V}$ 4.5 $\mu\text{V}/\text{V}$ 4.3 $\mu\text{V}/\text{V}$ 4.3 $\mu\text{V}/\text{V}$ 7.0 $\mu\text{V}/\text{V}$ 6.1 $\mu\text{V}/\text{V}$	Sourcing and measurement capability for the calibration of voltage instruments



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
DC/LF MULTIFUNCTION TRANSFER STANDARD SYSTEM (cont'd)			
DC CURRENT	100 μ A 1 mA 10 mA 100 mA 1 A 10 A	39 μ A/A 35 μ A/A 30 μ A/A 36 μ A/A 51 μ A/A 73 μ A/A	Sourcing and measurement capability for the calibration of current instruments
AC VOLTAGE	1 mV 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz 50 kHz 100 kHz 10 mV 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz 50 kHz 100 kHz 100 mV 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz 50 kHz 100 kHz	0.21 % 0.21 % 0.21 % 0.22 % 0.21 % 0.21 % 0.21 % 0.21 % 0.21 % 0.21 % 0.21 % 0.32 % 320 μ V/V 310 μ V/V 310 μ V/V 310 μ V/V 300 μ V/V 310 μ V/V 300 μ V/V 310 μ V/V 310 μ V/V 360 μ V/V 340 μ V/V 530 μ V/V 130 μ V/V 120 μ V/V 130 μ V/V 160 μ V/V 120 μ V/V 110 μ V/V 110 μ V/V 130 μ V/V 180 μ V/V 190 μ V/V 370 μ V/V	Sourcing and measurement capability for the calibration of voltage instruments



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	
AC VOLTAGE (cont'd)	1 V			
	10 Hz	49 μ V/V		
	20 Hz	51 μ V/V		
	30 Hz	46 μ V/V		
	40 Hz	41 μ V/V		
	55 Hz	41 μ V/V		
	300 Hz	39 μ V/V		
	1 kHz	45 μ V/V		
	10 kHz	41 μ V/V		
	20 kHz	40 μ V/V		
	30 kHz	40 μ V/V		
	50 kHz	47 μ V/V		
	100 kHz	78 μ V/V		
	300 kHz	180 μ V/V		
	500 kHz	470 μ V/V		
	1 MHz	0.16 %		
	10 V			
	10 Hz	55 μ V/V		
	20 Hz	48 μ V/V		
	30 Hz	53 μ V/V		
	40 Hz	41 μ V/V		
	55 Hz	40 μ V/V		
	300 Hz	40 μ V/V		
	1 kHz	45 μ V/V		
	19 V			
	1 kHz	39 μ V/V		
	10 kHz	36 μ V/V		
	20 kHz	38 μ V/V		
	30 kHz	46 μ V/V		
	50 kHz	57 μ V/V		
	100 kHz	91 μ V/V		
	300 kHz	160 μ V/V		
	500 kHz	710 μ V/V		
	1 MHz	0.14 %		
	100 V			
	10 Hz	55 μ V/V		
	20 Hz	51 μ V/V		
	30 Hz	50 μ V/V		
	40 Hz	52 μ V/V		
	55 Hz	48 μ V/V		
	300 Hz	44 μ V/V		
	1 kHz	42 μ V/V		
	10 kHz	43 μ V/V		
20 kHz	44 μ V/V			
30 kHz	45 μ V/V			
50 kHz	51 μ V/V			
100 kHz	100 μ V/V			
700 V				
50 kHz	190 μ V/V			
100 kHz	480 μ V/V			



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC VOLTAGE (cont'd)	1000 V 40 Hz 55 Hz 300 Hz 1 kHz 10 kHz 20 kHz 30 kHz	53 μ V/V 51 μ V/V 52 μ V/V 49 μ V/V 60 μ V/V 63 μ V/V 89 μ V/V	
AC CURRENT	100 μ A 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 1 mA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 10 mA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 100 mA 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 1 A 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz	190 μ A/A 190 μ A/A 180 μ A/A 180 μ A/A 190 μ A/A 180 μ A/A 170 μ A/A 390 μ A/A 140 μ A/A 150 μ A/A 130 μ A/A 110 μ A/A 110 μ A/A 110 μ A/A 110 μ A/A 170 μ A/A 140 μ A/A 130 μ A/A 120 μ A/A 110 μ A/A 110 μ A/A 110 μ A/A 110 μ A/A 170 μ A/A 140 μ A/A 130 μ A/A 130 μ A/A 110 μ A/A 110 μ A/A 110 μ A/A 110 μ A/A 170 μ A/A 180 μ A/A 180 μ A/A 170 μ A/A 140 μ A/A 140 μ A/A 130 μ A/A 180 μ A/A 300 μ A/A	Sourcing and measurement capability for the calibration of current instruments



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC CURRENT (cont'd)	10 A 10 Hz 20 Hz 30 Hz 40 Hz 55 Hz 300 Hz 1 kHz 5 kHz 10 kHz	370 μ A/A 380 μ A/A 380 μ A/A 360 μ A/A 360 μ A/A 370 μ A/A 390 μ A/A 640 μ A/A 0.12 %	
DC/LF AUTOMATED CALIBRATION SYSTEM			
DC VOLTAGE	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	15 μ V/V + 1.4 μ V 7.9 μ V/V + 2.6 μ V 7.8 μ V/V + 6.7 μ V 7.5 μ V/V + 110 μ V 9.0 μ V/V + 0.30 mV	Sourcing and measurement capability for the calibration of voltage instruments
DC CURRENT	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A	160 μ A/A + 2.3 nA 70 μ A/A + 12 nA 75 μ A/A + 120 nA 70 μ A/A + 1.2 μ A 160 μ A/A + 24 μ A 200 μ A/A + 180 μ A	Sourcing and measurement capability for the calibration of current instruments
DC RESISTANCE	0 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	10 m Ω (2-wire configuration) 50 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 35 $\mu\Omega/\Omega$ 630 $\mu\Omega/\Omega$ 250 $\mu\Omega/\Omega$ 500 $\mu\Omega/\Omega$	The zero uncertainty for 4-wire configurations will be lower than that for 2-wire configurations and will largely depend on the resolution of the instrument being calibrated.
AC VOLTAGE	2 mV to 200 mV 30 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	200 μ V/V + 11 μ V 160 μ V/V + 11 μ V 180 μ V/V + 11 μ V 540 μ V/V + 11 μ V	Sourcing and measurement capability for the calibration of voltage instruments
	200 mV to 2 V 30 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 30 kHz 30 kHz to 100 kHz	100 μ V/V + 47 μ V 100 μ V/V + 47 μ V 50 μ V/V + 47 μ V 180 μ V/V + 47 μ V	
	2 V to 20 V 30 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 30 kHz 30 kHz to 100 kHz	100 μ V/V + 0.47 mV 100 μ V/V + 0.47 mV 100 μ V/V + 0.47 mV 210 μ V/V + 0.47 mV	



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC VOLTAGE (cont'd)	20 V to 200 V 30 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	100 μ V/V + 4.7 mV 100 μ V/V + 4.7 mV 100 μ V/V + 4.7 mV 200 μ V/V + 4.7 mV 500 μ V/V + 4.7 mV	
AC CURRENT	200 V to 1000 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	150 μ V/V + 47 mV 150 μ V/V + 47 mV 210 μ V/V + 47 mV	Sourcing and measurement capability for the calibration of voltage instruments
	2 μ A to 200 μ A 40 Hz to 5 kHz	350 μ A/A + 200 nA	
	200 μ A to 2 mA 40 Hz to 5 kHz	360 μ A/A + 280 nA	
	2 mA to 20 mA 40 Hz to 5 kHz	550 μ A/A + 280 nA	
	20 mA to 200 mA 300 Hz to 1 kHz 1 kHz to 5 kHz	200 μ A/A + 300 μ A 300 μ A/A + 300 μ A	
PRESSURE			Methods consistent with EURAMET CG17.
Gas pressure (absolute)			
Calibration of pressure measuring instruments and gauges	3.5 kPa to 175 kPa 175 kPa to 700 kPa	Q[0.005 0 %, 2.2 Pa] Q[0.004 6 %, 2.2 Pa]	Calibration against a deadweight tester standard
Gas pressure (gauge)			
Calibration of pressure measuring instruments and gauges	-100 kPa to -20 kPa -20 kPa to 0 kPa 3.5 kPa to 175 kPa 175 kPa to 700 kPa 700 kPa to 3.5 MPa	0.004 3 % 70 Pa 0.004 3 % 0.004 3 % 0.005 7 %	Calibration against a deadweight tester standard
	-100 kPa to 0 kPa 0 to 40 kPa 40 kPa to 250 kPa 250 kPa to 3.5 MPa	Q[0.003 3 %, 63 Pa] Q[0.007 1 %, 9.5 Pa] Q[0.005 0 %, 70 Pa] Q[0.006 0 %, 0.74 kPa]	Calibration against digital pressure controller



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
PRESSURE (cont'd) Hydraulic pressure (gauge) Calibration of pressure measuring instruments and gauges	0.35 MPa to 0.69 MPa 0.69 MPa to 6.9 MPa 6.9 MPa to 69 MPa 0 MPa to 7 MPa 7 MPa to 35 MPa 35 MPa to 70 MPa	0.011 % 0.008 5 % 0.005 8 % Q[0.008 8%, 2.3 kPa] Q[0.005 8%, 15 kPa] Q[0.005 8%, 11 kPa]	Calibration against a deadweight tester standard Calibration against digital pressure gauges
FORCE Calibration of push strength testers in compression Calibration of force push pull Devices in compression only	150 N to 500 N 50 N to 500 N	6.0% of applied force 2.0 %	Calibrated using alignment guides
TEMPERATURE Temperature Indicators and/or recorders with temperature sensors	5 °C to 60 °C 60 °C to 120 °C	0.037 °C 0.057 °C	Calibration against a reference standard in stirred liquid baths
ROTATIONAL SPEED Optical Tachometers	45 RPM to 1000 RPM 1000 RPM to 100,000 RPM	0.011% + 0.050 RPM 0.011% + 0.50 RPM	Calibration against a reference standard
TORQUE Hand Torque Tools Hand Torque Tools	BS EN ISO 6789:2017 0.04 N·m to 1500 N·m For screwdrivers (Type II, Classes D, E and F) 1.0 N·m to 10 N·m BS EN ISO 6789:2003 (withdrawn & superseded) 0.04 N·m to 1500 N·m	1.0% of reading 1.0% of reading 1.0% of reading	The uncertainty quoted is both for the application of the calibration torque and the device being calibrated Calibrations may also be given in lb·in & lb·ft
END			



0246

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DE&S Deca

Issue No: 052 Issue date: 26 March 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$