


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0254 Accredited to ISO/IEC 17025:2017	Devonport Royal Dockyard Limited trading as Devonport Electrical and Nucleonic Calibration Facility	
	Issue No: 043 Issue date: 11 April 2025	
	Babcock International Group PC1409, Devonport Royal Dockyard Plymouth Devon PL1 4SG	Contact: Mr D Fisher Tel: +44 (0)1752 323805 Fax: +44 (0)1752 566357 E-Mail: David.Fisher@babcockinternational.com
Calibration performed at the above address only		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL CALIBRATION			Electrical calibrations are as a comparison against a reference standard
DC Resistance			
Generation			
Specific Values	1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	61 $\mu\Omega/\Omega$ 58 $\mu\Omega/\Omega$ 32 $\mu\Omega/\Omega$ 22 $\mu\Omega/\Omega$ 21 $\mu\Omega/\Omega$ 22 $\mu\Omega/\Omega$ 24 $\mu\Omega/\Omega$ 29 $\mu\Omega/\Omega$ 110 $\mu\Omega/\Omega$	
Measurement	0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω 2 G Ω to 100 G Ω	20 $\mu\Omega/\Omega + 0.20$ m Ω 20 $\mu\Omega/\Omega + 0.60$ m Ω 20 $\mu\Omega/\Omega + 0.60$ m Ω 15 $\mu\Omega/\Omega + 7.0$ m Ω 15 $\mu\Omega/\Omega + 65$ m Ω 25 $\mu\Omega/\Omega + 1.5$ Ω 150 $\mu\Omega/\Omega + 80$ Ω 200 $\mu\Omega/\Omega + 9.0$ k Ω 500 $\mu\Omega/\Omega + 1.0$ M Ω 2.0 %	
DC Voltage			
Generation	0 mV to 220 mV 220 mV to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	13 $\mu V/V + 2.0$ μV 5.0 $\mu V/V + 2.0$ μV 6.0 $\mu V/V + 4.0$ μV 6.0 $\mu V/V + 8.0$ μV 6.0 $\mu V/V + 0.10$ mV 6.0 $\mu V/V + 0.60$ mV	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
DC Voltage (continued)			
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	10 $\mu\text{V/V} + 2.0 \mu\text{V}$ 10 $\mu\text{V/V} + 2.0 \mu\text{V}$ 10 $\mu\text{V/V} + 3.0 \mu\text{V}$ 10 $\mu\text{V/V} + 40 \mu\text{V}$ 10 $\mu\text{V/V} + 100 \mu\text{V}$	
DC Current			
Generation	0 μA to 220 μA 220 μA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A	52 $\mu\text{A/A} + 10 \text{ nA}$ 15 $\mu\text{A/A} + 10 \text{ nA}$ 30 $\mu\text{A/A} + 100 \text{ nA}$ 37 $\mu\text{A/A} + 1.0 \mu\text{A}$ 100 $\mu\text{A/A} + 30 \mu\text{A}$ 190 $\mu\text{A/A} + 480 \mu\text{A}$	
Measurement	0 μA to 200 μA 200 μA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A 1 A to 10 A	50 $\mu\text{A/A} + 1.0 \text{ nA}$ 30 $\mu\text{A/A} + 5.0 \text{ nA}$ 40 $\mu\text{A/A} + 40 \text{ nA}$ 40 $\mu\text{A/A} + 1.0 \mu\text{A}$ 52 $\mu\text{A/A} + 10 \mu\text{A}$ 50 $\mu\text{A/A}$	
AC Voltage			
Generation	<p><i>40 Hz to 20 kHz</i> 10 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V 220 V to 750 V 750 V to 1100 V</p> <p><i>20 kHz to 50 kHz</i> 10 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V 220 V to 750 V</p> <p><i>50 kHz to 100 kHz</i> 10 mV to 220 mV 220 mV to 2.2 V 2.2 V to 22 V 22 V to 220 V</p> <p><i>100 kHz to 1 MHz</i> 220 mV to 2.2 V 2.2 V to 22 V</p>	<p>56 $\mu\text{V/V} + 10 \mu\text{V}$ 24 $\mu\text{V/V} + 20 \mu\text{V}$ 22 $\mu\text{V/V} + 70 \mu\text{V}$ 63 $\mu\text{V/V} + 2.0 \text{ mV}$ 110 $\mu\text{V/V} + 6.0 \text{ mV}$ 270 $\mu\text{V/V} + 11 \text{ mV}$</p> <p>82 $\mu\text{V/V} + 10 \mu\text{V}$ 27 $\mu\text{V/V} + 80 \mu\text{V}$ 22 $\mu\text{V/V} + 200 \mu\text{V}$ 91 $\mu\text{V/V} + 4.0 \text{ mV}$ 270 $\mu\text{V/V} + 11 \text{ mV}$</p> <p>160 $\mu\text{V/V} + 30 \mu\text{V}$ 56 $\mu\text{V/V} + 80 \mu\text{V}$ 52 $\mu\text{V/V} + 400 \mu\text{V}$ 200 $\mu\text{V/V} + 10 \text{ mV}$</p> <p>0.17 % + 1.0 mV 0.37 % + 9.0 mV</p>	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC Voltage (continued)			
Measurement	<i>20 Hz to 10 kHz</i> 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V <i>10 kHz to 100 kHz</i> 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V <i>55 Hz to 3 kHz</i> 200 V to 1000 V <i>3 kHz to 30 kHz</i> 200 V to 1000 V	270 $\mu\text{V/V} + 4.0 \mu\text{V}$ 680 $\mu\text{V/V} + 20 \mu\text{V}$ 260 $\mu\text{V/V} + 200 \mu\text{V}$ 230 $\mu\text{V/V} + 2.0 \text{ mV}$ 690 $\mu\text{V/V} + 20 \mu\text{V}$ 600 $\mu\text{V/V} + 200 \mu\text{V}$ 550 $\mu\text{V/V} + 2.0 \text{ mV}$ 700 $\mu\text{V/V} + 20 \text{ mV}$ 580 $\mu\text{V/V} + 10 \text{ mV}$ 940 $\mu\text{V/V} + 20 \text{ mV}$	
AC Current			
Generation	<i>50 Hz to 400 Hz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A <i>400 Hz to 1 kHz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A <i>1 kHz to 5 kHz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A <i>5 kHz to 10 kHz</i> 10 μA to 200 μA 0.2 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 220 mA to 2.2 A 2.2 A to 11 A	190 $\mu\text{A/A} + 20 \text{ nA}$ 150 $\mu\text{A/A} + 40 \text{ nA}$ 58 $\mu\text{A/A} + 0.40 \mu\text{A}$ 48 $\mu\text{A/A} + 4.0 \mu\text{A}$ 110 $\mu\text{A/A} + 40 \mu\text{A}$ 190 $\mu\text{A/A} + 170 \mu\text{A}$ 190 $\mu\text{A/A} + 50 \text{ nA}$ 150 $\mu\text{A/A} + 0.50 \mu\text{A}$ 58 $\mu\text{A/A} + 5.0 \mu\text{A}$ 54 $\mu\text{A/A} + 50 \mu\text{A}$ 110 $\mu\text{A/A} + 100 \mu\text{A}$ 250 $\mu\text{A/A} + 380 \mu\text{A}$ 700 $\mu\text{A/A} + 0.10 \mu\text{A}$ 350 $\mu\text{A/A} + 1.0 \mu\text{A}$ 530 $\mu\text{A/A} + 10 \mu\text{A}$ 640 $\mu\text{A/A} + 100 \mu\text{A}$ 410 $\mu\text{A/A} + 200 \mu\text{A}$ 390 $\mu\text{A/A} + 750 \mu\text{A}$ 0.18 % + 0.10 μA 0.12 % + 1.0 μA 0.18 % + 10 μA 0.22 % + 100 μA 0.090 % + 200 μA 0.060 % + 750 μA	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility
Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC Current (continued)			
Measurement	<i>55 Hz to 1 kHz</i> 20 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A <i>1 kHz to 5 kHz</i> 20 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A	850 μ A/A + 20 nA 850 μ A/A + 200 nA 850 μ A/A + 2.0 μ A 800 μ A/A + 20 μ A 800 μ A/A + 200 μ A 0.11 % + 20 nA 0.11 % + 200 nA 0.11 % + 2.0 μ A 0.10 % + 20 μ A 0.18 % + 400 μ A	
Capacitance	<i>1 kHz</i> 100 pF to 1 μ F	0.070 %	2 terminal measurements
Frequency			Averaged over 100 s
Generation			
Specific Values	100 kHz, 1 MHz, 5 MHz & 10 MHz	4.0 in 10^{12}	
	1 PPS	4.0 in 10^{11}	
Range values	1 Hz to 20 GHz	1.0 in 10^{11}	
Measurement	1 Hz to 20 GHz	1.0 in 10^{11} + 10 mHz	
Voltage Reflection Coefficient	<i>10 MHz to 18 GHz</i> 0 to 0.33	0.040	The CMC applies to 7 mm 50 Ω coaxial line fitted with Type N connectors; for other connector types the quoted uncertainties may be increased. Results may also be quoted in terms of VSWR or Return Loss with the uncertainty being converted to the corresponding units.
RF Attenuation	<i>10 MHz to 1 GHz</i> 0 dB to 10dB 10 dB to 20 dB 20 dB to 30 dB 30 dB to 40 dB 40 dB to 50 dB 50 dB to 60 dB 60 dB to 70 dB 70 dB to 80 dB	0.080 dB 0.10 dB 0.15 dB 0.18 dB 0.22 dB 0.25 dB 0.30 dB 0.35 dB	The CMCs apply to devices with input and output VSWR not exceeding 1.1: 1. For devices that exhibit higher values of VSWR the quoted uncertainties may be increased.



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
RF Power			
Specific value	50 MHz 1 mW	0.70 %	The CMC for measurement of 50MHz 1mW 50 Ω Reference sources with an output VSWR not exceeding 1.1 : 1, For devices that exhibit higher values of VSWR the quoted uncertainties may be increased.
Other values	-20 dBm to +20 dBm		
	10 MHz	2.5%	The CMCs for RF Power (Diode Sensors) apply to the measurement of the output power of a 50 Ω source with an output VSWR not exceeding 1.5:1. For devices that exhibit higher values of VSWR the quoted uncertainties may be increased.
	30 MHz	2.5%	
	50 MHz	2.3%	
	100 MHz	2.2%	
	200 MHz	2.2%	
	300 MHz	2.2%	
	400 MHz	2.2%	
	800 MHz	2.2%	
	1 GHz	2.2%	
	2 GHz	2.2%	
	4 GHz	2.2%	
	5 GHz	2.2%	
	6 GHz	2.2%	
	7 GHz	2.2%	
	8 GHz	2.5%	
	9 GHz	2.7%	
	10 GHz	2.5%	
	11 GHz	2.5%	
	12 GHz	2.6%	
	13 GHz	2.9%	
	14 GHz	3.0%	
	15 GHz	2.7%	
	16 GHz	2.8%	
	17 GHz	3.7%	
	18 GHz	5.1%	
	50 dBm to -20 dBm		
	10 MHz to 30 MHz	8.0 %	
	30 MHz to 8 GHz	3.5%	
	8 GHz to 18 GHz	4.5 %	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
RF Calibration Factor			
Nominal 1 mW	10 MHz and 30 MHz	1.6 %	The stated CMCs are for the measurement of RF Calibration Factor of power sensors in a 50 Ω coaxial system at a nominal level of 1 mW, with respect to a reference frequency of 50 MHz.
	100 MHz, 200 MHz, 300 MHz, 400 MHz, 800 MHz and 1000 MHz	1.4 %	
	2 GHz, 3 GHz and 4 GHz	1.7 %	
	5 GHz, 6 GHz, 7 GHz and 8 GHz	3.8 %	
	9 GHz, 10 GHz and 11 GHz	6.0 %	
	12 GHz, 13 GHz and 14 GHz	5.5 %	
	15 GHz, 16 GHz, 17 GHz and 18 GHz	5.3 %	
Nominal 1 μ W	10 MHz	1.72 %	The stated CMCs are for the measurement of RF Calibration Factor of power sensors in a 50 Ω coaxial system at a nominal level of 1 μ W, with respect to a reference frequency of 50 MHz.
	30 MHz, 100 MHz, 200 MHz, 300 MHz, 400 MHz, 800 MHz, 1 GHz, 2 GHz, 3 GHz and 4 GHz	1.5 %	
	5 GHz	1.6 %	
	6 GHz, 7 GHz and 8 GHz	3.0 %	
	9 GHz to 18 GHz	3.4 %	



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
RADIOLOGICAL CALIBRATION			
Air Kerma Rate	^{137}Cs : 3.7 μGyh^{-1} to 1.0 Gyh^{-1} ^{60}Co : 270 μGyh^{-1} to 6.1 mGyh^{-1} ^{241}Am : 10.6 μGyh^{-1} to 1.8 mGyh^{-1}	6.7 % 6.5 % 6.5 %	Calibration of air kerma/air kerma rate monitors using air kerma rates to national standards through a secondary standard dosimeter
Ambient Dose Equivalent Rate	^{137}Cs : 4.5 μSvh^{-1} to 1.20 Svh^{-1} ^{60}Co : 313 μSvh^{-1} to 7.1 mSvh^{-1} ^{241}Am : 18.4 μSvh^{-1} to 3.1 mSvh^{-1}	6.7 % 6.5 % 6.5 %	Calibration of ambient dose equivalent/dose rate monitors using air kerma rates to national standards through a secondary standard dosimeter and using appropriate coefficients given in ISO Standards for $\text{H}^*(10)$.
Personal Dose Equivalent Rate	^{137}Cs : 4.5 μSvh^{-1} to 1.20 Svh^{-1} ^{60}Co : 310 μSvh^{-1} to 7.0 mSvh^{-1} ^{241}Am : 20.0 μSvh^{-1} to 3.3 mSvh^{-1}	6.7 % 6.5 % 6.5 %	Calibration of electronic personal dosimeters using air kerma rates to national standards through a secondary standard dosimeter, and using appropriate coefficients given in ISO Standards for $\text{Hp}(10)$
Surface Contamination Monitor Response:			
Alpha (α) Contamination	Alpha-emitting nuclides: ^{241}Am , ^{238}Pu , ^{238}U	5.0 % to 22 % depending upon monitor type	Calibration process is completed as per the guidance of GPG 14
Beta (β) Contamination	Beta-emitting nuclides: $^{90}\text{Sr}/^{90}\text{Y}$, ^{36}Cl , ^{14}C , ^{60}Co , ^{137}Cs	5.0 % to 22 % depending upon monitor type	Calibration process is completed as per the guidance of GPG 14
END			



0254
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Devonport Royal Dockyard Limited trading as
Devonport Electrical and Nucleonic Calibration Facility**

Issue No: 043 Issue date: 11 April 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$