

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

	West Yorkshire Trading Standards Calibration Services, a part of West Yorkshire Joint Services	
0278	Issue No: 035 Issue date: 26 February 2025	
Accredited to ISO/IEC 17025:2017	Nepshaw Lane South Morley Leeds LS27 7JQ	Contact: Matthew Broome Tel: +44 (0)113 5353014 E-Mail: metrology@wyjs.org.uk Website: www.wyjs.org.uk/calibration
	Calibration performed at the above address only	

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
MASS	kg	g	NOTES
Nominal values <i>See Note 1</i>	1000 500 250 200 100 50	20 10 5.0 4.0 2.0 2.0	1 Intermediate values can be calibrated to an uncertainty proportional between those of the next higher and lower nominal values.
Nominal values <i>See Note 1</i>	g	Mg	2 Calibrations can be given in other units by arrangement.
	25 000 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.002 0.001	13.0 10 5.0 2.5 1.0 0.50 0.25 0.10 0.050 0.033 0.027 0.020 0.017 0.013 0.010 0.0080 0.0070 0.0050 0.0040 0.0030 0.0030 0.0020 0.0020 0.0020	3 Borda's Substitution Method
END			



0278
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

West Yorkshire Trading Standards Calibration Services, a part of West Yorkshire Joint Services

Issue No: 035 Issue date: 26 February 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$