

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0295</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p align="center">Haven Automation Limited</p> <p align="center">Issue No: 056 Issue date: 16 September 2025</p>	
	<p>Measurement House Kingsway Fforestfach Swansea Wales SA5 4EX</p>	<p>Contact: Mr David Gray Tel: +44 (0)1792 588722 Fax: +44 (0)1792 582624 E-Mail: mail@haven.co.uk Website: www.haven.co.uk</p>

Calibration performed at the above address only

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL	DC Resistance		
	Measurement		
	0 Ω to 20 Ω	25 $\mu\Omega/\Omega$ + 70 $\mu\Omega$	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
	20 Ω to 200 Ω	15 $\mu\Omega/\Omega$	
	200 Ω to 2 k Ω	10 $\mu\Omega/\Omega$	
	2 k Ω to 20 k Ω	12 $\mu\Omega/\Omega$	
	20 k Ω to 200 k Ω	20 $\mu\Omega/\Omega$	
	200 k Ω to 2 M Ω	25 $\mu\Omega/\Omega$	
	2 M Ω to 20 M Ω	0.010 %	
	20 M Ω to 200 M Ω	0.020 % + 15 k Ω	
	200 M Ω to 1 G Ω	0.060 % + 0.60 M Ω	
	Generation		Using multi-function calibrator.
	0 Ω to 40 Ω	0.018 % + 100 $\mu\Omega$	
	40 Ω to 400 Ω	0.0085 %	
	400 Ω to 4 k Ω	0.0055 %	
	4 k Ω to 40 k Ω	0.0075 %	
	40 k Ω to 400 k Ω	0.0095 %	
	400 k Ω to 4 M Ω	0.012 %	
	4 M Ω to 40 M Ω	0.030 %	
	40 M Ω to 400 M Ω	0.024 %	
DC Voltage	Measurement		
	0 mV to 200 mV	1.5 μV	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
	200 mV to 2 V	10 $\mu V/V$	
	2 V to 20 V	8.0 $\mu V/V$	
	20 V to 200 V	10 $\mu V/V$	
	200 V to 1 kV	15 $\mu V/V$	
	1 kV to 5 kV	1.2 %	
	Generation		Using multi-function calibrator.
	0 mV to 320 mV	0.0045 % + 2.5 μV	
	320 mV to 3.2 V	0.0060 %	
	3.2 V to 32 V	0.0070 %	
	32 V to 320 V	0.0090 %	
	320 V to 1050 V	0.0060 %	



0295
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Haven Automation Limited
Issue No: 056 Issue date: 16 September 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL (continued)			
DC Current			
Measurement	0 μ A to 200 μ A 0.2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	0.0065 % + 10 nA 0.0055 % + 10 nA 0.0080 % 0.015 %	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
Generation	0 μ A to 320 μ A 320 μ A to 3.2 mA 3.2 mA to 32 mA 32 mA to 320 mA 320 mA to 3 A 3 A to 10 A	0.012 % + 6.0 nA 0.021 % 0.026 % 0.040 % 0.040 % 0.042 %	Using multi-function calibrator.
AC Voltage			
Generation	40 Hz to 30 kHz 32 mV to 320 mV 320 mV to 320 V 40 Hz to 10 kHz 320 V to 750 V 750 V to 1050 V	 0.040 % 0.050 % 0.040 % 0.050 %	Using multi-function calibrator.
Measurement	40 Hz to 10 kHz 200 mV to 2 V 2 V to 20 V 20 V to 200 V 40 Hz to 10 kHz 20 mV to 200 mV 200 V to 1000 V 50 Hz 1 kV to 5 kV	 0.030 % 0.030 % 0.028 % 59 μ V 0.040 % 1.8 %	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
AC Current			
Generation	32 μ A to 320 mA 10 Hz to 110 Hz 110 Hz to 3 kHz 320 mA to 3 A 40 Hz to 110 Hz 110 Hz to 3 kHz 3 A to 10 A 40 Hz to 110 Hz 110 Hz to 3 kHz	 0.045 % 0.070 % 0.060 % 0.080 % 0.080 % 0.080 %	Using multi-function calibrator.



0295
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Haven Automation Limited
Issue No: 056 Issue date: 16 September 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL (continued)			
AC Current (continued)			
Measurement	<i>40 Hz to 1 kHz</i> 100 μ A to 200 μ A 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA <i>55 Hz to 300 Hz</i> 200 mA to 500 mA 500 mA to 2 A	0.035 μ A 0.030 % + 0.25 μ A 0.030 % + 2.5 μ A 0.030 % + 25 μ A 0.25 % 0.10 %	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
AC Resistance	<i>At 50 Hz</i> 0.05 Ω and 0.1 Ω 0.2 Ω , 0.5 Ω , 1 Ω , 5 Ω , 10 Ω , 50 Ω , 100 Ω , 500 Ω and 1 k Ω	0.40 % 0.30 %	For the calibration of the earth bond function on Portable Appliance Testers.
Capacitance			
Generation	1 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μ F 4 μ F to 40 μ F 40 μ F to 400 μ F 400 μ F to 4 mF 4 mF to 30 mF	0.30 % 0.20 % 0.20 % 0.20 % 0.20 % 0.20 % 0.35 % 0.35 %	
Frequency			
Generation	0.5 Hz to 200 kHz	0.0012 % + 0.010 Hz	
Temperature indicators, calibration by electrical simulation			
Base metal thermocouple	-50 $^{\circ}$ C to +1320 $^{\circ}$ C	0.30 $^{\circ}$ C	Including cold junction compensation
Noble metal thermocouple	-50 $^{\circ}$ C to +1800 $^{\circ}$ C	0.70 $^{\circ}$ C	Including cold junction compensation
Resistance thermometer (Pt 100)	-200 $^{\circ}$ C to 0 $^{\circ}$ C 0 $^{\circ}$ C to 250 $^{\circ}$ C 250 $^{\circ}$ C to 800 $^{\circ}$ C	0.0020 $^{\circ}$ C to 0.012 $^{\circ}$ C 0.012 $^{\circ}$ C to 0.025 $^{\circ}$ C 0.025 $^{\circ}$ C to 0.030 $^{\circ}$ C	
Cold junction compensation	21 $^{\circ}$ C to 25 $^{\circ}$ C	0.15 $^{\circ}$ C	Lab ambient temperature



0295
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Haven Automation Limited
Issue No: 056 Issue date: 16 September 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL (continued)			
Temperature simulators, calibration by electrical simulation			
Base metal thermocouple	-50 °C to +1320 °C	0.35 °C	Including cold junction compensation
Noble metal thermocouple	-50 °C to +1800 °C	0.90 °C	Including cold junction compensation
Resistance thermometer (Pt 100)	-200 °C to 0 °C 0 °C to 250 °C 250 °C to 800 °C	0.0020 °C to 0.012 °C 0.012 °C to 0.025 °C 0.025 °C to 0.030 °C	
Cold junction compensation	21 °C to 25 °C	0.15 °C	Lab ambient temperature
PRESSURE			
<u>Hydraulic pressure (gauge)</u>			
Calibration of pressure indicating instruments and gauges	552 kPa to 4.8 MPa 5.1 MPa to 110 MPa	0.0090 % + 0.15 kPa 0.010 %	Methods consistent with EURAMET CG17 1 Calibrations may be undertaken expressed in other units of pressure as required.
<u>Gas pressure (gauge)</u>			
Calibration of pressure indicating instruments and gauges	-90 kPa to -2.5 kPa 1.5 kPa to 2.5 kPa 2.5 kPa to 100 kPa 100 kPa to 690 kPa 690 kPa to 3.5 MPa	0.011 % 0.050 % 0.011 % 0.0060 % 0.0080 %	2 Calibration of pressure measuring devices with an electrical output may be undertaken. 3 Absolute pressures within these gauge ranges can be generated which will attract an additional measurement uncertainty of 90 Pa.
TEMPERATURE			
Resistance Thermometers	-50 °C to 0 °C 0 °C 0 °C to 230 °C 230 °C to 420 °C 420 °C to 650 °C	0.070 °C 0.060 °C 0.10 °C 0.10 °C 0.10 °C	Calibration performed within Liquid Baths
Thermocouples			
Base metal (Type J, K, T & N)	-50 °C to +650 °C	0.42 °C	



0295
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Haven Automation Limited
Issue No: 056 Issue date: 16 September 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
TEMPERATURE (continued)			
Electronic thermometers with sensors		As for sensor type	
Temperature loggers with integral probes	-20 °C to 0 °C 0 °C to 50 °C	0.070 °C 0.10 °C	
Metal Block calibrators	-50 °C to +300 °C 300 °C to 650 °C	0.20 °C 0.35 °C	Method consistent with Euramet CG13
Liquid Baths	-40 °C to +250 °C	0.15 °C	
Furnaces and ovens	50 °C to 600 °C	3.0 °C	Singlepoint monitoring probe only.
END			



0295
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Haven Automation Limited
Issue No: 056 Issue date: 16 September 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$