

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>0307</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>4ward Testing Ltd</h3> <p>Issue No: 054   Issue date: 18 July 2025</p>	
	<p>5 Hampers Common Industrial Estate Petworth West Sussex GU28 9NR</p>	<p>Contact: Mr Anthony Hanson Tel: +44 (0)1798 342240 Fax: +44 (0)1798 344482 E-Mail: anthony@4wardtesting.co.uk Website: www.4wardtesting.co.uk</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p><b>ADHESIVES</b></p> <p>Prepregs, pressure sensitive tapes, bitumen coated pipewraps and related materials exhibiting adhesive properties</p>	<p><u>Mechanical Tests</u></p> <p>Adhesion (Forces up to 1 kN)</p> <p><u>Environmental Tests</u></p> <p>Evaluation of resistance to chemical reagents</p>	<p>ASTM D1000-24 Methods A &amp; B Documented In-House Method No. M0032</p> <p>ASTM D543-21:Practice A: Procedures I and II</p>
<p><b>PLASTICS and PRODUCTS</b></p> <p>Non-rigid plastic sheet &lt;6mm thick</p> <p>Cured reinforced resins Non-cellular plastics</p> <p>Cast, hot moulded, cold moulded, laminated resinous products and plastic sheet</p>	<p><u>Mechanical Tests</u></p> <p>Tensile strength and elongation at ambient temperature (excluding modulus of elasticity) (Forces up to 1 kN)</p> <p>Tensile strength and elongation at ambient temperature (Forces up to 1 kN)</p> <p>Density (&gt; 1 gm/cm<sup>3</sup>)</p> <p>Evaluation of resistance to chemical reagents</p>	<p>ASTM D 412-16(2021) Method A ASTM D 638-22</p> <p>BS EN ISO 527-3:2018 BS 2782-3 Method 326E:1995 (Withdrawn)</p> <p>ASTM D792-20</p> <p>ASTM D543-21:Practice A: Procedures I and II</p>



0307  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**4ward Testing Ltd**  
**Issue No: 054 Issue date: 18 July 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
RUBBER and PRODUCTS	<u>Mechanical Tests</u>  Tensile strength and elongation at ambient temperature (Forces up to 245N)  <u>Environmental Tests</u>  Evaluation of resistance to chemical reagents	ASTM D412-16(2021) Method A    ASTM D543-21:Practice A: Procedures I and II
ADHESIVES and SEALANTS	<u>Mechanical Tests</u>  Elastic Recovery  Resistance to Flow  Tensile Properties, extension to break  Tensile properties at maintained extension  Adhesion/cohesion properties at constant temperature  Adhesion/cohesion properties at variable temperatures  Change in Mass and Volume  Tensile properties at maintained extension after immersion in water  Adhesion/Cohesion after immersion in water  Adhesion/Cohesion after exposure to Heat/Light/Water	BS EN ISO 7389: 2003  BS EN ISO 7390: 2003  BS EN ISO 8339: 2005  BS EN ISO 8340: 2005  BS EN ISO 9046: 2021  BS EN ISO 9047: 2003  BS EN ISO 10563: 2023  BS EN ISO 10590: 2005  BS EN ISO 10591: 2021  BS EN ISO 11431: 2002



0307  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**4ward Testing Ltd**  
**Issue No: 054 Issue date: 18 July 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ADHESIVES and SEALANTS (cont'd)  Silicone Sealants	<u>Microbiological Tests</u>  Determination of fungistatic effects	Documented in-house methods  MOO846 based on Method B described in: BS EN ISO 846:2019, BS EN ISO 846:1997 (withdrawn) In accordance with sample preparation/conditioning set out in BS EN 15651-3:2012
Medical Face Masks	<u>Mechanical Tests</u>  Breathability  Splash Resistance  <u>Microbiological Tests</u>  Bacterial Filtration Efficiency  Microbial cleanliness (bioburden)	M0121 based on BS EN 14683:2025 Clause 5.2.3 Annex C  BS EN 14683:2025 Clause 5.2.4; ISO 22609:2004  M0125 based on EN 14683:2025 Clause 5.2.2 Annex B  M0124 based on EN 14683:2025 Clause 5.2.5 Annex D
END		



0307  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**4ward Testing Ltd**  
**Issue No: 054 Issue date: 18 July 2025**

Testing performed at main address only

**Accreditation for the purpose of UK Approved Body Activity in accordance with UKCA Requirements and UKAS Publication GEN 5**

Directive / Regulation	Conformity Assessment Procedure / Module / Article	Category of Products or Individual Products	Essential Requirements: Product Specification / Properties / Standards
<b>Construction Products Regulation 2011</b> (retained EU law EUR 305/2011) as amended by the Construction Products (Amendment etc.) (EU Exit) Regulations 2019 and the Construction Products (Amendment etc.) (EU Exit) Regulations 2020	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for external walls - outdoor applications	BS EN 15651-1:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for glazing for uses in building construction	BS EN 15651-2:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for sanitary joints for uses in building construction	BS EN 15651-3:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for pedestrian walkways for uses in building construction	BS EN 15651-4:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (2/2) : - Sealants for joints (for uses subject to reaction to fire regulations)	BS EN 15651-1:2012 BS EN 15651-2:2012 BS EN 15651-3:2012 BS EN 15651-4:2012



0307  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**4ward Testing Ltd**  
**Issue No: 054 Issue date: 18 July 2025**

Testing performed at main address only

**Accreditation for the purpose of Notified Body Activity relating to the Northern Ireland market (CE + UKNI) taking into account EA-2/17**

Directive / Regulation	Conformity Assessment Procedure / Module / Article	Category of Products or Individual Products	Essential Requirements: Product Specification / Properties / Standards
<b>Construction Products Regulation 2011</b> (retained EU law EUR 305/2011) as amended by the Construction Products (Amendment etc.) (EU Exit) Regulations 2019 and the Construction Products (Amendment etc.) (EU Exit) Regulations 2020	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for external walls - outdoor applications	BS EN 15651-1:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for glazing for uses in building construction	BS EN 15651-2:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for sanitary joints for uses in building construction	BS EN 15651-3:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (1/2) : - Sealants for pedestrian walkways for uses in building construction	BS EN 15651-4:2012
	Annex V – Testing Laboratory (AVCP) System 3	2011/19/EU Sealants for non-structural use in joints in buildings and pedestrian walkways (2/2) : - Sealants for joints (for uses subject to reaction to fire regulations)	BS EN 15651-1:2012 BS EN 15651-2:2012 BS EN 15651-3:2012 BS EN 15651-4:2012

END