

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0310

Accredited to
ISO/IEC 17025:2017

RS Calibration a trading name of RS Components Ltd

Issue No: 063 Issue date: 20 August 2025

DPN 175
Birchington Road
Corby
Northamptonshire
NN17 5JF

Contact: Sean Smith
Tel: +44 (0)1536 405545
E-Mail: calibration.uk@rs-components.com
Website: <https://uk.rs-online.com/web/>

Calibration performed at the above address only

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL MEASUREMENTS			
DC RESISTANCE			
Specific values Generation	1 mΩ 10 mΩ 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ	120 μΩ/Ω 33 μΩ/Ω 25 μΩ/Ω 5.0 μΩ/Ω 9 μΩ/Ω 5.0 μΩ/Ω 5.8 μΩ/Ω 11 μΩ/Ω 28 μΩ/Ω 29 μΩ/Ω 39 μΩ/Ω 54 μΩ/Ω 300 μΩ/Ω	Known values of resistance for application to resistance measuring instruments. Specific values are those which fall within ± 0.5% of the stated values.
Measurement	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ	20 μΩ/Ω + 4.0 μΩ 11 μΩ/Ω + 14 μΩ 10 μΩ/Ω + 50 μΩ 10 μΩ/Ω + 0.50 mΩ 10 μΩ/Ω + 5.0 mΩ 10 μΩ/Ω + 50 mΩ 11 μΩ/Ω + 1.0 Ω 24 μΩ/Ω + 100 Ω 140 μΩ/Ω + 10 kΩ 0.18 % + 1.0 MΩ	Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased.
DC VOLTAGE			
Measurement	0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1 kV	6.0 μV/V + 0.10 μV 5.0 μV/V + 0.40 μV 5.0 μV/V + 4.0 μV 7.0 μV/V + 40 μV 7.0 μV/V + 500 μV	Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased.



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
DC CURRENT			
Generation	20 A to 500 A 500 A to 2500 A	590 μ A/A 490 μ A/A	Calibration of clamp-on ammeters and similar devices using multi-turn technique.
Measurement	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	14 μ A/A + 0.40 nA 14 μ A/A + 4.0 nA 17 μ A/A + 40 nA 56 μ A/A + 800 nA 220 μ A/A + 16 μ A 470 μ A/A + 400 μ A	Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased.
	2 A to 10 A 10 A to 20 A 20 A to 100 A	320 μ A/A 330 μ A/A 62 μ A/A	Using digital multimeter and shunt. Generation of these values may also be undertaken however the uncertainties may be increased.
AC VOLTAGE Specific Values			Using AC Measurement Standard.

Expanded uncertainty (k = 2) expressed in terms of (μ V/V + μ V)

Generation of these values may also be undertaken, with increased uncertainties and with the following limitations:

600 μ V to 220 V, 10 Hz to 1 MHz: A maximum volt-Hertz product of 2.2×10^7 applies.

220 V to 750 V: The applied frequency will be in the range 40 Hz to 100 kHz.

750 V to 1000 V: The applied frequency will be in the range 40 Hz to 30 kHz.

Voltage Range	10 Hz to 20 Hz	20 Hz to 40 Hz	40 Hz to 20 kHz	20 kHz to 50 kHz	50 kHz to 100 kHz	100 kHz to 300 kHz	300 kHz to 500 kHz	500 kHz to 1 MHz
600 μ V to 2.2 mV	1500 + 1.3	850 + 1.3	710 + 1.3	890 + 2.0	1200 + 2.5	1900 + 4.0	2100 + 8.0	3500 + 8.0
2.2 mV to 7 mV	710 + 1.3	390 + 1.3	270 + 1.3	370 + 2.0	510 + 2.5	990 + 4.0	1200 + 8.0	2800 + 8.0
7 mV to 22 mV	260 + 1.3	190 + 1.3	140 + 1.3	200 + 2.0	270 + 2.5	660 + 4.0	870 + 8.0	2500 + 8.0
22 mV to 70 mV	210 + 1.5	130 + 1.5	100 + 1.5	130 + 2.0	220 + 2.5	440 + 4.0	740 + 8.0	2300 + 8.0
70 mV to 220 mV	180 + 1.5	100 + 1.5	80 + 1.5	90 + 2.0	150 + 2.5	270 + 4.0	600 + 8.0	2300 + 8.0
220 mV to 700 mV	180 + 1.5	100 + 1.5	80 + 1.5	90 + 2.0	100 + 2.5	230 + 4.0	570 + 8.0	750 + 8.0
700 mV to 2.2 V	160 + 0	60 + 0	30 + 0	40 + 0	60 + 0	130 + 0	210 + 0	700 + 0
2.2 V to 7 V	160 + 0	60 + 0	30 + 0	40 + 0	70 + 0	150 + 0	320 + 0	940 + 0
7 V to 22 V	160 + 0	60 + 0	30 + 0	40 + 0	70 + 0	150 + 0	320 + 0	940 + 0
22 V to 70 V	160 + 0	60 + 0	30 + 0	50 + 0	80 + 0			
70 V to 220 V	160 + 0	60 + 0	30 + 0	60 + 0	90 + 0			
220 V to 700 V	160 + 0	80 + 0	40 + 0	110 + 0	390 + 0			
700 V to 1000 V	160 + 0	80 + 0	40 + 0	110 + 0	390 + 0			



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
AC VOLTAGE (continued) Other values Measurement			
	<i>10 Hz to 40 Hz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V	170 μ V/V + 4.0 μ V 140 μ V/V + 20 μ V 140 μ V/V + 200 μ V 140 μ V/V + 2.0 mV	Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased.
	<i>40 Hz to 100 Hz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	140 μ V/V + 4.0 μ V 110 μ V/V + 20 μ V 110 μ V/V + 200 μ V 110 μ V/V + 2.0 mV 140 μ V/V + 20 mV	
	<i>100 Hz to 2 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	130 μ V/V + 2.0 μ V 90 μ V/V + 20 μ V 90 μ V/V + 200 μ V 87 μ V/V + 2.0 mV 140 μ V/V + 20 mV	
	<i>2 kHz to 10 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	160 μ V/V + 4.0 μ V 130 μ V/V + 20 μ V 130 μ V/V + 200 μ V 130 μ V/V + 2.0 mV 140 μ V/V + 20 mV	
	<i>10 kHz to 30 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	400 μ V/V + 8.0 μ V 260 μ V/V + 40 μ V 260 μ V/V + 400 μ V 260 μ V/V + 4.0 mV 260 μ V/V + 40 mV	
	<i>30 kHz to 100 kHz</i> 1 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	900 μ V/V + 20 μ V 660 μ V/V + 200 μ V 660 μ V/V + 2.0 mV 660 μ V/V + 20 mV 670 μ V/V + 200 mV	
	<i>100 kHz to 300 kHz</i> 200 mV to 2 V 2 V to 20 V 20 V to 200 V	0.35 % + 2.0 mV 0.35 % + 20 mV 0.35 % + 200 mV	
	<i>300 kHz to 1 MHz</i> 200 mV to 2 V 2 V to 20 V	1.2 % + 20 mV 1.2 % + 200 mV	



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks		
AC CURRENT Generation	2 A to 250 A 50 Hz to 800 Hz	0.090 %	Calibration of clamp-on ammeters and similar devices using multi-turn technique.		
Measurement	250 A to 1750 A 50 Hz to 100 Hz	0.090 %	Using digital multimeter. Generation of these values may also be undertaken however the uncertainties may be increased.		
	1750 A to 2500 A 50 Hz to 60 Hz	0.090 %			
	10 Hz to 2 kHz 1 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	580 μ A/A + 20 nA 350 μ A/A + 200 nA 350 μ A/A + 2.0 μ A 340 μ A/A + 20 μ A 720 μ A/A + 200 μ A 950 μ A/A + 2.0 mA			
	2 kHz to 10 kHz 1 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	580 μ A/A + 20 nA 350 μ A/A + 200 nA 350 μ A/A + 2.0 μ A 340 μ A/A + 20 μ A 850 μ A/A + 200 μ A			
	50 Hz to 800 Hz 2 A to 20 A 10 A to 20 A 20 A to 50 A 50 A to 100 A	590 μ A/A 580 μ A/A 0.11 % 0.11 %			
	50 Hz to 60 Hz 50A to 100A	0.11 %			
	INDUCTANCE Generation	At 1 kHz: 1 mH 10 mH 100 mH 1 H		490 μ H/H 500 μ H/H 470 μ H/H 470 μ H/H	Known inductance values for the calibration of inductance measuring instruments



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
CAPACITANCE			
Generation	At 1 kHz: 1 nF 2 nF 3 nF 4 nF 5 nF 6 nF to 9 nF 10 nF 20 nF to 80 nF 90 nF 100 nF to 900 nF 1 μ F 2 μ F 3 μ F and 4 μ F 5 μ F and 9 μ F 10 μ F to 30 μ F 40 μ F and 50 μ F 60 μ F and 80 μ F 90 μ F and 100 μ F	0.092 % 0.064 % 0.058 % 0.055 % 0.053 % 0.052 % 0.050 % 0.037 % 0.037 % 0.043 % 0.12 % 0.12 % 0.098 % 0.12 % 0.19 % 0.19 % 0.19 % 0.19 %	Known capacitance values for the calibration of capacitance measuring instruments
FREQUENCY			
Measurement	10 Hz to 100 kHz 100 kHz to 1 MHz 1 MHz to 6 GHz	3.0 in 10^9 2.0 in 10^{10} 5.0 in 10^{10}	Using counter timer and off-air standard.
Timer and stopwatch calibrations	5 s to 99 999s	0.10 s	Manual calibration.
Optical Tachometers	25 rpm to 6000 rpm 6000 rpm to 30000 rpm 30000 rpm to 90000 rpm	0.059 rpm 0.084 rpm 0.19 rpm	Using light pulses synchronised to frequencies that correspond to the stated rpm values.
OSCILLOSCOPE CALIBRATION			
Horizontal deflection coefficients	500 ps to 10 ms	0.061 μ s/s	Using time markers. The uncertainty quoted will be particularly dependent on the horizontal resolution of the oscilloscope being calibrated.
Vertical deflection coefficients	6 mV to 60 mV 60 mV to 600 mV 600 mV to 60 V 60 V to 100 V 100 V to 120 V	0.12 % 0.05 % 0.03 % 0.12 % 0.12 %	Using chopped waveforms of known peak to peak amplitude. The uncertainty quoted will be particularly dependent on the vertical resolution of the oscilloscope being calibrated.



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
Bandwidth	<i>Input VSWR up to 1.2:1</i> 50 kHz to 550 MHz 550 MHz to 1 GHz 1 GHz to 1.5 GHz 1.5 GHz to 4.2 GHz <i>Input VSWR from 1.2:1 to 1.6:1</i> 50 kHz to 550 MHz 550 MHz to 1 GHz	0.15 dB 0.20 dB 0.31 dB/0.41 dB 0.44 dB 0.54 dB	Relative to a low frequency reference. The uncertainty quoted will be dependent on the vertical resolution of the oscilloscope being calibrated.
Rise and fall times	Using 150 ps edge Using 500 ps edge	21 ps 50 ps	Using fast rise pulses. The uncertainty quoted will be dependent on the vertical and horizontal resolution of the oscilloscope being calibrated.
Input resistance (DC)	50 Ω 1 M Ω	0.13 % 0.13 %	For values within 20% of the nominal values shown.
CALIBRATION OF MULTI-FUNCTION CALIBRATORS			
DC RESISTANCE			
Specific Values	0 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	20 $\mu\Omega + 4 \mu\Omega$ 11 $\mu\Omega/\Omega + 14 \mu\Omega$ 10 $\mu\Omega/\Omega + 50 \mu\Omega$ 10 $\mu\Omega/\Omega + 0.5 m\Omega$ 10 $\mu\Omega/\Omega + 5 m\Omega$ 10 $\mu\Omega/\Omega + 50 m\Omega$ 11 $\mu\Omega/\Omega + 1\Omega$ 24 $\mu\Omega/\Omega + 0.1 k\Omega$ 140 $\mu\Omega/\Omega + 10 k\Omega$ 1800 $\mu\Omega/\Omega + 1 M\Omega$	Using automated system. Generation of these parameters up to and including 100 kHz may also be undertaken but the uncertainties may be increased
DC VOLTAGE			
Specific Values	0 V 100 mV 1 V 10 V 19 V 100 V 1000 V	6.0 $\mu V/V + 0.1 \mu V$ 5.0 $\mu V/V + 0.4 \mu V$ 5.0 $\mu V/V + 4 \mu V$ 7.0 $\mu V/V + 40 \mu V$ 7.0 $\mu V/V + 500 \mu V$ 8.0 $\mu V/V$ 8.0 $\mu V/V$	



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
DC CURRENT Specific Values	0 μ A 100 μ A 1 mA 10 mA 100 mA 1 A 10 A	14 μ A/A + 0.5 nA 14 μ A/A + 4 nA 17 μ A/A + 40 nA 56 μ A/A + 800 nA 220 μ A/A + 16 μ A 470 μ A/A + 400 μ A 100 μ A/A	
AC VOLTAGE	<p>At 10 Hz, 20 Hz, 30 Hz, 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz, 1 kHz, 10 kHz and 20 kHz:</p> <p>1 mV 10 mV 100 mV</p> <p>At 30 kHz and 50 kHz:</p> <p>1 mV 10 mV 100 mV</p> <p>At 100 kHz:</p> <p>1 mV 10 mV 100 mV</p> <p>1 V and 10 V: 10 Hz, 20 Hz and 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz, 1 kHz, 10 kHz, 20 kHz and 30 kHz 50 kHz 100 kHz 300 kHz 500 kHz 1 MHz</p> <p>19 V: 1 kHz</p> <p>100 V: 10 Hz, 20 Hz and 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz, 1 kHz, 10 kHz, 20 kHz and 30 kHz 50 kHz 100 kHz</p> <p>1000 V: 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz</p>	<p>0.80 % 800 μV/V 170 μV/V</p> <p>0.80 % 810 μV/V 230 μV/V</p> <p>1.1 % 0.13 % 440 μV/V</p> <p>49 μV/V</p> <p>44 μV/V 80 μV/V 110 μV/V 270 μV/V 480 μV/V 0.11 %</p> <p>44 μV/V</p> <p>55 μV/V</p> <p>49 μV/V 80 μV/V 130 μV/V</p> <p>70 μV/V</p>	1 mV and 10 mV are not available at 10 Hz



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC CURRENT	100 µA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 1 mA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 10 mA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 100 mA: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 1 A: 10 Hz and 20 Hz 30 Hz 40 Hz, 50 Hz, 55 Hz, 60 Hz, 300 Hz and 1 kHz 5 kHz 10 A: 40 Hz, 55 Hz, 300 Hz and 1 kHz	230 µA/A 220 µA/A 210 µA/A 380 µA/A 190 µA/A 190 µA/A 180 µA/A 290 µA/A 190 µA/A 190 µA/A 180 µA/A 280 µA/A 190 µA/A 190 µA/A 180 µA/A 280 µA/A 230 µA/A 220 µA/A 190 µA/A 411 µA/A 800 µA/A	
ELECTRICAL SIMULATION OF TEMPERATURE (Base metal thermocouple indicators & PT100)			Application or measurement of DC voltages equivalent to those for the thermocouple types indicated, with cold junction compensation enabled.
K type	-200 °C to -190 °C -190 °C to -100 °C -100 °C to +1300 °C	0.18 °C 0.11 °C 0.086 °C	
T Type	-150 °C to -100 °C -100 °C to 0 °C 0 °C to 400 °C	0.11 °C 0.082 °C 0.065 °C	
J Type	-100 °C to 0 °C 0 °C to 1000 °C	0.083 °C 0.076 °C	
PT100	-200 °C to -50 °C -50 °C to 200 °C 200 °C to 300 °C 300 °C to 500 °C 500 °C to 600 °C 600 °C to 800 °C	1.4 m°C 2.5 m°C 5.0 m°C 7.4 m°C 10 m°C 11 m°C	



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
CALIBRATION OF 16TH/17TH EDITION TEST EQUIPMENT			Using dedicated calibration system.
Insulation Resistance	10 k Ω to 5 M Ω 5 M Ω to 100 M Ω 100 M Ω to 1 G Ω	0.16 % 1.2 % 1.4 %	
Continuity Resistance	900 m Ω to 2 Ω 2 Ω to 6 Ω 6 Ω to 20 Ω 100 Ω 1 k Ω	3.5 % 1.1 % 0.67 % 0.32 % 1.2 %	
Continuity Current	100 mA 200 mA 300 mA	2.2 % 1.9 % 1.7 %	
Insulation Voltage	50 V 100 V 250 V 500 V 1000 V	3.0 % 2.1 % 1.5 % 1.3 % 1.3 %	
Current on Insulation resistance function			
1000 V range	0.5 mA 1.0 mA	3.1 % 2.1 %	
500 V range	0.5 mA 1.0 mA	3.1 % 2.1 %	
Loop Resistance at 50 Hz	0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1000 Ω	10 % 6.7 % 3.5 % 1.2 % 1.0 % 0.87 % 0.87 %	
RCD Current at 50 Hz	10 mA to 30 mA 30 mA to 300 mA 300 mA to 2 A	2.1 % 1.9 % 1.6 %	
RCD Trip Time	20 ms to 40 ms 40 ms to 200 ms 200 ms to 390 ms 390 ms to 900 ms	4.8 % 2.4 % 0.48 % 0.90 %	



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
PAT Insulation Resistance	1 M Ω to 4 M Ω 4 M Ω to 10 M Ω	0.16 % 1.2 %	
PAT Earth Bond Resistance at 50 Hz	0.1 Ω 0.22 Ω 0.33 Ω 0.5 Ω 1 Ω 5 Ω 10 Ω 100 Ω 1000 Ω	8.0 % 4.1 % 3.0 % 2.3 % 1.6 % 1.0 % 0.94 % 0.87 % 0.87 %	
PAT Earth Bond Current at 50 Hz	0 mA to 100 mA 100 mA to 10 A 10 A to 30 A	2.1 % + 6 mA 1.7 % + 60 mA 1.7 % + 60 mA	
PAT Leakage Current Test	2.7 mA at 240 V 4.7 mA at 240 V 7.7 mA at 240 V	1.9 % 1.8 % 1.8 %	
PAT Flash Voltage Test At 50 Hz	1000 V (Class 1) 1500 V (Class 1) 1000 V (Class 2) 3000 V (Class 2)	5.8 % 5.4 % 5.8 % 5.0 %	
PAT Flash Current Test At 50 Hz	0.67 mA at 1000 V (Class 1) 1.00 mA at 1500 V (Class 1) 0.34 mA at 1000 V (Class 2) 1.00 mA at 3000 V (Class 2)	6.0 % 5.9 % 6.8 % 5.9 %	
TEMPERATURE CALIBRATION			Unless otherwise stated the calibration is performed by comparison with reference standards. Other units other than Celsius can be reported.
Temperature indicators and recorders with temperature sensor(s) – resistance	-20 $^{\circ}$ C to -10 $^{\circ}$ C -10 $^{\circ}$ C to +50 $^{\circ}$ C 50 $^{\circ}$ C to 100 $^{\circ}$ C 100 $^{\circ}$ C to 200 $^{\circ}$ C	0.042 $^{\circ}$ C 0.037 $^{\circ}$ C 0.055 $^{\circ}$ C 0.051 $^{\circ}$ C	Calibration performed within Liquid Baths.
Temperature indicators and recorders with temperature sensor(s) - thermocouple	-20 $^{\circ}$ C to -10 $^{\circ}$ C -10 $^{\circ}$ C to +50 $^{\circ}$ C 50 $^{\circ}$ C to 100 $^{\circ}$ C 100 $^{\circ}$ C to 200 $^{\circ}$ C	0.046 $^{\circ}$ C 0.040 $^{\circ}$ C 0.12 $^{\circ}$ C 0.21 $^{\circ}$ C	Calibration performed within Liquid Baths.



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
DEW POINT and RELATIVE HUMIDITY			Calibration by comparison with a reference dew point hygrometer and reference PRTs.
Dew-point Temperature probes in air and Temperature probes associated with hygrometers	1°C to 40°C 10 °C to 20 °C 20 °C to 25 °C 25 °C to 50 °C	0.33°C to 0.39°C 0.30 °C 0.30 °C 0.30 °C	Calibration performed within an air chamber.
Relative humidity instruments	Example conditions At 10 °C 53 %rh to 85 %rh At 23 °C 24 %rh to 85 %rh At 30 °C 16 %rh to 85 %rh At 60 °C 10 %rh to 37 %rh	Corresponding to above dew-point and temperature uncertainties 1.5 %rh to 2.3 %rh 0.63 %rh to 1.8 %rh 0.49 %rh to 1.8 %rh 0.41 %rh to 0.87 %rh	
DIMENSIONAL MEASUREMENTS Unless otherwise stated, the ranges are presented in millimetres (mm) and the uncertainties in micrometres (µm).			
MEASURING INSTRUMENTS AND MACHINES			
Micrometers			
External	BS 870:2008 0 to 600 Heads: (Zero) Setting, 0 to 25: (Zero) Setting, 25 to 600: Flatness of anvils: Parallelism of anvils:	2.0 between any two points 1.0 1.0 + (8.0 x length in m) 0.50 1.2	
Internal	BS 959:2008 0 to 600 Heads: (Zero) Setting and extension rods:	2.0 between any two points 1.0 + (8.0 x length in m)	
Depth	BS 6468:2008 0 to 300 Heads: (Zero) Setting: Base Flatness: Rod Flatness: Parallelism: Rod axis of rotation: Squareness of Face to spindle / rod axis: Rod axis to datum face:	2.0 between any two points 1.0 + (8.0 x length in m) 0.42 0.42 2.7 5.0 1.6 10	



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
Vernier, Digital and Dial Gauges			
Calipers	Based on BS 887:2008 0 to 600 Overall performance: Flatness: Parallelism, External: Parallelism, Internal: Squareness: Co-Planer error of Jaws: Width of internal jaws:	10 + (30 x length in m) 4.7 4.4 2.7 6.2 10 2.7	
Height gauges	Based on BS 1643:2008 0 to 600 Overall performance: Flatness: Parallelism: Depth of Jaw / Scriber:	10 + (30 x length in m) 2.7 3.7 2.7	
Depth gauges	Based on BS 6365:2008 0 to 300 Overall performance: Flatness / Straightness: Parallelism:	10 + (30 x length in m) 2.8 2.7	
Internal and external dial caliper and thickness gauges	Scale accuracy: 0 to 50 50 to 200 Parallelism:	1.6 10 1.6	Using procedures MLCP 12 or MLCP 51
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 100 Scale interval measurements: Discrimination:	1.1 1.6	
Road measuring wheels Derived calibration factor Diameter Circumference Counter distance	0.95 to 1.05 200 to 350 600 to 1050 29.9 to 30.1 revolutions	0.0014 0.12 mm 0.37 mm 0.030 m	Using procedure MLCP 52
LENGTH			
Length gauges, flat and spherical ended (excluding length bars)	BS 870:2008 0 to 600	1.0 + (8.0 x length in m)	
Feeler gauges	BS 957:2008 0.03 to 1	1.8	
Steel rules, engineers	0 to 1200	8.0 + (10 x length in metres)	Using procedure MLCP 57



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
PRESSURE CALIBRATION			
Gas pressure (gauge)			Methods consistent with EURAMET CG17
Calibration of pressure indicating instruments and gauges	-90 kPa to -3.5 kPa -3.5 kPa to 3.5 kPa 3.5 kPa to 190 kPa 190 kPa to 7 MPa 7 MPa to 10 MPa	0.004 6 % 16 Pa 0.004 6 % 0.007 3 % 1.3 kPa	Using deadweight tester. Using pressure controller Using deadweight tester. Using deadweight tester. Using pressure controller
Hydraulic pressure (gauge)			
Calibration of pressure indicating instruments and gauges	0.6 MPa to 6 MPa 6 MPa to 16 MPa 16 MPa to 40 MPa 40 MPa to 70 MPa 70 MPa to 100 MPa	0.020 % 6.8 kPa 61 kPa 0.020 % 78 kPa	Using deadweight tester. Comparison with pressure indicator.
END			



0310
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RS Calibration
a trading name of RS Components Ltd
Issue No: 063 Issue date 20 August 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$