


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION 0361</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Micro Precision (Yorkshire) Ltd</p> <p>Issue No: 047 Issue date: 21 November 2025</p>	
	<p>Unit 3 Shepcote Enterprise Park 2 3 Europa Drive Sheffield S9 1XT</p>	<p>Contact: Mr Michael Batty Tel: +44 (0)1142 730534 E-Mail: uk-contact@microprecision.com Website: www.microprecision.com</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
<p>ELECTRICAL CALIBRATION</p> <p>All electrical measurements are carried out using the method of direct comparison or transfer to laboratory reference standards unless otherwise determined in the remarks column. The measurement and generation headings in the first column declare the laboratory's ability to either measure outputs of submitted test items or to generate values as a stimulus for test items which measure.</p>			
DC Voltage			
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1100 V 1 kV to 4 kV 4 kV to 6 kV 6 kV to 8 kV 8 kV to 10 kV 10 kV to 15 kV 15 kV to 20 kV 20 kV to 25 kV 25 kV to 30 kV 30 kV to 35 kV 35 kV to 40 kV	9.0 $\mu\text{V/V} + 0.60 \mu\text{V}$ 5.0 $\mu\text{V/V}$ 4.0 $\mu\text{V/V}$ 7.0 $\mu\text{V/V}$ 7.0 $\mu\text{V/V}$ 0.49 % + 6.0 V 0.49 % + 6.1 V 0.61 % + 6.3 V 0.60 % + 6.6 V 0.72 % + 35 V 0.72 % + 35 V 0.72 % + 57 V 0.72 % + 68 V 0.72 % + 78 V 0.72 % + 97 V	These values can be generated for the calibration of measuring instruments, Outputs of instruments can be measured directly
Generation	0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V	13 $\mu\text{V/V} + 1.6 \mu\text{V}$ 8.0 $\mu\text{V/V}$ 9.0 $\mu\text{V/V}$ 11 $\mu\text{V/V}$ 12 $\mu\text{V/V}$	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC Voltage Measurement	1 mV to 12 mV 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 12 mV to 120 mV 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 120 mV to 1.2 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 1.2 V to 12 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 2 MHz 12 V to 120 V 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 120 V to 200 V 20 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	160 µV/V + 0.8 µV 100 µV/V + 0.8 µV 100 µV/V + 0.8 µV 160 µV/V + 0.8 µV 250 µV/V + 0.8 µV 0.63 % + 0.8 µV 110 µV/V 100 µV/V 100 µV/V 170 µV/V 250 µV/V 0.62 % 0.63 % 0.63 % 100 µV/V 100 µV/V 100 µV/V 160 µV/V 250 µV/V 280 µV/V 0.62 % 0.62 % 110 µV/V 100 µV/V 100 µV/V 100 µV/V 390 µV/V 410 µV/V 0.62 % 130 µV/V 110 µV/V 120 µV/V 250 µV/V 680 µV/V 690 µV/V 720 µV/V 190 µV/V 180 µV/V 160 µV/V 170 µV/V 420 µV/V	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC Voltage Measurement (cont'd)	200 V to 1100 V 20 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz	150 µV/V 140 µV/V 350 µV/V 790 µV/V	
	1 kV to 2 kV 50 Hz 2 kV to 4 kV 50 Hz 4 kV to 6 kV 50 Hz 6 kV to 7 kV 50 Hz	0.84 % + 8.4 V 0.84 % + 13 V 0.94 % + 18 V 0.97 % + 19 V	
	7 kV to 10 kV 50 Hz to 60 Hz 10 kV to 15 kV 50 Hz to 60 Hz 15 kV to 20 kV 50 Hz to 60 Hz 20 kV to 25 kV 50 Hz to 60 Hz	0.80 % + 40 V 0.80 % + 59 V 0.80 % + 87 V 0.80 % + 110 V	
AC Voltage Generation	10 Hz to 45 Hz 1 mV to 33 mV 33 mV to 33 V	0.060 % + 2.6 µV 0.030 %	
	45 Hz to 10 kHz 1 mV to 33 mV 33 mV to 330 V	0.050 % + 2.6 µV 0.010 %	
	10 kHz to 20 kHz 1 mV to 33 mV 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V	0.050 % + 2.6 µV 0.020 % 0.010 % 0.010 % 0.010 %	
	20 kHz to 50 kHz 1 mV to 33 mV 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V	0.050 % + 2.6 µV 0.020 % 0.010 % 0.010 % 0.030 %	
	50 kHz to 100 kHz 1 mV to 33 mV 33 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V	0.060 % + 2.6 µV 0.020 % 0.040 % 0.010 % 0.090 %	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC Voltage Generation (cont'd)	100 kHz to 500 kHz 1 mV to 33 mV 33 mV to 330 mV 330 mV to 3.3 V 330 V to 1020 V 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.13 % + 2.6 μV 0.20 % 0.20 % 0.010 % 0.010 % 0.030 %	
DC Current Measurement	0 μA to 1.2 μA 1.2 μA to 12 μA 12 μA to 120 μA 0.12 mA to 120 mA 120 mA to 200 mA 0.2 A to 2 A 2 A to 11 A	110 μA/A + 0.50 nA 23 μA/A + 0.50 nA 18 μA/A 16 μA/A 17 μA/A 65 μA/A 140 μA/A	
DC Current Generation	0 μA to 330 μA 0.33 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 0.33 A to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20.5 A 20.5 A to 150 A 150 A to 1025 A	110 μA/A + 3.8 nA 35 μA/A 33 μA/A 42 μA/A 190 μA/A 190 μA/A 350 μA/A 440 μA/A 0.61 % 0.60 %	For the calibration of clamp meters only
AC Current Measurement	10 Hz to 1 kHz 2 μA to 200 μA 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 0.2 A to 2 A 1 kHz to 20 kHz 0.12 mA to 12 mA 12 mA to 120 mA 120 mA to 1.05 A	240 μA/A + 17 nA 240 μA/A 260 μA/A 260 μA/A 650 μA/A 510 μA/A 630 μA/A 0.28 %	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC Current Measurement (cont'd)	1 kHz to 5 kHz 1.05 A to 2 A	0.16 %	
	20 Hz to 2 kHz 2 A to 11 A	660 μ A/A	
	2 kHz to 5 kHz 2 A to 11 A	0.12 %	
AC Current Generation	10 Hz to 20 Hz 30 μ A to 330 μ A	0.15 % + 60 nA	
	330 μ A to 3.3 mA	0.10 %	
	3.3 mA to 33 mA	0.14 %	
	33 mA to 330 mA	0.13 %	
	20 Hz to 45 Hz 30 μ A to 330 μ A	0.15 % + 60 nA	
	330 μ A to 3.3 mA	0.10 %	
	3.3 mA to 330 mA	0.060 %	
	10 Hz to 45 Hz 330 mA to 1.1 A	0.050 %	
	1.1 A to 3 A	0.04 %	
	45 Hz to 1 kHz 30 μ A to 330 μ A	0.14 % + 60 nA	
	330 μ A to 3.3 mA	0.10 %	
	3.3 mA to 330 mA	0.060 %	
	330 mA to 1.1 A	0.050 %	
	1.1 A to 3 A	0.040 %	
	1 kHz to 5 kHz 30 μ A to 330 μ A	0.24 % + 60 nA	
	330 μ A to 3.3 mA	0.12 %	
	3.3 mA to 33 mA	0.090 %	
	33 mA to 330 mA	0.070 %	
	330 mA to 1.1 A	0.090 %	
	1.1 A to 3 A	0.040 %	
	3 A to 11 A	0.23 %	
	11 A to 20.5 A	0.25 %	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC Current Generation (cont'd)	<i>5 kHz to 10 kHz</i> 30 μ A to 330 μ A 330 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA <i>45 Hz to 100 Hz</i> 3 A to 11 A 11 A to 20.5 A <i>100 Hz to 1 kHz</i> 3 A to 11 A 11 A to 20.5 A <i>10 A to 1025 A</i> 45 Hz to 65 Hz 65 Hz to 100 Hz	0.28 % + 60 nA 0.13 % 0.14 % 0.10 % 0.060 % 0.080 % 0.070 % 0.090 % 0.40 % 0.90 %	For the calibration of clamp meters only
DC Resistance Measurement	0 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 12 k Ω 12 k Ω to 120 k Ω 0.12 M Ω to 1.2 M Ω 1.2 M Ω to 12 M Ω 12 M Ω to 200 M Ω 200 M Ω to 2 G Ω	22 $\mu\Omega/\Omega$ + 1.0 $\mu\Omega$ 12 $\mu\Omega/\Omega$ 8.0 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 18 $\mu\Omega/\Omega$ 42 $\mu\Omega/\Omega$ 630 $\mu\Omega/\Omega$ 900 $\mu\Omega/\Omega$	
Generation (sourcing spot values)	0.0001 Ω 0.001 Ω 0.01 Ω 0.1 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω	660 $\mu\Omega/\Omega$ 290 $\mu\Omega/\Omega$ 180 $\mu\Omega/\Omega$ 80 $\mu\Omega/\Omega$ 28 $\mu\Omega/\Omega$ 28 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 12 $\mu\Omega/\Omega$ 9.0 $\mu\Omega/\Omega$ 9.0 $\mu\Omega/\Omega$ 6.0 $\mu\Omega/\Omega$ 7.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 9.5 $\mu\Omega/\Omega$	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
DC Resistance Generation (sourcing spot values cont'd)	100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	9.0 μΩ/Ω 9.5 μΩ/Ω 9.5 μΩ/Ω 15 μΩ/Ω 29 μΩ/Ω 53 μΩ/Ω 200 μΩ/Ω	
Range Values	0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 kΩ 1.1 kΩ to 3.3 kΩ 3.3 kΩ to 11 kΩ 11 kΩ to 33 kΩ 33 kΩ to 110 kΩ 110 kΩ to 330 kΩ 330 kΩ to 1.1 MΩ 1.1 MΩ to 3.3 MΩ 3.3 MΩ to 11 MΩ 11 MΩ to 33 MΩ 33 MΩ to 110 MΩ 110 MΩ to 330 MΩ 330 MΩ to 1.1 GΩ	71 μΩ/Ω + 60 μΩ 79 μΩ/Ω 34 μΩ/Ω 20 μΩ/Ω 12 μΩ/Ω 14 μΩ/Ω 12 μΩ/Ω 12 μΩ/Ω 12 μΩ/Ω 23 μΩ/Ω 18 μΩ/Ω 59 μΩ/Ω 64 μΩ/Ω 230 μΩ/Ω 300 μΩ/Ω 0.14 % 0.55 %	
DC Power	Voltage: 1 V to 1000 V Current: 10 mA to 20 A 10 mW to 20 kW	0.050 %	
	Voltage: 1 V to 1000 V Current: 1 A to 1000 A 1 W to 1000 kW	0.70 %	For the calibration of power clamp meters
AC Power	50 Hz to 1 kHz Voltage: 1 V to 1000 V Current: 10 mA to 20 A 10 mW to 20 kW	0.21 %	Active and reactive in phase (unity)
	50 Hz to 1 kHz Voltage: 1 V to 1000 V Current: 10 mA to 20 A 10 mW to 20 kW	0.60 %	Reactive power factor -1 to 1
	50 Hz to 1 kHz Voltage: 1 V to 1000 V Current: 1 A to 1000 A 1 W to 1000 kW	0.71 %	For the calibration of power clamp meters



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
Phase Angle	50 Hz to 1 kHz 0 ° to 360 °	0.18 °	
Frequency	10 MHz 1 Hz to 3 GHz	4.0 in 10 ¹² 5.0 in 10 ¹²	Source suitable for counters of sufficient resolution.
	0.1 Hz to 100 MHz 100 MHz to 3 GHz	5.0 in 10 ¹² + 10 mHz 5.0 in 10 ¹²	Measurement capability of stable sources.
Rotational Speed	30 RPM to 100 RPM 100 RPM to 1000 RPM 1000 RPM to 10 kRPM	0.070 RPM 0.70 RPM 1.0 RPM	Optical tachometers
Capacitance Generation	1 kHz 190 pF to 400 pF 0.4 nF to 1.1 nF 1.1 nF to 3.3 nF 3.3 nF to 11 µF 11 µF to 33 µF 33 µF to 110 mF	1.0 % 0.30 % 0.22 % 0.20 % 0.24 % 0.28 %	Values available for calibration of measuring devices.
Measurement	100 Hz to 1kHz 1 pF to 10 µF	0.12 %	Measurement of capacitors.
Temperature Indicators and simulators, calibration by electrical simulation			
Cold junction	21 °C to 25 °C	0.20 °C	For reporting CJ value in ambient conditions for electrical simulation of temperature.
Noble metal thermocouples	0 °C to 1820 °C	0.30 °C	Excluding cold junction compensation
	0 °C to 1820 °C	0.35 °C	Including cold junction compensation
Base metal thermocouples	- 200 °C to - 100 °C - 100 °C to + 1372 °C	0.25 °C 0.20 °C	Excluding cold junction compensation
	- 200 °C to - 100 °C - 100 °C to 120 °C - 120 °C to + 1372 °C	0.25 °C 0.30 °C 0.27 °C	Including cold junction compensation
Resistance sensors (Pt 100)	- 200 °C to + 800 °C	0.020 °C	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
17th Edition capability			
Insulation Resistance	10 kΩ to 5 MΩ 5 MΩ to 90 MΩ 90 MΩ to 1 GΩ 1 GΩ to 10 GΩ	0.070 % 0.36 % 1.2 % 1.4 %	
Insulation Resistance: Voltage measurement	50 V to 1 kV @ 0.5 mA or 1 mA	0.090 %	
Continuity resistance	20 mΩ to 1 Ω 1 Ω to 20 Ω 100 Ω 1 kΩ	1.9 % 1.5 % 0.2 % 0.2 %	
Continuity resistance current	0 mA to 320 mA @ 1 Ω	0.62 %	
Loop impedance	50 Hz 0.2 mΩ to 0.4 Ω 0.4 Ω to 0.8 Ω 0.8 Ω to 3 Ω 3 Ω to 8 Ω 8 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 1 kΩ	30 mΩ 31 mΩ 32 mΩ 33 mΩ 42 mΩ 210 mΩ 1.2 Ω	
RCD Trip Current	50 Hz 1 mA to 60 mA 60 mA to 3 A	0.64 % 0.36 %	
RCD Trip time	20 ms to 400 ms 400 ms to 5 s	0.80 ms 8.3 ms	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
PAT Testers			
Earth Bond resistance	0.2 Ω to 2 Ω	10 mΩ	
	2 Ω to 8 Ω	16 mΩ	
	8 Ω to 20 Ω	29 mΩ	
	20 Ω to 200 Ω	150 mΩ	
	200 Ω to 1 kΩ	1.6 Ω	
Earth bond current	0 mA to 300 mA	3.0 %	
	300 mA to 8 A	0.60 %	
	8A to 30 A	0.50 %	
Insulation resistance	10 kΩ to 5 MΩ	0.070%	
	5 MΩ to 90 MΩ	0.36 %	
	90 MΩ to 300 MΩ	1.2 %	
	300 MΩ to 1 GΩ	1.2 %	
	1 GΩ to 2 GΩ	1.4 %	
Leakage current	50 Hz		
	1 μA to 10 mA	1.6 %	
Load	50 Hz		
	0.13 kW	2.5 %	
Flash voltage	1 kV to 1.8 kV	2.5 %	
	2 kV to 3.6 kV	2.5 %	
Flash current	0.3 mA to 3 mA	4.0 %	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
DIMENSIONAL CALIBRATION			
Length			All linear calibrations may also be made in inch units.
Feeler gauges	BS 957:2008 0.02 to 1.00	2.0	
Gap Gauges (Plain parallel)	BS 969:2008 0.5 to 100 100 to 200 200 to 300	2.0 2.0 4.0	
Length Gauges, Flat and Spherical-ended (excluding length bars)	0 to 3000 Diameter:	1.0 + (8.0 x length in m)	By comparison with reference standards
Plain Plug Gauges (parallel) cylindrical setting standards and rollers	1 to 50 50 to 100 100 to 200 200 to 300	0.80 1.0 1.5 2.5	By comparison with reference standards
Plain ring gauges (parallel)	5 to 15 15 to 50 50 to 100 100 to 150 150 to 200 200 to 500	2.0 1.8 2.0 2.5 3.0 8.0	By comparison with reference standards
Measuring Instruments and Equipment			
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 50	1.0	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
DIMENSIONAL CALIBRATION			
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
Micrometers External	BS 870:2008 (and above) 0 to 1500 Zero Reading Flatness Parallelism Flats	Heads: 2.0 between any two points Setting and extension rods: 1.0 + (8.0 x length in m) 1.0 0.30 0.30	
Internal (including stick micrometers)	BS 959:2008 0 to 1000 Parallelism of Abutment Faces	Heads: 2.0 between any two points Setting and extension rods: 1.0 + (8.0 x length in m) 0.50	
Depth	BS 6468:2008 0 to 300 Flatness Base Mean Axial Error (Run Out) Type R Lateral Shake of Spindle Type S Squareness of Measuring Face to Rod Type R Parallelism of Measuring Face to Datum Type S Mean Axial Error (Run Out) Type R Using Projector Flatness of Rods	Heads: 2.0 between any two points Setting and extension rods: 1.0 + (8.0 x length in m) 0.003 0.009 0.009 0.003 0.003 0.003 0.010 0.0002	
Vernier caliper gauges	BS 887:2008 0 to 1000 Flatness Parallelism Squareness 30mm Jaws	Overall performance 10 + (30 x length in m) 3.0 4.0 7.0	
Vernier depth gauges	BS 6365:2008 0 to 600 Flatness Parallelism	Overall performance 10 + (30 x length in m) 4.0 4.0	
Vernier height gauges	BS 1643:2008 (withdrawn) ISO13225:2012 0 to 1000 Flatness Parallelism of measuring jaw to datum	Overall performance 10 + (30 x length in m) 4.0 4.0	



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AIR VELOCITY Calibration of anemometers and pitot tubes with a digital display	0.3 m/s to 0.8 m/s 0.8 m/s to 1.5 m/s 1.5 m/s to 3 m/s 3 m/s to 5 m/s 5 m/s to 6 m/s 6 m/s to 7 m/s 7 m/s to 9 m/s 9 m/s to 11 m/s 11 m/s to 21 m/s 21 m/s to 26 m/s 26 m/s to 30 m/s	0.14 m/s 0.17 m/s 0.22 m/s 0.32 m/s 0.50 m/s 0.50 m/s 0.65 m/s 0.69 m/s 0.83 m/s 1.3 m/s 1.4 m/s	Method by comparison using an open jet wind tunnel Calibration of devices up to 100 mm diameter may be undertaken
PRESSURE <u>Hydraulic pressure (gauge)</u> Calibration of pressure indicating instruments and gauges <u>Hydraulic pressure (absolute)</u> Calibration of pressure indicating instruments and gauges <u>Gas pressure (gauge)</u> Calibration of pressure indicating instruments and gauges <u>Gas pressure (absolute)</u> Calibration of pressure indicating instruments and gauges	140 kPa to 410 kPa 410 kPa to 4.1 MPa 4.1 MPa to 289 MPa 240 kPa to 510 kPa 510 kPa to 4.2 MPa 4.2 MPa to 289 MPa -90 kPa to -1.5 kPa 0 to 40 Pa 40 Pa to 200Pa 200 Pa to 240 Pa 240 Pa to 400 Pa 400 Pa to 800 Pa 800 Pa to 1.2 kPa 1.2 kPa to 1.5 kPa 1.5 kPa to 200 kPa 200 kPa to 6 MPa 6 MPa to 7.1 MPa 10 kPa to 80 kPa 80 kPa to 115 kPa 115 kPa to 300 kPa 300 kPa to 6.1 MPa 6.1 MPa to 7.2 MPa	Q [0.015 %, 55 Pa] 0.011 % 0.010 % Q [0.015 %, 56 Pa] Q [0.011 %, 12 Pa] Q [0.010 %, 12 Pa] Q [0.0095 %, 0.17 Pa] 2.6 Pa 2.7 Pa 2.8 Pa 3.1 Pa 4.3 Pa 5.8 Pa 7.3 Pa Q [0.0075 %, 0.17 Pa] Q [0.034 %, 7.0 Pa] 0.044 % Q [0.0095 %, 10 Pa] 10 Pa Q [0.0095 %, 10 Pa] Q [0.034 %, 12 Pa] Q [0.044 %, 10 Pa]	Methods consistent with EURAMET CG17 Calibration of pressure measuring devices with an electrical output may be undertaken.
END			



0361

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Micro Precision (Yorkshire) Ltd
Issue No: 047 Issue date: 21 November 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$