


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>0389</b></p> <p><b>Accredited to ISO/IEC 17025:2017</b></p>	<p align="center"><b>BAE Systems Marine Limited</b></p> <p align="center"><b>Issue No: 049    Issue date: 10 May 2025</b></p>	
	<p><b>Building B05, Cavendish Park Barrow-in-Furness Cumbria LA14 1AF</b></p>	<p><b>Contact: Mr David McBride</b>  <b>Tel: +44 (0)1229 - 447788 - 875432</b>  <b>E-Mail: <a href="mailto:calibration.commercial@baesystems.com">calibration.commercial@baesystems.com</a></b>  <b>Website: <a href="http://www.baesystems.com">www.baesystems.com</a></b></p>

**Calibration performed at the above address only**

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>PRESSURE</b>			Methods consistent with EURAMET CG3 and CG17
<u>Hydraulic pressure (gauge)</u>			
Determination of effective area of Dead Weight Testers	0.5 MPa to 140 MPa	0.0070 %	Using deadweight tester
Calibration of pressure indicating instruments and gauges	0.5 MPa to 140 MPa 140 MPa to 450 MPa	0.0070 % 0.010 % + $0.24 \times 10^{-6}$ /MPa	Using deadweight tester
<u>Gas pressure (gauge)</u>			
Determination of effective area of Dead Weight Testers	3.5 kPa to 700 kPa 700 kPa to 7 MPa	0.0036 % 0.0030 %	Using deadweight tester
Calibration of pressure indicating instruments and gauges	-90 kPa to +3.5 kPa 3.5 kPa to 700 kPa 700 kPa to 7 MPa 7 MPa to 20 MPa 20 MPa to 35 MPa 35 MPa to 70 MPa	27 Pa 0.0035 % 0.0035 % 25 kPa 35 kPa 60 kPa	Using pressure calibrator Using deadweight tester Using deadweight tester Using pressure calibrator Using pressure calibrator Using pressure calibrator
<u>Gas pressure (absolute)</u>			
Calibration of pressure indicating instruments and gauges.	3.5 kPa to 700 kPa 700 kPa to 7 MPa	0.0035 % + 1.3 Pa 0.0030 % + 1.3 Pa	Using deadweight tester



0389  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**BAE Systems Marine Limited**  
**Issue No: 049    Issue date: 10 May 2025**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
<b>TEMPERATURE</b>			
Resistance thermometers	-30 °C to 0 °C 0 °C to 50 °C 50 °C to 250 °C 250 °C to 300 °C	0.013 °C 0.020 °C 0.020 °C 0.040 °C	Calibration within liquid baths
	130 °C to 260 °C 260 °C to 520 °C 520 °C to 650 °C	0.80 °C 0.90 °C 0.90 °C	Calibration in a dry block
Thermocouples Base Metal	-30 °C to 300 °C 130 °C to 650 °C 650 °C to 1100 °C	0.16 °C 1.2 °C 4.4 °C	Calibration in liquid bath Calibration in a dry block
Electronic thermometers with sensors Analogue Digital	Range as per sensors	As for sensors plus  Half scale division One least significant digit	
<b>MASS</b>			
<u>Weights and Artefacts</u>	Nominal value g	mg	
	0.001	0.012	Intermediate values can be calibrated but at an appropriate uncertainty which may exceed the value interpolated from the next highest and lowest values.  Substitution method
	0.002	0.012	
	0.005	0.012	
	0.01	0.016	
	0.02	0.020	
	0.05	0.024	
	0.1	0.032	
	0.2	0.040	
	0.5	0.050	
	1	0.060	
	2	0.080	
	5	0.10	
	10	0.12	
	20	0.16	
	50	0.20	
	100	0.32	
	200	0.60	
	500	1.6	
	1000	3.2	
	2000	6.0	
	5000	16	
	10 000	32	
	20 000	60	
	25 000	150	
	30 000	150	



0389  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**BAE Systems Marine Limited**  
Issue No: 049    Issue date: 10 May 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
<b>ELECTRICAL</b>			Electrical calibrations are performed as a comparison against a reference standard unless otherwise stated
DC RESISTANCE			
Specific Values	1 mΩ 10 mΩ 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ	3.2 μΩ/Ω 0.99 μΩ/Ω 0.80 μΩ/Ω 0.30 μΩ/Ω 0.22 μΩ/Ω 0.14 μΩ/Ω 0.22 μΩ/Ω 0.32 μΩ/Ω 1.0 μΩ/Ω 2.1 μΩ/Ω 4.4 μΩ/Ω 7.2 μΩ/Ω 45 μΩ/Ω	Decade values using bridge referenced to standard resistor
Other values	10 μΩ to 1 mΩ 1 mΩ to 100 mΩ 100 mΩ to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 10 kΩ 10 kΩ to 100 MΩ	250 μΩ/Ω 5.0 μΩ/Ω 3.8 μΩ/Ω 0.80 μΩ/Ω 0.70 μΩ/Ω 0.80 μΩ/Ω 12 μΩ/Ω	Resistors suitable for oil immersion can be calibrated at specified temperatures from 18 °C to 26 °C Resistors suitable for high current can be calibrated at test currents up to 100 A
DC VOLTAGE			Sourcing and measurement capability for the calibration of voltage instruments
Standard Cell Values	1.018 V	0.92 μV/V	
Other Values	0.1 V 1 V 10 V 100 V 1000 V	8.5 μV/V 0.92 μV/V 0.29 μV/V 0.53 μV/V 0.70 μV/V	
Range Values	0 V to 1 mV 1 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 1100 V 1 kV to 20 kV	0.60 % + 0.10 μV 4.0 μV/V + 1.2 μV 4.0 μV/V + 1.2 μV 4.0 μV/V + 1.2 μV 5.0 μV/V 10 μV/V 0.060 % + 40 V	
Values for Temperature simulation			
Measurement	-10 mV to +120 mV	1.8 μV	
Generation	-10 mV to +120 mV	1.8 μV	
DC CURRENT	1 μA to 10 μA 10 μA to 1 A 1 A to 10 A 10 A to 100 A	35 μA/A 20 μA/A 25 μA/A 27 μA/A	Sourcing and measurement capability for the calibration of current instruments



0389  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**BAE Systems Marine Limited**  
**Issue No: 049    Issue date: 10 May 2025**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty    ( $k = 2$ )	Remarks
AC VOLTAGE  Specific Values	0.3 V 10 Hz, 20 Hz, 40 Hz, 300 Hz, 1 kHz, 10 kHz and 20 kHz 50 kHz and 100 kHz 300 kHz, 500 kHz and 1 MHz  1 V, 3 V, 10 V and 100 V 10 Hz, 20 Hz, 40 Hz, 300 Hz, 1 kHz, 10 kHz and 20 kHz 50 kHz and 100 kHz  30 V 10 Hz, 300 Hz, 1 kHz, 10 kHz and 20 kHz  1 V and 10 V 300 kHz, 500 kHz and 1 MHz  3 V 500 kHz and 1 MHz  300 V and 1000 V 10 Hz, 20 Hz, 40 Hz, 300 Hz, 1 kHz, 10 kHz and 20 kHz  300 V 50 kHz and 100 kHz  700 V 20 kHz, 50 kHz and 100 kHz	30 $\mu\text{V/V}$ 50 $\mu\text{V/V}$ 0.075 %  25 $\mu\text{V/V}$ 50 $\mu\text{V/V}$  25 $\mu\text{V/V}$  0.071 %  0.071 %  31 $\mu\text{V/V}$  60 $\mu\text{V/V}$  0.075 %	For the calibration of voltage measuring and generating equipment.
Other values	1 mV to 10 mV 50 Hz to 5 kHz 5 kHz to 20 kHz  10 mV to 31 mV 50 Hz to 5 kHz 5 kHz to 20 kHz  0.03 V to 0.1 V 10 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 1 MHz  0.1 V to 30 V 10 Hz to 20 kHz 20 kHz to 100 kHz  0.1 V to 10 V 100 kHz to 1 MHz	0.13 % 1.1 %  0.017 % 0.11 %  42 $\mu\text{V/V}$ 100 $\mu\text{V/V}$ 0.11 %  44 $\mu\text{V/V}$ 97 $\mu\text{V/V}$  0.11 %	For the calibration of voltage measuring and generating equipment.



**0389**  
Accredited to  
**ISO/IEC 17025:2017**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**BAE Systems Marine Limited**  
**Issue No: 049    Issue date: 10 May 2025**

**Calibration performed at main address only**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
AC VOLTAGE (continued)			
Other values (continued)	30 V to 340 V 10 Hz to 20 kHz 20 kHz to 100 kHz	52 $\mu$ V/V 150 $\mu$ V/V	For the calibration of voltage measuring and generating equipment.
	340 V to 1100 V 10 Hz to 20 kHz	58 $\mu$ V/V	
	340 V to 700 V 20 kHz to 100 kHz	0.075 %	
AC CURRENT	100 $\mu$ A to 400 mA 40 Hz to 1 kHz	130 $\mu$ A/A	By bridge measurement And transfer
	400 mA to 10 A 40 Hz to 1 kHz	190 $\mu$ A/A	
CAPACITANCE	At 100 Hz 10 $\mu$ F to 100 $\mu$ F 100 $\mu$ F to 10 mF	0.36 % 0.40 %	
	At 1 kHz: 1 pF to 10 pF 10 pF to 100 pF 100 pF to 100 nF 1 $\mu$ F to 10 $\mu$ F	0.20 % + 0.0020 pF 0.030 % 0.026 % 0.12 %	
	At 10 kHz 1 nF to 1 $\mu$ F	0.20 %	
	At 100 kHz 100 pF to 1 nF	0.38 %	
FREQUENCY MEASUREMENT	0.001 Hz to 0.01 Hz 0.01 Hz to 100 kHz 100 kHz to 500 MHz	0.35 % 2.0 in $10^{11}$ + 35 $\mu$ Hz 3.0 to 5.0 parts in $10^{11}$	
Specific Values	1 MHz, 5 MHz and 10 MHz	1.6 parts in $10^{12}$	
TIME INTERVAL	5 ms to 55 ms	0.34 %	
TEMPERATURE SIMULATION			
Type K thermocouples	-100 °C to 0 °C 0 °C to 1370 °C	0.13 °C 0.12 °C	Devices fitted with an internal cold junction.
PT 100	0 °C	0.028 °C	



**0389**  
Accredited to  
**ISO/IEC 17025:2017**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**BAE Systems Marine Limited**  
**Issue No: 049    Issue date: 10 May 2025**

**Calibration performed at main address only**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
<b>ACCELEROMETRY</b>  ACCELERATION TRANSDUCERS  Accelerometer Types: Piezo electric minimum sensitivity >1 pC/m/s <sup>2</sup> Integral electronic minimum sensitivity >1 mV/m/s <sup>2</sup>  All types with a nominal mass of up to 250 grams, a nominal Peak Acceleration of 0.98 ms <sup>-2</sup> to 98 ms <sup>-2</sup>	5 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz 100 Hz to 920 Hz 920 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz	1.6 % 1.3 % 1.0 % 1.2 % 1.8 % 2.3 % 2.8 % 4.5 %	Calibration method is by direct comparison against laboratory references  Transducer at 23°C
<b>LENGTH</b>			
<b>MEASURING INSTRUMENTS AND MACHINES</b>			
Micrometers - External  Ancillary Measurements  Calliper type gauges including dial and digital  Dial gauges and dial test indicators	0 mm to 25 mm 25 mm to 100 mm  Flatness Parallelism  ISO 13385-1:2019 Partial surface contact error (E ) 0 m to 150 mm  Shift error (S) – Internal measuring faces 20 mm to 50 mm  Shift error (S) – Crossed knife- edge internal measuring faces 5 mm  Shift error (S) – Depth or step measuring faces <50 mm  Line contact error 15 mm  0 mm to 50 mm	1.4 µm 1.6 µm  0.70 µm 1.0 µm  1.2 µm  5.4 µm  5.4 µm  4.6 µm  4.6 µm  13 µm	Calibration as BS 870:2008 Heads, between any two points  Calibration by comparison to length standards  The stated uncertainty has been calculated in accordance With ISO 14253-5 and relates to the test value uncertainty.  The uncertainty quoted Excludes contributions relating the instrument under test.  Calibration as BS 2795:1981 BS 907: 2008
END			



0389

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**BAE Systems Marine Limited**  
**Issue No: 049    Issue date: 10 May 2025**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$