


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>0446</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Stark Energy Services Limited</h3> <p>Issue No: 039      Issue date: 27 March 2026</p>	
	<p>Commissioners Road Medway City Estate Rochester Kent ME2 4GN</p>	<p>Contact: Mr Ravindu Rathnayaka Tel: +44 (0) 1634 715380 E-Mail: ravindu.rathnayaka@stark.co.uk</p>
<p><b>Calibration performed at the above address only</b></p>		

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
AC VOLTAGE	30 V to 500 V 45 Hz to 60 Hz	100 $\mu$ V/V	Using rms detection.
AC CURRENT	45 Hz to 60 Hz 10 mA to 50 mA 50 mA to 10 A 10 A to 120 A	100 $\mu$ A/A 100 $\mu$ A/A 100 $\mu$ A/A	Using rms detection.
FREQUENCY	10 Hz to 10 MHz 10 MHz to 100 MHz	0.10 $\mu$ Hz/Hz 1.0 $\mu$ Hz/Hz	Using frequency counter.
AC POWER / ENERGY  ACTIVE POWER / ENERGY	45 Hz to 60 Hz  <i>Cos <math>\phi = 1</math></i> 10 mA to 10 A 10A to 120 A  <i>Cos <math>\phi = 0.5</math> Lead to 0.5 Lag</i> 10 mA to 10 A 10 A to 120 A  <i>Cos <math>\phi = 0.25</math> Lead to 0.25 to Lag</i> 10 mA to 10 A 10 A to 120 A	  100 $\mu$ W/W or $\mu$ Whr/Whr 100 $\mu$ W/W or $\mu$ Whr/Whr  100 $\mu$ W/W or $\mu$ Whr/Whr 100 $\mu$ W/W or $\mu$ Whr/Whr  100 $\mu$ W/W or $\mu$ Whr/Whr 100 $\mu$ W/W or $\mu$ Whr/Whr	For voltages between 30 V and 500 V at frequencies between 45 Hz and 60 Hz using a phantom load technique. Single and Poly Phase devices may be calibrated.



0446  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

**Stark Energy Services Limited**  
Issue No: 039 Issue date: 27 March 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
AC POWER / ENERGY (continued)  REACTIVE POWER / ENERGY	<p><i>Sin <math>\phi = 1</math></i> 10 mA to 10 A 10 A to 120 A</p> <p><i>Sin <math>\phi = 0.5</math> Lead to 0.5 Lag</i> 10 mA to 10 A 10 A to 120 A</p> <p><i>Sin <math>\phi = 0.25</math> Lead to 0.25 Lag</i> 10 mA to 10 A 10 A to 120 A</p>	<p>100 <math>\mu</math>W/W or <math>\mu</math>Whr/Whr 100 <math>\mu</math>W/W or <math>\mu</math>Whr/Whr</p> <p>100 <math>\mu</math>W/W or <math>\mu</math>Whr/Whr 100 <math>\mu</math>W/W or <math>\mu</math>Whr/Whr</p> <p>100 <math>\mu</math>W/W or <math>\mu</math>Whr/Whr 100 <math>\mu</math>W/W or <math>\mu</math>Whr/Whr</p>	For voltages between 30 V and 500 V at frequencies between 45 Hz and 60 Hz using a phantom load technique. Single and Poly Phase devices may be calibrated.
END			



0446  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Stark Energy Services Limited**

**Issue No:** 039 **Issue date:** 27 March 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$