# **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



# Locations covered by the organisation and their relevant activities

## Laboratory locations:

Location details		Activity	Location code	
Address Longmead Business Centre Blenheim Road Epsom KT19 9QQ	Local contact: Helen Keogh	Volume Rotational speed Time Temperature	Perm	

# Site activities performed away from the locations listed above:

Location details		Activity	Location code
At customer premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Contact: Helen Keogh	Calibration of Non-Automatic Weighting Machines Volume Rotational speed Time Temperature	Site

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	Sartorius UK Limited		
0455 Accredited to ISO/IEC 17025:2017	Issue No: 034 Issue date: 12 May 2025		
	Calibration performed by the Organisation at the locations specified		

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
NON-AUTOMATIC WEIGHING MACHINES Self-indicating and semi-self- indicating instruments. See notes 1 and 2.	5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 100 g 200 g 500 g 100 g 200 g 500 g 100 g 200 g 500 g 100 s 200 g 500 s 100 s 200 g 500 s 100 s 200 s 500 s 200 s 500 s 200 s 500 s 200 s 500 s 200 s 500 s 200 s 500 s 200 s 3000 kg 3000 kg 3000 kg 3000 kg 3000 kg	0.007 mg 0.009 mg 0.011 mg 0.015 mg 0.018 mg 0.022 mg 0.029 mg 0.029 mg 0.037 mg 0.044 mg 0.059 mg 0.074 mg 0.093 mg 0.074 mg 0.093 mg 0.11 mg 0.37 mg 0.37 mg 0.33 mg 1.8 mg 3.7 mg 9.3 mg 1.8 mg 3.7 mg 9.3 mg 1.8 mg 200 mg 500 mg 10 g 20 g 50 g 100 g 200 g 300 g	Note 1 Weights are available in OIML class: E2 1 mg to 5 kg. max grouped load 11 kg. F1 10 kg to 20 kg. Max grouped load 60 kg. M1 10 kg to 250 kg. Max grouped load 3000 kg Note 2. Other loads within the overall listed range may also be used. Note 3.Method based on the requirements of Euramet guide cg-18	Site
VOLUME of liquids (See Notes 3 and 4)	Nominal value Single Channel Pipettes $0.2 \ \mu L to 10.0 \ \mu L$ $10.0 \ \mu L to 20.0 \ \mu L$ $20.0 \ \mu L to 100 \ \mu L$ $100 \ \mu L to 200 \ \mu L$ $200 \ \mu L to 500 \ \mu L$ $0.5 \ m L to 1 \ m L$ $1 \ m L to 2 \ m L$ $2 \ m L to 5 \ m L$ $5 \ m L to 10 \ m L$ Multi Channel Pipettes $0.5 \ \mu L to 10 \ \mu L$ $10.0 \ \mu L to 20.0 \ \mu L$ $20.0 \ \mu L to 200 \ \mu L$ $20.0 \ \mu L to 100 \ \mu L$ $100 \ \mu L to 200 \ \mu L$ $200 \ \mu L to 500 \ \mu L$ $0.5 \ m L to 1.25 \ m L$	0.055 μL 0.098 μL 0.31 μL 0.68 μL 1.5 μL 0.001 7 mL 0.004 6 mL 0.015 mL 0.026 μL 0.27 μL 1.2 μL 2.0 μL 0.004 6 mL	Note 3. For water delivered from piston and/or plunger operated volumetric apparatus using procedure as defined in ISO 8655:2022 (gravimetric method). Single Channel Pipettes and Multi-Channel Pipettes Note 4. Users requiring conformity to ISO 8655-6:2022 should note that this cannot be demonstrated using fewer than 10 readings.	Perm & Site

# Calibration and Measurement Capability (CMC)

1

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
0455 Accredited to ISO/IEC 17025:2017	Sartorius UK Limited Issue No: 034 Issue date: 12 May 2025			
Calibration performed by the Organisation at the locations specified				
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code

Instrument or Gauge		Uncertainty $(k = 2)$		Code
ROTATIONAL SPEED				Perm & Site
Centrifuges	100 rpm to 1000 rpm 1000 rpm to 5000 rpm 5000 rpm to 15000 rpm 15000 rpm to 26000 rpm	1.2 rpm 3.5 rpm 10.0 rpm 16.0 rpm	Calibration via Optical Tachometer	
TEMPERATURE				Perm & Site
Temperature controlled heat blocks	37 °C nominal	1.1 °C	For calibration of <i>Ortho</i> <i>Clinical Diagnostics</i> heat	
Temperature probes built in to centrifuges	0 °C to 10 °C	2.4 °C	devices.	
TIME				Perm & Site
Timers	30 s to 900 s	0.5 s	Calibration to a reference stopwatch	
END				



### Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### **Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$