

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0459 Accredited to ISO/IEC 17025:2017	Precision Balance Services Limited	
	Issue No: 022 Issue date: 19 November 2024	
	3 Atlas Court Atlas Road Hermitage Industrial Estate Coalville Leicestershire LE67 3FL	Contact: Mr Ian Hughes Tel: +44 (0)1530-834650 Fax: +44 (0)1530-834650 E-Mail: service@precisionbalance.co.uk Website: www.precisionbalance.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Location details	Activity	Location code
Address 3 Atlas Court Atlas Road Hermitage Industrial Estate Coalville Leicestershire LE67 3FL Local contact Mr Ian Hughes	Portable weighing machines (non-automatic)	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Sites The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. Contact Mr Ian Hughes	MASS Weighing machines (non-automatic)	S



0459
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Precision Balance Services Limited
Issue No: 022 **Issue date:** 19 November 2024

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
NON-AUTOMATIC WEIGHING MACHINES (self-indicating instruments) (From 1 mg to 500 kg)	20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 120 kg 200 kg 500 kg	0.0076 mg 0.010 mg 0.013 mg 0.015 mg 0.020 mg 0.025 mg 0.030 mg 0.038 mg 0.052 mg 0.067 mg 0.095 mg 0.17 mg 0.34 mg 0.86 mg 1.9 mg 2.9 mg 8.6 mg 17 mg 56 mg 160 mg 340 mg 5.1 g 18 g	1. Weights are available in OIML Class: E2 from 1 mg to 5 kg. Max grouped load 11 kg. F1 1 mg to 10 kg. Max grouped load 120 kg M1 5 kg to 20 kg. Max grouped load 500 kg. 2. Other loads within the overall listed range may also be used. 3. Method based on the requirements of Euramet guide cg-18	S and P
END				



0459
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Precision Balance Services Limited
Issue No: 022 Issue date: 19 November 2024

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$