


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0495 Accredited to ISO/IEC 17025:2017	Lambda Calibration Ltd Issue No: 059 Issue date: 13 September 2024	
	Units 11-13 Chorley Central Business Park Stump Lane Chorley Lancashire PR6 0BL	Contact: Denise Catterall Tel: +44 (0)125 724 4670 Fax: +44 (0)125 724 4671 E-Mail: mail@lambda-cal.co.uk Website: www.lambda-cal.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Lambda Calibration Ltd Units 11-13 Chorley Central Business Park Stump Lane Chorley Lancashire PR6 0BL Local contact Denise Catterall	Dimensional Electrical Pressure Torque Temperature	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises Denise Catterall	Dimensional Electrical Pressure	B



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
NOTES:				
Note 1: The uncertainty quoted is for the departure from flatness, straightness, parallelism or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration.				
Note 2: Single start, symmetrical thread forms only.				
Note 3: Functional test of size using setting plugs calibrated with an expended uncertainty of 3.0 μm				
Note 4: . Features and associated parts of these gauges can be measured to the uncertainties given for equivalent items listed in this schedule.				
Note 5: The uncertainty quoted is for the application of the calibration torque and does not take into account the characteristics of the device being calibrated.				
Note 6: Calibrations may also be given in lbf.in and lbf.ft				
Note 7: . Simple height gauges - vernier, dial and digital instruments designed only for measuring distances parallel to the beam.				
Note 8: Conformance statements cannot be made against specifications whose magnitudes are smaller than the specified CMC values				
Note 9: Class C uncertainties apply to the measurement of length of gauges by comparison with grade K standards of length of a similar material. Class C uncertainties apply to new and used grade 0, 1 and 2 gauges to BS 4311-1:2007 and BS EN ISO 3650:1999.				
Note 10: Ancillary measurements made for completeness of calibration. Best CMC's are dependent on methodology and range.				



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k=2)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH				
Gauge blocks		Class C (See Note 9)		A
Inch (Steel and Carbide)	BS 4311-1:2007 0.010 to 0.4 in 0.4 in to 1 in Size 2 in Size 3 in Size 4 in	C 3.0 4.0 5.0 μ inches 6.0 7.0		
Millimetre (Steel and Carbide)	BS EN ISO 3650:1999 0.5 to 10 10 to 25 Sizes 30, 40, 50 60, 70, 75 80, 90, 100	C 0.080 0.10 0.12 0.15 0.18		A
Gauge block accessories	As BS 4311-2:2009 0.1 to 12.5	0.30		A
Length gauges, flat and spherical ended (excluding Length Bars)	0 to 1000	1.0 + (5.0 x length m)	By comparison with reference standards	A
Plain plug gauges parallel, cylindrical setting standards and rollers.	1 to 50 diameter 50 to 100 100 to 300	0.80 1.0 1.5	By comparison with reference standards	A
Plain ring gauges (parallel)	0.7 to 10 diameter 10 to 50 50 to 100 100 to 300	1.0 0.80 1.0 2.5	By comparison with reference standards	A
Plain plug gauges (taper)				A
Parallel to 1 in 3 on diameter	3 to 200 diameter	3.0	By comparison with reference standards	
Plain ring gauges (taper)				A
Parallel to 1 in 3 on diameter	3 to 100 diameter 100 to 200 diameter	3.0 5.0	By comparison with reference standards	
Precision balls (steel, carbide, ceramic)	1 to 100 diameter	0.80	By comparison with reference standards	
Feeler Gauges	BS 957:2008.0.0025 to 1	1.0	By comparison with reference standards	A



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH (cont'd)				
Gap gauges (Plain parallel)	2 to 100 100 to 200 200 to 300	3.0 5.0 8.0	By comparison with reference standards	A
Paint thickness setting foils	0 to 8	1.0	By comparison with reference standards	A
Rule – steel	BS 4372:1968 0 to 1000 1000 to 2000	5.0 + (10 x length in m) 8.0 + (10 x length in m)		A
Linear Precision scales	0 to 400	5.0	By optical measurement 2 wire method	A
SCREW THREAD GAUGES				
Screw plug gauges (parallel) including check and setting plugs. See Note 2	1 to 100 diameter 100 to 150 diameter	3.0 on pitch 5.0 diameter		A
Screw plug gauges (taper) See Note 4	5 to 100 diameter 100 to 150 diameter	5.0 on pitch 8.0 diameter		
Screw ring gauges (parallel) See Note 2	1 to 75 diameter 75 to 150 diameter	5.0 on pitch 7.0 diameter	By comparison with reference standards	
Screw ring gauges (taper) See Note 2	5 to 100 diameter 100 to 150 diameter	7.0 on pitch 10.0 diameter		
Screw pitch Screw flank angle	0.2 to 8 0° to 52°	1.5 5.0 minutes of arc		
Screw thread adjustable caliper gauges (parallel) See Note 2	1 to 150	See note 3		A
Thread measuring cylinders	BS 5590:1978 and specials 0.1 to 5	0.50		A
Parallels	BS 906:1972 5 to 50 x 100 x 400	1.5 to 5.0		A
Vee blocks	BS 3731:1987 20 to 200	2.5 to 5.0		A
Receiver and position gauges, jigs, fixtures	0 to 600 x 600 x 600 0 to 700 x 450 x 250 (Using a coordinate measuring machine)	See note 4 17µm + 8µm/m	Documented in-house method	A



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
SCREW THREAD GAUGES				
Thread vee groove jaw blades	0.6 (40 T.P.I.) to 6.0 (4.5 T.P.I)	3.0	Documented in-house method	A
ANGLE				
Squares				A
Blade type	BS 939:2007 50 to 300 300 to 600	3.0 5.0		A
Cylindrical	75 to 300 300 up to 600	2.0 2.0 On squareness see Note 1		A
Block	50 to 300 300 to 600	3.0 5.0		A
Right angle and box angle plates	BS 5535:1878 50 to 600	Squareness: 3.0 + 1.0 per 100 mm Parallelism: 1.0 + 1.0 per 100 mm See Note 1		A
Sine bars and tables	BS 3064:1978 0 to 500 length	Linear dimensions 1.0 + (10 x length in m) Overall performance 3.0 seconds of arc		A
Clinometers	0° to 360°	10 seconds of arc	By comparison to reference standards	A
Electronic indicating levels	0 minutes of arc to 8 minutes of arc	1.0 % of range Minimum 0.40 seconds of arc		A
Spirit levels	BS 3509:1962 and BS 958:1968 5 seconds of arc to 60 minutes of arc nominal sensitivity	Mean sensitivity 10% of nominal Minimum 0.50 seconds of arc		A
Rotary tables and dividing heads, including inclinable versions	0 ° to 360 ° angular scales 0 to 600 Capacity	3.0 seconds of arc Linear measurements: 1.0 + (10 x length in m)	By comparison with angle standards. By comparison with end standards and datum planes.	A, B



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
MEASURING INSTRUMENTS				
Micrometers				
Micrometers				A
External	BS 870:2008 0 to 1000	Heads: 2.0		
Internal	BS 959:2008 0 to 900	Setting and extension rods		
Depth	BS 6468:2008 0 to 300	1.0 + (5.0 x length in m)		
Micrometer, 3 point bore	6 to 250	3.0 + (12 x length in m)	By comparison to reference standards	A
Micrometer Heads	BS 1734:1951 0 to 50	1.0		A
Height setting micrometer	0 to 300	Heads 1.5 between any two points Stepped column 2.5 Overall performance 3.0	By comparison to reference standards	A
Riser blocks for above	150 300	2.5 5.0	By comparison to reference standards	A
Vernier, digital electronic, dial caliper, height and depth gauges	BS 887:2008 0 to 1200 BS 1643:2008 0 to 1200 BS 6365:2008 0 to 600	Overall performance 10 + (30 x length in m)		A
	ISO 13385-1 2019 Partial surface contact error (E) 0 to 50 mm 50 to 100 mm 100 to 200 mm 200 to 300 mm 300 to 2000 mm	2.0 3.0 6.0 8.0 6 + (30 x length in m)	Calibration by comparison to length standards The stated uncertainty has been calculated in accordance With ISO 14253-5 and relates to the test value uncertainty.	A
	Shift error (S) internal jaws 3 to 50 mm	6.0	The uncertainty quoted Excludes contributions relating the instrument under test.	
	Shift error (S) depth and step 3 to 50 mm	6.0		
Height gauges - (Simple) including vernier, dial and digital types (See notes 7 and 8)	As BS EN ISO 13225:2012 0 to 1200	Length measurement error (E): 1 + (5 x length in metres)		A



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
MEASURING INSTRUMENTS (cont'd)				
Bevel protractors	As BS 1685:2008 0° to 360°	6.0 minutes of arc		A
Dial gauges and dial test indicators	BS 907:2008 and BS 2795:1981 0 to 50	1.0		A
Displacement transducers (linear)	0 to 100	0.08	By comparison to reference standards	A
Height gauges (digital and electronic)	0 to 1000	1.0 + (5 x length in m)	By comparison to reference standards	A
Straightedges Cast iron, Steel and Granite	BS 5204:Part 1:1975 BS 5204:Part 2:1977 0 to 2000	4.0 See note 1		A
Tool makers flats	63 to 150	0.15 See note 1	By comparison to reference standards	A
Cube moulds for cement and concrete	BS EN 12390-1 2012 100 to 150 x 150 x 150 BS EN 196-1 2005 160 x 40 x 40	17µm + 8µm/m		A
Profile projectors	5 to 100 magnifications 0 to 300 linear scales 0 ° to 360 ° angular scales	125 at the screen 4.0 3.0 minutes of arc	By comparison with end standards. Using optical methods. Using two point contact method.	A, B
Horizontal measuring machines	0 to 100	0.30 + (3.0 x L in m) Flatness of Anvils: 0.20 Parallelism of Anvils: 0.15	By comparison with end standards. Using optical methods. Using two point contact method.	A, B
FORM				
Surface plates				
Granite and Cast Iron	BS 817:2008 160 x 100 to 1600 x 1000	1.5 + (1.0 x diagonal in m)		A, B



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
ANCILLERY MEASUREMENTS				
Flatness		0.25	See note 10	A
Parallelism		0.5		
Squareness		3.0		
Straightness		$1.5 + (1.0 \times \text{length in m})$		
Angular		5 second of arc		

ELECTRICAL MEASUREMENTS: All values and uncertainties listed below are applicable for the calibration of both measurement instruments and for instruments with an output. the method used is by direct comparison unless otherwise stated in the remarks column. Uncertainties listed apply for the laboratory environment of 18°C to 22°C. Measurements can take place outside these limits but at increased uncertainties.

ELECTRICAL				
RESISTANCE	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω 2 G Ω to 20 G Ω	$\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega + 15 \mu\Omega$ 11 $\mu\Omega/\Omega + 20 \mu\Omega$ 9.5 $\mu\Omega/\Omega + 110 \mu\Omega$ 9.5 $\mu\Omega/\Omega + 600 \mu\Omega$ 9.5 $\mu\Omega/\Omega + 6.4 \text{ m}\Omega$ 9.7 $\mu\Omega/\Omega + 66 \text{ m}\Omega$ 12 $\mu\Omega/\Omega + 1.5 \Omega$ 24 $\mu\Omega/\Omega + 16 \Omega$ 87 $\mu\Omega/\Omega + 2.2 \text{ k}\Omega$ 240 $\mu\Omega/\Omega + 100 \text{ k}\Omega$ 0.16 % + 10 M Ω		A, B
Current carrying resistors DC	100 $\mu\Omega$ to 100 m Ω 1 m Ω to 100 m Ω 1 m Ω to 100 m Ω	0.035 % 0.050 % 0.13 %	50 A to 100 A 2 A to 5 A 10 A to 20 A	A
AC RESISTANCE	0.5 Ω to 2 k Ω 50 Hz	0.65 %		A, B
DC VOLTAGE	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1050 V 1 kV to 10 kV	7.1 $\mu\text{V}/\text{V} + 600 \text{ nV}$ 4.1 $\mu\text{V}/\text{V} + 800 \text{ nV}$ 4.1 $\mu\text{V}/\text{V} + 4.8 \mu\text{V}$ 6.3 $\mu\text{V}/\text{V} + 47 \mu\text{V}$ 6.5 $\mu\text{V}/\text{V} + 720 \mu\text{V}$ 0.050 %		A, B



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL (cont'd)				
DC CURRENT	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A 20 A to 100 A	15 μ A/A + 500 pA 16 μ A/A + 4.8 nA 17 μ A/A + 48 nA 54 μ A/A + 900 nA 200 μ A/A + 17 μ A 170 μ A/A + 61 μ A 440 μ A/A + 410 μ A 0.050 % + 0.58 mA		A, B
DC CURRENT Generation only	3.2 A to 105 A 105 A to 200 A 16 A to 160 A 160 A to 525 A 525 A to 1 kA	0.24 % + 10 mA 0.24 % + 46 mA 0.24 % + 6.4 mA 0.24 % + 52 mA 0.24 % + 230 mA	Suitable for Clamp Meters Using 10 turn coil Using 10 turn coil Using 50 turn coil Using 50 turn coil Using 50 turn coil	A, B
AC VOLTAGE	1 μ V to 200 mV 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	170 μ V/V + 4.2 μ V 170 μ V/V + 4.4 μ V 360 μ V/V + 5.0 μ V 720 μ V/V + 20 μ V		A, B
	200 mV to 2 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	130 μ V/V + 23 μ V 130 μ V/V + 24 μ V 240 μ V/V + 30 μ V 500 μ V/V + 100 μ V		A, B
	2 V to 20 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	130 μ V/V + 230 μ V 130 μ V/V + 230 μ V 240 μ V/V + 300 μ V 500 μ V/V + 1.0 mV		A, B
	20 V to 200 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	140 μ V/V + 2.3 mV 130 μ V/V + 2.4 mV 240 μ V/V + 4.0 mV 530 μ V/V + 11 mV		
	200 V to 300 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz	210 μ V/V + 71 mV 140 μ V/V + 25 mV 270 μ V/V + 44 mV		
	300 V to 1050 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz	210 μ V/V + 71 mV 170 μ V/V + 51 mV 340 μ V/V + 150 mV		
	1 kV to 10 kV @ 50 Hz	1.5 %		



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL (cont'd) AC CURRENT	1 nA to 200 μ A 10 Hz to 1 kHz 1 kHz to 5 kHz 200 μ A to 2 mA 10 Hz to 1 kHz 1 kHz to 10 kHz 2 mA to 20 mA 10 Hz to 1 kHz 1 kHz to 10 kHz 20 mA to 200 mA 10 Hz to 1 kHz 1 kHz to 10 kHz 200 mA to 500 mA 40 Hz to 3 kHz 3 kHz to 5 kHz 200 mA to 2 A 10 Hz to 1 kHz 1 kHz to 10 kHz 2 A to 4 A 40 Hz to 3 kHz 3 kHz to 5 kHz 2 A to 20 A 10 Hz to 1 kHz 1 kHz to 10 kHz 20A to 50A 50 Hz	380 μ A/A + 21 nA 600 μ A/A + 21 nA 380 μ A/A + 210 nA 360 μ A/A + 200 nA 360 μ A/A + 2.1 μ A 350 μ A/A + 2.0 μ A 380 μ A/A + 21 μ A 330 μ A/A + 20 μ A 0.12 % 0.15 % 670 μ A/A + 230 μ A 870 μ A/A + 210 μ A 0.16 % 0.27 % 920 μ A/A + 2.4 mA 0.26 % + 2.1 mA 0.37%		A, B
AC CURRENT Generation only	3.2 A to 32 A (10 to 100 Hz) 32 A to 200 A (10 to 100 Hz) 3.2 A to 32 A (100 to 440 Hz) 32 A to 200 A (100 to 440 Hz) 16 A to 160 A (10 to 100 Hz) 160 A to 1 kA (10 to 100 Hz) 10 A to 50 A @ 50 Hz	0.33 % + 6.0 mA 0.33 % + 100 mA 0.93 % + 31 mA 0.81 % + 290 mA 0.33 % + 320 mA 0.33 % + 520 mA 0.38 %	Suitable for Clamp Meters Using 10 turn coil Using 10 turn coil Using 10 turn coil Using 10 turn coil Using 50 turn coil Using 50 turn coil	A, B




0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL (cont'd) DISTORTION Distortion Factor	0.01 % to 0.1 % 0.2 V to 300 V 20 Hz to 20 kHz 20 kHz to 100 kHz 0.1 % to 1 % 0.2 V to 300 V 20 Hz to 20 kHz 20 kHz to 100 kHz 1 % to 30 % 0.2 V to 300 V 20 Hz to 20 kHz 20 kHz to 100 kHz	0.015 % distortion factor 0.029 % distortion factor 0.15 % distortion factor 0.29 % distortion factor 15 % of reading 30 % of reading	The capabilities for distortion factor relate to fundamental components in the frequency range 20 Hz to 100 kHz.	A, B
Spot frequency	1 kHz and 100 mV to 300 V 0.01 % to 0.1 % 0.1 % to 3 % 3 % to 30 %	0.0028 % distortion factor 3.1 % of reading 0.41 % of reading	Calibration of distortion Analyzers	A
Phase angle	0 ° to 360 ° 10 Hz to 10 kHz	0.22 °		



issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd

Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified


ELECTRICAL SIMULATION OF THERMOCOUPLES

UUT with RJC
Type:

B	+500 °C to 1820 °C
C	0 °C to 2320 °C
E	-200 °C to 1000 °C
J	-210 °C to -100 °C
J	-100 °C to 800 °C
J	800 °C to 1200 °C
K	-250 °C to -200 °C
K	-200 °C to -100 °C
K	-100 °C to 100 °C
K	100 °C to 1372 °C
L	-200 °C to 900 °C
N	-200 °C to -100 °C
N	-100 °C to 900 °C
N	900 °C to 1300 °C
R	0 °C to 200 °C
R	200 °C to 1767 °C
S	0 °C to 200 °C
S	200 °C to 1767 °C
T	-250 °C to -200 °C
T	-200 °C to -100 °C
T	-100 °C to 0 °C
T	0 °C to 400 °C

UUT without RJC
Type:

B	250 °C to +500 °C	0.520 °C
B	+500 °C to +1820 °C	0.260 °C
E	-200 °C to +1000 °C	0.021 °C
J	-200 °C to -80 °C	0.059 °C
J	-80 °C to +1200 °C	0.030 °C
K	-200 °C to -80 °C	0.082 °C
K	-80 °C to +1372 °C	0.039 °C
N	-200 °C to -80 °C	0.130 °C
N	-80 °C to +1300 °C	0.056 °C
R	0 °C to +20 °C	0.260 °C
R	+20 °C to +1767 °C	0.220 °C
S	0 °C to +20 °C	0.260 °C
S	+20 °C to +1767 °C	0.220 °C
T	-200 °C to -80 °C	0.082 °C
T	-80 °C to +400 °C	0.043 °C



issued by

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

PRT Simulation
At 1 mA

Spot Values - nominal

Pt100 (temperature equivalents)
BS EN 60751:2008

Intermediate Values

ELECTRICAL MEASUREMENT

Thermistor Simulation

Thermometer Readout
(Ref Junction Compensation)

Optical Tachometers



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
CAPACITANCE Generation only	0 to 350 Hz 0.5 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μ F 4 μ F to 40 μ F 40 μ F to 400 μ F 400 μ F to 4 mF 4 mF to 40 mF 350 Hz to 1.5 kHz 0.5 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μ F 4 μ F to 40 μ F 40 μ F to 400 μ F 400 μ F to 4 mF 4 mF to 40 mF	0.35 % + 18 pF 0.35 % + 35 pF 0.35 % + 190 pF 0.47 % + 2.0 nF 0.57 % + 19 nF 0.58 % + 190 nF 0.58 % + 2.0 μ F 1.2 % + 69 μ F 0.70 % +35 pF 0.70 % + 69 pF 0.70 % + 370 pF 0.93 % + 3.0 nF 1.2 % + 37 nF 1.2 % + 370 nF 1.2 % + 4.0 μ F 2.3 % +140 μ F	Suitable for the testing of capacitance measuring devices	A, B
BANDWIDTH	3 dB point with respect to set point 10 Hz to 50 kHz 50 kHz to 100 MHz 100 MHz to 250 MHz	0.60 % 2.4 % 5.2 %	Appropriate for calibration of oscilloscopes	A, B
TIMING MARKER	5 ns to 5 s per division	0.30 %	Appropriate for calibration of oscilloscopes	
FREQUENCY	10 MHz reference 100 mHz to 1 MHz 1 MHz to 20.1 GHz	1.0 in 10^9 1.5 in 10^9 + 1.0 μ Hz 1.5 in 10^9	Stable oscillators May be expressed as time (1/f) for repetitive measurements	A
Generation only	0.5 Hz to 250 MHz	0.29 $\mu\Omega/\Omega$	May be expressed as time (1/f) for repetitive measurements	A, B
Rise time	Nominal 1 ns 100 mV to 500 mV pulse rise and fall Nominal 100 ns 5 V to 50 V pulse rise and fall	90 ps 9 ns	In to 50 Ω In to 1 M Ω	
TIME Electronically Triggered Interval/Period Average	25 ns to 10 s	2.0 in 10^9 + (100 ns/No of periods)	Actual uncertainties quoted on certificate will include an allowance for the characteristics of the measured signal	A, B



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
FREQUENCY (cont'd)				
Time Interval (A-A Event)	100 ns to 10^9 s	2.0 in 10^9 + 100 ns		A, B
Time Interval (A-B Event)	100 ns to 10^9 s	2.0 in 10^9 + 100 ns		
Time Interval	0.1 ms to 10 s	1.8 %	Appropriate for the calibration of RCD testers	A, B
TIME Mechanically Triggered	over 1 second	50 ms		A, B
POWER				
At unity power factor ± 1	10 μ W to 10.5 kW @ 40 Hz to 1 kHz	0.20 %		
	4 kW to 21 kW @ 50 Hz	0.47 %		
AC Power Factor	0 to unity capacitive or Inductive	0.034		A, B
DC power	10 μ W to 10.5 kW	0.060 %		
TORQUE				
Hand torque tools (including drivers)	To BS EN ISO 6789:2003 (withdrawn) 0.2 N·m to 1500 N·m	1.5 %		A
	BS EN ISO 6789:2017 0.2 N·m to 0.5 N·m 0.5 N·m to 1.0 N·m 1.0 N·m to 10 N·m 10 N·m to 1500 N·m	3.1 % 1.6 % 0.60 % 0.50 %	By comparison with reference transducers See Notes 5 & 6	



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
PRESSURE <u>Hydraulic pressure (gauge)</u> Calibration of pressure indicating instruments and gauges, pressure relief valves and switches Calibration of Piezoelectric pressure transducers at quasi-static pressures Pressure equivalent calibration of Dead Weight Testers (pressure balance supplied with an associated mass set) <u>Gas pressure (gauge)</u> Calibration of pressure indicating instruments and gauges, pressure relief valves and switches Pressure equivalent calibration of Dead Weight Testers (pressure balance supplied with an associated mass set) Permanent lab only <u>Gas pressure (absolute)</u> Calibration of pressure indicating instruments and gauges	400 kPa to 6 MPa 6 MPa to 26 MPa 26 MPa to 260 MPa 260 MPa to 372 MPa 400 kPa to 260 MPa 260 MPa to 372 MPa 400 kPa to 6 MPa 6 MPa to 26 MPa 26 MPa to 260 MPa 260 MPa to 372 MPa -95 kPa to -3.5 kPa -3.5 kPa to -1.5 kPa -1.5 kPa to -240 Pa -240 Pa to -24 Pa -24 Pa to 24 Pa 24 Pa to 240 Pa 240 Pa to 1.5 kPa 1.5 kPa to 5 kPa 5 kPa to 50 kPa 50 kPa to 2.5 MPa 2.5 MPa to 8 MPa 8 MPa to 45 MPa 0.1 Pa to 1 Pa 1 Pa to 10 Pa 10 Pa to 60 Pa 60 Pa to 100 Pa 100 Pa to 1 kPa 1 kPa to 131 kPa	0.009% + 12 Pa 0.0080% 0.0090% 0.002% + 0.60 x10 ⁻⁶ /MPa 0.36 % 0.37 % 0.0090 % + 170 Pa 0.0080 % + 1.7 kPa 0.0090 % + 8.2 kPa 0.020 % + 16 kPa + 0.60 x10 ⁻⁶ /MPa 0.010 % 0.011 % 0.63 Pa 0.35 % + 0.16 Pa 0.59 % + 0.044 Pa 0.35 % + 0.16 Pa 0.63 Pa 0.0080 % 0.0070 % 0.0055 % 0.0080 % 16 kPa 0.055 Pa 0.080 Pa 0.31 Pa 0.46 Pa 5.0 Pa 6.0 Pa	Calibration of devices with an electrical output may be undertaken Calibration of devices with a charge output may be undertaken Absolute pressure calibrations can be undertaken using associated barometric pressure measurement correction. The uncertainties quoted will be increased by 6 Pa.. Pressure indicating instruments and gauges without a pressure port can be calibrated.	A,B A,B A A, B A A, B



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
TEMPERATURE				
Radiation thermometers (pyrometers)	-30 °C to -10 °C -10 °C to +15 °C 15 °C to 20 °C (ambient) 20 °C (ambient) to 100 °C 100 °C to 200 °C 200 °C to 300 °C 300 °C to 400 °C 400 °C to 500 °C	1.65 °C 1.25 °C 0.8 °C 0.6 °C 0.8 °C 1.0 °C 1.3 °C 1.6 °C	By comparison with a reference blackbody source or radiation thermometer.	A
Blackbody sources	-30 °C to -10 °C -10 °C to +15 °C 15 °C to 20 °C (ambient) 20 °C (ambient) to 100 °C 100 °C to 200 °C 200 °C to 300 °C 300 °C to 400 °C 400 °C to 500 °C	1.65 °C 1.25 °C 0.8 °C 0.6 °C 0.8 °C 1.0 °C 1.3 °C 1.6 °C	By comparison with a reference blackbody source or radiation thermometer.	A
Resistance Thermometers (4 Wire)	- 80 °C to - 50 °C -50 °C to 5 °C 0 °C (Ice Point) 0.01 °C (Triple Point of Water) 20 °C (Ambient) 5 °C to 80 °C 80 °C to 160 °C 160 °C to 260 °C 260 °C to 425 °C 425 °C to 660 °C	0.013 °C 0.011 °C 0.007 °C 0.003 °C 0.010 °C 0.013 °C 0.015 °C 0.016 °C 0.032 °C 0.055 °C	3 and 2 wire PRT can also be calibrated but will have an increased uncertainty. By comparison in a range of metal and liquid media baths	A
Temperature Indicators and/or recorders with temperature sensors	- 80 °C to- 50 °C - 50 °C to 5 °C 0 °C (Ice Point) 0.01 °C (Triple Point of Water) 20 °C (Ambient) 5 °C to 80 °C 80 °C to 160 °C 160 °C to 260 °C 260 °C to 425 °C 425 °C to 660 °C	0.012 °C 0.010 °C 0.006 °C 0.003 °C 0.010 °C 0.012 °C 0.014 °C 0.015 °C 0.031 °C 0.054 °C	By comparison in a range of metal and liquid media baths	A



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurment Uncertainty ($k=2$)	Remarks	Location Code
TEMPERATURE (cont'd)				
Block Calibrators and small liquid baths	- 80 °C to 0 °C 0 °C to 150 °C 150 °C to 250 °C 250 °C to 425 °C 425 °C to 660 °C 660 °C to 1100 °C 1100 °C to 1200 °C	0.0344 °C 0.041 °C 0.053 °C 0.076 °C 0.11 °C 1.0 °C 1.9 °C	By comparison to a reference PRT or thermocouple.	A
Noble Metal Thermocouples (Thermocouples with cold junctions)	0 °C (Ice Point) 20 °C (Ambient) 20 °C to 260 °C 260 °C to 660 °C 660 °C to 1100 °C 1100 °C to 1200 °C	0.12 °C 0.11 °C 0.11 °C 0.35 °C 1.5 °C 2.3 °C	Thermocouples without a cold junction will have an increased uncertainty. By comparison in a range of metal and liquid media baths	A
Base metal Thermocouples	0 °C (Ice Point) 20 °C (Ambient) -80 °C to 260 °C 260 °C to 425 °C 425 °C to 660 °C 660 °C to 1100 °C 1100 °C to 1200 °C	0.062 °C 0.069 °C 0.070 °C 0.15 °C 0.15 °C 1.5 °C 2.3 °C	By comparison in a range of metal and liquid media baths	A
Extension or compensating cable Base Metal	0 °C (Ice Point) 20 °C (Ambient) -25 °C to 200 °C	0.026 °C 0.040 °C 0.040 °C	By comparison in a range of metal and liquid media baths	A
Noble Metal	0 °C (Ice Point) 20 °C (Ambient) 20 °C to 200 °C	0.11 °C 0.10 °C 0.10 °C		
Thermistors	0°C (Ice Point) 0.01°C (Triple Point of Water) Ambient (nominally 20°C) -50°C to 5°C 5°C to 125°C	0.007°C 0.003°C 0.010°C 0.010°C 0.013°C		A
END				



0495
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Lambda Calibration Ltd
Issue No: 059 Issue date: 13 September 2024

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$