


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0513</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Charnwood Instrumentation Services Limited</b>	
	<b>Issue No: 029</b>	<b>Issue date: 24 December 2024</b>
	<b>81 Park Road Coalville Leicestershire LE67 3AF</b>	<b>Contact: Mr Tony Cox Tel: +44 (0)1530 510615 E-Mail: sales@charnwoodinstruments.co.uk Website: www.charnwoodinstruments.co.uk</b>
<b>Calibration performed at the above address only</b>		

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
TEMPERATURE			By comparison with reference Platinum Resistance Thermometers in a fluid bath unless otherwise stated
Liquid-in-glass thermometers (see note)	- 80 °C to 255 °C	0.020 °C*	Total immersion thermometers which can be immersed to at least 150 mm
	- 80 °C to 255 °C 265 °C to 410 °C	0.050 °C* 0.10 °C*	All other thermometers  *plus 1/10 scale division
Resistance thermometers (Pt 100)	- 80 °C to 255 °C 265 °C to 410 °C	0.020 °C 0.10 °C	
Thermometers with indicators	- 80 °C to 410 °C	as for resistance thermometers:	
Temperature calibration in air	5 °C to 50 °C	0.14 °C	Suitable for temperature probes built in to humidity instruments
HUMIDITY			By comparison with reference hygrometer and Platinum Resistance Thermometers
Humidity instruments	10 %rh to 95 %rh 5 °C to 20 °C	2.0 %rh	
	10 %rh to 95 %rh 20 °C to 25 °C	1.9 %rh	
	10 %rh to 90 %rh 25 °C to 50 °C	1.8 %rh	
DENSITY			
Hydrometers	500 kg/m <sup>3</sup> to 2000 kg/m <sup>3</sup>	0.050 kg/m <sup>3</sup>	By Cuckow's method
END			



0513  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Charnwood Instrumentation Services Limited**

**Issue No:** 029    **Issue date:** 24 December 2024

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$