


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	Air Surveys Ltd	
	Issue No: 019	Issue date: 01 February 2022
	2 College Mews St Ann's Hill Wandsworth London SW18 2SJ	Contact: Mr R Thomas Tel: +44 (0)208 871 4577 E-Mail: airsurveysltd@aol.com Website: www.airsurveysltd.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 2 College Mews St Ann's Hill Wandsworth London SW18 2SJ	Local contact Mr R Thomas Tel: +44 (0)208 871 4577 Fax: +44 (0)208 874 2044 Email: airsurveysltd@aol.com Website: airsurveysltd.com	Health and Hygiene Head Office Asbestos – All Support Functions <u>Support Functions:</u> Quality Management including document control, auditing and quality control	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



0553
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Air Surveys Ltd

Issue No: 019 **Issue date:** 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method APP 7 based on HSG 248	B
	Fibre counting	Documented In-House Method APP 7, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248)	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method APP 7, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method APP 15 based on HSG 248	
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method APP 8 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	B A
END			