

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0595

Accredited to
ISO/IEC 17025:2017

Common Services Agency - Public Services Delivery Scotland

Issue No: 023 Issue date: 15 June 2026

Gyle Square
1 South Gyle Crescent
Edinburgh
EH12 9EB

Contact: Mr N McLean
Tel: +44 (0) 131 275 6562
Email: Norman.McLean@nhs.scot
Website: <http://www.hfs.scot.nhs.uk/validation>

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Unit 5 Kirkhill Business Unit Park Court Broxburn West Lothian EH52 6EE</p> <p>Local contact Mr Norman McLean Tel: +44 (0)131-275 6562 Email: Norman.McLean@nhs.scot Website: www.hfs.scot.nhs.uk/validation</p>	<p>Electrical Pressure Temperature</p>	P
<p>Address The Jack Copland Centre 52 Research Ave North Herriot Watt Research Park Edinburgh EH14 4BE</p> <p>Local contact Mr Sam Nicholson Tel: +44 (0)131-314 5509 Mobile: +44(0)7964840319 Email: sam.nicholson@nhs.scot Website: www.scotblood.co.uk</p>	<p>Temperature</p>	SNBTS

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p>Local contact Mr Norman McLean Contact details as above</p>	<p>Electrical Pressure Temperature</p>
	<p>Local contact Mr Sam Nicholson Contact details as above</p>	<p>Temperature</p>



0595
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Common Services Agency – Public Services Delivery Scotland

Issue No: 023 Issue date: 15 June 2026

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL				
Calibration of temperature indicators by electrical simulation.				
Base metal thermocouples	0 °C to 200 °C	0.69 °C	with cold junction compensation	P and S
Time Interval	10 s to 1 hour	2.1 s		P and S
PRESSURE				
Calibration by comparison with a reference device				
Gas pressure (absolute)	3 kPa to 400 kPa	0.35 kPa	Direct calibration and measurement	P and S
	3 kPa to 400 kPa	2.2 kPa	Transfer calibration and measurement	
Gas pressure (gauge)	-90 kPa to 300 kPa	0.35 kPa	Direct calibration and measurement	P and S
	-90 kPa to 300 kPa	2.2 kPa	Transfer calibration and measurement	
TEMPERATURE				
Calibration by comparison with a reference device				
Thermometers with sensors, digital and analogue	0 °C to 200 °C	0.052 °C	Fluid Bath	P
	0 °C to 200 °C	0.090 °C	Dry block	P and S
Temperature controlled autoclaves, washer disinfectors, media preparators, incubators, sterilizers, ovens, environmental chambers, fridges/refrigerators and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	0 °C to 200 °C	0.24 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	P and S
Temperature controlled environmental chambers, fridges/refrigerators fitted with sensors	0 °C to 37 °C	0.31 °C	Calibration by comparison with reference dry block, monitoring probe only	SNBTS
END				



0595
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Common Services Agency – Public Services Delivery Scotland

Issue No: 023 Issue date: 15 June 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$