

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0601 Accredited to ISO/IEC 17025:2017	Electronic Temperature Instruments Limited Issue No: 035 Issue date: 10 September 2025	
	Unit 1 Teville Industrials Worthing West Sussex BN14 8NW	Contact: Mr G Hills Tel: +44 (0)1903 202151 Fax: +44 (0)1903 202445 E-Mail: sales@etiltd.co.uk Website: www.etiltd.co.uk
Calibration performed at the above address only		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
TEMPERATURE			Including calculated units. Calibration by comparison with reference Platinum Resistance Thermometers in a stirred liquid bath above -80 °C and in a dry block below -80 °C For probe length of greater than 120 mm, shorter probes will have an increased uncertainty
Platinum Resistance Thermometers	-100 °C to -80 °C	0.14 °C to 0.18 °C 0.16 °C to 0.19 °C 0.20 °C to 0.23 °C	4 Wire Configuration 3 Wire Configuration 2 Wire Configuration
	-80 °C to +250 °C	0.07 °C 0.10 °C 0.19 °C	4 Wire Configuration 3 Wire Configuration 2 Wire Configuration
Thermistor Thermometers	-40 °C to +150 °C	0.06 °C	At fixed temperatures and may differ using manufactures tables
Resistance Sensors with Indicators	-100 °C to -80 °C -80 °C to -38 °C -38 °C to +156 °C +156 °C to +250 °C	0.16 °C 0.060 °C 0.027 °C 0.060 °C	
Thermocouple Sensors with Indicators	-100 °C to -80 °C -80 °C to +250 °C	0.19 °C 0.15 °C	
Temperature data loggers with integrated sensors	-80 °C to +231 °C	0.06 °C	Comparison by stirred bath liquid immersion. Following manufacturers data sheets for maximum operating temperature.



0601
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Electronic Temperature Instruments Limited
Issue No: 035 **Issue date:** 10 September 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL			Calibration is performed by the comparison with a reference instrument unless otherwise stated.
DC Resistance			
Measurement	0 Ω to 200 Ω 200 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω	18 m Ω 36 m Ω 0.4 Ω 4.0 Ω 65 Ω	Including test caps and resistance simulators, used to calibrate resistance thermometers (thermistor and Pt100)
Sourcing	0 Ω to 200 Ω 200 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω	20 m Ω 40 m Ω 0.4 Ω 4.0 Ω 70 Ω	
DC Voltage			
Measurement	0 mV to 100 mV	5.0 μ V	
Temperature indicators and simulators, calibration by electrical simulation			
Base metal thermocouple			Including cold junction compensation
Type K	-200 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -20 $^{\circ}$ C -20 $^{\circ}$ C to +200 $^{\circ}$ C +200 $^{\circ}$ C to +500 $^{\circ}$ C +500 $^{\circ}$ C to +800 $^{\circ}$ C +800 $^{\circ}$ C to +1300 $^{\circ}$ C	0.42 $^{\circ}$ C 0.22 $^{\circ}$ C 0.18 $^{\circ}$ C 0.17 $^{\circ}$ C 0.17 $^{\circ}$ C 0.20 $^{\circ}$ C	
Type J	-200 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -20 $^{\circ}$ C -20 $^{\circ}$ C to +200 $^{\circ}$ C +200 $^{\circ}$ C to +500 $^{\circ}$ C +500 $^{\circ}$ C to +800 $^{\circ}$ C +800 $^{\circ}$ C to +1300 $^{\circ}$ C	0.40 $^{\circ}$ C 0.22 $^{\circ}$ C 0.19 $^{\circ}$ C 0.17 $^{\circ}$ C 0.17 $^{\circ}$ C 0.17 $^{\circ}$ C	
Type T	-200 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -20 $^{\circ}$ C -20 $^{\circ}$ C to +200 $^{\circ}$ C +200 $^{\circ}$ C to +400 $^{\circ}$ C	0.41 $^{\circ}$ C 0.23 $^{\circ}$ C 0.19 $^{\circ}$ C 0.16 $^{\circ}$ C	
Type E	-200 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to -20 $^{\circ}$ C -20 $^{\circ}$ C to +200 $^{\circ}$ C +200 $^{\circ}$ C to +500 $^{\circ}$ C +500 $^{\circ}$ C to +800 $^{\circ}$ C +800 $^{\circ}$ C to +1300 $^{\circ}$ C	0.35 $^{\circ}$ C 0.20 $^{\circ}$ C 0.19 $^{\circ}$ C 0.15 $^{\circ}$ C 0.14 $^{\circ}$ C 0.15 $^{\circ}$ C	



0601
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Electronic Temperature Instruments Limited
Issue No: 035 **Issue date:** 10 September 2025

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (<i>k</i> = 2)	Remarks
Temperature indicators and simulators, calibration by electrical simulation (cont'd)			
Noble metal thermocouple	+500 °C to +1768 °C	0.70 °C	Including cold junction compensation
Resistance thermometer (Pt100)	-200 °C to +250 °C +250 °C to +850 °C	0.05 °C 0.10 °C	
Resistance thermometer (Pt1000)	-200 °C to 0 °C 0 °C to +850 °C	0.015 °C 0.182 °C	
HUMIDITY			
Relative Humidity	<i>Performed at 0 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 10 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 20 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 25 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 30 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 40 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 50 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh <i>Performed at 60 °C</i> 10 %rh to 50 %rh 50 %rh to 90 %rh	0.36 %rh to 0.99 %rh 0.99 %rh to 1.74 %rh 0.20 %rh to 0.72 %rh 0.75 %rh to 1.25 %rh 0.20 %rh to 0.64 %rh 0.64 %rh to 1.10 %rh 0.20 %rh to 0.62 %rh 0.62 %rh to 1.06 %rh 0.20 %rh to 0.62 %rh 0.62 %rh to 1.07 %rh 0.20 %rh to 0.63 %rh 0.62 %rh to 1.09 %rh 0.20 %rh to 0.64 %rh 0.64 %rh to 1.11 %rh 0.20 %rh to 0.63 %rh 0.63 %rh to 1.09 %rh	Calibration by comparison with reference dew point hygrometer and reference Platinum Resistance Thermometers in an air chamber
Temperature in Air	0 °C to 20 °C 20 °C to 25 °C 25 °C to 60 °C	0.26 °C 0.18 °C 0.26 °C	
END			



0601
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Electronic Temperature Instruments Limited
Issue No: 035 **Issue date:** 10 September 2025

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$