


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0615 Accredited to ISO/IEC 17025:2017	Consolidated Medical Industries Limited Trading as TME Calibration and Testing Issue No: 041 Issue date: 01 May 2026	
	CMI House 9 Milbanke Court Milbanke Way Bracknell RG12 1RP	Contact: Mr A Hodgson Tel: +44 (0)1344 621378 E-Mail: sales@cmihealthcare.co.uk Website: www.cmihealthcare.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address CMI House 9 Milbanke Court Milbanke Way Bracknell RG12 1RP Local contact Mr A Hodgson	Pressure Temperature Time interval	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Pressure Temperature Time interval	Site



0615
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Consolidated Medical Industries Limited
Trading as TME Calibration and Testing
Issue No: 041 Issue date: 01 May 2026

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Temperature sensors with indicators				Lab
Resistance sensors	-40 °C to +200 °C -40 °C to +150 °C	0.030 °C 0.046 °C	By comparison in a liquid bath By comparison in a dry block	
Thermocouple sensors	-40 °C to +50 °C 50 °C to 200 °C	0.10 °C 0.17 °C	By comparison in a liquid bath	
Temperature block calibrators	-40 °C to +150 °C 150 °C to 200 °C	0.046 °C 0.14 °C	Using PRT sensors	
Temperature sensors with indicators				Site
Resistance sensors	-40 °C to +140 °C	0.11 °C	By comparison in a dry block	
Thermocouple sensors	-40 °C to +140 °C	0.18 °C	By comparison in a dry block	
Temperature controlled autoclaves, media preparators, incubators, ovens, environmental chambers, fridges/refrigerators and freezers (inclusive of associated indicators, controllers and recorders)	-40 °C to +200 °C	0.50 °C	Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping Calibrations can be carried out as part of the performance and safety tests for sterilizers as prescribed in the following standards: BS 2646:1993:Part 5:Section 3 BS 3970:1990 EN 285:2015	
TIME INTERVAL				
Stopwatch / Timers Relating to the timer functions of chamber and autoclaves	10 s to 12 hrs	2.0 s	Calibration by comparison	Lab & site



0615
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Consolidated Medical Industries Limited
Trading as TME Calibration and Testing
Issue No: 041 Issue date: 01 May 2026

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
PRESSURE			Methods consistent with EURAMET CG17.	
<u>Gas pressure gauge</u>				
Calibration of Pressure indicating instruments and gauges	-96 kPa to 30 kPa	23 Pa	By comparison with a digital pressure standard.	Lab
	30 kPa to 700 kPa	0.14 kPa	By comparison with a deadweight tester.	
	-96 kPa to +100 kPa	0.34 kPa	By comparison with a digital pressure standard.	Site
	100 kPa to 400 kPa	0.40 kPa		
	400 kPa to 700 kPa	0.64 kPa		
<u>Gas pressure absolute</u>				
Calibration of Pressure indicating instruments and gauges	3.5 kPa to 130 kPa	20 Pa	By comparison with a digital pressure standard.	Lab
	130 kPa to 800 kPa	0.14 kPa	By comparison with a deadweight tester.	
	35 kPa to 200 kPa	0.34 kPa	By comparison with a digital pressure standard.	Site
	200 kPa to 500 kPa	0.40 kPa		
	35 kPa to 800 kPa	0.64 kPa		
END				



0615
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Consolidated Medical Industries Limited
Trading as TME Calibration and Testing
Issue No: 041 Issue date: 01 May 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$