


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0630</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>The Tintometer Limited</h3> <p>Issue No: 019 Issue date: 05 February 2026</p>	
	<p>Lovibond House Sun Rise Way Amesbury SP4 7GR</p>	<p>Contact: Dr P Clarke Tel: +44 (0)1980 664800 Fax: +44 (0)1980 625412 E-Mail: support@lovibond.uk Website: www.lovibond.com</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
SPECTRAL TRANSMITTANCE	<i>Wavelength 380 nm to 780 nm</i> T ≤ 20% T > 20%	(0.013T + 0.044) % (0.0032T + 0.24) %	All calibrations are against reference transmittance standards
Luminous transmittance		0.0030	Measured at intervals 1, 2, 5 or 10 nm ASTM D1500
ISO 2049 A.2.2 Daylight Filter	<i>Wavelength 380 nm to 780 nm</i> T ≤ 20%: T > 20%	(0.013T + 0.044) % (0.0032T + 0.24) %	Compliance derivation in accordance with A.2.2
ABSORBANCE	<i>Wavelength 380 nm to 780 nm</i>	T % uncertainty transformed to absorbance	T % to absorbance = - log ₁₀ (T % / 100)
Glass colour reference materials			
CIE tristimulus values derived from spectral transmittance data	<i>Wavelength 380 nm to 780 nm</i> X Y Z x y z L* a* b*	0.30 0.30 0.30 0.00060 0.00060 0.00080 0.20 0.20 0.20	Mathematical derivation in accordance with CIE Document 15.2 and ASTM E308-08
UCS Judd rgb derived from CIE tristimulus values	<i>Wavelength 380 nm to 780 nm</i> r g b	0.0010 0.0010 0.0010	
ASTM colour Saybolt	0.5 to 8 + 30 to - 16	0.10 1.0	Mathematical derivation in accordance with ASTM D6045
Gardner colour	1 to 18	0.10	Mathematical derivation in accordance with ASTM D6166



0630
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Tintometer Limited
Issue No: 019 Issue date: 05 February 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
Liquid colour reference materials			
CIE tristimulus values	<i>Wavelength 380 nm to 780 nm</i>		
	Y	0.35	Mathematical derivation in accordance with CIE Document 15.2 and ASTM E308-08
	x	0.00070	
	y	0.00060	
ASTM colour Saybolt	0.5 to 8 + 30 to -16	0.10 1.0	Mathematical derivation in accordance with ASTM D6045
Gardner colour	1 to 18	0.10	Mathematical derivation in accordance with ASTM D6166
Platinum-Cobalt Colour	0 to 500	1.0	Mathematical derivation based on CIE tristimulus values
END			



0630
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

The Tintometer Limited
Issue No: 019 **Issue date:** 05 February 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$