# **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



## Locations covered by the organisation and their relevant activities

## Laboratory locations:

Location details		Activity	Location code
Address 11 Frensham Road Norwich NR3 2BT	<b>Local contact:</b> Mr P K Clark Tel: +44 (0)1603 279557	Electrical Calibration Accelerometer calibration Mass Calibration NAWI Calibration	A
Address 5 Octavian Way, Team Valley Trading Est Gateshead NE11 0HZ	<b>Local contact:</b> Mr John Fryer Tel: +44 (0)191 4875951	Electrical Calibration Pressure Calibration	В

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
O654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Org	anisation at the locations	specified	
	Calibration and Measu	rement Capability (CM	1C)	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration on against laboratory references unl	of both measuring instrumeess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The
ELECTRICAL MEASUREMENTS DC VOLTAGE			Calibrations are performed as a direct comparison against a reference standard	A
Generation	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1100 V 1 kV to 2 kV 2 kV to 3 kV 3 kV to 4 kV 4 kV to 5 kV	$\begin{array}{c} 0.55 \ \mu V \\ 1.4 \ \mu V \\ 6.6 \ \mu V \\ 44 \ \mu V \\ 630 \ \mu V \\ 8.0 \ m V \\ 0.45 \ \% + 7.0 \ V \\ 0.45 \ \% + 8.0 \ V \\ 0.45 \ \% + 9.0 \ V \\ 0.45 \ \% + 11 \ V \end{array}$	For the calibration of measuring instruments	
Measurement	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV 1 kV to 2 kV 2 kV to 3 kV 3 kV to 4 kV 4 kV to 5 kV	1.0 $\mu$ V 2.0 $\mu$ V 10 $\mu$ V 98 $\mu$ V 1.3 mV 14 mV 0.67 % + 9.0 V 0.67 % + 11 V 0.67 % + 14 V 0.67 % + 16 V	For instruments with a voltage output	
DC RESISTANCE				A
Sourcing - Spot Values	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω	110 μΩ 210 μΩ 270 μΩ 540 μΩ 1.2 mΩ	For the calibration of measuring instruments	
	190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ	2.5 mΩ 8.7 mΩ 17 mΩ 81 mΩ 170 mΩ		
	100 kΩ 190 kΩ 1 MΩ 1.9 MΩ	1.1 Ω 2.4 Ω 16 Ω 48 Ω		
	10 ΜΩ 19 ΜΩ 100 ΜΩ	460 Ω 1.1 kΩ 12 kΩ		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue dat	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Org	anisation at the locations	s specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references un	of both measuring instrum less otherwise stated in the	ents and for instruments with an e remarks column.	n output. The
DC RESISTANCE (continued)				А
Sourcing - Range values	0Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ	2.0 mΩ 5.0 mΩ 31 mΩ	For the calibration of measuring instruments	
	1 kΩ to 10 kΩ 10 kΩ to 100 kΩ	310 mΩ 3.0 Ω		
	100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 33 MΩ	37 Ω 2.0 kΩ 11 kΩ		
	33 MΩ to 110 MΩ 110 MΩ to 330 MΩ 330 MΩ to 1.1 GΩ	60 kΩ 1.0 MΩ 17 MΩ		
Measurement	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ	250 μΩ 2.0 mΩ 14 mΩ 140 mΩ 1.0 Ω	For instruments with a resistance output	
	100 kΩ to 1 MΩ 1 MΩ to 10 GΩ 10 GΩ to 100 GΩ	22 Ω 0.71 % 1.0 %		
DC CURRENT				А
Generation	0 μA to 10 μA 10 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA	7.4 nA 12 nA 49 nA 450 nA 6.0 μA	For the calibration of measuring instruments	
	100 mA to 1 A 1 A to 3 A 3 A to 5 A 5 A to 10 A 10 A to 20 A 20 A to 30 A	110 μA 1.0 mA 4.0 mA 6.0 mA 12 mA 570 mA		
	16 A to 160 A 160 A to 1000 A	25 mA 580 mA	Current clamp calibration using a multi turn coil	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025		
	Calibration performed by the Org	anisation at the locations	s specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrum ess otherwise stated in the	ents and for instruments with a eremarks column.	n output. The
DC CURRENT (continued)				А
Measurement	0 μA to 10 μA 10 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 2 A 2 A to 10 A 10 A to 20 A 20 A to 30 A	2.0 nA 4.0 nA 34 nA 340 nA 5.0 μA 150 μA 2.4 mA 25 mA 330 mA 490 mA	For instruments which generate current	
AC VOLTAGE				A
Generation	1 Hz to 10 Hz 10 μV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 7 V	60 μV 270 μV 1.9 mV 9.2 mV	For the calibration of measuring instruments	
	<b>10</b> $\mu$ <b>V to 2.2 mV</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	5.3 μV 4.9 μV 4.9 μV 4.9 μV 4.9 μV 4.9 μV 5.2 μV 7.1 μV 14 μV 27 μV 30 μV		
	2.2 mV to 22 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 500 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	11 μV 7.1 μV 6.9 μV 6.9 μV 6.9 μV 10 μV 19 μV 38 μV 59 μV 92 μV		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Orga	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration on against laboratory references unl	of both measuring instrumess otherwise stated in the	ents and for instruments with an remarks column.	output. The
AC VOLTAGE (continued)				А
Generation (continued)	22 mV to 220 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	76 μV 31 μV 23 μV 23 μV 23 μV 23 μV 23 μV 39 μV 99 μV 190 μV 380 μV 740 μV	For measuring instruments	
	<b>220 mV to 2.2 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	660 μV 250 μV 120 μV 120 μV 120 μV 120 μV 120 μV 180 μV 250 μV 950 μV 2.8 mV 4.7 mV		
	2.2 V to 22 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	6.6 mV 2.5 mV 1.1 mV 1.1 mV 1.1 mV 1.1 mV 1.8 mV 2.3 mV 7.2 mV 28 mV 42 mV		
	<b>22 V to 220 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40Hz to 500 Hz 500 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	66 mV 25 mV 14 mV 14 mV 14 mV 14 mV 22 mV 41 mV		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
Accredited to				
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrumess otherwise stated in the	ents and for instruments with ar remarks column.	output. The
AC VOLTAGE (continued)				А
Generation (continued)	<b>220 V to 1100 V</b> 55 Hz to 1 kHz	110 mV	For measuring instruments	
	1 kV to 2 kV 50 Hz to 60 Hz 2 kV to 3 kV 50 Hz to 60 Hz 3 kV to 4 kV 50 Hz to 60 Hz 4 kV to 5 kV 50 Hz to 60 Hz	0.67 % + 12 V 0.67 % + 16 V 0.67 % + 20 V 0.67 % + 24 V		
Measurement	<b>30 μV to 10 mV</b> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 1 MHz	8.0 μV 5.0 μV 6.0 μV 13 μV 60 μV 160 μV	For instruments with a voltage output	
	<b>10 mV to 100 mV</b> 1 Hz to 40 Hz 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	15 μV 13 μV 20 μV 38 μV 100 μV 360 μV 2.3 mV		
	<b>100mV to 1V</b> 1 Hz to 40 Hz 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	140 μV 120 μV 200 μV 390 μV 970 μV 3.6 mV 24 mV		
	1 V to 10 V 1 Hz to 40 Hz 40 Hz to 1 kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	2.0 mV 1.0 mV 2.0 mV 4.0 mV 10 mV 36 mV 240 mV		
	<b>10V to 100V</b> 1 Hz to 40 Hz 40 Hz to 1kHz 1kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100kHz to 300 kHz	38 mV 36 mV 37 mV 51 mV 150 mV 480 mV		

	Scl United Ki 2 Pine Trees, Chert	hedule of Accu issued by ngdom Accreo sey Lane, Staines-u	reditation ditation Service	IR, UK
O654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
	Calibration performed by the Org	anisation at the locations	s specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	elow are applicable for the calibration son against laboratory references un	of both measuring instrum less otherwise stated in the	ents and for instruments with a e remarks column.	an output. The
VOLTAGE (continued) Measurement (continued)	<b>100V to 1 kV</b> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	520 mV 490 mV 720 mV 1.4 V 3.5 V	For instruments with a voltage output.	A
	1 kV to 2 kV 30 Hz to 200 Hz 2 kV to 3 kV 30 Hz to 200 Hz 3 kV to 4 kV 30 Hz to 200 Hz 4 kV to 5 kV 30 Hz to 200 Hz	0.45 % + 8.0 V 0.45 % + 11 V 0.45 % + 13 V 0.45 % + 16 V		
AC CURRENT Generation	<b>10 μA to 220 μA</b> 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	38 nA 38 nA 88 nA 360 nA	For the calibration of measuring instruments	A
	<b>220 µA to 2.2 mA</b> 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	320 nA 320 nA 660 nA 3.6 μA		
	2.2 mA to 22 mA 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	29 μΑ 29 μΑ 55 μΑ 290 μΑ		
	<b>22 mA to 220 mA</b> 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	3.0 μΑ 3.0 μΑ 5.7 μΑ 34 μΑ		
	<b>220 mA to 2.2 A</b> 55 Hz to 400 Hz 400 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	660 μΑ 660 μΑ 1.3 mA 18 mA		
	<b>10 Hz to 1 kHz</b> 2 A to 3 A 3 A to 5 A 5 A to 10 A	7.0 mA 9.0 mA 14 mA		
	10 A to 20 A 50 Hz to 400 Hz 400 Hz to 1 kHz 20 A to 30 A	310 mA 290 mA		
	50 Hz to 400 Hz 400 Hz to 1 kHz	430 mA 420 mA		

Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
v are applicable for the calibration of against laboratory references unle	of both measuring instrume ess otherwise stated in the	ents and for instruments with an remarks column.	output. The
<b>30 A to 40 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	560 mA 540 mA		A
<b>10 A to 50 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	680 mA 670 mA		
1 <i>0 Hz to 100 Hz</i> 10 A to 32 A 16 A to 160 A 160 A to 1000 A	9.0 mA 35 mA 580 mA	Calibration of current clamps using a multi turn coil	
<b>20 nA to 10 μA</b> 20 Hz to 45 Hz 45 Hz to 1 kHz	23 nA 14 nA	For instruments which generate current	
<b>10 μA to 100 μA</b> 20 Hz to 45 Hz 45 Hz to 1 kHz	210 nA 110 nA		
100 μA to 1 mA 20 Hz to 45 Hz 45 Hz to 20 kHz 20 kHz to 50 kHz	2.0 μΑ 1.0 μΑ 5.0 μΑ		
<b>I mA to 10 mA</b> 20 Hz to 45 Hz 45 Hz to 20 kHz 20 kHz to 50 kHz	20 μΑ 9.0 μΑ 51 μΑ		
<b>10 mA to 100 mA</b> 20 Hz to 45 Hz 45 Hz to 20 kHz 20 kHz to 50 kHz	200 μΑ 95 μΑ 510 μΑ		
1 <b>00 mA to 1 A</b> 20 Hz to 45 Hz 45 Hz to 20 kHz 20 kHz to 50 kHz	2.0 mA 1.0 mA 12 mA		
<b>I A to 10 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	180 mA 140 mA		
<b>10 A to 20 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	270 mA 250 mA		
2 <b>0 A to 30 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	370 mA 360 mA		
	United Kir 2 Pine Trees, Cherts Calibrat Issue N Ibration performed by the Orga Range are applicable for the calibration of against laboratory references unler 0 A to 400 A 0 Hz to 400 Hz 00 Hz to 1 kHz 0 Hz to 400 Hz 00 Hz to 1 kHz 0 Hz to 400 Hz 00 Hz to 100 HZ 0 A to 50 A 0 D A to 100 µA 0 Hz to 45 Hz 5 Hz to 1 kHz 00 µA to 100 µA 0 Hz to 45 Hz 5 Hz to 1 kHz 00 µA to 100 mA 0 Hz to 45 Hz 5 Hz to 20 kHz 0 kHz to 50 kHz to 50 kHz 0 kHz to 50 kHz to 5	United Kingdom Accrec 2 Pine Trees, Chertsey Lane, Staines-upCalibration, Maintenance Issue No: 044 Issue dateIssue No: 044 Issue dateIlibration performed by the Organisation at the locationsRangeExpanded Measurement Uncertainty $(k = 2)$ are applicable for the calibration of both measuring instrume against laboratory references unless otherwise stated in the0 A to 40 A 0 Hz to 400 Hz560 mA 560 mA0 A to 50 A 0 Hz to 400 Hz560 mA 670 mA0 A to 50 A 0 Hz to 1 kHz9.0 mA 570 mA0 A to 50 A 0 Hz to 400 Hz9.0 mA 580 mA0 A to 50 A 0 Hz to 1000 HZ 0 Hz to 1000 A9.0 mA 580 mA0 A to 10 HZ 0 Hz to 45 Hz23 nA 5 Hz to 1 kHz0 D A to 100 µA 0 Hz to 45 Hz 5 Hz to 1 kHz20 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 1 kHz20 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz20 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz20 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz51 µA0 mA to 10 mA 0 Hz to 45 Hz 5 Hz to 20 kHz90 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz10 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz10 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz10 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz10 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz10 µA 5 0 µA0 Hz to 45 Hz 5 Hz to 20 kHz10 µA 5 0 µA0 Hz to 400 Hz 5 Hz to 20 kHz10 µA 5 0 µA0	United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HI         Calibration, Maintenance and Repair Ltd Issue No: 044         Issue date: 07 May 2025         Interview of the organisation at the locations specified         Remarks         Remarks         Remarks         Calibration of both measuring instruments and for instruments with an against laboratory references unless otherwise stated in the remarks column.         D A to 40 A 0 Hz to 400 Hz         680 mA 540 mA         D A to 50 A 0 Hz to 400 Hz         OHz to 100 HZ         D A to 50 A 0 Hz to 400 Hz         OHz to 100 HZ         D A to 50 A 0 Hz to 400 Hz         OHz to 100 HZ         D A to 50 A 0 Hz to 400 Hz       Calibration of current clamps using a multi tum coil         D A to 50 A 0 Hz to 100 HZ       Calibration of current clamps using a multi tum coil         D A to 50 A 0 Hz to 45 Hz       Calibration of current clamps using a multi tum coil         D A to 100 HZ       Calibration of current clamps using a multi tum coil         D A to 50 Hz       Calibration of current clamps using a multi tum coil     <

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017				
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	output. The
AC CURRENT (continued) Measurement (continued)	<b>30 A to 40 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	480 mA 470 mA		А
	<b>40 A to 50 A</b> 50 Hz to 400 Hz 400 Hz to 1 kHz	590 mA 580 mA		
FREQUENCY Generation	0.5 Hz to 5.4 GHz	2.1 in 10 <sup>9</sup>	For the calibration of measuring instruments May be reported as events per unit time	A
	1 kHz to 27 GHz	0.24 µHz/Hz	For instruments outputting frequency	
Measurement	0.5 Hz to 5.4 GHz 5.4 GHz to 20 GHz	2.1 in 10 <sup>9</sup> 4.0 MHz	For the calibration of instruments with frequency outputs. May be reported as events per unit time	
TIME and FREQUENCY				А
Elapsed time			Mechanical timers / stop watches	
Single event	10 s to 12 hrs	0.23 s		
Revolutions Per Minute	60 RPM to 1000 RPM 1000 RPM to 30000 RPM 30000 RPM to 96000 RPM	0.37 RPM 1.6 RPM 23 RPM		

	Sch United Ki	nedule of Accr issued by ngdom Accreo	reditation ditation Service	r uk
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrumers otherwise stated in the	ents and for instruments with ar remarks column.	n output. The
CAPACITANCE				А
Measurement For capacitance sources	<b>1 kHz</b> 0.1 pF to 1 pF 1 pF to 10 pF 10 pF to 100 pF 100 pF to 1000 pF	0.0052 pF 0.014 pF 0.11 pF 1.1 pF	For capacitance sources	
	<b>1 MHz</b> 0.1 pF to 1 pF 1 pF to 10 pF 10 pF to 100 pF 100 pF to 1000 pF	0.0031 pF 0.014 pF 0.11 pF 1.3 pF		
Simulated For capacitance meters	<b>10 Hz to 10 kHz</b> 220 pF to 400 pF 400 pF to1.1 nF	13 pF 16 pF		
	<b>10 Hz to 3 kHz</b> 1.1 nF to 3.3 nF	27 pF		
	<i>10 Hz to 1 kHz</i> 3.3 nF to 11 nF 11 nF to 33 nF	39 pF 180 pF		
	<b>10 Hz to 1 kHz</b> 33 nF to 110 nF 110 nF to 330 nF	380 pF 1.0 nF		
	<b>10 Hz to 600 Hz</b> 330 nF to 1.1 μF	5.0 nF		
	<b>10 Hz to 300 Hz</b> 1.1 μF to 3.3 μF	12 nF		
	<b>10 Hz to 150 Hz</b> 3.3 μF to 11 μF	39 nF		
	<b>10 Hz to 120 Hz</b> 11 μF to 33 μF	140 nF		
	<b>10 Hz to 80 Hz</b> 33 μF to 110 μF	610 nF		

	Sch United Ki 2 Pine Trees, Cherts	nedule of Accr issued by ngdom Accred sey Lane, Staines-u	editation ditation Service	R, UK
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrum ess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The
CAPACITANCE (continued)				А
For capacitance meters	<b>0 Hz to 50 Hz</b> 110 μF to 330 μF	2.0 µF		
	<b>0 Hz to 20 Hz</b> 330 μF to 1.1 mF	6.0 µF		
	<i>0 Hz to 6 Hz</i> 1.1 mF to 3.3 mF	2.0 μF		
	<i>0 Hz to 2 Hz</i> 3.3 mF to 11 mF	60 µF		
	<i>0 Hz to 0.6 Hz</i> 11 mF to 33 mF	280 µF		
	<i>0 Hz to 0.2 Hz</i> 33 mF to 110 mF	1.0 mF		
CALIBRATION OF INSULATION TESTERS				A
Insulation Resistance	10 kΩ to 20 kΩ 20 kΩ to 40 kΩ 40 kΩ to 100 kΩ 100 kΩ to 200 kΩ 200 kΩ to 1 MΩ 1 MΩ to 2 MΩ 2 MΩ to 10 MΩ 10 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 1 GΩ 1 GΩ to 2 GΩ 2 GΩ to 10 GΩ	46 Ω 93 Ω 230 Ω 460 Ω 2.3 kΩ 7.0 kΩ 35 kΩ 120 kΩ 1.2 MΩ 5.8 MΩ 23 MΩ 120 MΩ	Up to 1100 V	
Insulation test voltage Nominal	50 to 400 V 400 to 1 kV	6.1 V 18 V	The test voltage will normally be measured with a 1 mA load.	
Continuity Resistance	$\begin{array}{l} 0.2 \ \Omega \ to \ 30 \ \Omega \\ 30 \ \Omega \ to \ 100 \ \Omega \\ 100 \ \Omega \ to \ 200 \ \Omega \\ 200 \ \Omega \ to \ 500 \ \Omega \\ 500 \ \Omega \ to \ 1 \ k\Omega \\ 1 \ k\Omega \ to \ 2 \ k\Omega \\ 2 \ k\Omega \ to \ 5 \ k\Omega \\ 5 \ k\Omega \ to \ 10 \ k\Omega \end{array}$	80 mΩ 310 mΩ 0.67 Ω 1.70 Ω 2.60 Ω 4.80 Ω 17 Ω 26 Ω		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
	Calibration performed by the Org	anisation at the locations	specified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty $(k = 2)$	Remarks	Location Code	
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration on against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The	
Temperature indicators, calibration by electrical simulation				A	
Type K thermocouple	-200 °C to +1372 °C	0.73 °C	Including reference junction compensation		
Type T thermocouple	-250 °C to +400 °C	0.65 °C	Including reference junction compensation		
Type N thermocouple	-200 °C to +1300 °C	0.46 °C	Including reference junction compensation		
Type S thermocouple	0 °C to 1767 °C	0.77 °C	Including reference junction compensation		
Type J thermocouple	-210 °C to +1200 °C	0.23 °C	Including reference junction compensation		
Type E thermocouple	-250 °C to +1000 °C	0.26 °C	Including reference junction compensation		
Resistance thermometer (Pt 100)	-200°C to +800°C	0.05°C			
Calibration of Oscilloscopes				А	
Voltage deflection	1 mV to 25 mV 25 mV to 110 mV 110 mv to 2.2 V 2.2 V to 6.6 V	120 μV 370 μV 6.5 mV 19 mV	Into 50 Ω		
	1 mV to 25 mV 25 mV to 110 mV 110 mv to 2.2 V 2.2 V to 11 V 11 V to 130 V	63 μV 120 μV 1.3 mV 6.5 mV 76 mV	Into 1 MΩ		
Edge verification – Amplitude	4.5 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 2.5 V	2.5 mV 4.4 mV 28 mV 82 mV	Into 50 Ω 1 kHz to 10 kHz		
Time markers	20 ns to 50 ns 50 ns to 500 ns 500 ns to 5 $\mu$ s 5 $\mu$ s to 50 $\mu$ s 50 $\mu$ s to 500 $\mu$ s 500 $\mu$ s to 20 ms 20 ms to 500 ms 500 ms to 5 s	1.5 ps 15 ps 150 ps 1.50 ns 15.0 ns 580 ns 15 μs 380 μs			
Rise time	1 kHz to 2 MHz 2 MHz to 10 MHz	300 ps 350 ps			

Accredited to ISO/IEC 17025:2017 Measured Quantity Instrument or Gauge	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK         Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025         Calibration performed by the Organisation at the locations specified         Expanded Measurement Uncertainty (k = 2)       Content         Remarks       Location Code			
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrumers of both measuring instrumers of the stated in the	ents and for instruments with ar remarks column.	n output. The
OSCILLOSCOPES (Continued) Bandwidth Level flatness 50 kHz ref Nominal 3 V	To 100 MHz To 300 MHz To 600 MHz To 1.1 GHz	130 mV 160 mV 250 mV 300 mV	CMC is with respect to relative amplitude of level flatness this will be quoted in frequency terms when stating bandwidth uncertainty as it will vary from device to device	A
ACCELEROMETRY Accelerometer calibration Charge	Sensitivity range: 0.3 pC/g to 1000 pC/g (0.03 pC/ms <sup>2</sup> to 100 pC/ms <sup>2</sup> ) 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz Sensitivity range: 0.04 pC/g to 0.3 pC/g (0.004 pC/ms <sup>2</sup> to 0.03 pC/ms <sup>2</sup> )	3.0 % 1.2 % 0.85 % 0.90 %	Uncertainties for accelerometry sensitivity calibrations at ambient temperature applies for masses up to 40 grams. Nominal acceleration ms <sup>-2</sup> : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.3 to 100.0	
Devices with integral electronics Voltage (mV/ms <sup>2</sup> )	0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup> 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	2.0 % 0.83 % 0.85 % 0.90 %	Nominal acceleration 0.3 ms <sup>-2</sup> to 100.0 ms <sup>-2</sup> : Nominal acceleration ms <sup>-2</sup> : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.3 to 100.0	
Resistive/Capacitive	0.004 mV/ms <sup>2</sup> to 0.03 mV/ms <sup>2</sup> 20 Hz to 10 kHz 0.01 mV/ms <sup>2</sup> to 100 mV/ms <sup>2</sup> 1 Hz to 2 Hz 2 Hz to 5 Hz 5 Hz to 5 kHz 5 kHz to 10 kHz	3.0 % 2.0 % 0.83 % 0.85 % 0.90 %	Nominal acceleration ms <sup>-2</sup> : 1 Hz: 0.10 3 Hz: 0.20 4 Hz to 10 kHz: 0.3 to 100.0	
Vibration Acceleration Velocity Displacement Velocity transducers	10 Hz to 5 kHz 0.1g to 10 g 3 mm/s to 2000 mm/s 0.2μm to 100 mm 5 Hz to 10 kHz 1.5 mm/s to 3000 mm/s	0.40 % 0.40 % 0.40 % 0.90 %	Reference to transfer standard Accelerometer	

Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
low are applicable for the calibration son against laboratory references unl	of both measuring instrumers of both measuring instrumers of the stated in the	ents and for instruments with an eremarks column.	n output. The
			A
-60 °C to 100 °C 100 °C to 180 °C	1.7 ºC 3.2 ºC	For accelerometer calibration.	
0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup> 20 Hz to 630 Hz 0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup>	3.3 %	At a nominal 1 g <sub>n</sub> This uncertainty applies for masses up to 60 grams, larger masses can be calibrated but at greater	
20 Hz to 630 Hz	4.5 %	uncertainties.	
Frequency 160 Hz 10 ms <sup>2</sup> nominal	1.5 %		
10mV to 10V, 5 Hz to 10KHz Up to 25g	2.4 %	Note indicated outputs on Bench Calibrators for acceleration are normally indicated in "g".	
10mV to 10V, 10 Hz to 1KHz Up to 1000 m/sec pk	2.4 %		
10 mV to 10V, 10 Hz to 250 Hz Up to 10 mm pk to pk	2.4 %		
20 Hz to 20 kHz 10 mV to 750 V 0 % to 1 % 1 % to 100 %	0.20 % distortion 0.47 % distortion	Bandpass filter between 20 Hz and 50 kHz	
	Sch United Ki 2 Pine Trees, Cherts Calibra Issue Calibration performed by the Org Range Iow are applicable for the calibration son against laboratory references und -60 °C to 100 °C 100 °C to 180 °C 0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup> 20 Hz to 630 Hz 0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup> 20 Hz to 630 Hz 0.03 mV/ms <sup>2</sup> to 1000 mV/ms <sup>2</sup> 20 Hz to 630 Hz Frequency 160 Hz 10 ms <sup>2</sup> nominal 10mV to 10V, 5 Hz to 10KHz Up to 25g 10mV to 10V, 10 Hz to 1KHz Up to 1000 m/sec pk 10 mV to 10V, 10 Hz to 250 Hz Up to 100 m/sec pk 10 mV to 10V, 10 Hz to 250 Hz Up to 10 mp k to pk 20 Hz to 20 kHz 10 mV to 750 V 0 % to 1% 1 % to 100 %	Schedule of Accre issued by United Kingdom Accrea 2 Pine Trees, Chertsey Lane, Staines-u         Calibration, Maintenance Issue No: 044 Issue date         Calibration performed by the Organisation at the locations         -60 °C to 100 °C       1.7 °C         100 °C to 180 °C       1.7 °C         20 Hz to 630 Hz       3.3 %         0.03 mV/ms² to 1000 mV/ms²       2.0 Hz         20 Hz to 630 Hz       1.5 %         10mV to 10V, 5 Hz to 10KHz Up to 1000 m/sec pk       2.4 %         10mV to 10V, 10 Hz to 1250 Hz Up to 10mm pk to pk       2.4 %	Schedule of Accreditation Issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3H         Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025         Calibration Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025         Calibration performed by the Organisation at the locations specified         Remarks         Calibration of both measuring instruments and for instruments with arison against laboratory references unless otherwise stated in the remarks column.         -60 °C to 100 °C 100 °C to 100 °C 100 °C to 100 °C 100 °C to 100 °C 20 Hz to 630 Hz 20 Hz to 630 Hz 20 Hz to 630 Hz 10 ms' nominal 10 ms' nominal 10 ms' nominal 10 ms' to 1000 mV/ms² 20 Hz to 630 Hz 10 ms' nominal 10 ms' to 100, 5 Hz to 10KHz Up to 200 msee pk 10 m00 to 10V, 10 Hz to 1KHz Up to 1000 msee pk 10 m00 to 10V, 10 Hz to 1KHz Up to 1000 msee pk 10 m00 to 10V, 10 Hz to 250 Hz Up to 100 msee pk 10 m00 to 10V, 10 Hz to 250 Hz Up to 100 ms kep k 2.4 % 10 mV to 10V, 10 Hz to 250 Hz Up to 100 msee pk 12 CHz to 20 Hz 10 mm pk to pk 12 CHz to 100 ms 10 CHz to 100 m

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
		Expanded		
Measured Quantity Instrument or Gauge	Range	Measurement Uncertainty $(k = 2)$	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrumess otherwise stated in the	ents and for instruments with ar remarks column.	output. The
ACCELERATION TRANSDUCERS - SHOCK CALIBRATION				A
Piezoelectric and Piezoresistive and voltage types Transducer at nominal 23 °C	Sensitivity $0.001 \text{ pC/ms}^2$ to 100 pC/ms <sup>2</sup> $0.001 \text{ mV/ms}^2$ to 10 mV/ms <sup>2</sup> Within the range of: $2 \text{ ms}^2$ to 1000 ms <sup>2</sup>	3.0 %	Calibration of charge sensitivity by comparison with a reference (precision grade) transducer The transducer to be calibrated must have a mass of no more than 40 grams.	
Vibration meters and analysers	<i>Frequency 10 Hz to 1 kHz</i> Range 0.1 ms <sup>2</sup> to 1 ms <sup>2</sup>	2.0 %		
Shock and Impulse hammers Sensitivity	Half Sine Wave Frequency 500 Hz to 20 kHz Force 0.2 N to 450 N pC/N mV/N	3.3 % 3.3 %	Sensitivity measured in pC/N or mV/N dependant on hammer type	
Spring hammers				
Impact Energy Imparted from Spring Operated Impact Test Apparatus - as specified in BS EN 60068-2-75:1997 and IEC 60068-2-75 1997	0.1 Joule to 1.0 Joule	0.015 Joule	Calibrations can be given in Joule or Newton Meter units.	
HELICOPTER BLADE BALANCING TEST SETS				
Strobe frequency				
	1 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 150 Hz 150 Hz to 200 Hz	1.2 Hz 1.4 Hz 1.6 Hz 1.7 Hz		
MASS	500 g 200 g 100 g	9 mg 7 mg 7 mg	Calibrated using Borda substitution method.	A
NON-AUTOMATIC WEIGHING MACHINES	100 g 200 g 500 g	19 mg 19 mg 20 mg	Weights are available in OIML Class: M1 from 1 g to 500 g Calibration method in line with the requirements of Euramet gauge cg-18	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Cal	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
	Calibration performed by the	e Organisation at the locations sp	pecified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
Values and uncertainties listed I method used is by direct compa	below are applicable for the calibi rison against laboratory reference	ration of both measuring instruments es unless otherwise stated in the re	s and for instruments with marks column.	an output. The	
	Gates	head Laboratory			
ELECTRICAL					
Generation	$\begin{array}{l} 0 \text{ mV to } 50 \text{ mV} \\ 50 \text{ mV to } 100 \text{ mV} \\ 100 \text{ mV to } 220 \text{ mV} \\ 0.22 \text{ V to } 0.5 \text{ V} \\ 0.5 \text{ to } 1 \text{ V} \\ 1 \text{ V to } 2.2 \text{ V} \\ 2.2 \text{ V to } 5 \text{ V} \\ 5 \text{ V to } 11 \text{ V} \\ 11 \text{ V to } 22 \text{ V} \\ 22 \text{ V to } 50 \text{ V} \\ 50 \text{ V to } 100 \text{ V} \\ 100 \text{ V to } 220 \text{ V} \\ 220 \text{ V to } 500 \text{ V} \\ 500 \text{ V to } 1100 \text{ V} \\ 500 \text{ V to } 1100 \text{ V} \\ \end{array}$	1.1 $\mu$ V 1.4 $\mu$ V 2.5 $\mu$ V 4.9 $\mu$ V 9.4 $\mu$ V 20 $\mu$ V 47 $\mu$ V 0.10 mV 0.20 mV 0.54 mV 1.0 mV 2.3 mV 6.5 mV 14 mV			
Measurement	0 mV 0 mV to 50 mV 50 mV to 100 mV 100 mV to 200 mV 200 mV to 500 mV 500 mV to 1 V 1 V to 2 V 2 V to 5 V 5 V to 10 V 10 V to 20 V 20 V to 50 V 50 V to 100 V 100 V to 200 V 200 V to 500 V 500 V to 1050 V	0.65 µV 0.86 µV 1.1 µV 2.0 µV 2.1 µV 3.6 µV 7.0 µV 18 µV 34 µV 69 µV 0.30 mV 0.53 mV 1.1 mV 2.8 mV 5.5 mV			

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	output. The
DC RESISTANCE Sourcing - Spot Values	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 190 kΩ 190 kΩ 190 MΩ 19 MΩ 100 MΩ	$\begin{array}{c} 58 \ \mu\Omega \\ 0.13 \ m\Omega \\ 0.24 \ m\Omega \\ 0.38 \ m\Omega \\ 0.68 \ m\Omega \\ 2.3 \ m\Omega \\ 4.4 \ m\Omega \\ 17 \ m\Omega \\ 33 \ m\Omega \\ 0.16 \ \Omega \\ 0.31 \ \Omega \\ 1.8 \ \Omega \\ 3.5 \ \Omega \\ 27 \ \Omega \\ 53 \ \Omega \\ 0.53 \ k\Omega \\ 1.2 \ k\Omega \\ 15 \ k\Omega \end{array}$		В
DC RESISTANCE				В
Sourcing - Range values	0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1 kΩ	0.51 mΩ 1.1 mΩ 3.6 mΩ 11 mΩ 36 mΩ		
	1.1 kΩ to 3.3 kΩ 3.3 kΩ to 11 kΩ 11 k Ω to 33 kΩ 33 kΩ to 110 kΩ	110 mΩ 0.36 mΩ 1.1 Ω 3.6 Ω		
	110 kΩ to 330 kΩ 330 kΩ to 1.1 MΩ 1.1 MΩ to 3.3 MΩ 3.3 MΩ to 11 MΩ 11 MΩ to 33 MΩ	12 Ω 41 Ω 230 Ω 1.7 kΩ 9.6 kΩ		
	33 MΩ to 110 MΩ 110 MΩ to 330 MΩ 330 MΩ to 1.1 GΩ	64 kΩ 1.1 MΩ 19 MΩ		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibr Issue	ration, Maintenance ar e No: 044 Issue date:	nd Repair Ltd 07 May 2025	
	Calibration performed by the O	rganisation at the locations sp	pecified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed b method used is by direct compar	elow are applicable for the calibratic rison against laboratory references u	on of both measuring instrument unless otherwise stated in the re	ts and for instruments with an emarks column.	output. The
DC RESISTANCE Measurement	$\begin{array}{c} 0 \ \Omega \\ 0 \ \Omega \ \text{to} \ 0.5 \ \Omega \\ 0.5 \ \Omega \ \text{to} \ 1 \ \Omega \\ 1 \ \Omega \ \text{to} \ 2 \ \Omega \\ 2 \ \Omega \ \text{to} \ 5 \ \Omega \\ 2 \ \Omega \ \text{to} \ 5 \ \Omega \\ 5 \ \Omega \ \text{to} \ 10 \ \Omega \\ 10 \ \Omega \ \text{to} \ 20 \ \Omega \\ 20 \ \Omega \ \text{to} \ 50 \ \Omega \\ 50 \ \Omega \ \text{to} \ 100 \ \Omega \\ 100 \ \Omega \ \text{to} \ 200 \ \Omega \\ 100 \ \Omega \ \text{to} \ 200 \ \Omega \\ 0.2 \ \text{k} \ \Omega \ \text{to} \ 100 \ \Omega \\ 100 \ \Omega \ \text{to} \ 200 \ \Omega \\ 0.5 \ \text{k} \ \Omega \ \text{to} \ 1 \ \text{k} \ \Omega \\ 1 \ \text{k} \ \Omega \ \text{to} \ 5 \ \text{k} \ \Omega \\ 5 \ \text{k} \ \Omega \ \text{to} \ 10 \ \text{k} \ \Omega \\ 10 \ \text{k} \ \Omega \ \text{to} \ 20 \ \text{k} \ \Omega \\ 20 \ \text{k} \ \Omega \ \text{to} \ 100 \ \text{k} \ \Omega \\ 100 \ \text{k} \ \Omega \ \text{to} \ 200 \ \text{k} \ \Omega \\ 100 \ \text{k} \ \Omega \ \text{to} \ 50 \ \text{k} \ \Omega \\ 0.5 \ \text{M} \ \Omega \ \text{to} \ 100 \ \text{k} \ \Omega \\ 100 \ \text{k} \ \Omega \ \text{to} \ 5 \ \text{M} \ \Omega \\ 100 \ \text{k} \ \Omega \ \text{to} \ 5 \ \text{M} \ \Omega \\ 100 \ \text{k} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 100 \ \text{M} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 50 \ \text{M} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 50 \ \text{M} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 50 \ \text{M} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 100 \ \text{M} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 100 \ \text{M} \ \Omega \ \text{to} \ 50 \ \text{M} \ \Omega \\ 100 \ \text{M} \ \Omega \ \text{to} \ 100 \ \text{M} \ \Omega \ $	$\begin{array}{c} 7.5 \ \mu\Omega \\ 13 \ \mu\Omega \\ 21 \ \mu\Omega \\ 43 \ \mu\Omega \\ 92 \ \mu\Omega \\ 0.14 \ m\Omega \\ 0.25 \ m\Omega \\ 0.58 \ m\Omega \\ 1.1 \ m\Omega \\ 2.2 \ m\Omega \\ 5.5 \ m\Omega \\ 11 \ m\Omega \\ 21 \ m\Omega \\ 5.5 \ m\Omega \\ 11 \ m\Omega \\ 21 \ m\Omega \\ 5.5 \ m\Omega \\ 11 \ m\Omega \\ 22 \ \Omega \\ 0.62 \ \Omega \\ 1.1 \ \Omega \\ 0.22 \ \Omega \\ 0.62 \ \Omega \\ 1.1 \ \Omega \\ 2.2 \ \Omega \\ 7.8 \ \Omega \\ 13 \ \Omega \\ 26 \ \Omega \\ 0.18 \ \kappa\Omega \\ 0.26 \ \kappa\Omega \\ 0.52 \ \kappa\Omega \\ 14 \ \kappa\Omega \\ 19 \ \kappa\Omega \\ 0.34 \ M\Omega \end{array}$		В
Generation	0 $\mu$ A to 50 $\mu$ A 50 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 220 $\mu$ A 0.22 mA to 0.5 mA 0.5 mA to 1 mA 1 mA to 2.2 mA 2.2 mA to 5 mA 5 mA to 10 mA 10 mA to 22 mA 22 mA to 50 mA 22 mA to 50 mA 50 mA to 100 mA 100 mA to 220 mA 0.22 A to 0.5 A 0.5 A to 1 A 1 A to 2.2 A 2.2 A to 3 A 3 A to 11 A 11 A to 20.5 A	12 nA 13 nA 19 nA 37 nA 71 nA 0.15μA 0.37 μA 0.71 μA 1.5 μA 4.2 μA 8.2 μA 8.2 μA 8.2 μA 0.12 mA 0.12 mA 0.24 mA		В

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Orga	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The
DC CURRENT (continued) Generation	20 A to 30 A 30 A to 110 A 110 A to 200 A 200 A to 550 A 550 A to 1025 A	150 mA 500 mA 930 mA 2.5 A 4.7 A	Current clamp calibration using a multi turn coil	В
Measurement AC VOLTAGE Generation	0 $\mu$ A to 5 $\mu$ A 5 $\mu$ A to 10 $\mu$ A 10 $\mu$ A to 20 $\mu$ A 20 $\mu$ A to 50 $\mu$ A 50 $\mu$ A to 50 $\mu$ A 50 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 200 $\mu$ A 0.2 mA to 0.5 mA 0.5 mA to 1 mA 1 mA to 2 mA 2 mA to 5 mA 5 mA to 10 mA 10 mA to 20 mA 20 mA to 50 mA 50 mA to 100 mA 100 mA to 200 mA 0.2 A to 0.5 A 0.5 A to 1 A 1 A to 2 A 2 A to 5 A 5 A to 10 A 10 A to 20 A 20 A to 30 A <b>0.22 mV to 0.5 mV</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz	0.57 nA 0.63 nA 1.2 nA 1.1 nA 1.4 nA 2.9 nA 9.1 nA 13 nA 26 nA 0.11 µA 0.18 µA 0.35 µA 3.7 µA 6.8 µA 14 µA 0.14 mA 0.19 mA 0.39 mA 1.5 mA 2.8 mA 14 mA 20 mA 5.8 µV 5.8 µV		В
	500 kHz to 1 MHz	46 μV		

Accredited to ISO/IEC 17025:2017	Schedule of Accreditation issued by         United Kingdom Accreditation Service         2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK         Calibration, Maintenance and Repair Ltd         Issue No: 044       Issue date: 07 May 2025         Calibration performed by the Organisation at the locations specified			
Instrument or Gauge	Range	Measurement Uncertainty ( $k = 2$ )	Remarks	
method used is by direct comparis	son against laboratory references unl	ess otherwise stated in the	remarks column.	
AC VOLTAGE Generation (continued)	0.5 mV to 1 mV 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 500 kHz 200 kHz to 500 kHz 200 kHz to 100 Hz 200 kHz to 100 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 20 kHz 20 kHz to 500 kHz 10 kHz to 20 kHz 20 kHz to 500 kHz 100 kHz to 200 kHz 20 kHz to 500 kHz 200 kHz to 500 kHz 200 kHz to 100 Hz 100 kHz to 200 kHz 200 kHz to 1 MHz 220 kHz to 500 kHz 500 kHz to 1 MHz 220 kHz to 500 kHz 500 kHz to 1 MHz 200 kHz to 500 kHz 500 kHz to 1 MHz 200 kHz to 500 kHz 500 kHz to 1 MHz 200 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 KHz 100 kHz to 200 kHz 200 kHz to 500 kHz 200 kHz to 500 kHz 200 kHz to 100 Hz 100 kHz to 200 kHz 200 kHz to 500 kHz 200 kHz to 100 Hz 100 Hz to 10 Hz 100 Hz to 10 Hz 100 Hz to 10 Hz 200 kHz to 500 kHz 200 kHz to 500 kHz 200 kHz to 100 Hz 100 Hz to 10 Hz 100 Hz to 10 Hz 100 Hz to 10 Hz 200 kHz to 500 kHz 500 kHz to 100 Hz 100 Hz to 10 Hz 100 Hz to 10 Hz 200 kHz to 500 kHz 200 kHz to 100 Hz 100 Hz to 10 Hz 200 kHz to 500 kHz 200 kHz to 100 KHz 100 kHz to 200 kHz 200 kHz to 100 Hz 100 kHz to 200 kHz 200 kHz to 100 kHz 100 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 100 kHz 100 kHz to 100	5.9 $\mu$ V 5.8 $\mu$ V 9.3 $\mu$ V 17 $\mu$ V 35 $\mu$ V 47 $\mu$ V 6.0 $\mu$ V 5.8 $\mu$ V 7.8 $\mu$ V 7.0 $\mu$ V 7.1 $\mu$ V 7.2 $\mu$ V 7.2 $\mu$ V		В

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with an remarks column.	output. The
AC VOLTAGE Generation (continued)	<b>10 mV to 22 mV</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 50 mV 10 Hz to 20 HZ 20 kHz to 100 Hz 100 Hz to 20 HZ 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 10 kHz 100 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 20 kHz to 500 kHz 50 kHz to 100 mV 10 Hz to 20 Hz 200 kHz to 500 kHz 500 kHz to 1 MHz <b>50 mV to 100 mV</b> 10 Hz to 20 Hz 20 hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 10kHz to 20 kHz 20 kHz to 500 kHz 500 kHz to 100 Hz 100 Hz to 1 kHz 10kHz to 20 kHz 20 kHz to 500 kHz 50 kHz to 100 kHz 100 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz <b>100 mV to 220 mV</b> 10 Hz to 20 Hz 20 kHz to 100 Hz 100 kHz to 100 Hz 100 kHz to 100 Hz 100 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz <b>100 mV to 20 mV</b> 10 Hz to 20 Hz 20 kHz to 500 kHz 500 kHz to 100 Hz 100 kHz to 20 kHz 20 kHz to 500 kHz 500 kHz to 100 Hz 100 kHz to 20 kHz 20 kHz to 500 kHz 500 kHz to 100 Hz 100 kHz to 20 kHz 20 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 kHz to 200 kHz 500 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 100 kHz 100	17 $\mu$ V 9.3 $\mu$ V 7.6 $\mu$ V 7.6 $\mu$ V 7.6 $\mu$ V 7.6 $\mu$ V 13 $\mu$ V 26 $\mu$ V 37 $\mu$ V 57 $\mu$ V 0.13 mV 39 $\mu$ V 13 $\mu$ V 13 $\mu$ V 13 $\mu$ V 13 $\mu$ V 13 $\mu$ V 13 $\mu$ V 24 $\mu$ V 62 $\mu$ V 72 $\mu$ V 0.11 mV 0.24 mV 72 $\mu$ V 30 $\mu$ V 17 $\mu$ V 17 $\mu$ V 17 $\mu$ V 17 $\mu$ V 17 $\mu$ V 17 $\mu$ V 13 $\mu$ V 0.11 mV 0.24 mV 0.13 mV 0.13 mV 0.21 mV 0.43 mV 0.23 mV 0.28 mV 0.92 mV		В

UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017 Measured Quantity	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK         Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025         Calibration performed by the Organisation at the locations specified         Expanded			
Instrument or Gauge           Values and uncertainties listed be method used is by direct comparison	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with an remarks column.	Code
AC VOLTAGE Generation (continued)	<b>220 mV to 0.5 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 500 kHz 200 kHz to 500 kHz 200 kHz to 200 Hz 200 kHz to 1 MHz <b>0.5 V to 1 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz <b>1 V to 2 .2 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 100 Hz to 100 Hz 20 kHz to 500 kHz 500 kHz to 100 Hz 100 Hz to 1 kHz 1 kHz to 100 kHz 100 kZ to 500 kHz 500 kHz to 100 Hz 200 kHz to 500 kHz 200 kHz to 100 Hz 100 hz to 1 MHz <b>2.2 V to 5 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 kHz to 100 Hz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 Hz 100 Hz to 100 Hz 100 Hz to 100 Hz 100 Hz to 100 Hz 200 kHz to 500 kHz 200 kHz to 500 kHz	0.37 mV 0.11 mV 52 µV 52 µV 52 µV 52 µV 0.19 mV 0.33 mV 0.33 mV 0.83 mV 1.8 mV 0.70 mV 0.10 mV 0.10 mV 0.10 mV 0.10 mV 0.10 mV 0.10 mV 0.10 mV 0.10 mV 0.34 mV 0.58 mV 1.5 mV 3.0 mV 1.5 mV 0.22 mV 0.36 mV 0.72 mV 1.1 mV 6.2 mV 3.7 mV 1.1 mV 0.52 mV 0		В

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
O654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maintenance No: 044 Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	output. The
AC VOLTAGE Generation (continued)	<b>5</b> V to 10 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 500 kHz 50 kHz to 100 kHz 100 kHz to 200 kHz 200 kHz to 100 Hz 100 V to 22 V 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 10 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 100 kHz 100 Hz to 2 Hz 200 kHz to 500 kHz 500 kHz to 1 MHz <b>22 V to 50 V</b> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 100 Hz to 1 kHz 1 kHz to 100 kHz 50 kHz to 100 KHz 100 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 KHz 100 Hz to 1 kHz 1 kHz to 100 kHz 10 kHz to 20 Hz 20 kHz to 50 kHz 50 kHz to 100 kHz 10 kHz to 20 Hz 20 kHz to 50 kHz 50 kHz to 100 KHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 Hz 10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 10 kHz 10 kHz to 20 kHz 50 kHz to 10 kHz 50 kHz to 1	7.0 mV 2.1 mV 1.0 mV 1.0 mV 1.0 mV 1.0 mV 1.0 mV 1.6 mV 3.3 mV 7.2 mV 17 mV 36 mV 15 mV 2.2 mV 2.2 mV 2.2 mV 2.2 mV 2.2 mV 3.6 mV 7.1 mV 15 mV 36 mV 77 mV 37 mV 11 mV 5.8 mV 5.8 mV 5.8 mV 5.8 mV 5.8 mV 5.8 mV 5.8 mV 70 mV 21 mV 11 mV 11 mV 12 mV 29 mV 70 mV 29 mV 70 mV		В

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
libration performed by the Orga	anisation at the locations	specified		
Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
are applicable for the calibration of against laboratory references unle	of both measuring instrume ess otherwise stated in the	ents and for instruments with an remarks column.	output. The	
<b>00 V to 220 V</b> 0 Hz to 20 Hz 0 Hz to 40 Hz 0 Hz to 100 Hz 00 Hz to 1 kHz	0.15 V 46 mV 23 mV 23 mV 23 mV		В	
kHz to 10 kHz 0 kHz to 20 kHz 0 kHz to 50 kHz 0 kHz to 100 kHz	23 mV 23 mV 64 mV 0.15 V			
<b>20 V to 500 V</b> 0 Hz 0 Hz to 500 Hz 00 Hz to 1 kHz	57 mV 57 mV 57 mV			
<b>00 V to 1100 V</b> 0 Hz 0 Hz to 500 Hz 00 Hz to 1 kHz	0.12 V 0.12 V 0.12 V			
<b>mV to 5 mV</b> 0 Hz to 2 kHz kHz to 10 kHz. 0 kHz to 30 kHz 0 kHz to 100 kHz 00 kHz to 300 kHz 00 kHz to 1 MHz	3.0 μV 3.3 μV 3.3 μV 18 μV 58 μV 0.12 mV		В	
<b>mV to 10 mV</b> 0 Hz 0 Hz to 2 kHz kHz to 10 kHz. 0 kHz to 30 kHz 0 kHz to 60 kHz 00 kHz to 100 kHz 00 kHz to 1 MHz	3.0 μV 4.2 μV 4.9 μV 5.0 μV 36 μV 35 μV 0.12 mV 0.23 mV			
	Range         are applicable for the calibration of against laboratory references unleaded of the construction of against laboratory references unleaded of the construction of the consthe construction of the consthe construction	RangeExpanded Measurement Uncertainty $(k = 2)$ are applicable for the calibration of both measuring instrume regainst laboratory references unless otherwise stated in the <b>0 V to 220 V</b> PHz to 20 Hz0.15 V 0.15 V 0.15 V 1Hz to 40 Hz1Hz to 40 Hz23 mV 23 mV0 Hz to 1 kHz23 mV 23 mV0 Hz to 10 kHz23 mV 23 mV0 Hz to 10 kHz23 mV 23 mV0 Hz to 10 kHz23 mV 23 mV0 Hz to 500 Hz57 mV 57 mV0 V to 500 V HZ to 500 Hz57 mV 57 mV0 V to 100 V HZ to 500 Hz0.12 V 0.12 V 0.12 V1 Hz0.12 V 0.12 V1 Hz to 500 Hz0.12 V 0.12 V0 V to 1100 V Hz to 500 Hz0.12 V 0.12 V1 Hz to 500 Hz0.12 V 0.12 V0 Hz to 1 kHz3.0 $\mu$ V sHz to 500 Hz0 Hz to 10 kHz3.3 $\mu$ V sHz to 30 kHz1 Hz to 2 kHz3.0 $\mu$ V sHz to 300 kHz0 kHz to 100 kHz18 $\mu$ V 0.12 mV0 KHz to 10 kHz3.0 $\mu$ V sHz to 30 kHz1 Hz0.12 mV1 Hz0.23 mV	Range     Expanded Measurement Uncertainty (k = 2)     Remarks       are applicable for the calibration of both measuring instruments and for instruments with an rgainst laboratory references unless otherwise stated in the remarks column.     Image: Column 2015 V V to 220 V       0 V to 220 V     0.15 V       Hz to 20 Hz     0.15 V       Hz to 20 Hz     0.15 V       Hz to 20 Hz     0.15 V       Hz to 100 Hz     23 mV       0 V to 220 V     0.15 V       Hz to 100 Hz     23 mV       Viktz to 20 KHz     23 mV       Viktz to 20 KHz     23 mV       Viktz to 20 KHz     23 mV       Viktz to 50 KHz     64 mV       Viktz to 50 KHz     64 mV       Viktz to 50 KHz     64 mV       Vitz to 500 V     57 mV       0 V to 500 V     0 Hz       Hz     0.12 V       Hz to 50 Hz     0.12 V       0 V to 50 V     0.12 V       Hz     0.12 V       0 Hz to 1 kHz     3.0 µV       Hz to 20 KHz     3.0 µV       Hz to 10 KHz     3.0 µV       Hz to 20 KHz     3.0 µV       Hz to 20 KHz     3.0 µV       Hz to 20 KHz     3.0 µV	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
	Calibration performed by the Orga	anisation at the locations	specified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
Values and uncertainties listed bel method used is by direct comparis	ow are applicable for the calibration on against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with an remarks column.	output. The	
AC VOLTAGE Measurement (continued)	<b>10 mV to 50 mV</b> 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 2 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 6 MHz 6 MHz to 8 MHz 8 MHz to 10 MHz <b>50 mV to 100 mV</b> 10 Hz 10 Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 MHz <b>50 mV to 0.5 V</b> 10Hz 10 Hz 10 Hz 2 MHz to 2 MHz 2 MHz to 8 MHz 8 MHz to 10 MHz <b>0.1 V to 0.5 V</b> 10Hz 10 kHz to 300 kHz 300 kHz to 100 kHz 10 kHz to 300 kHz 300 kHz to 100 kHz 10 kHz to 300 kHz 300 kHz to 100 kHz 10 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 kHz to 10 MHz <b>0.5 V to 1 V</b> 10Hz 10 Hz to 2 kHz 2 kHz to 10 MHz <b>0.5 V to 1 V</b> 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 300 kHz 300 kHz to 10 MHz <b>0.5 V to 1 V</b> 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 300 kHz 300 kHz to 10 MHz <b>0.5 V to 1 V</b> 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 300 kHz 300 kHz to 10 0 kHz 10 kHz to 300 kHz 300 kHz to 10 0 kHz 10 kHz to 300 kHz 300 kHz to 10 0 kHz 10 kHz to 2 MHz 2 kHz to 10 kHz 10 kHz to 300 kHz 300 kHz to 10 0 kHz 10 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 2 MHz 2 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 3 MHz to 2 MHz 4 MHz to 8 Mz 3	6.1 μV 5.7 μV 7.9 μV 14 μV 34 μV 0.14 mV 0.69 mV 1.1 mV 2.7 mV 5.1 mV 5.2 mV 9.5 mV 11 10 μV 15 μV 27 μV 68 μV 0.25 mV 1.3 mV 1.9 mV 4.9 mV 10 mV 19 mV 4.9 mV 10 mV 19 mV 50 μV 4.7 μV 73 μV 0.13 mV 0.34 mV 1.4 mV 6.3 mV 1.4 mV 6.3 mV 1.1 mV 92 mV 91 μV 90 μV 0.14 mV 0.62 mV 0.12 mV 12 mV 18 mV 12 mV 18 mV 19 mV		В	

UKAS CALIBRATION 0654	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd			
Accredited to	Issue	No: 044 Issue date	e: 07 May 2025	
ISO/IEC 17025:2017	Calibration performed by the Org	anisation at the locations	specified	
			·	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The
AC VOLTAGE Measurement (continued)	1 V to 5 V 10Hz 10 Hz to 2 kHz 2 kHz to 10 kHz. 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz 5 V to 10 V 10Hz 10 Hz to 2 kHz 2 kHz to 100 kHz 10 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz 10 V to 50 V 10Hz 10 K to 50 V 10Hz 10 kHz to 300 kHz 300 kHz to 100 KHz 10 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 50 V to 100 V 10 Hz to 2 kHz 2 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 50 V to 100 V 10 Hz to 2 kHz 2 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 100 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz 300 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 100 kHz	0.49 mV 0.48 mV 0.73 mV 1.3 mV 3.3 mV 14 mV 60 mV 0.10 V 0.26 V 0.49 V 0.90 V 0.90 V 0.90 mV 1.4 mV 2.7 mV 6.3 mV 25 mV 0.12 V 0.18 V 0.48 V 1.0 V 1.8 V 6.0 mV 5.6 mV 6.7 mV 16 mV 39 mV 0.22 V 0.88 V 11 mV 13 mV 28 mV 71 mV 0.43 V 1.4 V 73 mV 0.16 V 0.38 V 0.13 V 0.28 V 0.70 V		В

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
	Calibration performed by the Org	anisation at the locations	specified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume less otherwise stated in the	ents and for instruments with ar remarks column.	output. The	
AC CURRENT Generation	<b>10 μA to 50 μA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>50 μA to 100 μA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>100 μA to 220 μA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>0.22 mA to 0.5 mA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>0.5 mA to 1 mA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>1 mA to 2.2 mA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>1 mA to 2.2 mA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>1 mA to 2.4 mA</b> 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz <b>1 k</b> Hz to 5 kHz <b>5 k</b> Hz to 10 kHz	38 nA         26 nA         71 nA         0.16 μA         57 nA         30 nA         0.10 μA         0.24 μA         0.11 μA         48 nA         0.19 μA         0.47 μA         48 nA         47 nA         0.58 μA         1.2 μA         0.49 μA         0.19 μA         1.0 μA         2.4 μA         1.1 μA         0.41 μA         1.9 μA         1.0 μA         2.5 μA         1.0 μA         2.4 μA		В	
	20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	11 μΑ 4.1 μΑ 19 μΑ 47 μΑ			

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
	Calibration performed by the Org	anisation at the locations	specified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code	
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The	
AC CURRENT Generation (continued)	22 mA to 50 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 50 mA to 100 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 100 mA to 220 mA 20 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.22 A to 0.5 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.5 A to 1 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 0.5 A to 1 A 20 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	25 μA 11 μA 70 μA 0.16 mA 49 μA 21 μA 99 μA 0.24 mA 0.11 mA 46 μA 0.19 mA 0.47 mA 0.47 mA 0.62 mA 0.87 mA 0.99 mA 1.2 mA		В	
	<ul> <li>1 A to 2.2 A</li> <li>20 Hz to 1 kHz</li> <li>1 kHz to 5 kHz</li> <li>5 kHz to 10 kHz</li> <li>2.2 A to 3 A</li> <li>10 Hz to 45 Hz</li> <li>45 Hz to 1 kHz</li> <li>1 kHz to 5 kHz</li> <li>5 kHz to 100 Hz</li> <li>3 A to 11 A</li> <li>45 Hz to 100 Hz</li> <li>100 Hz to 1 kHz</li> <li>1 kHz to 5 kHz</li> <li>11 A to 20.5 A</li> <li>45 Hz to 100 Hz</li> <li>100 Hz to 1 kHz</li> <li>1 kHz to 5 kHz</li> <li>45 Hz to 100 Hz</li> <li>100 Hz to 1 kHz</li> <li>1 kHz to 5 kHz</li> <li>45 Hz to 100 Hz</li> <li>20 A to 30 A</li> </ul>	1.9 mA 2.2 mA 2.6 mA 6.2 mA 2.1 mA 21 mA 87 mA 8.0 mA 13 mA 0.38 mA 29 mA 36 mA 710 mA	Calibration of current clamps using a multi turn coil		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
CALIBRATION 0654 Accredited to ISO/IEC 17025:2017					
	Calibration performed by the Org	anisation at the locations	specified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration on against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	n output. The	
AC CURRENT Generation (continued)	<b>30 A to 110 A</b> 45 Hz to 100 Hz	520 mA	Calibration of current clamps using a multi turn	В	
	<b>110 A to 200 A</b> 100 Hz to 400 Hz	960 mA			
	<b>200 A to 550 A</b> 45 Hz to 100 Hz	2.5 A			
Measurement	<b>550 A to 1025 A</b> 45 Hz to 100 Hz	4.8 A			
Weasurement	<b>0 μΑ to 5 μΑ</b> 10 Hz to 55 Hz 55 Hz to 10 kHz	13 nA 25 nA			
	<b>5 μΑ to 10 μΑ</b> 10 Hz to 55 Hz 55 Hz to 10 kHz	24 nA 34 nA			
	<b>10 μA to 20 μA</b> 10 Hz to 55 Hz 55 Hz to 10 kHz	47 nA 52 nA			
	<b>20 μA to 50μA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 30 kHz	23 nA 35 nA 48 nA 53 nA			
	<b>50 μA to 100 μA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 20 kHz 20 kHz to 30 kHz	36 nA 63 nA 88 nA 91 nA			
	<b>100 μA to 200 μA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	67 nA 0.12 μA 0.17 μA			
	<b>0.2 mA to 0.5 mA</b> 10Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	0.19 μΑ 0.32 μΑ 0.44 μΑ			
	<b>0.5 mA to 1 mA</b> 10 Hz to 2 kHz. 2 kHz to10 kHz 10 kHz to 30 kHz	0.34 μA 0.62 μA 0.86 μA			

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
	Calibration performed by the Org	anisation at the locations	specified		
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code	
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume less otherwise stated in the	ents and for instruments with an remarks column.	output. The	
AC CURRENT Measurement (continued)	<b>1 mA to 2 mA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	0.66 μΑ 1.2 μΑ 1.7 μΑ		В	
	<b>2 mA to 5 mA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	1.9 μΑ 3.2 μΑ 4.4 μΑ			
	<b>5 mA to 10 mA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	3.4 μΑ 6.2 μΑ 8.6 μΑ			
	<b>10 mA to 20 mA</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	6.6 μΑ 12 μΑ 17 μΑ			
	20 mA to 50 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	19 μΑ 31 μΑ 45 μΑ			
	50 mA to 100 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	34 μΑ 61 μΑ 86 μΑ			
	100 mA to 200 mA 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	66 μΑ 0.12 mA 0.17 mA			
	0.2 A to 0.5 A 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	0.23 mA 0.36 mA 0.47mA			
	0.5 A to 1 A 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	0.38 mA 0.66 mA 0.92 mA			
	<b>1 A to 2 A</b> 10 Hz to 2 kHz. 2 kHz to 10 kHz 10 kHz to 30 kHz	0.74 mA 1.3 mA 1.8 mA			
	<b>2 A to 5 A</b> 10 Hz to 10 kHz	5.0 mA			

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK					
Accredited to	Calibra Issue	Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025				
	Calibration performed by the Org	anisation at the locations	specified			
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code		
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume less otherwise stated in the	ents and for instruments with ar remarks column.	n output. The		
AC CURRENT Measurement (continued)	<b>5 A to 10 A</b> 10 Hz to 10 kHz	10 mA		В		
	10 A to 20 A 10 Hz to 10 kHz 20 A to 30 A 10 Hz to 2 kHz	19 mA 24 mA				
	2 kHz to 10 kHz	32 mA		D		
Generation	10 MHz	3.2 in 10 <sup>9</sup>	Reference	D		
Measurement	10 Hz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 50 MHz 50 MHz to 600 MHz 600 MHz to 6 GHz	16 mHz 170 mHz 1.7 Hz 17 Hz 4.8 Hz 35 Hz				
Generation	1 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 20 MHz	2.3 μHz/Hz 1.2 μHz/Hz 0.012 μHz/Hz	May be reported as events per unit time			
Elapsed time			Mechanical timers / stop			
Single event	1 s to 12 hrs	0.05 s	watches			
Revolutions Per Minute Simulation	6 RPM to 600 RPM 600 RPM to 6000 RPM 6000 RPM to 120000 RPM	0.064 RPM 0.065 RPM 0.066 RPM	Optical			
CAPACITANCE				В		
Simulated generation	1 kHz 220 pF to 400 pF 400 pF to 1.1 nF 1.1 nF to 3.3 nF 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF	12 pF 13 pF 22 pF 34 pF 100 pF 320 pF 970 pF				

UKAS CALIBRATION 0654	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
Accredited to ISO/IEC 17025:2017			-	
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty $(k = 2)$	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume less otherwise stated in the	ents and for instruments with an remarks column.	output. The
CAPACITANCE (continued) Simulated generation (continued)	<b>100 Hz</b> 0.33 $\mu$ F to 1.1 $\mu$ F 1.1 $\mu$ F to 3.3 $\mu$ F 3.3 $\mu$ F to 11 $\mu$ F 11 $\mu$ F to 33 $\mu$ F 33 $\mu$ F to 110 $\mu$ F 110 $\mu$ F to 330 $\mu$ F 330 $\mu$ F to 1.1 mF 1.1 mF to 3.3 mF 3.3 mF to 11 mF 11 mF to 33 mF 33 mF to 110 mF	3.4 nF 10 nF 34 nF 160 nF 600 nF 1.8 μF 5.8 μF 18 μF 58 μF 290 μF 1.4 mF		В
Measurement	0 to 0.5 nF 0.5 nF to 1 nF 1 nF to 2 nF 2 nF to 5 nF 5 nF to 10 nF 10 nF to 20 nF 20 nF to 50 nF 50 nF to 100 nF 100 nF to 200 nF 0.2 $\mu$ F to 0.5 $\mu$ F 0.5 $\mu$ F to 1 $\mu$ F 1 $\mu$ F to 2 $\mu$ F 2 $\mu$ F to 5 $\mu$ F 50 $\mu$ F to 100 $\mu$ F 100 $\mu$ F to 200 $\mu$ F 0.2 mF to 0.5 mF 0.5 mF to 1 mF 1 mF to 2 mF 2 mF to 5 mF 5 mF to 10 mF 10 mF to 20 mF 20 mF to 50 mF 50 mF to 100 mF 100 mF to 200 mF 20 mF to 200 mF	2.0 pF 2.7 pF 5.1 pF 6.1 pF 10 pF 20 pF 35 pF 60 pF 0.12 nF 0.28 nF 0.50 nF 1.1 nF 3.5 nF 5.5 nF 10 nF 45 nF 76 nF 0.15 µF 0.45 µF 0.76 µF 1.5 µF 5.0 µF 8.7 µF 17 µF 45 µF 8.7 µF		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
UKAS CALIBRATION 0654 Accredited to ISO/IEC 17025:2017				
	Calibration performed by the Org	anisation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both measuring instrume ess otherwise stated in the	ents and for instruments with ar remarks column.	output. The
CALIBRATION OF INSULATION TESTERS				В
Insulation Resistance	10 kΩ to 190 kΩ 200 kΩ 400 kΩ 800 kΩ 1.9 MΩ 2 MΩ 4 MΩ 8 MΩ 19 MΩ 20 MΩ 40 MΩ 80 MΩ 190 MΩ 200 MΩ 400 MΩ 800 MΩ 1.9 GΩ	5.8 kΩ 6.1 kΩ 6.2 kΩ 6.4 kΩ 7.0 kΩ 9.4 kΩ 15 kΩ 28 kΩ 58 kΩ 0.13 MΩ 0.24 MΩ 0.47 MΩ 0.58 MΩ 1.8 MΩ 3.4 MΩ 6.7 MΩ 13 MΩ	Up to 1350 V	В
Insulation test voltage Nominal	0 to 10 V 10 V to 100 V 100 V to 1100 V	0.10 V 0.26 V 2.6 V	The test voltage will normally be measured with a 1 mA load.	В
Continuity Resistance	100 mΩ to 2 Ω 4 Ω 8 Ω 19 Ω 20 Ω 40 Ω 80 Ω 190 Ω 200 Ω 400 Ω 800 Ω 1.9 k Ω 2 kΩ 4 kΩ 8 kΩ 10 kΩ	$\begin{array}{c} 30 \text{ m}\Omega \\ 33 \text{ m}\Omega \\ 35 \text{ m}\Omega \\ 55 \text{ m}\Omega \\ 0.12 \Omega \\ 0.20 \Omega \\ 0.24 \Omega \\ 0.74 \Omega \\ 1.1 \Omega \\ 2.0 \Omega \\ 2.4 \Omega \\ 8.1 \Omega \\ 11 \Omega \\ 20 \Omega \\ 24 \Omega \end{array}$		В

	Sch United Kin 2 Pine Trees, Cherts	nedule ngdom <sub>sey Lane,</sub>	of Accr issued by Accreo Staines-u	editation ditation Service pon-Thames, TW18 3H	R, UK
0654 Accredited to ISO/IEC 17025:2017	Calibra Issue	tion, Maiı No: 044	ntenance Issue date	and Repair Ltd e: 07 May 2025	
	Calibration performed by the Org	anisation at	the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expa Measu Uncertair	nded rement nty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both meas ess otherwise	uring instrum stated in the	ents and for instruments with ar e remarks column.	n output. The
Temperature indicators, calibration by electrical simulat	ion	Excluding		By injection of equivalent DC voltages	В
Type K thermocouple	-270 °C to -200 °C -200 °C to -50 °C -50 °C to 1370 °C	1.6 °C 0.10 °C 0.10 °C	2.8 °C 0.17 °C 0.13 °C		
Type J thermocouple	-210 °C to -200 °C -200 °C to 0 °C 0 °C to 1200 °C	0.10 °C 0.10 °C 0.10 °C	0.23 °C 0.21 °C 0.14 °C		
Type E thermocouple	-270 °C to -200 °C -200 °C to 0 °C 0 °C to 1000 °C	0.69 °C 0.10 °C 0.10 °C	0.95 °C 0.13 °C 0.12 °C		
Type T thermocouple	-270 °C to -200 °C -200 °C to 0 °C 0 °C to 400 °C	1.1 °C 0.10 °C 0.10 °C	7.2 °C 0.48 °C 0.22 °C		
Type N thermocouple	-270 °C to -200 °C -200 °C to -100 °C -100 °C to 800 °C 800 °C to 1300 °C	3.7 °C 0.13 °C 0.10 °C 0.10 °C	4.3 °C 0.18 °C 0.13 °C 0.12 °C		
Type R thermocouple	-50 °C to 0 °C 0 °C to 150 °C 150 °C to 400 °C 400 °C to 1768 °C	0.30 °C 0.22 °C 0.15 °C 0.12 °C	0.34 °C 0.25 °C 0.19 °C 0.17 °C		
Type S thermocouple	-50 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 1768 °C	0.28 °C 0.21 °C 0.16 °C 0.13 °C	0.31 °C 0.24 °C 0.20 °C 0.17 °C		
Resistance thermometer (Pt 100)	-200 °C to -80 °C -80 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 400 °C 400 °C to 630 °C 630 °C to 800 °C	0.058 °C 0.058 °C 0.081°C 0.10 °C 0.12 °C 0.14 °C 0.27 °C		By injection of equivalent DC resistance	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK				
O654 Accredited to ISO/IEC 17025:2017	Calibra Issue Calibration performed by the Org	tion, Main No: 044	ntenance Issue date	and Repair Ltd e: 07 May 2025	
Measured Quantity Instrument or Gauge	Range	Expa Measu Uncertair	nded rement hty $(k = 2)$	Remarks	Location Code
Values and uncertainties listed be method used is by direct comparis	low are applicable for the calibration son against laboratory references unl	of both meas ess otherwise	uring instrum e stated in the	ents and for instruments with ar eremarks column.	output. The
Temperature simulators, calibration by electrical measur	ement			By measurement of equivalent DC voltages	В
Type K thermocouple	-270 °C to -200 °C -200 °C to -50 °C -50 °C to 1370 °C	Excluding CJC 1.2 °C 0.10 °C 0.10 °C	Including CJC 2.6 °C 0.17 °C 0.13 °C		
Type J thermocouple	-210 °C to -200 °C -200 °C to 0 °C 0 °C to 1200 °C	0.10 °C 0.10 °C 0.10 °C	0.23 °C 0.21 °C 0.14 °C		
Type E thermocouple	-270 °C to -200 °C -200 °C to 0 °C 0 °C to 1000 °C	0.54 °C 0.10 °C 0.10 °C	0.85 °C 0.13 °C 0.12 °C		
Type T thermocouple	-270 °C to -200 °C -200 °C to 0 °C 0 °C to 400 °C	0.86 °C 0.10 °C 0.10 °C	7.2 °C 0.47 °C 0.22 °C		
Type N thermocouple	-270 °C to -200 °C -200 °C to -100 °C -100 °C to 800 °C 800 °C to 1300 °C	2.9 °C 0.11 °C 0.10 °C 0.10 °C	3.6 °C 0.16 °C 0.13 °C 0.12 °C		
Type R thermocouple	-50 °C to 0 °C 0 °C to 150 °C 150 °C to 400 °C 400 °C to 1768 °C	0.24 °C 0.17 °C 0.12 °C 0.10 °C	0.29 °C 0.22 °C 0.17 °C 0.15 °C		
Type S thermocouple	-50 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 1768 °C	0.22 °C 0.17 °C 0.13 °C 0.11 °C	0.26 °C 0.21 °C 0.17 °C 0.16 °C		
Resistance thermometer (Pt 100)	-200 °C to 0 °C 0 °C to 850 °C	0.0086 °C 0.020 °C		By measurement of equivalent DC resistance	

UKAS CALIBRATION 0654	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Calibration, Maintenance and Repair Ltd Issue No: 044 Issue date: 07 May 2025			
Accredited to ISO/IEC 17025:2017				
	Calibration performed by the Or	ganisation at the locations	s specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks	Location Code
Values and uncertainties listed b method used is by direct compar	elow are applicable for the calibration ison against laboratory references ur	n of both measuring instrum nless otherwise stated in the	ents and for instruments with an remarks column.	n output. The
Calibration of Oscilloscopes				В
Voltage deflection	1 mV to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 6 V	Q [0.29 %, 47 μV] Q [0.29 %, 48 μV] Q [0.29 %, 92 μV] Q [0.29 %, 0.21 mV]	Into 50 Ω	
	1 mV to 25 mV 25 mV to 110 mV 110 mV to 2.2 V 2.2 V to 11 V 11 V to 130 V	Q [0.12 %, 46 V] Q [0.12 %, 46 µV] Q [0.12 %, 51 µV] Q [0.12 %, 0.22 mV] Q [0.12 %, 0.47 mV]	Into 1 MΩ	
Time markers	2 ns to 500 ns 500 ns to 500 µs 500 µs to 50 ms 50 ms to 5 s	Q [0.082 %, 0.12 ps] Q [0.082 %, 0.12 ns] Q [0.058 %, 12 ns] Q [0.59 %, 1.2 μs]		
PRESSURE			Methods consistent with	В
Pneumatic Pressure (gauge)			Absolute pressures can be generated within these gauge pressure ranges. This will attract an additional uncertainty of 10 Pa.	
Calibration of pressure indicating instruments and gauges	- 95 kPa to -10 kPa -10 kPa to -3.5 kPa -3.5 kPa to -1.5 kPa 1.5 kPa to 3.5 kPa 3.5 kPa to 10 kPa 10 kPa to 20 kPa 20 kPa to 2.5 MPa 2.5 MPa to 14 MPa	0.0087 % 0.0097 % Q [0.014 %, 0.50 Pa] Q [0.013 %, 0.50 Pa] 0.0078 % 0.0066 % 0.0063 % 0.0079 %	Sensors with an electrical output can be calibrated	
Hydraulic Pressure (gauge) Calibration of pressure indicating instruments and gauges	0.6 MPa to 6 MPa 6 MPa to 120 MPa	0.011 % 0.010 %		
	E	END		



## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where *q* is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: Q[a, b] =  $[a^2 + b^2]^{1/2}$