

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0661 Accredited to ISO/IEC 17025:2017	S.J. Electronics Ltd trading as John Godrich Issue No: 018 Issue date: 02 June 2023	
	3 Vernon Court Henson Way Kettering Northants NN16 8PX	Contact: Jonathan Kail Tel: +44 (0)1584 873153 Fax: +44 (0)1584 872424 E-Mail: JonKail@johngodrich.co.uk Website: www.johngodrich.co.uk

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Humidity, Temperature, Textile Testing Machines	Any



0661
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

S.J. Electronics Ltd trading as John Godrich
Issue No: 018 Issue date: 02 June 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
RELATIVE HUMIDITY				Any
Humidity controlled chambers (inclusive of associated indicators, controllers and recorders, all with sensors within the specified parameters and ranges)	20 °C to 70 °C 10 %rh to 98 %rh	3.7 %rh	Calibration by comparison with a reference hygrometer and temperature probes	
	70 °C to 80 °C 5 %rh to 65 %rh 65 %rh to 90 %rh	3.7 %rh 5.0 %rh		
TEMPERATURE				Any
Temperature controlled fridges/refrigerators, freezers, ovens and environmental chambers (inclusive of associated indicators, controllers and recorders, all with sensors within the specified parameters and ranges)	- 70 °C to + 300 °C	0.50 °C	Calibration by comparison with a reference temperature probes	
			Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping	
TEXTILE TESTING MACHINES				Any
Wascators			Calibration by comparison with a reference instruments	
Temperature	25 °C to 90 °C	0.60 °C		
Linear	0 mm to 200 mm	1.0 mm		
END				



0661
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

S.J. Electronics Ltd trading as John Godrich
Issue No: 018 Issue date: 02 June 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$