


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0665</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Health and Safety Executive - Science Engineering and Analysis Division</p> <p>Issue No: 062 Issue date: 07 May 2026</p>	
	<p>Testing & Monitoring Service Health and Safety Executive Science Division Harpur Hill Buxton Derbyshire SK17 9JN</p>	<p>Contact: Darren Musgrove Tel: +44 (0)203 028 1923 E-Mail: darren.musgrove@hse.gov.uk Website: www.hsl.gov.uk</p>

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ASBESTOS FIBRES IN AIR	<p><u>Health and Hygiene</u></p> <p>Fibre counting</p>	<p>Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021</p> <p>Documented In-House Method FT OP101 , Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248</p>
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<p>Identification of:</p> <ul style="list-style-type: none"> Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite 	<p>Documented In-House Method No. OP106 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248</p>



0665
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Health and Safety Executive - Science Engineering and Analysis
Division**

Issue No: 062 Issue date: 07 May 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
FIBRES	<u>Health and Hygiene</u> (cont'd)	International and National Standards and In-House Methods as listed using Transmission Electron Microscopy (TEM) with Energy Dispersive X-Ray Analysis, (EDXA), Electron Diffraction (ED) and Image Analysis Techniques
Asbestos Fibres in Air	Identification, Counting, Dimensional and Mass Measurement	1) ISO 10312:2019 using TEM, ED and EDXA 2) ISO 13794:2019 using TEM, ED and EDXA
Asbestos Fibres in Liquids	Identification, Counting, Dimensional and Mass Measurement	1) ISO 13794:2019 using TEM, ED and EDXA 2) Method 100.1, EPA-600/4-83-043:2019 using TEM, ED and EDXA. 3) Method 100.2, EPA-600/R-94/134:2002 using TEM, ED and EDXA
Asbestos Fibres in Solids, including: <ul style="list-style-type: none"> • Minerals • Soils, and • Asbestos Products 	Identification, Counting, Dimensional and Mass Measurement	Documented in-house method FT OP254 based on ISO 13794:2019 using TEM, ED and EDXA
Inorganic and other Fibres on Prepared Samples from Air, Liquids and Solids	Identification, Counting, Dimensional and Mass Measurement	Documented in-house method FT OP280 based on ISO 13794:2019 using TEM, ED and EDXA
Asbestos and other Fibres in Lung Tissue Sections, Blocks and BAL Fluids	Identification, Counting and Dimensional Measurement	Documented in-house method FT OP257 based on ISO 13794:2019 using TEM, ED and EDXA



0665
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Health and Safety Executive - Science Engineering and Analysis
Division**

Issue No: 062 Issue date: 07 May 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS	<u>Health and Hygiene (cont'd)</u>	
Urine	Measurement of: Aluminium Antimony Cadmium Chromium Cobalt Copper Lead Mercury Nickel Thallium	Documented in-house method BM OP01 by ICP/MS
	Measurement of: Creatinine	Documented in-house method No SOP 23 using a colorimetric method
	Measurement of: Isocyanate metabolites (HDA, TDA, IPDA AND MDA)	Documented in house method OT OP33 by GC-MS
Blood	Measurement of: Cadmium Lead	Documented in-house method BM OP01 using ICP/MS

END