# **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( <i>k</i> = 2)	Remarks
AIR VELOCITY			
Calibration of Anemometers and Pitot Tubes by comparison	0.1 m/s to 1 m/s 1 m/s to 2 m/s 2 m/s to 5 m/s 5 m/s to 10 m/s 10 m/s to 20 m/s 20 m/s to 30 m/s 30 m/s to 40 m/s 40 m/s to 50 m/s	0.10 % + 0.035 m/s 0.10 % + 0.030 m/s 0.10 % + 0.13 m/s 0.25 % + 0.13 m/s 0.25 % + 0.11 m/s 0.25 % + 0.14 m/s 0.25 % + 0.24 m/s 0.25 % + 0.35 m/s	Usable wind tunnel diameter 100 mm. Calibrations by comparison with a master instrument within a characterised airstream. CMC above 30 m/s may. increase for UUT > 3" diameter
PRESSURE			Methods consistent with EURAMET CG17.
Gas pressure (absolute)			
Calibration of pressure indicating instruments and gauges	3.5 kPa to 131 kPa	0.0050 % + 10 Pa	NOTE: Absolute pressure calibration can be carried out using associated barometric pressure measurement. The uncertainty values given below will be increased by 11 Pa.
Gas pressure (gauge)			,
Calibration of pressure indicating instruments and gauges	-95 kPa to 0 Pa 0 Pa to 500 Pa 500 Pa to 2.5 kPa 2.5 kPa to 8 kPa 8 kPa to 24 kPa 24 kPa to 500 kPa 500 kPa to 4 MPa	0.0050 % + 100 Pa 0.030 % + 0.15 Pa 0.030 % + 0.15 Pa 0.20 % + 0.95 Pa 0.20 % + 2.7 Pa 0.0050 % + 100 Pa 0.0050 % + 1.0 kPa	Calibration of devices with an electrical output may be undertaken
Hydraulic pressure (gauge)			
Calibration of pressure indicating instruments and gauges	0.7 MPa to 2.8 MPa 2.8 MPa to 28 MPa	0.030 % + 1.7 kPa 0.020 % + 1.7 kPa	

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK Airflow Measurements Ltd Issue No: 032 Issue date: 14 July 2025				
CALIBRATION 0690 Accredited to ISO/IEC 17025:2017					
	Calibration performed at main address only				
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks		
ELECTRICAL CALIBRATION: Electrical values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison against laboratory standards unless otherwise stated in the remarks column.					
DC VOLTAGE					
	0 V to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1 kV	6.5 μV 41 μV/V + 2.0 μV 36 μV/V + 53 μV 47 μV/V + 740 μV 47 μV/V + 10 mV	Direct measurement capability suitable for calibrating the outputs of devices submitted for calibration.		
	0 V to 100 mV 100 mV to 1 V 1 V to 20 V 20 V to 200 V 200 V to 1 kV	0.0020 % + 5.0 μV 0.0010 % + 15 μV 0.0010 % + 120 μV 0.0010 % + 7.0 mV 0.0010 % + 24 mV	These are source values available for the calibration of measuring equipment.		
DC RESISTANCE					
	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 10 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ	$\begin{array}{c} 130 \; \mu\Omega/\Omega \; + \; 4.7 \; m\Omega \\ 130 \; \mu\Omega/\Omega \; + \; 12 \; m\Omega \\ 200 \; \mu\Omega/\Omega \; + \; 120 \; m\Omega \\ 130 \; \mu\Omega/\Omega \; + \; 1.3 \; m\Omega \\ 110 \; \mu\Omega/\Omega \; + \; 13 \; \Omega \\ 950 \; \mu\Omega/\Omega \; + \; 37 \; \Omega \\ 0.90 \; \% \; + \; 1.1 \; k \; \Omega \end{array}$	Direct measurement capability suitable for calibrating the outputs of devices submitted for calibration.		
	100 μΩ 1 mΩ 5 mΩ 10 mΩ 100 mΩ 1 Ω	0.10 μΩ 0.31 μΩ 1.3 μΩ 1.0 μΩ 20 μΩ 430 μΩ	These are source values available for the calibration of measuring equipment using standard resistors		
	1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ	0.016 % 0.0051 % 0.0027 % 0.0020 % 0.0020 % 0.016 % 0.0031 % 0.022 % 0.062 %	These are source values available for the calibration of measuring equipment using multi- function calibrator		
DC CURRENT					
	0 A to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 2 A	510 μΑ/Α + 240 nA 510 μΑ/Α + 6.5 μΑ 1000 μΑ/Α + 160 μΑ 1200 μΑ/Α + 680 μΑ	Direct measurement capability suitable for calibrating the outputs of devices submitted for calibration.		

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 0690 Accredited to	lss	Airflow Measurements Ltd Issue No: 032 Issue date: 14 July 2025		
150/IEC 17025:2017	Calibration perform	ned at main address only		
Measured Quantity	Range	Expanded Measurement	Remarks	

Measured Quantity Instrument or Gauge	Range	Uncertainty $(k = 2)$	Remarks
DC CURRENT (continued)			
	0 A to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	0.0020 % + 25 nA 0.0020 % + 100 nA 0.0010 % + 5.0 μA 0.0010 % + 10 μA 0.0050 % + 40 μA 0.010 % + 600 μA	These are source values available for the calibration of measuring equipment.
	10 A to 100 A 100 A to 1000 A	0.080 % + 1.0 mA 0.080 % + 1.0 mA	For calibration of current clamps using a muilti turn coil.
AC VOLTAGE			
	5 mV to 100 mV 100 mV to 1 V 45 Hz to 1 kHz 100 mV to 1 V 1kHz to 30 kHz 1 V to 10 V 45 Hz to 1 kH 1 V to 10 V 1 kHz to 30 kH 10 V to 100 V 45 Hz to 1 kHz 10 V to 100 V 1 kHz to 30 kHz 100 V to 750 V 45 Hz to 10 kHz	660 μV/V + 42 μV 620 μV/V + 310 μV 620 μV/V + 320 μV/ 400 μV/V + 37 mV 1200 μV/V + 5.1 mV 480 μV/V + 260 mV 620 μV/V + 58 mV 660 ppm + 240 mV	Direct measurement capability suitable for calibrating the outputs of devices submitted for calibration.
	45 Hz to 5 kHz: 1 mV to 200 mV 200 mV to 2 V 10 V to 20 V 20 V to 200 V 200 V to 1000 V	0.0050 % + 70 μV 0.010 % + 820 μV 0.010 % + 12 mV 0.010 % + 120 mV 0.010 % + 450 mV	These are source values available for the calibration of measuring equipment.
AC RESISTANCE	At 50 Hz: 0 Ω to 0.3 Ω 0.3 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ	0.10 % + 20 mΩ 0.10 % + 25 mΩ 0.20 % + 200 mΩ 0.30 % + 250 mΩ	Calibration of loop impedance meters.
	45 Hz to 1 kHz:		
	1 μA to 1 A 1 A to 2 A	0.26 % + 400 μA 0.47 % + 12 mA	Direct measurement capability suitable for calibrating the outputs of devices submitted for calibration.
	45 Hz to 1 kHz: 25 μA to 200 μA 200 μA to 2 mA 2 mA to 20 mA 20 mA to 20 mA 200 mA to 2 A 2 A to 20 A	0.020 % + 500 nA 0.015 % + 4.0 μA 0.010 % + 40 μA 0.015 % + 500 μA 0.015 % + 5.0 mA 0.020 % + 15 mA	These are source values available for the calibration of measuring equipment.
	45 Hz to 400 Hz 10 A to 100 A 100 A to 1000 A	0.10 % + 15 mA 0.10 % + 120 mA	For calibration of current clamps.



Instrument or Gauge	Range		Uncertainty $(k = 2)$		Remarks
RCD TRIP CURRENT	0 mA to 300 mA		0.30 % + 5.0 mA		
RCD TRIP TIME	0 s to 400 ms 400 ms to 1 s		0.30 % + 5.0 ms 0.30 % + 10 ms		
FREQUENCY	1 Hz to 100 kHz 100 kHz, 1 MHz, 5 MHz and 10 MHz		100 μHz 1.0 part in 10 <sup>7</sup>		
TIME INTERVAL	1 µs 5 µs 20 µs 500 µs 1 ms 5 ms 10 ms 50 ms 100 ms		0.020 µs 0.10 µs 11 µs 0.020 ms 0.10 ms 0.20 ms 1.1 ms 2.1 ms		Repetitive signals suitable for calibrating oscilloscope time bases.
Elapsed time, single event					
Manually triggered devices Electronically triggered devices	0 s to 24 Hours 1 s to 24 hours		3.5 μs/s + 2.0 ms 3.5 μs/s	5	
ELECTRICAL CALIBRATION OF TEMPERATURE INDICATORS AND CALIBRATORS	Туре	Range °C	Without Cold Junction Compensation	Including Cold Junction Compensation	Methods consistent with Euramet CG-11
Thermocouple indicators	т	-240 to -100 -100 to 0 0 to 400	0.70 °C 0.70 °C 0.60 °C	0.80 °C 0.80 °C 0.80 °C	
	К	-200 to -100 -100 to 0 0 to 1370	0.80 °C 0.70 °C 0.60 °C	0.90 °C 0.80 °C 0.80 °C	
	S	0 to 1700	0.80 °C	0.90 °C	
	R	-50 to 0 0 to 1700	1.2 °C 0.70 °C	1.3 °C 0.90 °C	
	Ν	-250 to 0 0 to 1300	0.62 °C 0.60 °C	0.80 °C 0.80 °C	
	J	-180 to 0 0 to 700	0.60 °C 0.60 °C	0.80 °C 0.80 °C	
	E B	0 to 800 0 to 1800	0.60 °C 0.80 °C	0.80 °C 0.90 °C	
RTD indicators	Pt 100	-200 to +800	0.050 °C		
Reference junction compensation	15 °C to 25 °C		0.50 °C		
END					



# Schedule of Accreditation issued by

**United Kingdom Accreditation Service** 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

## **Airflow Measurements Ltd**

Issue No: 032 Issue date: 14 July 2025

Accredited to ISO/IEC 17025:2017

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC). which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant nonrepeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: Q[a, b] =  $[a^2 + b^2]^{1/2}$