

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|   |   |  |
|---|---|--|
|  <p><b>UKAS</b><br/>CALIBRATION</p> <p>0764</p> <p>Accredited to<br/><b>ISO/IEC 17025:2017</b></p> | <p><b>Klipspringer Limited</b></p> <p>Issue No: 025 Issue date: 24 July 2025</p>  |  |
|   | <p><b>Foxtail House</b><br/><b>Foxtail Road</b><br/><b>Ransomes Industrial Estate</b><br/><b>Ipswich</b><br/><b>IP3 9RX</b></p> | <p><b>Contact: Mr T A Carlyon</b><br/><b>Tel: +44 (0)1473 741500</b><br/><b>Fax: +44 (0)1473 747200</b><br/><b>E-Mail: service@klipspringer.co.uk</b><br/><b>Website: www.klipspringer.com</b></p> |
| <p><b>Calibration performed at the above address only</b></p>   |   |  |

### Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge                        | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ )                              | Remarks   |
|---|--|--|---|
| <b>ELECTRICAL</b>   |  |  | Comparison with reference instrument                      |
| Temperature indicators, calibration by electrical simulation    |  |  |   |
| Base metal thermocouples<br>Type K and T                        | -20 °C to +250 °C  | 0.30 °C  | including cold junction compensation                      |
| Resistance thermometers<br>(Pt 1000)                            | -18 °C to + 120 °C   | 0.20 °C  |   |
| Temperature simulators,<br>calibration by electrical simulation |  |  |   |
| Base metal thermocouples<br>Type K and T                        | -20 °C to +250 °C  | 0.30 °C  | including cold junction compensation                      |
| Resistance thermometers<br>(Pt 1000)                            | -40 °C to +250 °C  | 0.20 °C  |   |
| DC Resistance   |  |  |   |
| Measurement   | 200 $\Omega$ to 2 k $\Omega$<br>2 k $\Omega$ to 20 k $\Omega$<br>20 k $\Omega$ to 1 M $\Omega$ | 0.024 % + 15 m $\Omega$<br>0.026 % + 0.11 $\Omega$<br>0.049 % + 5.8 $\Omega$ |   |
| <b>TEMPERATURE</b>  |  |  |   |
| Temperature loggers and indicators with sensors                 | -30 °C to +150 °C  | 0.080 °C   | Comparison in liquid bath                                 |
| Resistance thermometers<br>(Pt 1000)                            | -30 °C to +150 °C  | 0.20 °C  |   |
| Liquid in glass thermometers                                    | -30 °C to +150 °C  | 0.13 °C + 1/5 scale division   | Partial immersion thermometers                            |
| Block calibrators   | -18 °C to +100 °C  | 0.27 °C  | Calibration of ECMP4-8-12 Lazaport Multipoint Calibrators |
|   | +75 °C   | 0.35 °C  | ECMP1 Lazaport Singlepoint                                |



0764  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Klipspringer Limited**  
Issue No: 025 Issue date: 24 July 2025

Calibration performed at main address only

| Measured Quantity<br>Instrument or Gauge                   | Range                              | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  |
|--|------------------------------------|---|--|
| HUMIDITY<br><br>Relative humidity indicators and recorders | 10 %rh to 85 %rh<br>20 °C to 23 °C | 2.2 %rh   | Comparison in humidity and temperature controlled chamber<br><br>Relative humidity can also be reported in terms of $a_w$ (water activity) |
| Temperature probes incorporated into humidity indicators   | 20 °C to 23 °C                     | 0.31 °C   |  |
| END  |                                    |   |  |



0764  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Klipspringer Limited**  
Issue No: 025 Issue date: 24 July 2025

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$