Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

0766

Accredited to ISO/IEC 17025:2017

Michell Instruments Limited

Issue No: 021 **Issue date:** 04 August 2026

Unit 1A
Crompton Fields
Crompton Way
Crawley
West Sussex
RH10 9EE

Contact: Mr Gareth Thompson
Tel: +44 (0)1293 571000
Fax: +44 (0)1293 571008
E-Mail: uk.service.instruments@dwyeromega.com
Website: www.dwyeromega.com/en-us/

Calibration performed at the above address only

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
DEW POINT			
Using a chilled mirror hygrometer			
Dew-point	- 60 ° to - 40 °C - 40 °C to 60 °C 60 °C to 70 °C 70 °C to 95 °C	0.14 °C 0.11 °C 0.12 °C 0.25 °C	Calibrations can be performed on instruments with an electrical output
TEMPERATURE			
Temperature sensors incorporated in humidity instruments	-60 °C to 0 °C 0 °C to 70 °C 70 °C to 150 °C	0.08 °C to 0.06 °C 0.05 °C 0.07 °C to 0.16 °C	Calibrations in air
Platinum Resistance Thermometers and Sensors with Indicators	-70 °C to 200 °C	0.019 °C	Calibration in fluid
	0.01 °C	0.003 °C	Triple point of water
RELATIVE HUMIDITY			
Relative humidity	Example conditions	Corresponding to above dew-point and temperature uncertainties	By comparison with dew-point hygrometer and Platinum Resistance Thermometers
	At 0 °C: 2 %rh to 10 %rh 10 %rh to 98 %rh	 0.20 %rh 0.20 %rh to 1.0 %rh	
	At 23 °C: 2 %rh to 10 %rh 10 %rh to 98 %rh	 0.20 %rh 0.20 %rh to 0.80 %rh	
	At 70 °C: 2 %rh to 98 %rh	 0.20 %rh to 0.60 %rh	
	At 95 °C 30 %rh to 98 %rh	 0.40 %rh to 1.0 %rh	

0766
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Michell Instruments Limited
Issue No: 021 **Issue date:** 04 August 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
RELATIVE HUMIDITY (continued)			
Using unsaturated salts capsules	*Temperature range 23 °C ± 2 °C:*		
	5 %rh	0.50 %rh	
	10 %rh	0.50 %rh	
	11 %rh	0.50 %rh	
	20 %rh	0.60 %rh	
	35 %rh	0.70 %rh	
	50 %rh	1.1 %rh	
	65 %rh	1.0 %rh	
	75 %rh	1.3 %rh	
	80 %rh	1.3 %rh	
	95 %rh	1.3 %rh	
END			

Michell Instruments Limited

Issue No: 021 **Issue date:** 04 August 2026

0766
Accredited to
ISO/IEC 17025:2017

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.
Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$