


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |   |  |
|--|---|--|
| <br><b>0773</b><br><br>Accredited to<br><b>ISO/IEC 17025:2017</b> | <b>TORUS Measurement Systems Ltd</b><br><br>Issue No: 030    Issue date: 24 June 2025 |  |
|  | <b>Nedge Hill Science Park</b><br>Telford<br>Shropshire<br>TF3 3AJ                    | <b>Contact: Ms Denise Ball</b><br>Tel: +44 (0)1952 210020<br>E-Mail: <a href="mailto:denise.ball@industrialphysics.com">denise.ball@industrialphysics.com</a><br>Website: <a href="http://www.industrialphysics.com">www.industrialphysics.com</a> |

**Calibration performed by the Organisations at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details  |                                     | Activity    | Location code |
|---|-------------------------------------|-------------|---------------|
| <b>Address</b><br>Nedge Hill Science Park<br>Telford<br>Shropshire<br>TF3 3AJ | <b>Local contact</b><br>Denise Ball | Dimensional | A             |

#### Site activities performed away from the locations listed above:

| Location details   |                                     | Activity    | Location code |
|--|-------------------------------------|-------------|---------------|
| The location must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer | <b>Local contact</b><br>Denise Ball | Dimensional | B             |



0773  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TORUS Measurement Systems Ltd**  
**Issue No: 030 Issue date: 24 June 2025**

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge  | Range  | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code    |
|---|--|--|---|---------------------|
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED  |  |  |   |                     |
| <b>LENGTH</b><br><br>Plain plug gauges (parallel) cylindrical setting standards and rollers.<br><br>Plain plug gauges, taper.<br><br>Receiver and position gauges, jigs and fixtures. | 1 to 3 diameter<br>3 to 100<br>100 to 150<br>150 to 300<br><br>0 to 100<br><br>0 to 1500 x 900 x 800   | 0.79 on diameter<br>1.0<br>1.5<br>2.3<br><br>3.3<br><br>2.2 + (2.3 x length in m)<br>3.5 + (2.2 x length in m) | By comparison with reference standards<br><br>By comparison with reference standards<br><br>Using a CMM<br>Using first principles<br><br>Documented in-house methods using first principles and/or a CMM. The stated uncertainties are based on the calibration of an ideal item under ideal conditions. The reported uncertainties will reflect the item calibrated and the conditions at the time of the calibration. | A<br><br>A<br><br>A |
| Beverage Masters  | 1 to 200 height / diameter / dome depth<br><br>1 to 4 flange width<br><br>2 to 4 curl opening<br><br>10 to 90 curl diameter<br><br>20 to 70 inner curl dia<br><br>0.1 to 1.5 thickness<br><br>10 to 70 reform diameter<br><br>1 to 5 reform depth<br><br>30 to 100 master punch ironing ring diameter<br><br>10 to 70 internal reform diameter | 3.3<br><br>2.2<br><br>2.4<br><br>1.6<br><br>5.0<br><br>0.80<br><br>5.9<br><br>12<br><br>2.4<br><br>12          | Documented in-house methods using the following equipment: Reference standards, electronic probe, surface table, vertical and horizontal length measuring machines, probe comparator, dedicated fixtures, sine table, master scanner and standard metrology holding equipment.  | A                   |



0773  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TORUS Measurement Systems Ltd**  
Issue No: 030 Issue date: 24 June 2025

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range   | Expanded<br>Measurement<br>Uncertainty ( $k = 2$ )                                   | Remarks   | Location<br>Code   |
|--|---|--|---|--------------------|
| Beverage Masters Length<br>Blocks  | 300   | 1.4 + (6.6 x length in m)  | Comparison to reference<br>standards  | A                  |
| Height gauges (complex)<br>Tesa electronic height<br>gauges, for E, B and R            | 0 to 1000<br>Length measurement error, E<br><br>Bidirectional length<br>measurement error, B  | 1.6 + (7.0 x length in m)<br><br>1.6 + (7.0 x length in m)                           | BS EN ISO13225:2012<br>comparison to length<br>standards  | A & B<br><br>A & B |
| Height gauge setting masters.  | 25  | 2.0  | Comparison to reference<br>standards  | A                  |
| Performance verification of co-<br>ordinate measuring machines<br>(CMM's)              | Length measurement –<br>$E_L$<br>0 to 1500<br>(longest diagonal)<br><br>Single stylus probing test:<br>$P_{Form.Sph.1x25:SS:Tact}$<br>$P_{Size.Sph.1x25:SS:Tact}$                             | 0.78 + (1.1 x length in m)<br><br>0.37<br>0.46                                       | ISO 10360-2:2009<br>using end standards<br><br>ISO 10360-5:2020<br>Using a 10 to 50 mm test<br>sphere   | A & B              |
| Performance verification of<br>non-cartesian co-ordinate<br>measuring machines (CMM's) | Single stylus probing test<br>$P_{FTU}$<br><br>Length measurement -<br>$E_{Bi}$<br>$E_{Uni}$<br>0 to 1545 diameter<br><br>Probing measurement -<br>$P_{Size.SPH.1x25}$<br>$P_{Form.SPH.1x25}$ | 0.17<br><br>2.9 + (1.3 x length in m)<br>2.9 + (1.3 x length in m)<br><br>2.2<br>2.2 | ISO 10360-5:2010<br>(withdrawn) Using a 10 to<br>50 mm test sphere<br><br>ISO 10360-12:2016<br>using end standards.<br>Test value uncertainties<br>based on ISO/TS<br>23165:2006<br><br>ISO 10360-12:2016<br>using a 10 mm to 51 mm<br>diameter test sphere.<br>Test value uncertainties<br>based on ISO/TS<br>17865:2016 | A & B              |
|  | Articulated location<br>measurement -<br>$L_{Dia.5x5:Art}$  | 2.2  | ISO 10360-12:2016<br>using a 10 mm to 51 mm<br>diameter test sphere.<br>Test value uncertainties<br>based on ISO/TS<br>17865:2016   |                    |
| END  |   |  |   |                    |



0773

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TORUS Measurement Systems Ltd**  
**Issue No: 030 Issue date: 24 June 2025**

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$