

issued by

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Eurotherm Limited

Issue No: 020 Issue date: 13 September 2023

**Faraday Close
Durrington
Worthing
West Sussex
BN13 3PL**

Contact: Mr Warwick Vercoe
Tel: +44 (0)1903 268500
E-Mail: warwick.vercoe@watlow.com
Website: www.eurotherm.com

Locations covered by the organisation and their relevant activities

| Location details | | Activity | Location code |
|--|--|------------|----------------------|
| Address Faraday Close Durrington Worthing West Sussex BN13 3PL | Local contact Mr Warwick Vercoe | Electrical | Permanent Laboratory |

| Location details | | Activity | Location code |
|---|--|---|------------------|
| Customers' sites or premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Local contact Mr Warwick Vercoe | Electrical Pressure Temperature Time | Site Calibration |



0778
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Eurotherm Limited
Issue No: 020 Issue date: 13 September 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|----------------------------------|--|---|------------------|
| ELECTRICAL | | | | |
| All electrical measurements are carried out using the method of direct comparison or transfer to laboratory reference standards unless otherwise determined in the remarks column. | | | | |
| Temperature indicators and simulators, calibration by electrical simulation | | | | Site Calibration |
| Base metal thermocouples | | | including cold junction compensation | |
| Type T | -100 °C to +400 °C | 0.40 °C | | |
| Type N | -100 °C to +1300 °C | 0.50 °C | | |
| Type K | -100 °C to +1300 °C | 0.40 °C | | |
| Type J | -100 °C to +1200 °C | 0.40 °C | | |
| Type E | -200 °C to +1000 °C | 0.50 °C | | |
| Noble metal thermocouples | | | including cold junction compensation | |
| Type S | 400 °C to 1760 °C | 0.95 °C | | |
| Type R | 400 °C to 1760 °C | 0.95 °C | | |
| Type B | 500 °C to 1820 °C | 1.5 °C | | |
| Pt 100 indicators | -100 °C to +400 °C | 0.17 °C | | |
| Pt100 simulators | -100 °C to +400 °C | 0.17 °C | | |
| DC VOLTAGE | | | | |
| Generation | 0 mV to 100 mV 100 mV to 10 V | 12.5 µV 2.5 mV | | |
| Measurement | 0 mV to 100 mV 100 mV to 10 V | 20 µV 2.5 mV | | |
| DC CURRENT | | | | |
| Generation | 0 mA to 20 mA | 10 µA | | |
| Measurement | 0 mA to 20 mA | 10 µA | | |
| DC RESISTANCE | | | | |
| Generation | 0 Ω to 1 kΩ | 0.35 Ω | | |
| TIME | | | | |
| Timers | 10 s to 200 hr | 3.0 s | Including absolute time | |



0778
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Eurotherm Limited
Issue No: 020 Issue date: 13 September 2023

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|---|--|---|--|------------------|
| <p>PRESSURE</p> <p><u>Gas Pressure (Gauge)</u></p> <p>Calibration of pressure indicating instruments and gauges</p> <p>TEMPERATURE</p> <p>Electronic thermometers with sensors</p> <p>Platinum resistance thermometers 3-wire and 4-wire</p> <p>Type T Thermocouples</p> <p>Temperature controlled ovens, furnaces and autoclaves</p> | <p>-80 kPa to +2.07 MPa</p> <p>-40 °C to -20 °C -20 °C to +100 °C 100 °C to 140 °C</p> <p>-40 °C to -20 °C -20 °C to +100 °C 100 °C to 140 °C</p> <p>-40 °C to -20 °C -20 °C to +100 °C 100 °C to 140 °C</p> <p>20 °C to 650 °C 650 °C to 1100 °C 1100 °C to 1200 °C</p> | <p>2.5 kPa</p> <p>0.40 °C 0.40 °C 0.40 °C</p> <p>0.65 °C 0.45 °C 0.65 °C</p> <p>0.75 °C 0.60 °C 0.75 °C</p> <p>2.1 °C 2.7 °C 4.2 °C</p> | <p>Methods consistent with EURAMET CG17</p> <p>Results may be expressed in other units of pressure as required</p> <p>Calibration of pressure measuring devices with an electrical output may be undertaken</p> <p>Calibration by comparison with reference thermometers</p> <p>Calibration of temperature measuring devices with an electrical output may be undertaken</p> <p>Multiple point measurements Time dependent profiling spatial, survey(ing) or mapping</p> | Site Calibration |



0778
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Eurotherm Limited
Issue No: 020 Issue date: 13 September 2023

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|-------------------------------|--|---------|----------------------|
| DC VOLTAGE | | | | Permanent Laboratory |
| Generate | | | | |
| | 0 V to 100 mV | 4.0 μ V | | |
| | 100 mV to 300 mV | 7.0 μ V | | |
| | 300 mV to 1 V | 12.5 μ V | | |
| | 1 V to 3 V | 31 μ V | | |
| | 3 V to 10 V | 131 μ V | | |
| | 10 V to 30 V | 350 μ V | | |
| | 30 V to 100 V | 3.0 mV | | |
| Measure | | | | |
| | 0 V to 100 mV | 3.1 μ V | | |
| | 100 mV to 1 V | 29 μ V | | |
| | 1 V to 10 V | 288 μ V | | |
| | 10 V to 100 V | 3.1 mV | | |
| DC CURRENT | | | | |
| Generate | | | | |
| | 0 A to 100 μ A | 30 nA | | |
| | 100 μ A to 300 μ A | 60 nA | | |
| | 300 μ A to 1 mA | 200 nA | | |
| | 1 mA to 3 mA | 300 nA | | |
| | 3 mA to 10 mA | 1.2 μ A | | |
| | 10 mA to 30 mA | 2.7 μ A | | |
| | 30 mA to 100 mA | 12 μ A | | |
| | 100 mA to 300 mA | 27 μ A | | |
| | 300 mA to 1 A | 195 μ A | | |
| DC CURRENT (cont'd) | | | | |
| Measure | | | | |
| | 0 A to 100 μ A | 4.0 nA | | |
| | 100 μ A to 1 mA | 110 nA | | |
| | 1 mA to 10 mA | 1.8 μ A | | |
| | 10 mA to 100 mA | 8.0 μ A | | |
| | 100 mA to 1 A | 219 μ A | | |
| Resistance | | | | |
| Generate | | | | |
| | 0 Ω to 10 Ω | 10 m Ω | | |
| | 10 Ω to 100 Ω | 20 m Ω | | |
| | 100 Ω to 1 k Ω | 40 m Ω | | |
| Measure | | | | |
| | 0 Ω to 1 Ω | 3.1 μ Ω | | |
| | 1 Ω to 10 Ω | 3.2 μ Ω | | |
| | 10 Ω to 100 Ω | 5.0 μ Ω | | |
| | 100 Ω to 1 k Ω | 35 μ Ω | | |
| | 1 k Ω to 10 k Ω | 340 μ Ω | | |



0778
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Eurotherm Limited
Issue No: 020 Issue date: 13 September 2023

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|--|--|---|---------|----------------------|
| Temperature simulation | | | | Permanent Laboratory |
| Source and Measure | | | | |
| Thermocouples | | | | |
| Type T | -200 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C | 0.49 °C 0.19 °C 0.13 °C 0.11 °C | | |
| Type S | 0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1400 °C 1400 °C to 1767 °C | 0.37 °C 0.29 °C 0.29 °C 0.36 °C | | |
| Type R | 0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1400 °C 1400 °C to 1767 °C | 0.45 °C 0.28 °C 0.26 °C 0.32 °C | | |
| Type N | -200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C | 0.32 °C 0.18 °C 0.15 °C 0.14 °C 0.21 °C | | |
| Thermocouples (cont'd) | | | | |
| Type K | -200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C | 0.26 °C 0.15 °C 0.13 °C 0.21 °C 0.32 °C | | |
| Type J | -210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C | 0.21 °C 0.13 °C 0.11 °C 0.14 °C 0.18 °C | | |
| Type E | -250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C | 0.40 °C 0.13 °C 0.11 °C 0.13 °C 0.17 °C | | |
| Type B | 600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1550 °C 1550 °C to 1820 °C | 0.35 °C 0.28 °C 0.24 °C 0.21 °C | | |
| Cold junction measurement in thermocouple generate mode | | 125 m°C | | |



0778
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Eurotherm Limited
Issue No: 020 **Issue date:** 13 September 2023

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | |
|--|---|---|---------|--|
| PT 100 Source | -200 °C to -80 °C -80 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 400 °C 400 °C to 630 °C 630 °C to 800 °C | 40 m°C 40 m°C 55 m°C 70 m°C 80 m°C 95 m°C 0.18 °C | | |
| PT 100 Measure | -200 °C to 0 °C 0 °C to 232 °C 232 °C to 660 °C | 5.0 m°C 7.0 m°C 14 m°C | | |
| END | | | | |



0778
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Eurotherm Limited
Issue No: 020 Issue date: 13 September 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$