


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

 <p><b>0801</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<p align="center"><b>University of Salford</b></p> <p align="center"><b>Issue No: 021      Issue date: 13 July 2023</b></p> <table border="1"> <tr> <td data-bbox="403 432 826 678"> <p><b>School of Science, Engineering &amp; Environment</b></p> <p><b>Newton Building G31</b></p> <p><b>The Crescent</b></p> <p><b>Salford</b></p> <p><b>M5 4BR</b></p> </td><td data-bbox="826 432 1477 678"> <p><b>Contact: Ms C Lomax</b></p> <p><b>Tel: +44 (0)161 295 3030 / (0)161 295 3319</b></p> <p><b>E-Mail: C.Lomax1@salford.ac.uk</b></p> <p><b>Website:</b> <a href="https://acoustictesting.salford.ac.uk/services/acoustic-calibration/">https://acoustictesting.salford.ac.uk/services/acoustic-calibration/</a></p> </td></tr> </table>	<p><b>School of Science, Engineering &amp; Environment</b></p> <p><b>Newton Building G31</b></p> <p><b>The Crescent</b></p> <p><b>Salford</b></p> <p><b>M5 4BR</b></p>	<p><b>Contact: Ms C Lomax</b></p> <p><b>Tel: +44 (0)161 295 3030 / (0)161 295 3319</b></p> <p><b>E-Mail: C.Lomax1@salford.ac.uk</b></p> <p><b>Website:</b> <a href="https://acoustictesting.salford.ac.uk/services/acoustic-calibration/">https://acoustictesting.salford.ac.uk/services/acoustic-calibration/</a></p>
<p><b>School of Science, Engineering &amp; Environment</b></p> <p><b>Newton Building G31</b></p> <p><b>The Crescent</b></p> <p><b>Salford</b></p> <p><b>M5 4BR</b></p>	<p><b>Contact: Ms C Lomax</b></p> <p><b>Tel: +44 (0)161 295 3030 / (0)161 295 3319</b></p> <p><b>E-Mail: C.Lomax1@salford.ac.uk</b></p> <p><b>Website:</b> <a href="https://acoustictesting.salford.ac.uk/services/acoustic-calibration/">https://acoustictesting.salford.ac.uk/services/acoustic-calibration/</a></p>		
<p align="center"><b>Calibration performed at the above address only</b></p>			

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>Sound calibrators &amp; pistonphones</b>			
Sound pressure level of pistonphones	110 dB to 125 dB	0.07 dB	Verification of Pistonphones according to IEC 60942:2003 (Withdrawn) Annex B Using IEC Type WS2P or LS2P microphones.
Frequency of pistonphones	250 Hz 1000 Hz	0.012 % 0.012 %	
Sound pressure level of sound calibrators	85 dB to 125 dB	0.07 dB	
Frequency of sound calibrators	250 Hz 1000 Hz	0.012 % of reading 0.012 % of reading	Verification of Sound Calibrators according to IEC 60942:2003 (Withdrawn) Annex B Using IEC Type WS2P or LS2P microphones.
Distortion		0.22 %	
<b>Microphones</b>			
Pressure sensitivity of IEC type LS2/WS2 microphones	250 Hz 1000 Hz	0.09 dB	Using a sound calibrator
Pressure sensitivity of microphones to BS EN 61094-5:2016	20 Hz to <50 Hz 50 Hz to <100 Hz 100 Hz to 10 kHz >10 kHz to 12.5 kHz >12.5 kHz to 16 kHz >16 kHz to 20 kHz	0.11 dB 0.09 dB 0.08 dB 0.09 dB 0.12 dB 0.15 dB	Using an LS2 reference microphone
<b>Sound level meters</b>			
Sound Level Meters	BS 7580:Part 1:1997 (Withdrawn)	See remarks	Verification of Type 1 Sound level meters originally manufactured in accordance with BS EN 60651:1994 (Withdrawn) BS EN 60804:1994 (Withdrawn) and for which required correction factors are known and agreed.



**0801**  
Accredited to  
**ISO/IEC 17025:2017**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

**University of Salford**  
**Issue No: 021    Issue date: 13 July 2023**

**Calibration performed at main address only**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>Sound level meters</b> (cont'd)			
Sound Level Meters	BS EN 61672-3: 2006 (Withdrawn) as modified by UKAS TPS 49 Edition 2: June 2009	See remarks	Class 1 & 2 Sound Level Meters originally manufactured in accordance with IEC 61672- 1:2002 (Withdrawn) and for which required correction factors are known and agreed
Filters - sound level meter based			
Octave band	16Hz to 16kHz	0.18 dB	Filters originally manufactured in accordance with IEC 61260:1995 (Withdrawn)(BS EN 61260:1996 (Withdrawn)) in combination with a sound level meter
One-third octave band	20 Hz to 20 kHz		
<b>Reverberation time</b>	125 Hz 1 kHz 10 kHz  Decay times: 300 ms 1 s 3 s  10 s	0.01 s @ 125 Hz, 1 & 10 kHz 0.02 s @ 125 Hz, 1 & 10 kHz 0.03 s @ 125 Hz & 1 kHz, 0.05 s @ 10 kHz 0.12 s @ 125 Hz & 1 kHz 0.17 s @ 10 kHz	Verification of specific RT modules on sound level meters using a transfer standard (electrical signals with a continuous decay)
<b>Tapping Machines</b>	In support of BS EN ISO 140-7: 1998 (Withdrawn); BS EN ISO 10140-5:2010+A1:2014 (Withdrawn), BS EN ISO 10140- 5:2021, BS EN ISO 16283- 2:2015 (Withdrawn) and BS EN ISO 16283-2:2020 Annex A  Velocity 0.70 m/s to 1.00 m/s  Mass 480 g to 520 g  Time 50 ms to 150 ms  Diameter 25 mm to 35 mm  Radius of curvature 300 mm to 700 mm  Angle of fall 0° to 0.6°	0.03 m/s  0.20 g  0.20 ms  0.01 mm  10 mm  0.10 °	Suitable to support the verification of Tapping Machines
<b>END</b>			



0801  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

**University of Salford**  
**Issue No:** 021    **Issue date:** 13 July 2023

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$