


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

 <p><b>0801</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<h3>University of Salford</h3> <p><b>Issue No: 024      Issue date: 12 May 2026</b></p>	
	<p><b>School of Science, Engineering &amp; Environment</b>  <b>Newton Building G31</b>  <b>The Crescent</b>  <b>Salford</b>  <b>M5 4BR</b></p>	<p><b>Contact: Ms C Lomax</b>  <b>Tel: +44 (0)161 295 3030 / (0)161 295 3319</b>  <b>E-Mail: C.Lomax1@salford.ac.uk</b>  <b>Website:</b>  <a href="https://acoustictesting.salford.ac.uk/services/acoustic-calibration/">https://acoustictesting.salford.ac.uk/services/acoustic-calibration/</a></p>
<p><b>Calibration performed at the above address only</b></p>		

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>Sound calibrators &amp; pistonphones</b>			
Sound pressure level of pistonphones	110 dB to 125 dB	0.07 dB	Suitable to support the verification of Pistonphones according to IEC 60942:2003 (Withdrawn) Annex B; IEC 60942:2018 Annex B Using IEC Type WS2P or LS2P microphones.
Frequency of pistonphones	250 Hz 1000 Hz	0.012 % of reading 0.012 % of reading	
Distortion	0.01 % to 4 %	0.22 %	
Sound pressure level of sound calibrators	85 dB to 125 dB	0.07 dB	Suitable to support the verification of Sound Calibrators according to IEC 60942:2003 (Withdrawn)Annex B,IEC 60942:2018 Annex B Using IEC Type WS2P or LS2P microphones.
Frequency of sound calibrators	250 Hz 1000 Hz	0.012 % of reading 0.012 % of reading	
Distortion	0.01 % to 4 %	0.22 %	
<b>Microphones</b>			
Pressure sensitivity of IEC type LS2/WS2 microphones to BS EN 61094-5:2016	20 Hz to 50 Hz	0.11 dB	Using an LS2 reference microphone in comparison coupler
	50 Hz to 100 Hz 100 Hz to 2 kHz	0.08 dB 0.06 dB	
	2 kHz to 10 kHz	0.08 dB	Using LS2 reference microphone in a reverberation field
	10 kHz to 12.5 kHz	0.09 dB	
	12.5 kHz to 16 kHz	0.12 dB	
	16 kHz to 20 kHz	0.15 dB	
	2 kHz to 10 kHz	0.06 dB	LS2 or WS2 converted to LS2 configuration calibrated using LS2 reference microphone in comparison coupler
	10 kHz to 12.5 kHz	0.08 dB	
	12.5 kHz to 16 kHz	0.11 dB	
	16 kHz to 20 kHz	0.13 dB	



0801  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

**University of Salford**  
**Issue No: 024 Issue date: 12 May 2026**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>Microphones cont.</b>			
Pressure sensitivity of microphones to BS EN 61094-5:2016	20 Hz to 50 Hz 50 Hz to 100 Hz 100 Hz to 10 kHz 10 kHz to 12.5 kHz 12.5 kHz to 16 kHz 16 kHz to 20 kHz	0.11 dB 0.09 dB 0.08 dB 0.09 dB 0.12 dB 0.15 dB	Using an LS2 reference microphone
Pressure sensitivity of IEC type LS2/WS2 microphones	250 Hz 1000 Hz	0.09 dB	Using a sound calibrator
<b>Sound Level Meters</b>			
	BS 7580:Part 1:1997 (Withdrawn)	See remarks	Suitable to support the verification of Type 1 Sound level meters originally manufactured in accordance with BS EN 60651:1994 (Withdrawn) BS EN 60804:1994 (Withdrawn) and for which required correction factors are known and agreed.
	BS EN 61672-3: 2006 (Withdrawn) as modified by UKAS TPS 49 Edition 3: December 2022	See remarks	Suitable to support the verification of Class 1 & 2 Sound Level Meters originally manufactured in accordance with IEC 61672-1:2002 (Withdrawn) and for which required correction factors are known and agreed
	BS EN 61672-3: 2013	See remarks	Suitable to support the verification of Class 1 & 2 Sound Level Meters originally manufactured in accordance with IEC 61672-1:2013 and for which required correction factors are known and agreed
<b>Filters - sound level meter based</b>			
Octave band	16Hz to 16kHz	0.18 dB	Filters originally manufactured in accordance with IEC 61260:1995 (Withdrawn)(BS EN 61260:1996 (Withdrawn)) in combination with a sound level meter
One-third octave band	20 Hz to 20 kHz		
<b>Reverberation Time</b>	<b>Calibration and Measurement Capability in seconds of decay times over the frequency ranges shown</b>		
	<b>125 Hz</b>	<b>1 kHz</b>	<b>10 kHz</b>
300 ms	0.01 s	0.01 s	0.01 s
1 s	0.02 s	0.02 s	0.02 s
3 s	0.03 s	0.03 s	0.05 s
10 s	0.12 s	0.12 s	0.17 s
			(electrical signals with a continuous decay) Suitable to support the verification of specific RT modules on sound level meters using a transfer standard



0801  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

**University of Salford**  
**Issue No: 024 Issue date: 12 May 2026**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>Tapping Machines</b> Velocity: Mass: Time: Diameter: Radius of Curvature: Angle of fall:	0.70 m/s to 1.00 m/s 480 g – 520 g 50 ms – 150 ms 25 mm to 35 mm 300 mm to 700 mm 0° to 0.6°	0.01 m/s 0.20 g 0.20 ms 0.01 mm 10 mm 0.10°	In support of BS EN ISO 140-7: 1998 (Withdrawn); BS EN ISO 10140-5:2010+A1:2014 (Withdrawn), BS EN ISO 10140-5:2021, BS EN ISO 16283-2:2015 (Withdrawn) and BS EN ISO 16283-2:2020 Annex A to support the verification of Tapping Machines
END			



0801  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames. TW18 3HR

**University of Salford**  
**Issue No: 024 Issue date: 12 May 2026**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$