

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|----------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------|
|  0816 Accredited to ISO/IEC 17025:2017 | Caltech Service Co Ltd Issue No: 021 Issue date: 18 May 2023 | |
| | Unit 5, Poplar Drive Poplar Industrial Estate Moor Lane Witton Birmingham B6 7AD | Contact: Mr P Westwood Tel: +44 (0)121 331 4445 Fax: +44 (0)121 331 4447 E-Mail: Paul@caltechservice.co.uk |

Calibration performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

| Location details | Activity | Location code | |
|-------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------|----------------------------------------------------------------------|
| Address Unit 5, Poplar Drive Poplar Industrial Estate Moor Lane Witton Birmingham B6 7AD | Local contact Mr P Westwood Tel: +44 (0)121 331 4445 Fax: +44 (0)121 331 4447 Email: Paul@caltechservice.co.uk | Electrical Temperature Pressure | Calibrations performed at Permanent Laboratory are denoted: P |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code | |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------|------------------------|------------------------------------------------------------------------------------------|
| The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Contact Mr P Westwood Tel: +44 (0)121 331 4445 Fax: +44 (0)121 331 4447 Email: Paul@caltechservice.co.uk | Electrical Pressure | Calibrations performed away from the Permanent Laboratory, on site are denoted: S |



0816
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Caltech Service Co Ltd
Issue No: 021 Issue date: 18 May 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------|------------------------------------------------------|------------------|
| ELECTRICAL MEASUREMENT | | | | |
| DC Voltage | 0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V | 0.031 mV 0.10 mV 0.60 V 6.0 mV 0.50 mV | Measurement with reference sources and meters. | P & S |
| AC Voltage | 10 Hz to 50 kHz 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000V | 0.050 mV 0.60 mV 6.2 mV 60 mV 1.0 V | | |
| DC Current | 0 mA to 100 mA 100 mA to 1 A 1 A to 10 A | 0.12 mA 0.34 mA 3.0 mA | | |
| AC Current | 1 kHz 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A | 0.13 mA 0.30 mA 3.0 mA | | |
| DC Resistance | 0 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 100 M Ω | 0.055 Ω 0.21 Ω 14 Ω 28 Ω 290 Ω | | |



0816
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Caltech Service Co Ltd
Issue No: 021 Issue date: 18 May 2023

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------|-----------------------------------------------------------------------------|------------------|
| ELECTRICAL GENERATION | | | | P & S |
| DC Voltage | 0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V | 0.07 mV 0.11 mV 1.0 mV 6.0 mV 70 mV | These values can be generated for the calibration of measuring instruments. | |
| AC Voltage | 10 Hz to 60 kHz 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000V | 0.070 mV 0.90 mV 7.0 mV 60 mV 1.0 V | | |
| DC Current | 0 mA to 100 mA 100 mA to 1 A 1 A to 10 A | 0.12 mA 0.38 mA 3.10 mA | | |
| AC Current | 1 kHz 10 mA to 100 mA 100 mA to 1 A 1 A to 10 A | 0.17 mA 0.32 mA 3.0 mA | | |
| DC Resistance | 0 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 100 M Ω | 0.12 Ω 0.27 Ω 1.2 Ω 5.4 Ω 7.0 Ω | | |



0816
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Caltech Service Co Ltd
Issue No: 021 Issue date: 18 May 2023

Calibration performed by the Organisation at the locations specified

| Measured Quantity Instrument or Gauge | Range | Expanded Measurement Uncertainty ($k = 2$) | Remarks | Location Code |
|-----------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------|----------------------------------------------------|-----------------------------------------------------------------------|------------------|
| Electrical calibration of temperature indicators, controllers and recorders for the following sensors: | | | | |
| Noble metal thermocouples | -50 °C to 0 °C 0 °C to 1768 °C | 1.3 °C (lab) 0.9 °C (lab) 1.2 °C (site) | with cold junction compensation | P & S |
| Base metal thermocouples | -200 °C to +1372 °C | 0.5 °C (lab) 1.0 °C (site) | with cold junction compensation | P & S |
| Resistance thermometers (PT100) | -200 °C to +800 °C | 0.3 °C | | P & S |
| PRESSURE | | | Methods consistent with EURAMET CG17. | |
| Hydraulic Pressure (Gauge) Calibration of pressure indicating instruments and gauges | 0.55 to 5.5 MPa 5.5 to 55 MPa | 0.023 % 0.023 % | | P |
| | 0 to 4 MPa 4 MPa to 25 MPa 25 to 60 MPa | 29 kPa 61 kPa 140 kPa | Including crack pressure calibration of pressure relief valves. | S |
| Gas Pressure (Gauge) Calibration of pressure indicating instruments and gauges | 0 to 200 kPa 200 kPa to 700 kPa 700 kPa to 2 MPa 2 MPa to 4 MPa | 0.30 kPa 0.48 kPa 2.2 kPa 1.8 kPa | including crack pressure calibration of pressure relief valves. | P & S |
| TEMPERATURE | | | | |
| Resistance thermometers and digital indicators with probes | 0 °C to 100 °C 100 °C to 800 °C | 0.32 °C 1.0 °C | With dry and liquid media dependent on range | P |
| Thermocouples, noble metal | 100 °C to 500 °C 500 °C to 1000 °C 1000 °C to 1200 °C | 1.7 °C 1.6 °C 1.9 °C | Calibration within horizontal furnace | P |
| Thermocouples, base metal | 100 °C to 1000 °C 1000 °C to 1200 °C | 2.5 °C 2.7 °C | Calibration within Horizontal furnace | P |
| END | | | | |



0816
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Caltech Service Co Ltd
Issue No: 021 Issue date: 18 May 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$