

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0819

Accredited to  
ISO/IEC 17025:2017

### TDC Calibration - a trading division of TDC Aberdeen Ltd

Issue No: 027 Issue date: 13 October 2025

23 Kirkton Avenue  
Pitmedden Industrial Estate  
Dyce  
Aberdeen  
AB21 0BF

Contact: Miss Alexis Ross  
Tel: +44 (0)1224 710033  
Fax: +44 (0)1224 710099  
E-Mail: alexis.ross@tdcaberdeen.co.uk  
Website: www.tdcaberdeen.co.uk

Calibration performed at the above address only

#### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>ELECTRICAL MEASUREMENTS</b>			
<b>DC VOLTAGE</b>			
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	10 $\mu$ V/V + 1.6 $\mu$ V 7.5 $\mu$ V/V + 1.7 $\mu$ V 7.4 $\mu$ V/V + 2.8 $\mu$ V 13 $\mu$ V/V + 50 $\mu$ V 15 $\mu$ V/V + 500 $\mu$ V	Using digital multimeter.
	1 kV to 10 kV 10 kV to 40 kV	1.5 % 2.0 %	Using high voltage divider.
<b>DC CURRENT</b>			
Measurement	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A	120 $\mu$ A/A + 500 pA 120 $\mu$ A/A + 5.0 nA 120 $\mu$ A/A + 50 nA 120 $\mu$ A/A + 1.2 $\mu$ A 240 $\mu$ A/A + 25 $\mu$ A	
Generation	1 A to 3A 3 A to 11 A 11 A to 20.5 A	450 $\mu$ A/A + 50 $\mu$ A 600 $\mu$ A/A + 600 $\mu$ A 0.12 % + 870 $\mu$ A	Using multifunction calibrator.
	20 A to 150 A 150 A 1000 A	0.30 % + 0.020 A 0.30 % + 0.060 A	For calibration of current clamps and similar devices, using multi-turn coil arrangement.
<b>DC RESISTANCE</b>			
Measurement	0 $\Omega$ to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 200 $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$ 200 M $\Omega$ to 1 G $\Omega$	23 $\mu$ $\Omega$ / $\Omega$ + 23 $\mu$ $\Omega$ 15 $\mu$ $\Omega$ / $\Omega$ + 70 $\mu$ $\Omega$ 12 $\mu$ $\Omega$ / $\Omega$ + 700 $\mu$ $\Omega$ 12 $\mu$ $\Omega$ / $\Omega$ + 7 m $\Omega$ 15 $\mu$ $\Omega$ / $\Omega$ + 70 m $\Omega$ 26 $\mu$ $\Omega$ / $\Omega$ + 1.6 $\Omega$ 47 $\mu$ $\Omega$ / $\Omega$ + 10 $\Omega$ 390 $\mu$ $\Omega$ / $\Omega$ + 10 k $\Omega$ 0.37 % + 100 k $\Omega$	





0819  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TDC Calibration - a trading division of TDC Aberdeen Ltd**  
**Issue No: 027 Issue date: 13 October 2025**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
ELECTRICAL MEASUREMENTS (cont'd)			
AC CURRENT			
Measurement	2 $\mu$ A to 200 $\mu$ A 50 Hz to 1 kHz 1 kHz to 5 kHz	380 $\mu$ A/A + 20 nA 700 $\mu$ A/A + 20 nA	Using digital multimeter.
	200 $\mu$ A to 2 mA 50 Hz to 1 kHz 1 kHz to 5 kHz	380 $\mu$ A/A + 0.23 $\mu$ A 700 $\mu$ A/A + 0.23 $\mu$ A	
	2 mA to 20 mA 50 Hz to 1 kHz 1 kHz to 5 kHz	360 $\mu$ A/A + 2.3 $\mu$ A 700 $\mu$ A/A + 2.3 $\mu$ A	
	20 mA to 200 mA 40 Hz to 1 kHz 1 kHz to 5 kHz	380 $\mu$ A/A + 23 $\mu$ A 700 $\mu$ A/A + 23 $\mu$ A	
	200 mA to 1 A 40 Hz to 1 kHz 1 kHz to 5 kHz	800 $\mu$ A/A + 460 $\mu$ A 0.24% + 920 $\mu$ A	
Generation	45 Hz to 1 kHz 1 A to 3 A	0.070 % + 120 $\mu$ A	Using multifunction calibrator.
	45 Hz to 400 Hz 3 A to 11 A 11 A to 20.5 A	0.080 % + 2.3 mA 0.14 % + 6.0 mA	
	45 Hz to 400 Hz 20 A to 150 A 150 A to 1000 A	0.33 % + 0.020 A 0.35 % + 0.10 A	For calibration of current clamps and similar devices, using multi- turn coil arrangement.
CAPACITANCE	3 nF to 10 $\mu$ F	0.60 % + 50 pF	Simulated capacitance suitable for the calibration of capacitance meters.
FREQUENCY	1 Hz to 3 GHz	7.5 in $10^{12}$	Also suitable for calibration of timer counters averaged over a minimum of 1000 seconds gate time.
TIME INTERVAL			
Elapsed time	0 s to 1000 s 0 s to 1000 s	5.0 $\mu$ s 0.30 s	Electronically triggered events Manually triggered events



0819  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TDC Calibration - a trading division of TDC Aberdeen Ltd**  
**Issue No: 027 Issue date: 13 October 2025**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
OSCILLOSCOPE CALIBRATION			
Vertical deflection	0 V to $\pm 33$ V	0.30 % + 120 $\mu$ V	Using square wave signals.
Horizontal deflection	2 ns to 1 ms 1 ms to 100 ms 100 ms to 5 s	0.0030 % 0.015 % 0.60 %	Using pulse signals.
Bandwidth	50 kHz to 300 MHz 10 mV to 100 mV 100 mV to 5.5 V	3.3 % 2.0 %	Using levelled sine wave generator. The results are reported in terms of the frequency at which the -3 dB point is obtained.
Additional measurements in support of 17 <sup>th</sup> Edition type test equipment.			
RCD Current	At 50 Hz: 10 mA to 2.5 A	1.2 % + 1.6 mA	
RCD Time	20 ms to 5 s	1.2 % + 1.2 ms	
Earth bond & Loop resistance	20 m $\Omega$ 50 m $\Omega$ 100 m $\Omega$ 350 m $\Omega$ 500 m $\Omega$ 900 m $\Omega$ 1.7 $\Omega$ 4.7 $\Omega$ 9 $\Omega$ 17 $\Omega$ 47 $\Omega$ 90 $\Omega$ 170 $\Omega$ 470 $\Omega$ 900 $\Omega$ 1.7 k $\Omega$	14.6 m $\Omega$ 15.8 m $\Omega$ 15.7 m $\Omega$ 19.5 m $\Omega$ 20.4 m $\Omega$ 30.3 m $\Omega$ 1.2 $\Omega$ 1.2 $\Omega$ 1.2 $\Omega$ 1.2 $\Omega$ 1.4 $\Omega$ 1.4 $\Omega$ 1.8 $\Omega$ 3.6 $\Omega$ 6.8 $\Omega$ 13.4 $\Omega$	
Earth bond current	10 mA to 4 A 4 A to 40 A	1.2 % + 14 mA 1.2 % + 140 mA	
1 m $\Omega$ to 960 m $\Omega$	10 mA to 0.3 A 0.3 A to 3 A	0.3 % + 1.3 mA 0.3 % + 100 mA	
1.7 $\Omega$ to 1.7 k $\Omega$			
ELECTRICAL SIMULATION OF TEMPERATURE			
Resistance thermometer (Pt 100)	-200 $^{\circ}$ C to +800 $^{\circ}$ C	0.080 $^{\circ}$ C	By resistance simulation.



0819  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TDC Calibration - a trading division of TDC Aberdeen Ltd**  
**Issue No: 027 Issue date: 13 October 2025**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
Temperature simulators, calibration by electrical simulation			
Type K thermocouple	-200 °C to +1372 °C	0.13 °C	Excluding cold junction compensation.
Type K thermocouple	-200 °C to +1372 °C	0.05 °C	Including cold junction compensation.
Cold junction compensation	At ambient temperature of 20 °C ± 3 °C	0.15 °C	Other thermocouple types may be calibrated, the temperature equivalent to $\mu\text{V}$ will be calculated in accordance with the prevailing BS EN 60584-1 tables
Temperature indicators, calibration by electrical simulation		As per simulators plus resolution and stability of the device being calibrated.	
<b>PRESSURE MEASUREMENTS</b>			Methods consistent with EURAMET CG17.
Gas pressure (gauge)			
Calibration of pressure indicating instruments and gauges	- 80 kPa to -10 kPa -10 to -1.5 kPa -1.5 kPa to +1.5 kPa 1.5 kPa to 10 kPa 10 kPa to 100 kPa 100 kPa to 2.5 MPa	0.012 % 0.013 % 15 Pa 0.010 % 0.011 % 0.006 5 %	Pressure instruments with an electrical output may be calibrated.
Gas pressure (absolute)			
Calibration of pressure indicating instruments and gauges	20 kPa to 90 kPa 90 kPa to 200 kPa 200 kPa to 2.6 MPa	0.012 + 59 Pa 0.011 % + 59 Pa 0.006 5 % + 59 Pa	
Hydraulic pressure (gauge)			
Calibration of pressure indicating instruments and gauges	600 kPa to 6.05 MPa 6.0 MPa to 120 MPa 120 MPa to 206.9 MPa	0.012 % + 104 Pa 0.013 % + 104 Pa 0.012 % + 58 kPa	
<b>TEMPERATURE MEASUREMENTS</b>			Calibration performed within Liquid Baths
Temperature indicators and recorders, with temperature sensors	-25 °C to 0 °C 0 °C to 100 °C 100 °C to 140 °C 140 °C to 400 °C 400 °C to 650 °C	0.18 °C 0.10 °C 0.070 °C 0.36 °C 0.28 °C	For immersion depth of 130 mm; shorter probes can be calibrated but with increased uncertainty.
END			



0819  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TDC Calibration - a trading division of TDC Aberdeen Ltd**  
**Issue No: 027 Issue date: 13 October 2025**

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$