


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>10456</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>CMR Controls Ltd</b>	
	Issue No: 011    Issue date: 21 December 2025	
	22 Repton Court Repton Close Basildon SS13 1LN United Kingdom	Contact: Ian Doswell Tel: +44 (0) 1268 287222 E-Mail: sales@cmr.uk Website: www.cmr.co.uk
<b>Calibration performed at the above address only</b>		

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
PRESSURE  Calibration of digital pressure indicating instruments and sensors	0 Pa to 100 Pa 100 Pa to 1 kPa 1 kPa to 3 kPa 3 kPa to 10 kPa	Q[0.070 %, 0.35 Pa] Q[0.090 %, 0.27 Pa] Q[0.055 %, 0.42 Pa] Q[0.030 %, 2.2 Pa]	Methods consistent with EURAMET CG17.  Calibration using primary pressure transfer standards. n.b. The laboratory only calibrates instruments manufactured by CMR Controls Ltd.  Calibration of pressure measuring devices with an electrical output may be undertaken.
END			



10456  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**CMR Controls Ltd**  
**Issue No: 011 Issue date: 21 December 2025**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$