Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

10460

Accredited to ISO/IEC 17025:2017

Nature's Consultancy Ltd

Issue No: 007 **Issue date:** 08 July 2026

Suite S7
Bates Business Centre
Church Road
Harold Wood
Romford
Essex
RM3 0JF

Contact: Mr F Kadia
Tel: +44 (0) 1708 985746
E-Mail: falgun@naturesconsultancy.co.uk
Website:www.naturesconsultancy.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Suite S7 Bates Business Centre Church Road Harold Wood Romford Essex RM3 0JF	**Local contact** Mr F Kadia Tel: +44 (0) 1708 985746 Email: falgun@naturesconsultancy.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Heath and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C

10460

Accredited to ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Nature's Consultancy Ltd

Issue No: 007 **Issue date:** 08 July 2026

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
	Health and Hygiene	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
ASBESTOS FIBRES IN AIR	Sampling of air for fibre counting	Documented In-House Method DOC/002 based on HSG 248)	A, B, C
	Fibre counting	Documented In-House Method DOC/002, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method DOC/002, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
END			