

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 10460 Accredited to ISO/IEC 17025:2017	Nature's Consultancy Ltd Issue No: 006 Issue date: 20 February 2023	
	Suite S7 Bates Business Centre Church Road Harold Wood Romford Essex RM3 0JF	Contact: Mr F Kadia Tel: +44 (0) 7828 989817 E-Mail: falgun@naturesconsultancy.co.uk Website: www.naturesconsultancy.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Suite S7 Bates Business Centre Church Road Harold Wood Romford Essex RM3 0JF	Local contact Mr F Kadia Tel: +44 (0) 07828 989817 Email: falgun@naturesconsultancy.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Testing Laboratories	Health and Hygiene	C



10460

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Nature's Consultancy Ltd
Issue No: 006 Issue date: 20 February 2023

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method DOC/002 based on HSG 248)	A, B, C
	Fibre counting	Documented In-House Method DOC/002, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method DOC/002, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B

END