


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |  |  |
|--|--|--|
| <br><b>10547</b><br>Accredited to<br>ISO/IEC 17025:2017 | <b>Medipure Ltd</b>  |  |
|  | <b>Issue No:</b> 003 <b>Issue date:</b> 14 March 2024                                  |  |
|  | <b>Erw Las</b><br>Colomendy Industrial Estate<br>Denbigh<br>LL16 5TA<br>United Kingdom | <b>Contact:</b> James Daly<br><b>Tel:</b> +44 (0) 1745 814518<br><b>E-Mail:</b> enquiries@medipure.co.uk<br><b>Website:</b> www.medipure.co.uk |
| Testing performed at the above address only  |  |  |

### DETAIL OF ACCREDITATION

| Materials/Products tested  | Type of test/Properties measured/Range of measurement                              | Standard specifications/ Equipment/Techniques used  |
|--|--|---|
| WATERS -<br><br>Washer-disinfector final rinse water (mains fed, steam condensate and reverse osmosis) | <u>Microbiological Testing</u>   | Documented In -House methods  |
|  | Environmental Mycobacteria   | WI 038 & WI 040 (confirmation by acid fast staining) to meet the requirements of HTM 01-06 Part E 2016, WHTM01-06 and NICFPP  |
|  | <i>Pseudomonas aeruginosa</i>  | WI 020 using membrane filtration and Pseud CN agar with incubation at 35°C) to meet the requirements of HTM 01-06 Part E 2016, WHTM01-06 and NICFPP   |
| ENDOSCOPE SURROGATE DEVICES  | Total Aerobic Colony Count   | WI 020 using membrane filtration and Tryptone Soya Agar with incubation at 30°C) to meet the requirements of HTM 01-06 Part E 2016, WHTM01-06 and NICFPP  |
|  | Evaluation of Bacterial Contamination levels as a measure of disinfection efficacy | WI 004 & 005 assessment of contamination levels of <i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i> , <i>Mycobacterium terrae</i> and <i>Candida albicans</i> levels on PTFE tubing before and after a Washer Disinfector Cycle at customer premises, based on BS EN ISO 15883-4:2018 6.6.2 and Annex B, plus HTM 01-06 Part E 2016 |
| END  |  |   |