


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION</p> <p><b>10741</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<p><b>AVL United Kingdom Ltd</b></p> <p><b>Issue No: 012    Issue date: 14 October 2025</b></p>	
	<p><b>Avon House</b> <b>Hartlebury Trading Estate</b> <b>Hartlebury</b> <b>DY10 4JB</b> <b>United Kingdom</b></p>	<p><b>Contact: Mr Nick Barker</b> <b>Tel: +44 (0) 1299 254600</b> <b>E-Mail: nick.barker@avl.com</b> <b>Website: www.avl.com</b></p>
<p><b>Calibration performed by the Organisation at the locations specified</b></p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p>Unit 1 Viggen Way Ansty Business Park Coventry CV7 9RE</p> <p style="text-align: center;"><b>Local contact</b> As above</p>	<p>Pressure calibration Temperature calibration Humidity calibration Flow Calibration</p>	P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer</p>	<p>Pressure calibration Temperature calibration Humidity calibration Flow Calibration</p>	S



10741

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### AVL United Kingdom Ltd

Issue No: 012 Issue date: 14 October 2025

Calibration performed by the Organisation at the locations specified

#### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>PRESSURE</b>			Methods consistent with EURAMET CG17	
Calibration of differential pressure measuring instruments	0 kPa to $\pm 0.2$ kPa (0.2 kPa to 2 kPa) (2 kPa to 20 kPa)	Q [0.77 %, 1.8 Pa) Q [0.38 %, 1.1 Pa) Q [0.35 %, 3.1 Pa)		P&S
Calibration of gauge pressure measuring instruments	-100 kPa to 0 kPa 0 kPa to 350 kPa	Q [0.078 %, 120 Pa) Q [0.039 %, 120 Pa)		P&S
	-100 kPa to 0 kPa 0 kPa to 1 MPa	Q [0.053 %, 58 Pa) Q [0.052 %, 140 Pa)		
	-100 kPa to 0 kPa 0 kPa to 2 MPa	Q [0.058 %, 92 Pa) Q [0.043 %, 361 Pa)		
Gas pressure absolute: Calibration of pressure indicating instruments and gauges by comparison	60 kPa to 120 kPa	42 Pa		P & S
<b>FLOW</b>			By comparison with a reference meter.	
Mass flow controller, air	20 L/min to 200 L/min	0.58 %		S
Critical flow venturi, air	1 m <sup>3</sup> /min to 30 m <sup>3</sup> /min	0.54 %		S
Gas divider flow nozzle at 3.2 bar using N <sub>2</sub> gas	20 mL/min to 200 mL/min 100 mL/min to 1000 mL/min 1 L/min to 10 L/min	0.30 %		P
Mass flow meter, air	1 m <sup>3</sup> /min to 2.8 m <sup>3</sup> /min 2.8 m <sup>3</sup> /min to 14 m <sup>3</sup> /min 14 m <sup>3</sup> /min to 30.2 m <sup>3</sup> /min 30.2 m <sup>3</sup> /min to 48 m <sup>3</sup> /min	0.35 % 0.35 % 0.34 % 0.33 %	Volume flowrates referred to 0 °C, 101325 Pa	P
<b>TEMPERATURE</b>				
Rigid Bodied temperature sensor and indicator calibration	1 °C to 95 °C	0.20 °C	By comparison in a liquid bath	P & S
Electronic thermometers with indicators and data-loggers	5 °C to 40 °C	0.22 °C	By comparison in a dry block calibrator	P & S



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**AVL United Kingdom Ltd**

**Issue No: 012 Issue date: 14 October 2025**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
HUMIDITY Relative humidity	Performed at 20 °C 5 %rh to 10 %rh 10 %rh to 20 %rh 20 %rh to 30 %rh 30 %rh to 40 %rh 40 %rh to 50 %rh 50 %rh to 60 %rh 60 %rh to 80 %rh 80 %rh to 95 %rh	0.54 %rh to 0.72 %rh 0.72 %rh to 0.89 %rh 0.89 %rh to 0.93 %rh 0.93 %rh to 1.01 %rh 1.01 %rh to 1.09 %rh 1.09 %rh to 1.33 %rh 1.33 %rh to 1.76 %rh 1.76 %rh to 2.43 %rh	By comparison with a reference hygrometer in a portable chamber	P & S
END				



Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

#### AVL United Kingdom Ltd

Issue No: 012 Issue date: 14 October 2025

Calibration performed by the Organisation at the locations specified

### Appendix - Calibration and Measurement Capabilities

#### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$