

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



10883

Accredited to
ISO 15189:2012

UK Health Security Agency, An executive agency of the Department of Health and Social Care

Issue No: 006 Issue date: 17 April 2025

Leeds UKHSA Regional Laboratory
Centre for Laboratory Medicine
St James University Hospital
Beckett Street
Leeds
LS9 7TF

Contact: Hannah Armitage
Tel: +44 (0)113 392 8796
E-Mail: hannah.armitage1@nhs.net
Website: www.leedsth.nhs.uk/services/pathology

Testing performed at the above address only

Site activities performed away from the locations listed above:

Location details		Activity
Leeds General Infirmary Old Medical School Great George Street Leeds LS1 3EX		Specimen Reception



10883
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

UK Health Security Agency,
An executive agency of the Department of Health and Social Care

Issue No: 006 Issue date: 17 April 2025

Testing performed at main address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN TISSUES AND FLUIDS	<u>Microbiological examination activities for the purpose of clinical diagnosis</u>	In-house documented procedures incorporating manufacturer's instructions where relevant
Faeces		Manual culture of C difficile from faeces SOP: MIC-UKHSA-CDRN-EP-1
<i>Clostridioides difficile</i> Bacterial Isolates (cultured in-house as above or received from external sources)	<i>Clostridioides difficile</i> ribotyping	Bacterial DNA extraction using QIAamp 96 DNA QIAcube HT kit for QIAcube HT PCR amplification of 16-23S intergenic spacer regions. QIAgility liquid handling robot Verti 96 Well Thermo Cycler Genetic Analyser AB 3500 XL SOP: MIC-UKHSA-CDRN-EP-1
<i>Clostridioides difficile</i> Bacterial Isolates (cultured in-house as above or received from external sources)	<i>Clostridioides difficile</i> typing	Bacterial DNA extraction using QIAamp 96 DNA QIAcube HT kit for QIAcube HT PCR amplification of tandem repeat loci QIAgility liquid handling robot Verti 96 Well Thermo Cycler Multiple locus variable number tandem repeat analysis (MLVA) via Genetic Analyser AB 3500 XL SOP: MIC-UKHSA-CDRN-EP-2
END		