


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING 1102</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Government Analysts Laboratory</h3> <p>Issue No: 039    Issue date: 16 January 2026</p>	
	<p><b>Ballakermeen Road</b> <b>Douglas</b> <b>Isle of Man</b> <b>IM1 4BR</b></p>	<p><b>Contact: Stephen Desmond</b> <b>Tel: +44 (0)1624 642250</b> <b>Fax: +44 (0)1624 642222</b> <b>E-Mail: stephen.desmond@gov.im</b> <b>Website: www.gov.im/gov.labs</b></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ALCOHOLIC BEVERAGES	<u>Chemical Tests</u>  Alcoholic strength	Documented In-House Methods based on British and International Standards  Documented In-House Method No 048 based on BS 733:Part 2:1987 and ISO 3507:1999 using density, obscuration
DAIRY PRODUCTS	<u>Chemical Tests</u>	Documented In-House Methods based on British Standards
Cheese	Fat	Documented In-House Method No 005 based on ISO 1735:2004 using Acid hydrolysis/solvent extraction/gravimetric weighing
	Nitrogen	Documented In-House Method No 004 based on BS 770:Part 8:1987 using Kjeldahl digestion
Cream	Fat	Documented In-House Method No 002 based on IDF 16C:2008/ BS EN ISO 2450:2008 using Alkaline hydrolysis/solvent extraction/gravimetric weighing
	Total Solids	Documented In-House Method No 001 based on BS ISO 6731:2010 IDF 21:2010 using Oven drying followed by gravimetric weighing



1102  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Government Analysts Laboratory**  
**Issue No: 039 Issue date: 16 January 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
DAIRY PRODUCTS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on British Standards
Ice Cream	Fat	Documented In-House Method No 003 based on BS EN ISO 23318:2022 using Alkaline hydrolysis/solvent extraction/gravimetric weighing
	Total Solids	Documented In-House Method No 006 based on BS ISO 3728:2004 using Oven drying followed by gravimetric weighing
Milk	Acidity	Documented In-House Method No 020 based on BS 1741:Section 10.1:1989 using Titration
	Freezing Point Depression	Documented In-House Method No 019 based on ISO 5764:2009 [IDF 108:2009] using Cryoscope
	Antibiotics	Documented In-House Method No 016 based on Technical data sheet, Delvotest SP™ Gist-Brocades NV using Colorimetric bacteria inhibition
	Fat	Documented In-House Method No 010 based on IDF 1D:1996 using Alkaline hydrolysis/solvent extraction/gravimetric weighing
	Phosphatase	Documented In-House Method No M10 based on SI 1989/2383 Schedule 5, Part IV using Colorimetry
	Total Solids	Documented In-House Method No 009 based on BS ISO 6731:2010 and IDF 21B:1987 using Oven drying followed by gravimetric weighing



1102  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Government Analysts Laboratory**  
**Issue No: 039 Issue date: 16 January 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
MEAT AND MEAT PRODUCTS	<u>Chemical Tests (cont'd)</u>  Ash  Fat  Hydroxyproline  Moisture  Nitrogen	Documented In-House Methods based on British Standards  Documented In-House Method No 012 based on BS 4401-1:1998 using ashing followed by gravimetric weighing  Documented In-House Method No 014 based on BS 4401:Part 4:1970 using Acid hydrolysis/solvent extraction/ gravimetric weighing  Documented In-House Method No 081 based on BS 4401-11:1995 using microwave digestion followed by UV-visible spectrophotometry  Documented In-House Method No 013 based on BS 4401-3:1997 using Oven drying followed by gravimetric weighing  Documented In-House Method No 015 based on BS 4401:Part 2:1980 using Kjeldahl digestion, copper catalyst



1102  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Government Analysts Laboratory**  
**Issue No: 039 Issue date: 16 January 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
FOODS AND FOOD PRODUCTS Unspecified	<u>Microbiological Tests</u>  Enumeration:  <i>Escherichia coli</i> , presumptive  Coagulase Positive <i>Staphylococcus</i>  Total aerobic organisms at 30°C  Detection:  <i>Escherichia coli</i> O157:H7	Documented In-House Methods based on British Standards  Method No M3 based on BS ISO 7251:2005 + A1:2023 using Multiple Tube  Method No M13 using Staph Brilliance 24 and confirmation with Staphylase  Method No M1 based on BS EN ISO 4833-1:2013 + A1:2022 using Plate Count  Documented In-House Method No M25 based on AOAC Official Method No. 2000 using the 20 hour Reveal method, with confirmation by latex agglutination (presence/absence in 25 g)
DAIRY PRODUCTS Unspecified	Enumeration:  <i>Escherichia coli</i> , presumptive	Method No M6 based on ISO 7251:2005 using Multiple Tube
WATERS  Potable, Surface and Ground	<u>Chemical Tests and Physical Tests</u>  Alkalinity (total)  Ammonium  Chloride, Nitrate, Nitrite, Phosphate and Sulphate	Documented In-House Methods based on procedures in the HMSO Portable series "Method for the Examination of Waters and Associated Materials", specified by ISBN and by year of publication  In-House Method No 028 ISBN 011 751 601 5, 1981, using titrimetry  In-House Method No 031 ISBN 011 751 613 9, 1981, filtration/colorimetric  In-House Method No 229 ISBN 011 752 331 3, 1990 using Ion chromatography



1102  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Government Analysts Laboratory**  
**Issue No: 039 Issue date: 16 January 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS (cont'd) Potable, Surface and Ground (cont'd)	<u>Chemical Tests and Physical Tests (cont'd)</u>	Documented In-House Methods based on procedures in the HMSO Portable series "Method for the Examination of Waters and Associated Materials", specified by ISBN and by year of publication
	Free Chlorine	In-House Method No 040 using HACH test kit using Colorimetric
	Conductivity	In-House Method No 027 ISBN 011 751 428 4, 1978 using Potentiometric
	pH	In-House Method No 021 ISBN 011 752 084 5, 1988 using Potentiometric
Potable Water	Aluminium, Antimony, Arsenic, Barium, Boron, Cadmium, Chromium, Copper, Iron, Lead, Manganese, Nickel and Zinc	Documented In-House Method No 209 using ICP-MS based on instrument manufacturer procedures
Surface and Ground	Biological oxygen demand	1) In-House Method No 024 ISBN 011 752 212 0, 1989 using Winkler iodometric  2) In-House Method No 025 ISBN 011 752 212 0, 1989 oxygen probe with and without ATU suppression
	Fluoride	In-House Method No 229 using Ion Chromatography
	Suspended solids	In-House Method No 050 ISBN 011 751 787 9, 1984, filtration/gravimetric weighing



1102  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Government Analysts Laboratory**  
**Issue No: 039 Issue date: 16 January 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WASTE WATERS</p> <p>Untreated and Treated Trade and Domestic Effluents</p>	<p><u>Chemical Tests and Physical Tests</u></p> <p>Ammonium</p> <p>Biological Oxygen Demand</p> <p>Conductivity</p> <p>pH</p> <p>Suspended solids</p>	<p>Documented In-House Methods based on procedures in the HMSO Potable series "Methods for the Examination of Waters and Associated Materials", specified by ISBN and by year of publication</p> <p>In-House Method No 031 ISBN 011 751 613 9, 1981, filtration/colorimetric</p> <p>1) In-House Method No 024 ISBN 011 752 212 0, 1989 using Winkler iodometric</p> <p>2) In-House Method No 025 ISBN 011 752 212 0, 1989, using oxygen probe with/without ATU suppression</p> <p>In-House Method No 027 ISBN 011 751 428 4, 1978 using Potentiometric</p> <p>In-House Method No 021 ISBN 011 752 084 5, 1988 using Potentiometric</p> <p>In-House Method No 050 ISBN 011 751 787 9, 1984, filtration/gravimetric weighing</p>
<p>WATERS (Drinking including Bottled Waters, Saline and Recreational)</p>	<p><u>Microbiological Tests</u></p> <p>Enumeration:</p> <p>Coliforms, confirmed</p> <p>Enterococci</p>	<p>Documented In-House Methods based on The Microbiology of Drinking Water (MDW) Series as indicated</p> <p>Method No M 20 using Colilert™ using MPN based on MDW (2016) part 4D</p> <p>Documented in-house method No M22 using membrane filtration based on MDW (2012) Part 5</p>



1102  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Government Analysts Laboratory**  
**Issue No: 039 Issue date: 16 January 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS (Drinking including Bottled Waters, Saline and Recreational) – (cont'd)	<u>Microbiological Tests (cont'd)</u>  Enumeration (cont'd)  <i>Escherichia coli</i> , confirmed  Total aerobic organisms at 22 °C and 37 °C	Documented In-House Methods based on The Microbiology of Drinking Water (MDW) Series as indicated  Method No M20 using Colilert™ using MPN based on MDW (2016) Part 4D  Method No M9 using Plate count based on MDW (2020) Part 7A
WATERS (Drinking)	Detection and enumeration of <i>Legionella</i> spp.	Method No M24 based on ISO 11731:2017 using filtration with washing, using GVPC. Confirmation using BYCE [Matrix A; Procedure 5]
END		