


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>1181</b> Accredited to ISO/IEC 17025:2017</p>	<b>Northumbrian Water Ltd</b> <b>Issue No: 138 Issue date: 13 November 2025</b>	
	<b>Northumberland Dock Road</b> <b>Howdon</b> <b>Wallsend</b> <b>Tyne and Wear</b> <b>NE28 0QD</b>	<b>Contact: Mrs S Alexander</b> <b>Tel: +44 (0)870 320 4567</b> <b>Fax: +44 (0)191 296 8560</b> <b>E-Mail: sarah.alexander@nwl.co.uk</b> <b>Website: www.nwss-labs.co.uk</b>
<b>Testing performed by the Organisation at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Northumberland Dock Road Wallsend Tyne and Wear NE28 0QD  <b>Local contact</b> Ms Christine Pratt Mr Rick Kirkpatrick  Tel: +44 (0)191-296 8500 Email: Chris.Pratt2@nwl.co.uk / Rick.Kirkpatrick@nwl.co.uk	<b>Support Functions:</b> Quality System Quality Audit administration  <b>Sampling and Testing:</b> Environmental Analysis Operation of flexible scope	Howdon
<b>Address</b> Horsley Laboratory Water Lane Horsley Newcastle-upon-Tyne NE15 0PA  <b>Local contact</b> Mr Ian Lumsdon  Tel: +44 (0)191-296 8500 Email: Ian.Lumsdon@nwl.co.uk	Water Sampling	Horsley
<b>Address</b> Levenmouth Laboratory Levenmouth Waste Water Treatment Works Methilhaven Fife KY8 3LA  <b>Local contact</b> Mr Rick Kirkpatrick  Tel: +44 (0)191-296 8500 Email: Rick.Kirkpatrick@nwl.co.uk	Environmental Analysis	Levenmouth
<b>Address</b> Chelmsford Laboratory Hanningfield Treatment Works South Hanningfield Chelmsford Essex CM3 8BD  <b>Local contact</b> Mr Bob Wilson  Tel: +44 (0) 191-296 8500 Email: Bob.wilson@nwl.co.uk	Environmental Analysis Water Microbiology	Chelmsford



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
<b>Address</b> Horsley Laboratory Water Lane Horsley Newcastle-upon-Tyne NE15 0PA <b>Local contact</b> Mr Bob Wilson  Tel: +44 (0) 191-296 8500 Email: Bob.Wilson@nwl.co.uk	Environmental Analysis Water Microbiology	Horsley
<b>Address</b> Leat House Pattinson Road District 15 Washington NE38 8LB <b>Local contact</b> Mr Ian Lumsdon  Tel: +44 (0)191-296 8500 Email: Ian.Lumsdon@nwl.co.uk	Waste water sampling	Site - Trade effluents Site - Water Waste
<b>Address</b> Horsley Laboratory Water Lane Horsley Newcastle-upon-Tyne NE15 0PA <b>Local contact</b> Mr Ian Lumsdon  Tel: +44 (0)191-296 8500 Email: Ian.Lumsdon@nwl.co.uk	Raw and drinking water sampling	Site - Water clean
<b>Address</b> Chelmsford Laboratory Hanningfield Treatment Works South Hanningfield Chelmsford Essex CM3 8BD <b>Local contact</b> Mr Ian Lumsdon  Tel: +44 (0)191 2968500 Email: Ian.Lumsdon@nwl.co.uk	Raw and drinking water sampling	Site - Water clean
<b>Address</b> Levenmouth Laboratory Levenmouth Waste Water Treatment Works Methilhaven Fife KY8 3LA <b>Local contact</b> Mr Rick Kirkpatrick  Tel: +44 (0)191-296 8500 Email: Rick.Kirkpatrick@nwl.co.uk	Untreated Sewage and Treated Sewage Effluent Sampling	Site - MACS



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

**Site activities performed away from the locations listed above:**

Location details	Activity	Location code
Waste water treatment works	Sampling (teams from Washington)	Site - Water waste
Customer premises - Trade effluents	Sampling (teams from Washington)	Site - Trade effluent
Water treatment works (to include trade effluent to controlled water), service reservoirs and commercial, industrial and domestic premises	Sampling (teams from Horsley and Chelmsford)	Site - Water clean
Commercial, industrial and domestic premises	Sampling (teams from Horsley)	Site - Water Customer
Waste Water treatment works – untreated sewage and treated sewage effluent to MACS requirements	Sampling (teams from Levenmouth)	Site - MACS

**This schedule is ordered as follows:**

**Section 1**

Laboratory Flexible scope of accreditation to ISO/IEC 17025, the Drinking Water Testing Specification (DWTS) and the EA MCERTS: performance standard for organisations undertaking sampling and chemical testing of water (waters)

**Section 2**

Accreditation to ISO/IEC 17025 and the Drinking Water Testing Specification (DWTS)

**Section 3**

Accreditation to ISO/IEC 17025 and the EA MCERTS: performance standard for organisations undertaking sampling and chemical testing of water (waters)

**Section 4 - Levenmouth**

Accreditation to ISO/IEC 17025 and the SEPA MACS Performance Standard MACS-WAT-01 Sampling and chemical testing of water

**Section 5 - Levenmouth**

Accreditation to ISO/IEC 17025 only

**Section 6**

Accreditation to ISO/IEC 17025 only

**Note: accreditation to MCERTS (waters), MACS (waters) and DWTS automatically confers an equivalent accreditation to ISO/IEC 17025**



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<b>SECTION 1 – Flexible Scope</b>			
	<p>The laboratory holds a flexible scope of accreditation for inorganic and organic chemistry test methods managed by the Howdon Laboratory in accordance with TI347 covering the following:</p> <ul style="list-style-type: none"> <li>• Incorporation of additional determinands (Howdon only) or matrices (Howdon, Horsley and Chelmsford) to existing accredited methods</li> <li>• Authorising the use of replacement equipment for existing methods (Howdon, Horsley and Chelmsford)</li> <li>• Development of new methods for matrix types and using techniques and instruments that appear on the fixed scope (Howdon only)</li> </ul> <p>Please contact the laboratory for details of the individual determinands and matrices that can be analysed.</p>	<p>Meeting the requirements of The Drinking Water Testing Specification (DWTS) and the requirements of the Environment Agency MCERTS Performance Standard - sampling and chemical testing of untreated sewage, sewage effluent and trade effluent as applicable</p>	<p>Howdon Horsley Chelmsford</p>
<b>END of SECTION 1</b>			



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<b>SECTION 2 – Sampling and Analysis to DWTS specification</b>			
<p>WATERS, drinking, surface and groundwaters</p> <p>Distribution samples including:-</p> <p>Consumer taps Standpipes Commercial and industrial premises Service reservoirs Water treatment works Raw waters including: Impounding reservoirs Spring supplies Boreholes</p>	<p>Sampling and Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614)</p> <p>Sampling with subsequent analysis by an ISO/IEC 17025 and DWTS Accredited Laboratory for Chemical and Microbiological testing (including cryptosporidium)</p>	<p><b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b></p> <p>Documented In-House Methods L1 and L17 for sampling based on:</p> <p>a) Standing Committee of Analysts Methods b) "The Microbiology of Drinking Water 2010" c) DWI SOP for the monitoring of Cryptosporidium oocysts to satisfy water supply (water quality) regulations 2000:Part 1</p>	<p>Site-Water Clean</p>
<p>WATERS, drinking, surface and groundwaters</p> <p>Consumer taps Standpipes Commercial and industrial premises Raw waters including: Impounding reservoirs Spring supplies Boreholes</p>	<p><u>Physical and Chemical Tests</u></p> <p>Free Residual Chlorine Total Residual Chlorine</p>	<p>Documented In-House Methods using:</p> <p>Method L13 using a Colorimetric Hach SL 1000 Meter</p>	<p>Site-Water Clean</p>
<p>WATERS, drinking, surface and groundwaters</p> <p>Consumer taps Standpipes Commercial and industrial premises Raw waters including: Impounding reservoirs Spring supplies Boreholes</p>	<p>Temperature Measurements</p>	<p>Method L7 using Digital Thermometers</p>	<p>Site-Water Clean</p>
<p>Consumer taps, Commercial and industrial premises</p>	<p>Sampling with subsequent analysis by an ISO/IEC 17025 and DWTS Accredited Laboratory for Chemical and Microbiological testing</p>	<p>Method L1</p>	<p>Site-Water Customer</p>



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
WATERS, drinking, surface and groundwaters	<u>Physical and Chemical Tests</u> (cont'd)	Documented In-House Methods using:	
Consumer taps, Commercial and industrial premises	Free Residual Chlorine Total Residual Chlorine	Method L13 using a Colorimetric Hach SL 1000 Meter	Site-Water Customer
Consumer taps, Commercial and industrial premises	Temperature Measurements	Method L7 using Digital Thermometers	Site-Water Customer
WATERS – Drinking and Raw Waters from household & commercial taps	Sampling and on-site testing For the purpose of the Enforcement of The Private Water Supplies (England) Regulations 2016 for Chemical & Microbiological Testing	Documented In-House Method L1	Site-Water Customer
Raw (Ground and Surface) and Drinking Waters	<u>Inorganic Testing:</u> Ammonium Chloride Soluble reactive Phosphate (as Phosphorus) Nitrite Total Oxidised Nitrogen Nitrate (calculated)	Method HY280 using automated colorimetry	Howdon
Raw (Ground and Surface), and Drinking Waters	Ammonium Chloride Soluble reactive Phosphate (as Phosphorus) Nitrite Total Oxidised Nitrogen Nitrate (calculated)	Method HY280 using Automated Colorimetry	Chelmsford
Drinking Waters	Turbidity	Method HY272 by Turbidimeter	Horsley
Raw (Ground and Surface), and Drinking Waters	Turbidity	Method HY272 by Turbidimeter	Chelmsford



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
Raw (Ground and Surface) and Drinking Waters	<u>Inorganic Testing:</u> (cont'd) Colour (filtered)	Method HY266 using Spectrophotometry	Howdon
Raw (Ground and Surface) and Drinking Waters	Dry Residues	Method HY212 using gravimetry	Howdon
Raw (Ground and Surface) and Drinking Waters	Suspended solids	Method HY217 using gravimetry	Howdon
Raw (Ground and Surface) and Drinking Waters	Total dissolved solids	Method HY225 using gravimetry	Howdon
Raw (Ground and Surface) and Drinking Waters	Total organic carbon (TOC) Dissolved organic carbon (DOC)	Method HY210 using catalytic oxidation followed by Infra-red detection	Howdon
Raw (Ground and Surface) and Drinking Waters	Bromate Bromide Chlorate	Method HY 281 using IC-MS	Howdon
Raw (Ground and Surface) and Drinking Waters	Turbidity	Method HY272 using turbidity Meter	Howdon
Raw (Ground and Surface) and Drinking Waters	Total cyanide	Method HY243 using colorimetry	Howdon
Raw (Ground and Surface) and Drinking Waters	pH Fluoride Conductivity Alkalinity Carbonate (calculated) Bicarbonate (calculated)	Method HY283 using Automated ion selective electrode Automated non selective Electrode Automated titration	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	Howdon
Raw (Ground and Surface) and Drinking Waters	<u>Inorganic Testing:</u> (cont'd)  Metals: Aluminium Barium Iron Manganese Phosphorous Potassium Sodium Total Boron Soluble Boron Calcium Magnesium Total Hardness Sulphate Silica as SiO <sub>2</sub>		
Raw (Ground and Surface Water) and Drinking Water	Metals: Antimony Arsenic Cadmium Chromium Copper Lead Nickel Mercury Selenium Silver Zinc	Method HY267 using ICP-MS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
Raw (Ground and Surface) and Drinking Waters	<u>Organic Testing:</u> Trichloromethane Trichloroethene Tribromomethane Tetrachloroethene Tetrachloromethane Dichlorobromomethane Dibromochloromethane Benzene 1,2 Dichloroethane Vinyl Chloride	Method HY277 using Headspace GC/MS	Howdon
Raw (Ground and Surface) and Drinking Waters	Total THMs Total Ethenes (total of those measured)	Method HY277 by Calculation	Howdon
Raw (Ground and Surface) and Drinking Waters	<u>Polyaromatic Hydrocarbons:</u> Anthracene Fluoranthene Benzo(b)fluoranthene* Benzo(k)fluoranthene* Benzo(a)pyrene Indeno(1,2,3-cd)pyrene* Benzo(ghi)perylene*	Method HY 123 using GC-MS	Howdon
Raw (Ground and Surface) and Drinking Waters	Total PAH (total of those marked *)	Method HY 123 by calculation	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
Raw (Ground and Surface) and Drinking Waters	<u>Organic Testing:</u> (cont'd)  <u>Phenyl Urea Herbicides and Conazoles:</u> Metamitron (41394-05-2) Metoxuron (19937-59-8) Carbetamide (16118-49-3) Monuron (150-68-5) Carbofuran (1563-66-2) Tebuthiuron (34014-18-1) Chlortoluron (15545-48-9) Methabenzthiazuron (18691-97-9) Isoproturon (34123-59-6) Diuron (330-54-1) Linuron (330-55-2) Epoxiconazole (133855-98-8) Tebuconazole (107534-96-3) Prothioconazole (178928-70-6) Propiconazole(60207-90-1) Carbendazim (10605-21-7) Dimethenamid (163515-14-8) Ethofumesate,(26225-79- 6) Triallate(2303- 17-5) Prosulfcarb (52888-80-9) Flufenacet(142459-58-3 ) Oxamyl (23135-22-0)	Method HY270 using LC-MS-MS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
Raw (Ground and Surface) and Drinking Waters	<u>Organic Testing:</u> (cont'd) Metaldehyde	Method HY127 using direct injection LC-MS	Howdon
Raw (Ground and Surface) and Drinking Waters	<u>Alkylphenols:</u> 4-tert-octylphenol, nonylphenol, 4-n-octylphenol 4-n-nonylphenol	Method 239 using solid phase extraction and GC-MS	Howdon
Raw (Ground and Surface) and Drinking Waters	<u>Organochlorine pesticides:</u> Dichlobenil Pentachlorobenzene Tecnazene Trifluralin alpha-BHC Hexachlorobenzene beta-BHC gamma-BHC Chlorthalonil Alachlor Heptachlor Chlorthal-Dimethyl (DCPA) Isodrin Aldrin alpha-Endosulfan beta-Endosulfan Endrin Dieldrin P,p-DDT Heptachlor-EXO-Epoxyde Heptachlor-ENDO-Epoxyde Heptachlor-epoxydes	Method HY275 using GC-MSMS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
Raw (Surface and Ground Waters) and Drinking Waters	<u>Organic Testing:</u> (cont'd)  <u>Acidic Herbicides, specifically:</u> Clopyralid MCPP MCPA Dichlorprop Bromoxynil Triclopyr Pentachlorophenol (PCP) Fenoprop MCPB Bentazone Benazolin 2,4-D 2,4,5-T 2,4-DB Propyzamide Ioxynil Picloram Chloridazon Lenacil Asulam Dicamba Fluorxypyr Quinmerac	Method HY271 using online enrichment followed by LC-MS/MS	Howdon
Raw (Ground and Surface) and Drinking Waters	Chlormequat	Method HY118 by LC-MS/MS	Howdon
Raw (Ground and Surface) and Drinking Waters	Glyphosate AMPA	Method HY121 using Off-line FMOC-CL derivitisation followed by direct injection LC-MS/MS	Howdon
Surface Water, Drinking Water	Taste & Odour Compounds; Geosmin 2-Methylisoborneol (2-MIB)	HY268 by GCMSMS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS, drinking, surface and groundwaters (cont'd)</p> <p>Raw (Surface and Ground Waters) and Drinking Water</p>	<p>Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)</p> <p><u>Organic Testing:</u> (cont'd)</p> <p><u>Organophosphorus Pesticides, Triazine Herbicides, Fungicides and Insecticides:</u>            Dimethoate            Cypermethrin            Propachlor            Simazine            Atrazine            Propazine            Trietazine            Terbutylazine            Metribuzin            Diazinon            Phorate            Propetamphos            Pirimicarb            Metalaxyl            Prometryn            Fenpropidin            Terbutryn            Chlorpyrifos            Pendimethalin            Metazachlor            Chlorfenvinphos            Malathion            Fenpropimorph            Diflufenican</p>	<p><b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b></p> <p>Method HY264 using solid phase extraction and GC-MS</p>	<p>Howdon</p>



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>	
Raw (Ground and Surface), and Drinking Waters	<u>Sensory Tests</u>  Quantitative Taste Quantitative Odour	Method 5 based on SCA "The determination of taste and odour in drinking waters (2014)" using assessed panel	Horsley Chelmsford
	<u>Microbiological Tests</u>  Enumeration:	Documented In-House Methods based on methods published in "The Microbiology of Drinking Water (MODW)"	
Drinking water, bottled water (Emergency Water, water from main supply into a bottle), raw (Ground and Surface) water	Total viable count at 22 °C and 37 °C	Method HY003 using pour plates based on MODW part 7 (2020)	Horsley Chelmsford
Drinking water, bottled water (Emergency Water, water from main supply into a bottle), raw (Ground and Surface) water	Coliforms and <i>Escherichia coli</i> (presumptive and confirmed)	Method HY093 using MLGA based on MODW part 4 (2016)	Horsley
Drinking water, Raw (Ground Water & Surface Water), Bottled water (Emergency Water, water from main supply into a bottle)	Coliforms and <i>E coli</i>	HY073 using Colilert based on MODW part 4 (2016)	Horsley
Drinking water, bottled Water (Emergency Water, water from main supply into a bottle), Raw (Ground and Surface) Water	Coliforms and <i>E coli</i>	HY073 using Colilert based on MODW part 4 (2016)	Chelmsford



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code		
WATERS, drinking, surface and groundwaters (cont'd)	Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)	<b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b>			
	<u>Microbiological Tests</u> (cont'd)				
	Enumeration: (cont'd)				
Drinking water, bottled water (Emergency Water, water from main supply into a bottle), raw (Ground and Surface) water	Clostridium perfringens (presumptive and confirmed)			Method HY019 using membrane filtration TSC Agar based on MODW part 6 (2021)	Horsley
Drinking water, raw (Ground and Surface) water	Clostridium perfringens (presumptive and confirmed)			Method HY019 using membrane filtration TSC Agar based on MODW part 6 (2021)	Chelmsford
Drinking water, bottled water (Emergency Water, water from main supply into a bottle)	Enumeration of sulphite reducing clostridia			Method HY087 using membrane filtration based on based on MODW part 6 (2021)	Horsley
Drinking water, bottled water (Emergency Water, water from main supply into a bottle), raw (Ground and Surface) water	Enterococci (presumptive and confirmed)			Method HY004 using membrane filtration based on MODW part 5 (2012)	Horsley
Drinking water, raw (Ground and Surface) water	Enterococci (presumptive and confirmed)	Method HY004 using membrane filtration based on MODW part 5 (2012)	Chelmsford		
Drinking water, bottled water (Emergency Water, water from main supply into a bottle)	<i>Pseudomonas aeruginosa</i> (presumptive and confirmed)	Method HY072 using membrane filtration based on MODW part 8 (2015)	Horsley		



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS, drinking, surface and groundwaters (cont'd)</p> <p>Raw (Ground and Surface), Drinking water</p>	<p>Analysis for the purpose of enforcement of The Water Supply (Water Quality) Regulations 2016 (SI 2016/614) (cont'd)</p> <p><u>Microbiological Tests</u> (cont'd)</p> <p>Isolation and enumeration of <i>Cryptosporidium</i> oocysts</p>	<p><b>Methodology meeting the requirements of The Drinking Water Testing Specification (DWTS)</b></p> <p>CM001 using the Filta-Max xpress Method based on 'the Microbiology of Drinking Water (2010) - Part 14 - Methods for the isolation, identification and enumeration of <i>Cryptosporidium</i> oocysts and <i>Giardia</i> cysts'</p>	<p>Horsley</p>

**END of SECTION 2**



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<b>SECTION 3 – Sampling and Analysis to MCERTS (Water)</b>			
WASTE WATERS	<u>Sampling with subsequent analysis by an ISO/IEC 17025 and MCERTS (waters) Accredited Laboratory</u>	<b>Documented In-House Method to meet the requirements of the Environment Agency MCERTS Performance Standard - sampling and chemical testing of untreated sewage, sewage effluent and trade effluent</b>	
Untreated sewage Treated sewage effluent Trade effluent (to controlled water)	Sampling with subsequent analysis by an ISO/IEC 17025 and MCERTS waters Accredited Laboratory for Chemical Testing	Documented In-House Methods for sampling (Manual and Automatic Methods)	
Untreated sewage Treated sewage effluent Trade effluent (to controlled water)		Method L4 using manual spot Sampling	Site-Water Waste Site-Water Clean Site Water Customer
Trade effluent (to controlled water)	On-site measurement of free and total residual chlorine	Method L25 using a colorimetric Hach SL1000 Meter	Site-Water Clean
		<b>Documented in house method to meet the requirements of The Urban Waste Water Treatment (England &amp; Wales) Regulations 1994</b>	
Treated sewage effluent		Method L18 using refrigerated automatic sampling equipment	Site-Water Clean
Untreated sewage Treated sewage effluent Trade effluent (to controlled water)		Method L18 using refrigerated automatic sampling equipment	Site-Water Waste



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code		
WASTE WATERS (cont'd)	<u>Chemical Testing</u>	<b>Documented In-House Method to meet the requirements of the Environment Agency MCERTS Performance Standard - sampling and chemical testing of untreated sewage, sewage effluent and trade effluent</b>			
	<u>Inorganic Testing</u>				
Untreated sewage, treated sewage effluent and trade effluent	Biochemical Oxygen Demand			In-house method H042 Manual and Robotic	Howdon
Untreated sewage, treated sewage effluent and trade effluent	Chemical Oxygen Demand			In-house method H171	Howdon
Untreated sewage, treated sewage effluent and trade effluent	pH			In-house method H031	Howdon
Untreated sewage, treated sewage effluent and trade effluent	Suspended solids			In-house method H035	Howdon
Untreated sewage, treated sewage effluent and trade effluent	Aluminium Cadmium Copper Iron Manganese Nickel Zinc			In-house method H225 by ICP-MS	Howdon
Untreated sewage, treated sewage effluent and trade effluent	Phosphorus			In-house method H230 by ICP-OES	Howdon
Treated sewage effluent	Ammonium Nitrate Nitrite Total Oxidised Nitrogen (TON)			In-house method H219 by discrete colorimetric analyser	Howdon





1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<b>SECTION 4 – Sampling and Analysis to SEPA MACS (Water) and ISO/IEC 17025 LEVENMOUTH laboratory</b>			
WASTEWATERS to SEPA MACS scheme	<u>Sampling with subsequent analysis by an ISO/IEC 17025 and SEPA MACS Accredited Laboratory</u>	<b>Documented In-House Methods to meet the requirements of the SEPA MACS Standard - Sampling and chemical testing of water</b>	
Wastewater: Crude Sewage (untreated sewage influent) Final Effluent (treated sewage effluent)	Sampling of wastewater for Subsequent chemical testing	Spot sampling in-house method LEV 006 Automated sampling in-house Method LEV 007	Levenmouth Site - MACS
WASTEWATERS to SEPA MACS scheme	<u>Inorganic Testing</u>		
Crude Sewage (untreated sewage influent) Final Effluent (treated sewage effluent)	Biochemical Oxygen Demand	H042 using oxygen meter Manual	Levenmouth
Crude Sewage (untreated sewage influent) Final Effluent (treated sewage effluent)	Chemical Oxygen Demand	H171 using colorimetry	Levenmouth
Final Effluent (treated sewage effluent)	Suspended Solids	H035 using Gravimetry	Levenmouth
Final Effluent (treated sewage effluent)	pH	Method LEV003 using pH meter	Levenmouth
<b>END of SECTION 4</b>			



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<b>SECTION 5 –Analysis to ISO/IEC 17025 only LEVENMOUTH laboratory</b>			
<p>WASTEWATERS</p> <p>Crude Sewage (untreated sewage influent)</p> <p>Crude Sewage (untreated sewage influent)</p>	<p><u>Inorganic Testing</u></p> <p>Suspended Solids</p> <p>pH</p>	<p>Documented In-House Methods to meet the requirements of ISO/IEC 17025</p> <p>H035 using Gravimetry</p> <p>Method LEV003 using pH meter</p>	<p>Levenmouth</p> <p>Levenmouth</p>

**END of SECTION 5**



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<b>SECTION 6 Sampling and Analysis to ISO/IEC 17025</b>			
SEDIMENTS/ SLUDGES	<u>Chemical Tests</u>	Documented In-House Methods	
Sludge, ashes, soils, sediments and dusts	<u>Poly Chlorinated Dibenzo-p-Dioxins (PCDD):</u> 2,3,7,8-TCDD 1,2,3,7,8-PeCDD 1,2,3,6,7,8-HxCDD 1,2,3,4,7,8-HxCDD 1,2,3,7,8,9-HxCDD 1,2,3,4,6,7,8-HpCDD OCDD  <u>Poly Chlorinated Dibenzo Furans (PCDF):</u> 2,3,7,8-TCDF 1,2,3,7,8-PeCDF 2,3,4,7,8-PeCDF 1,2,3,4,7,8-HxCDF 1,2,3,6,7,8-HxCDF 2,3,4,6,7,8-HxCDF 1,2,3,7,8,9-HxCDF 1,2,3,4,6,7,8-HpCDF 1,2,3,4,7,8,9-HpCDF OCDF	Method O036 using solvent extraction followed by HRGC/HRMS	Howdon
Sludge, ashes, soils, sediments and dusts	<u>Polychlorinated biphenyls:</u> Congener 77 Congener 81 Congener 105 Congener 114 Congener 118 Congener 123 Congener 126 Congener 156 Congener 157 Congener 167 Congener 169 Congener 189	Method O084 using HRGC/HRMS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>ATMOSPHERIC POLLUTANTS AND EFFLUENTS - STACK GAS SAMPLES</p> <p>Filters Probe and impinger rinses (toluene, acetone and water) XAD-2 resin trap</p>	<p><u>Chemical Tests (cont'd)</u></p> <p><u>Poly Chlorinated Dibenzo-p-Dioxins (PCDD):</u> 2,3,7,8-TCDD 1,2,3,7,8-PeCDD 1,2,3,6,7,8-HxCDD 1,2,3,4,7,8-HxCDD 1,2,3,7,8,9-HxCDD 1,2,3,4,6,7,8-HpCDD OCDD</p> <p><u>Poly Chlorinated Dibenzo Furans (PCDF):</u> 2,3,7,8-TCDF 1,2,3,7,8-PeCDF 2,3,4,7,8-PeCDF 1,2,3,4,7,8-HxCDF 1,2,3,6,7,8-HxCDF 2,3,4,6,7,8-HxCDF 1,2,3,7,8,9-HxCDF 1,2,3,4,6,7,8-HpCDF 1,2,3,4,7,8,9-HpCDF OCDF</p>	<p>Documented In-House Methods based on the following national, international and other recognised standards</p> <p>BS EN 1948-2:2006 and BS EN 1948-3:2006. Extraction followed by High Resolution Gas Chromatography/High Resolution Mass Spectrometry (GCHRMS) (SOP O055)</p>	Howdon
<p>Filters Probe and impinger rinses (toluene, acetone and water) XAD-2 resin trap</p>	<p><u>Polychlorinated Biphenyls(PCBs):</u> Congener 77 Congener 81 Congener 105 Congener 114 Congener 118 Congener 123 Congener 126 Congener 156 Congener 157 Congener 167 Congener 169 Congener 189</p>	<p>BS EN 1948-4:2010+A1:2013. Extraction followed by GCHRMS (SOP O084)</p>	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p><b>ATMOSPHERIC POLLUTANTS AND EFFLUENTS – DIGESTER GASES</b></p> <p>Digester gases including biomethane Collected on Tenax TA sorbent tube</p> <p>Digester gases including biomethane Collected in Tedlar bags</p>	<p><u>Chemical Tests</u> (cont'd)</p> <p><u>Siloxanes:</u> Octamethyltrisiloxane Octamethylcyclotetrasiloxane Decamethyltetrasiloxane Decamethylcyclopentasiloxane Dodecamethylpentasiloxane Dodecamethylcyclohexasiloxane</p> <p><u>Bulk and Trace components:</u> Methane Ethane Ethylene Propane Carbon dioxide Oxygen Nitrogen Hydrogen Hydrogen sulphide</p>	<p>Documented In-House Methods</p> <p>Method O108 using Thermal Desorption and Ion Trap GC-MS</p> <p>Method O109 using GCFID, GCTCD and GCFPD</p>	<p>Howdon</p> <p>Howdon</p>



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Groundwater, Surface water, Landfill Leachate, Trade Effluent, Treated Sewage Effluent	Alkalinity	H109 using autotitrator	Howdon
Landfill Leachate, Untreated Sewage, Treated sewage effluent, Trade Effluent, Groundwater	Bromide Chloride Fluoride Sulphate	H090 using Ion Chromatography	Howdon
Groundwater, Surface water, Landfill Leachate Trade Effluent, Treated Sewage Effluent	Formaldehyde - free - total	H040 using spectrophotometry H039 using spectrophotometry	Howdon
Groundwater, Untreated Sewage, Treated Sewage Effluent & Trade Effluent	pH	H031 by pH meter	Howdon
Groundwater, Surface Water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent	Boron	H228 by ICP-OES	Howdon
Groundwater, Surface Water, Landfill Leachate, Trade Effluent, Untreated Sewage, Treated Sewage Effluent	Mercury	H108 using Cold Vapour Atomic Fluorescence Spectrometry	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Groundwater, Surface Water, Landfill Leachate, Untreated Sewage, Trade Effluent, Treated Sewage Effluent	Settleable solids	H036 using Gravimetry	Howdon
Treated sewage effluent, Untreated Sewage & trade effluent	Surfactants - anionic	H227 using Manual Extraction and Colorimetry	Howdon
Untreated Sewage & trade effluent	- non-ionic	H235 using Colorimetry	Howdon
Groundwater, Surface water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent	Suspended solids	H035 using Gravimetry	Howdon
Groundwater, Surface water, Landfill Leachate, Trade Effluent, Untreated Seage, & Treated Sewage Effluent	<u>Elements:</u> Calcium Magnesium Potassium Sodium Hardness – total (by calculation)	H231 using ICP-OES	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Groundwater, Surface water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent to Sewer, Trade Effluent to controlled Water	<u>Elements:</u> Aluminium Antimony Arsenic Barium Cadmium Cobalt Chromium Copper Iron Lithium Manganese Molybdenum Nickel Lead Selenium Tin Titanium Thallium Vanadium Zinc	H225 using ICP-MS	Howdon
Groundwater, Surface water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent to Sewer	Beryllium	H225 using ICP-MS	Howdon
Groundwater, Surface water, Untreated Sewage, Treated Sewage Effluent, Trade Effluent, Landfill Leachate	Phosphorus	H230 using ICP-OES	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Trade Effluent, Landfill Leachate, Surface Water, Ground Water, Untreated Sewage	Cyanide – Total	H236 using Segmented Flow Colorimetry	Howdon
Groundwater, Surface water, Landfill Leachate Trade Effluent, Treated Sewage Effluent	Sulphides	H041 using spectrophotometry	Howdon
Groundwater, Surface water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent	Biochemical oxygen demand	H042 using oxygen meter Manual and Robotic	Howdon
Groundwater, Surface water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent	Chemical oxygen demand	H171 using colorimetry	Howdon
Groundwater, Surface water, Landfill Leachate, Trade Effluent, Treated Sewage Effluent	Conductivity	H032 using meter/autotitrator	Howdon
Untreated Sewage, Treated Sewage Effluent and Trade Effluent, raw (ground and surface), Landfill Leachate	Elements: Silver	H233 using ICPOES	Howdon
Trade Effluent, Landfill Leachate, Surface Water, Ground Water, Untreated Sewage	Phenolic compounds (as Phenol)	H237 using Segmented Flow Colorimetry	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Groundwater, Surface water, Drinking Water (non regulatory), Treated Sewage Effluent	<u>Poly Chlorinated Dibenzo-p-Dioxins (PCDD):</u> 2,3,7,8-TCDD 1,2,3,7,8-PeCDD 1,2,3,6,7,8-HxCDD 1,2,3,4,7,8-HxCDD 1,2,3,7,8,9-HxCDD 1,2,3,4,6,7,8-HpCDD OCDD	Method O036 using solvent extraction followed by HRGC/HRMS	Howdon
	<u>Poly Chlorinated Dibenzo Furans (PCDF):</u> 2,3,7,8-TCDF 1,2,3,7,8-PeCDF 2,3,4,7,8-PeCDF 1,2,3,4,7,8-HxCDF 1,2,3,6,7,8-HxCDF 2,3,4,6,7,8-HxCDF 1,2,3,7,8,9-HxCDF 1,2,3,4,6,7,8-HpCDF 1,2,3,4,7,8,9-HpCDF OCDF		
Groundwater, Surface Water, Landfill Leachate, Untreated Sewage, Treated Sewage Effluent, Trade Effluent	<u>Anions and Cations:</u> Ammonia Nitrate Nitrite Total Oxidised Nitrogen (TON) Dissolved ortho phosphate Total phosphate	In-house method H219 by discrete colorimetric analyser	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Groundwater, Surface Water, Treated sewage effluent, Drinking Water (non regulatory) Trade Effluent	<u>Organochlorine Pesticides:</u> Aldrin BHC-alpha BHC-beta BHC-gamma DDE-op DDE-pp DDT-pp Dieldrin Endosulphan-alpha Endosulphan-beta Endrin Hexachlorobenzene Isodrin TDE-op TDE-pp Tecnazene Trifluralin	Method O052 using GCMS	Howdon
Ground Water, Surface Water, Treated Sewage Effluent	DDT-op	Method O052 using GCMS	
Groundwater, Surface Water, Treated sewage effluent, Drinking Water (non regulatory), Trade Effluent	<u>Polychlorinated Biphenyls:</u> Congener 28/31 Congener 52 Congener 101 Congener 118 Congener 138 Congener 153 Congener 180	Method O052 using GCMS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Ground water, Surface water, Drinking Water (non regulatory), Treated Sewage Effluent	<u>Polychlorinated Biphenyls:</u> Congener 77 Congener 81 Congener 105 Congener 114 Congener 118 Congener 123 Congener 126 Congener 156 Congener 157 Congener 167 Congener 169 Congener 189	Method O084 using HRGC/HRMS	Howdon
Groundwater, Surface water, Treated Sewage Effluent	<u>Triazines &amp; Organophosphorus Pesticides:</u> Atrazine Azinphos-ethyl Azinphos-methyl Chlorfenvinphos Diazinon Dichlorvos Dimethoate Fenitrothion Malathoin Propazine Propetamphos Trietazine Simazine	Method O110 by GC-MSMS	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Groundwater, Surface water, Trade effluent, Treated sewage effluent, Drinking Water (non-regulatory)	<u>Alcohols:</u> Methanol Ethanol Acetone Propan-2-ol Butan-2-ol Methyl ethyl ketone Butan-1-ol Isobutanol Propan-1-ol Methylisobutylketone	Method O015 using Headspace-GC	Howdon
Groundwater, Surface water, Drinking Water (non regulatory), Trade effluent, Treated Sewage Effluent	Total Petroleum Hydrocarbons	Method O082 using GC-FID	Howdon
Groundwater, Surface water, Drinking Water (non regulatory), Trade effluent, Treated Sewage Effluent	Aliphatic Mineral Oil fractionation	Method O082 using GC-FID	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS (cont'd)</p> <p>Groundwater, Drinking Water (non regulatory)</p> <p>Treated Sewage Effluent, Drinking Water (non-regulatory), Trade Effluent, Surface Water</p>	<p><u>Chemical Tests</u> (cont'd)</p> <p><u>Polynuclear Aromatic Hydrocarbons:</u>            Acenaphthene            Acenaphthylene            Anthracene            Benz(a)anthracene            Benzo(b)fluoranthene            Benzo(k)fluoranthene            Benzo(a)pyrene            Benzo(ghi)perylene            Chrysene            Dibenz(a,h)anthracene            Fluoranthene            Fluorene            Indeno(1,2,3,c,d)pyrene            Naphthalene            Phenanthrene            Pyrene            Total PAH</p> <p><u>Organotin Compounds:</u>            Dibutyl tin (DBT)            Tributyl Tin (TBT)            Tetrabutyl Tin (TeBT)            Triphenyl Tin (TPT)</p>	<p>Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers</p> <p>Method O044 using GCMS</p> <p>Method O106 using GC-MS-MS</p>	<p>Howdon</p> <p>Howdon</p>



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**  
**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS (cont'd)</p> <p>Raw Water (Ground and Surface), Trade Effluent, Treated Sewage, Landfill Leachate</p>	<p><u>Chemical Tests</u> (cont'd)</p> <p><u>Volatile Organic Compounds:</u>            1,1-Dichloroethene            Dichloromethane            Trans-1,2-Dichloroethene            MTBE            1,1-Dichloroethane            Cis-1,2-Dichloroethene            2,2-Dichloropropane            Bromochloroethane            Trichloromethane            THF            1,2-Dichloroethane            1,1,1-Trichloroethane            1,1-Dichloropropene            Carbon Tetrachloride            Benzene            Trichloroethene            1,2-Dichloropropane            Dibromomethane            Cis-1,3-Dichloropropene            1,1,2-Trichloroethane            Tetrachloroethene            1,3-Dichloropropane            Dibromochloromethane            1,2-Dibromoethane            Chlorobenzene            1,1,1,2-Tetrachloroethane            Ethylbenzene            m+p-xylenes            o-xylene            Styrene            Bromoform            Isopropylbenzene            1,1,2,2-Tetrachloroethane            Bromobenzene</p>	<p>Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers</p> <p>Method O102 using GC-MS</p>	<p>Howdon</p>



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No:** 138 **Issue date:** 13 November 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)	Documented In-House Methods based on 'Methods for the Examination of Water and Associated Materials', Department of the Environment Standing Committee of Analysts, Methods developed by instrument manufacturers	
Raw Water (Ground and Surface, Trade Effluent, Treated Sewage, Landfill Leachate (cont'd))	<u>Volatile Organic Compounds:</u> (cont'd) 2-Chlorotoluene 4-Chlorotoluene 1,3,5-Trimethylbenzene Tert-Butylbenzene 1,2,4-Trimethylbenzene Sec-Butylbenzene 1,3-Dichlorobenzene Isopropyltoluene 1,4-Dichlorobenzene 1,2-Dichlorobenzene n-Butylbenzene 1,2-Dibromo-3-Chloropropane 1,2,4-Trichlorobenzene Hexachlorobutadiene Naphthalene Toluene MIBK 1,2,3-Trichlorobenzene Bromodichloromethane Trans-1,3-Dichloropropene 1,2,3-Trichloropropane n-Propylbenzene	Method O102 by GC-MS	Howdon
Treated sewage effluent and trade effluent	2,4,6-Trinitrophenol 4-Nitrophenol 2,4-Dinitrophenol 2,6-Dinitrophenol 2-Nitrophenol Nitrobenzene	In house method O042 by HPLC-UV	Howdon



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS, waste and groundwaters</p> <p>Boreholes (Ground Water &amp; Landfill Boreholes)</p>	<p><u>Physical and Chemical Tests</u></p> <p>Volume of Water in Boreholes</p>	<p>Documented In-House Methods using:</p> <p>Method L15 using Electrochemical Water Detector</p>	<p>Site Water Waste Site-Water Customer</p>
<p>WATERS</p> <p>Ground Water, Surface Water, Landfill Leachate</p> <p>Ground Water, Surface Water, Landfill Leachate, Trade Effluent, Treated Sewage, Untreated Sewage</p> <p>Ground Water, Surface Water, Trade Effluent, Treated Sewage, Untreated Sewage</p> <p>Ground Water, Surface Water, Landfill Leachate, Untreated Sewage, Treated Sewage, Trade Effluent</p>	<p><u>Chemical Tests</u></p> <p>Dissolved Oxygen</p> <p>Chemical testing</p> <p>Chemical testing</p> <p>Temperature measurements</p>	<p>Documented In-House Methods:</p> <p>Method L10 using Electrochemical Sensor</p> <p>Method L4 by manual spot sampling</p> <p>Method L4 by manual spot sampling</p> <p>Method L7 using Calibrated thermometer</p>	<p>Site-Water Customer</p> <p>Site-Water Customer</p> <p>Site-Water Waste</p> <p>Site-Water Waste Site-Water Customer</p>



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS	<u>Microbiological Tests</u>	Documented In-House Methods based on methods published in "The Microbiology of Drinking Water (MODW)"	
Bottled water	Enumeration:  Total viable count at 22 °C and 37 °C	Method HY003 using pour plates based on MODW part 7 (2020)	Horsley
Bottled water	Coliforms and <i>Escherichia coli</i> (presumptive and confirmed)	Method HY093 using MLGA based on MODW part 4 (2016)	Horsley
Bottled water	<i>Clostridium perfringens</i> (presumptive and confirmed)	Method HY019 using membrane filtration TSC Agar based on MODW part 6 (2021)	Horsley
Bottled water	Enumeration of sulphite reducing clostridia	Method HY087 using membrane filtration based on based on MODW part 6 (2021)	Horsley
Bottled water	Enterococci (presumptive and confirmed)	Method HY004 using membrane filtration based on MODW part 5 (2012)	Horsley
Bottled water	<i>Pseudomonas aeruginosa</i> (presumptive and confirmed)	Method HY072 using membrane filtration based on MODW part 8 (2015)	Horsley
Final Treated and Recreational Pool Water	Isolation and enumeration of <i>Cryptosporidium</i> oocysts	CM001 using the Filta-Max xpress Method based on "the Microbiology of Drinking Water (2010) - Part 14 - Methods for the isolation, identification and enumeration of <i>Cryptosporidium</i> oocysts and <i>Giardia</i> cysts"	Horsley



1181

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Microbiological Tests</u> (cont'd)	Documented In-House Methods based on methods published in "The Microbiology of Drinking Water (MODW)"	
Purified water and swimming pool water	Enumeration and Identification: (cont'd)  Coliforms - faecal <i>Escherichia coli</i> (presumptive and confirmed)	Method HY093 using MLGA based on MODW part 4 (2016)	Horsley
Purified water and swimming pool water, Endoscope Waters, Process Water (non-drinking tap water)	<i>Pseudomonas aeruginosa</i> (presumptive and confirmed)	Method HY072 using membrane filtration based on MODW part 8 (2015)	Horsley
Wastewater - Effluent: Swimming pool Water	Enterococci (presumptive and confirmed)	HY004 using membrane filtration based on MODW part 5 (2012)	Horsley
Swimming pool Water	Coliforms and <i>E coli</i>	HY073 using Colilert based on MODW part 4 (2016)	Chelmsford
Swimming pool water	Total viable count at 22 °C and 37 °C	Method HY003 using pour plates based on MODW part 7 (2020)	Horsley Chelmsford
Endoscope Rinse Waters	Total Viable Count Aerobic heterotrophic mesophilic microorganisms at 30°C	Method HY 097 using membrane filtration and culture on R2A agar in duplicate based on Health Technical Memorandum 01-06; Decontamination of flexible endoscopes (HTM01-06)	Horsley
Purified water	Total viable counts at 32.5 °C	HY023 using membrane filtration and R2A agar based on Eu Pharma 2005	Horsley
Purified water	Detection of Endotoxins	CM008 using LAL Gel Clot Technique	Horsley



1181

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Northumbrian Water Ltd**

**Issue No: 138 Issue date: 13 November 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>WATERS (cont'd)</p> <p>Process Waters (hot and cold water systems), Man-made Recreational Waters (swimming pools, spas, Hot-tubs), Hydrotherapy Waters</p> <p>Raw (Ground and Surface) water</p> <p>Process Waters – Tanked hot and cold water systems and non-regulatory drinking water</p> <p>Treated Sewage Effluent, Recreational Water (saline bathing water)</p>	<p><u>Microbiological Tests</u> (cont'd)</p> <p>Enumeration and Identification: (cont'd)</p> <p>Isolation, enumeration and identification of <i>Legionella</i> spp and <i>Legionella pneumophila</i></p> <p>Algal counts and identification of Algal spp (abundance species)</p> <p>Total Viable Counts at 22°C and 37°C</p> <p><i>Escherichia Coli</i> enumeration</p>	<p>Documented In-House Methods based on methods published in “The Microbiology of Drinking Water (MODW)”</p> <p>Method LEG 01 using membrane filtration and selective media</p> <p>BIO 1 using sedimentation with Lugol’s iodine</p> <p>Method HY011 using spread plate technique based on MoDW Part 7 (2020)</p> <p>Method HY002 IDEXX Quantitray Standard Method Based on MoSS Part 3 (2024), MoREW Part 3 (2016)</p>	<p>Horsley</p> <p>Chelmsford</p> <p>Horsley</p> <p>Horsley</p>
<p>WATERS Process water, Drinking water (non regulatory) and recreational water</p> <p>Potable waters and Non-potable waters from domestic hot and cold water systems, Calorifiers, water dispensers, cold water tanks, humidifiers/water spray, Spa Pools, TMV/Mixer taps, instantaneous water heaters and showers/spray taps</p>	<p><u>Sampling with subsequent analysis by an ISO/IEC 17025 Accredited Laboratory For:</u></p> <p>Sampling for the purpose of facilities management and as a standalone practice for microbiological testing including <i>Legionella</i></p>	<p>Documented In-House Methods for sampling</p> <p>Documented In-House Methods L1 and L26 based on BS 7592:2008, BS 8554:2015 and The Microbiology of Drinking Water part 2 2010</p>	<p>Site-Water Customer</p>
END			