


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	Envirochem Analytical Laboratories Ltd	
	Issue No: 044 Issue date: 16 April 2025	
	12 The Gardens Broadcut Fareham Hampshire PO16 8SS	Contact: Mr R Mirzaians Tel: +44 (0)1329 287777 Fax: +44 (0)1329 287755 E-Mail: office@envirochem.co.uk Website: www.envirochem.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 12 The Gardens Broadcut Fareham Hampshire PO16 8SS	Local contact Mr R Mirzaians	Health and Hygiene Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



1227
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Envirochem Analytical Laboratories Ltd

Issue No: 044 Issue date: 16 April 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method No. 3.02 based on HSG 248	B
	Fibre counting	Documented In-House Method No. 3.01, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method No. 3.03, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for asbestos identification	Documented In-House Method No. 2.02 based on HSG248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No. 2.01 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG248	A
ASBESTOS IN SOILS – The Identification of Asbestos fibres in bulk samples of Soil, <i>specifically: Soil</i>	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No. 2.04 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A



1227
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Envirochem Analytical Laboratories Ltd
Issue No: 044 Issue date: 16 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN SOILS – The Identification and Quantification of Asbestos fibres in bulk samples of Soil, <i>specifically: Soil</i>	<u>Health and Hygiene</u> (cont'd) Identification and Quantification of Asbestos content of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021 Documented In-House Method No. 2.04 for identification using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248. Documented In-House Method No. 2.04 for quantification of asbestos.	A
AIRBORNE INDUSTRIAL POLLUTANTS	Determination of Inhalable/Respirable Aerosols (Weighing of filters)	MDHS 14/4:June 2014 by Documented In-House Method, Procedure No 4.02 by gravimetry	A
PAINT	Lead	Procedure 5.18, 6.08 using ICP-OES	A
WATERS Raw Waters, River Waters and Effluents Raw waters, River Waters, Effluents and Leachates	<u>Chemical Tests</u> Bromide Chloride Fluoride Nitrate Sulphate pH Metals, specifically: Arsenic Barium Cadmium Calcium Chromium Cobalt Copper Iron Lead	Documented In-House Methods Procedure 6.02 Procedure 6.03 Procedure 6.08 using ICP-OES	A A A



1227
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Envirochem Analytical Laboratories Ltd
Issue No: 044 Issue date: 16 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd) Raw waters, River Waters, Effluents and Leachates (cont'd)	<u>Chemical Tests</u> (cont'd) Metals, specifically: (cont'd) Mercury Potassium Magnesium Manganese Sodium Nickel Thallium Zinc	Documented In-House Methods Procedure 6.08 using ICP-OES	A
WASTE WATERS Trade Effluent and Treated Sewage Effluent	Total Suspended Solids	Procedure MET 6.10 by gravimetry based on EPA method 160,2	A
SOILS	<u>Chemical Tests</u> Chloride Sulphate pH Metals: Arsenic Barium Beryllium Cadmium Chromium Cobalt Copper Lead Manganese Mercury Molybdenum Nickel Selenium Tin Zinc	Documented In-House Method to meet the requirements of the Environment Agency MCERTS Performance Standard - Chemical Testing of Soil Procedure 5.06, 6.02 Procedure 5.06, 6.03 Procedure 5.18, 6.08 by ICP-OES	A A A



1227
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Envirochem Analytical Laboratories Ltd
Issue No: 044 Issue date: 16 April 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS	<u>Microbiological Tests</u>		
Potable water (non-regulatory), and recreational water (pools and spas)	Enumeration of: Coliforms and <i>Escherichia coli</i> - presumptive Coliforms and <i>E. coli</i> - confirmed	Procedure MET 10.16 by membrane filtration using MLGA Procedure MET 10.17 based on Microbiology of Drinking Water part 4B 2016	A A
Potable water (non-regulatory), recreational water (pools and spas) and process water	Aerobic Colony Count at 22°C and 37°C	Procedure MET 10.04 by pour plate technique, based on the Microbiology of Drinking Water part 7, 2020	A
Potable water (non-regulatory), and recreational water (pools and spas)	<i>Pseudomonas aeruginosa</i> , presumptive and confirmed	Procedure MET 10.07 by membrane filtration, based on the Microbiology of Drinking Water part 8, 2015	A
Potable Water (non-regulatory), process water and recreational water (pools and spas)	Detection and enumeration of: <i>Legionella</i> species	Procedure MET 10.03 by membrane filtration based on ISO 11731:2017 using filtration with washing, using GVPC, Identification latex agglutination (Matrix A & B: Procedure 8, 9 & 10, Media C)	A
	Identification of: <i>Legionella pneumophila</i> serogroup 1 <i>Legionella pneumophila</i> serogroups 2 – 14 <i>Legionella</i> species (other than <i>L. pneumophila</i>)	Procedure MET 10.03 using latex agglutination	A
END			