


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING</p> <p>1366</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>ITW Ltd trading as Avery Weigh-Tronix</p> <p>Issue No: 034 Issue date: 07 August 2025</p>	
	<p>Avery Weigh-Tronix Foundry Lane Smethwick West Midlands B66 2LP</p>	<p>Contact: Mr Jim Cameron Tel: +44 (0) 121 568 1478 E-Mail: jcameron@awtx-itw.com Website: www.averyweigh-tronix.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
LOAD CELLS UP TO 50 TONNE CAPACITY	<p>Determination of load cell error, repeatability error and temperature effect on minimum dead load output</p> <p>Determination of creep error</p> <p>Determination of minimum dead load output return</p> <p>Determination of barometric pressure effects</p> <p>Determination of humidity effects</p> <p>Additional tests for load cells equipped with electronics</p>	<p>OIML R60 Edition 2021 (E) OIML R60 Edition 2017 (E) [Withdrawn]Accuracy Class C and D up to a number of verification intervals consistent with the measurement capabilities.</p>
<p>END</p>		